A-A-51225
9 December 1985

COMMERCIAL ITEM DESCRIPTION

SIMULATOR, ARRHYTHMIA, ECG

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a ECG/Arrhythmia simulator suitable for use in testing functions of arrhythmia monitoring systems.

Salient characteristics:

Shall be a portable, battery operated, simulator capable of testing functions of arrhythmia monitoring systems. Shall be capable of simulating a normal rhythm and 17 common arrhythmias for the purpose of testing the recognition capabilities and proper operation of rate alarms and detection protocol. The following waveform selections shall be provided:

a. VPBI
b. Multifocal run
c. Missed beat
d. Fusion beat
e. VBP 2
f. Couplet (pair)
g. Run
h. R on T wave
i. Asystole
j. Ventricular Fibrillation
k. Atrial Fibrillation
l. Heart block
m. Regular rhythm
n. Paced rhythm
o. Tachycardia
p. Bradycardia
q. Ventricular tachycardia
r. Bigeminy

The instrument shall provide for 12 lead connections. Shall be capable of providing outputs of 1 mV and 1V into lead II. The maximum weight shall be 1 lb. (0.45 kg).

Workmanship. The simulator shall be free from defects ehich detract from its appearance or impair its serviceability.

Unit. Each (EA). One simulator, as specified, constitutes one unit.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

FSC 6910

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

AMSC N/A

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

The following National Stock Number is covered by this document:

6910-01-186-8436

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract/or order.

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS

PGC 59797

Project No. 6910-0436
Location: H1-9-9

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.