A-A-51226
20 December 1985

COMMERCIAL ITEM DESCRIPTION

LINER, SURGICAL SUCTION CANISTER

The General Services Administration has authorized the use of this commerical item description.

This commercial item description covers one type of liner, surgical suction canister, flexible, plastic, 2000 ml capacity, disposable with lid and internal float shut-off valve, and used with 2000 ml cylindrical suction canister 4 by 14 inches (10.2 by 35.6 cm) long.

Salient characteristics:

Shall be a minimum 2000 ml flexible, clear plastic closed system disposable liner, with non-removable lid and internal float shut-off valve, suitable for use with 2000 ml cylindrical suction canister 4 inch (10.2 cm) diameter by approximately 14 inches (35.6 cm) long.

Shall be compatible with 2000 ml canisters that are universally acceptable to any ceiling, wall or portable suction system.

Shall provide adequate asepsis in collecting and disposing of contents.

The vacuum side of the surgical suction canister liner shall include one (1) tube with an in-line aerosol filter.

The aerosol filter shall provide a barrier between the reusable canister and the disposable liner.

The aerosol filter translucent body shall enable visual identification of some types of contamination.

The lid shall have one port marked "patient" for patient tubing and the opposite port shall be marked "vacuum" for vacuum tubing.

The liner shall be able to withstand 4 PSI when inflated with air.

The aerosol filter shall cause no more than a minimal air flow resistance.

FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

AMSC N/A

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.

Tests.

Every liner shall be preassembled, then tested for leaks at 4 PSI before packaging and shipment Aerosol filter shall be tested at airflows that are representative of clinical vacuum suctions and to assure minimal air flow resistance

Workmanship. The surgical suction canister liner shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PC). Ten packages of five liners, as specified, constitute one unit of issue.

Contractor certification The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract

Regulatory requirements.

Federal Food, Drug and Cosmetic Act If the product covered by this document has been determined by the U S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions handling etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity and shall conform to applicable carrier's rules and regulations Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.

The following National Stock Number is covered by this document:

6515-01-074-5063

Ordering data Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

MILITARY INTERESTS

Custodians

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS

PGC 59773

Project No. 6515-0304
TYPED "C-5063"

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.