A-A-51230
29 November 1985

COMMERCIAL ITEM DESCRIPTION

TEST KIT, OCCULT BLOOD DETERMINATION

The General Services Administration has authorized the use of this commercial item description.

The commercial item description covers Test Kit, Occult Blood Determination suitable to test for occult blood in the feces.

Salient characteristics:

Shall be a fully prepared Guaiac Test for occult blood in the feces. Shall be suitable for use in detecting the presence of occult blood in feces by means of a guaiac impregnated paper in a cardboard covering. Each kit shall be capable of performing 100 occult blood tests.

Shall be labeled for IN VITRO DIAGNOSTIC USE.

Each kit shall be capable of performing 100 occult blood tests, and shall contain the following components:

a. 100 guaiac impregnated test slides in a suitable cardboard covering.

b. A sufficient amount of material to produce 100 positive controls.

c. A sufficient amount of denatured alcohol and hydrogen peroxide as a developing solution for a minimum of 100 tests in a suitable container.

d. 100 wooden spatulas.

Sensitivity test. The finished test slides shall develop a definite blue circle or blue ring within 30 seconds when 20 random slides are tested as specified below:

Transfer approximately 25 mg of dried beef blood into a 500 ml volumetric flask and add approximately 200 ml of purified water. Shake the flask until the blood has dissolved, dilute to volume with purified water, and mix well. Transfer one drop of this solution to the impregnated paper. Add 2 drops of the developer solution to the reverse side of the paper, in the area wetted by the blood solution.

Not less than 32 months of the expiration dating period shall remain at time of delivery to the Government.

Workmanship. Kit shall be free from defects which detract from its appearance or impair its serviceability.

FSC 6550

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.

Unit. Each (EA). One kit, consisting of 100 guaiac impregnated test slides, 100 controls, developing solution and 100 wooden spatulas, as specified, constitutes one unit.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, ?rhandling, etc., involving shipment from the supply source to the receiving activity,?r plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

For information only. The following product(s), as of the date of this document, is/are purported by the following companies to meet the salient characteristics of the item(s) described herein:

Ames: Hema-Chek
Smith, Kline Diagnostics: Hemoccult Slide

NOTE: The following National Stock Number is covered by this document:

6550-00-165-6538

Ordering data. (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.

COMMERCIAL ITEM DESCRIPTION A-A-51230

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

Project No. 6550-0031

Location H2-23-37

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.