A-A-51238
29 November 1985

COMMERCIAL ITEM DESCRIPTION

TEST KIT, ANAEROBE IDENTIFICATION

This Commercial Item Description covers a testing system for the performance of 21 biochemical tests for the identification of anaerobes.

Salient characteristics:

Shall be a test strip holding a preset battery of 20 microtubes containing dehydrated media, for the identification of anaerobes from a single colony of organisms.

Shall contain 20 different differential media for performing the following 21 biochemical tests:

1.	Indole	11.	Glycerol	21. Catalase
2.	Urease	12.	Cellobiose	
3.	Glucose	13.	Mannose	
4.	Mannitol	14.	Melezitose	
5.	Lactose	15.	Raffinose	
6.	Saccharose	16.	Sorbitol	
7.	Maltose	17.	Rhamnose	
8.	Salicin	18.	Trehalose	
9.	(d+) Xylose	19.	Gelatin	
10.	(1+) Arabinose	20.	Esculin	

The dehydrated media shall be reconstituted by adding a bacterial suspension formed by the inoculation of an anaerobe basal medium.

The microtubes on the test strip shall be able to be inoculated under aerobic conditions before the test strip is placed in an anaerobic system for incubation.

Shall be supplied with an ampule of anaerobe basal medium and an incubation tray with a lid.

Shall be supplied with a package insert (instructions), a differential chart and report sheet.

After a test strip has been prepared and inoculated in accordance with the manufacturer's instructions, organism identification shall be facilitated by using the differentiation chart, or a reference manual that contains an interpretation scheme for use in identifying clinically significant anaerobic bacteria.

Not less than 16 months of the expiration dating period shall remain at the time of delivery to the Government.

Workmanship. The anaerobic identification test kit shall be free from defects which detract from its appearance or impair its serviceability.

FSC 6550

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:53Z
Check the source to verify that this is the current version before use.

Unit. Each (EA). One set consisting of twenty-five sets of media together with twenty-five incubation covers with lids, twenty-five report sheets and package insert, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Tests shall be conducted on the test strip media, to check the positive and negative biochemical reactions of the media to different viable cultures that include the following:

1.	Clostridium histolyticum	(ATTC 19401)
2.	Bacteroides ovatus	(ATCC 8483)
3.	Propionibacterium acnes	(ATCC 11827)
4.	Clostridium sordellii	(ATCC 9714)
5.	Clostridium perfringens	(ATCC 13124)

The tests shall be conducted in accordance with the manufacturer's instructions. The biochemical reactions shall be recorded and compared with a suitable reference covering expected biochemical reactions.

The sampling for testing shall be in accordance with commercial practice.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:53Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

The following National Stock Number is covered by this document:

6550-01-139-7505

The following National Stock Number covers a reference manual for use in identifying clinically significant anaerobic bacteria.

6550-01-139-7482

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS

PGC 52863

Project No. 6550-0036
Loc. S3, 7-25

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.