A-A-51242
22 November 1985

COMMERCIAL ITEM DESCRIPTION

BURNER, GAS, LABORATORY, (Bunsen, Automatic)

The General Services Administration has authorized the use of this commercial item description in preference to Military Specification MIL-B-36274A.

This commercial item description covers an automatic Bunsen burner, adjustable for use with manufactured, natural, mixed (natural and manufactured) gas, or bottled liquid petroleum gas such as propane, butane, suitable for use in a dental laboratory.

Salient characteristics:

The burner shall be constructed with independently controlled pilot and main flames, and shall be adjustable for use with manufactured, natural, mixed (manufactured and natural) gas, or bottled liquid petroleum gas such as propane or butane. The burner shall consist of the following:

Base housing
Valve actuating housing
"ON-OFF" platform
Detachable burner
Draft guard
Detachable nozzle for each gas
Waxing cup

The pilot flame shall burn continuously and shall be adjustable by means of a pilot needle valve.

The main flame shall be normally off, and shall ignite only when the spring valve mechanism is activated by pressure on the "ON-OFF" platform. The main flame shall be instantly off when pressure on the "ON-OFF" platform is removed. A continuous flame shall be obtained by depressing and slightly rotating the "ON-OFF" platform in either direction, and shall be instantly off when the platform is returned to the released position.

The base housing shall be "D" shaped and shall contain a needle valve with an exterior nozzle orifice at one end and a knob at the other end for regulating the main flame. Gas input to the housing shall be through a serrated inlet connection which shall accommodate 1/4 to 5/16 inch inside diameter tubing. The base housing shall also contain a removable metal spring valve mechanism, which when closed shall prevent the flow of gas to the main flame needle valve.

FSC 6640

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.

The valve actuating housing shall contain the valve actuating mechanism, and shall be securely joined to the top of the base housing. A diaphragm fitted between the two housings shall permit the opening and closing of the spring valve mechanism by the valve actuating mechanism, and shall prevent leakage of gas from the base housing to the atmosphere.

The "ON-OFF" platform shall be flat, mounted on top of the valve actuating housing, and attached to the valve actuating mechanism. Depression and release of the platform shall cause the valve actuating mechanism to open and close the spring valve mechanism.

The detachable burner shall contain the main gas passage, the pilot gas passage, pilot needle valve, and air intake adjustment. The burner shall be attached to the base housing in such a manner that the main gas passage shall be opposite the exterior nozzle orifice of the main flame needle valve, and shall be held securely to the base housing by means of a thumbscrew.

The draft guard, with mounting bracket, shall be attached to the burner and shall be rotatable in order to protect the pilot and main flames against draft from any direction.

The waxing cup shall be 2-1/2 inch diameter and 1/2 inch depth. The cup shall be mountable and rotatable upon a stud provided on the burner, and shall have a loop heat conductor to control the consistency of the wax. The loop heat conductor shall be independently rotatable to permit adjustment for the desired degree of heat transmitted to the waxing cup.

Instructions for maintenance and use shall be furnished with each Bunsen burner.

Workmanship. The Bunsen burner shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One Bunsen burner, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Inspections and tests. Inspections and tests shall be conducted in accordance with commercial practice to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for inspections and tests. Sampling for inspections and tests shall be conducted in accordance with commercial practice. The unit of product for test purposes shall be one Bunsen burner.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

The following National Stock Number is covered by this document:

6640-00-510-6075

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.

For information only. The following product, as of the date of this document, is purported by the following company to meet the salient characteristics of the item dscribed herein:

Teledyne Hanau Touch-O-Matic 77-0 002803

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVILIAN AGENCY COORDINATING ACTIVITY:

Va-OSS

55650

Project No. 6640-1334

Location S3-8-14

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.