A-A-51244
3 September 1986

COMMERCIAL ITEM DESCRIPTION

TUBE, CENTRIFUGE, Conical Bottom

The General Services Administration has authorized the use of this commercial item description in preference to Federal Specification NNN-T-780B.

This commercial item description covers requirements for 15 milliliter (ml) capacity conical bottom glass centrifuge tubes, suitable for use in laboratory procedures.

Salient characteristics:

Shall be 15 milliliter (ml) capacity centrifuge tubes in the following types: type I (graduated) and type II (ungraduated).

Shall be cylindrical in shape with conical bottoms and tops on open ends that are finished with beaded rims. The conical tips shall be convex or straight sided. The wall shall be essentially of uniform thickness with some thickening at the end of the tip being acceptable.

Shall be fabricated of borosilicate glass conforming to ASTM Standard E-438, Standard Specification for Glasses in Laboratory Apparatus.

Shall be dimensioned in accordance with the following:

Overall length	118 millimeters (mm) ± 2 mm
Outside diameter (open end)	19.25 mm +1.00 mm
Outside diameter (below bead)	16.50 to 17.25 mm
Length of taper from tip	41 ± 6 mm
Outside base diameter at tapered end	6.5 mm ± 0.5 mm

Type I centrifuge tube shall be graduated in 0.1 ml increments from 0.1 to 15 ml. Main graduations shall have Arabic numerals at the whole values from 1 to 15 ml. Intermediate graduations for the 0.5 ml values from 0.5 to 14.5 ml may be of equal length with the main graduations or slightly shorter. All other graduations shall be shorter than the main and intermediate graduations. All markings shall be permanent and legible.

AMSC N/A FSC 6640

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202640-26T08:53Z
Check the source to verify that this is the currentversion before use.

Type I centrifuge tube graduations shall have the following volumetric accuracies:

From 0	to 0.2	ml	±	0.05 ml
From 0.2	to 1	ml	±	0.1 ml
From 1	to 3	ml	±	0.15 ml
From 3	to 15	ml	±	0.20 ml

The volumetric accuracies (type I centrifuge) shall be tested in accordance with ASTM E-542 for Calibration of Volumetric Ware. Volumetric certification in the form of a certificate of quality may be furnished in lieu of individual testing, provided the lot/sample identities can be maintained. The certificate shall include the manufacturer's lot number, lot size, sample size, and test method.

Reference standard. Where standard is referenced herein, the issue of that standard in effect on the date of solicitation shall be used to determine compliance with stated requirements.

Workmanship. The centrifuge tubes shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Package (PG). One package containing twelve tubes, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:53Z
Check the source to verify that this is the current version before use.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection characteristics. Unless otherwise specified, sampling for examination shall be conducted at inspection level S-2, AQL 2.5, and sampling for tests (when applicable) shall be conducted at inspection level S-2, AQL 1.0.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics. Each centrifuge tube shall be tested to demonstrate its capability for proper operation.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:53Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

For information only. The following product, as of the date of this document, is purported by the following companies to meet the salient characteristics of the item described herein:

For type I centrifuge tube: Corning Cat. No. 8080-15 and Kimble Cat. No. 45164-15 or Kimble Cat. No. 45165-15

For type II centrifuge tube: Corning Cat. No. 8060-15 and Kimble Cat. No. 45160-15

The following National Stock Numbers are covered by this document:

For type I - 6640-00-414-2000
For type II - 6640-00-414-3000

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS

05600

Project No. 6640-1163

Location: CID-centrifuge-tube-2000

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:53Z
Check the source to verify that this is the current version before use.