COMMERCIAL ITEM DESCRIPTION

BURET, LABORATORY, (Straight)

The General Services Administration has authorized the use of this commercial item description in preference to Federal Specification NNN-B-789.

This commercial item description covers one type of precision laboratory buret.

Salient characteristics:

Shall be a straight volumetric buret suitable for use in performing precision laboratory techniques. Shall be of the following characteristics listed in TABLE I.

TABLE I

Size	Capacity (ml)	Accuracy (ml)	Graduation Intervals (ml)	Numbered Intervals
1	10	± 0.02	0.05	0.5
2	25	± 0.03	0.1	1.0
3	50	± 0.05	0.1	1.0

Shall be fabricated of a borosilicate glass conforming to Type 2, Class A of ASTM E 438. Shall be filled with a permanently attached leak-proof stopcock and glass delivery tip. The stopcock shall be fabricated of glass or TFE fluorocarbon and shall have a retaining device. The glass plug shall be in conformance with ASTM E-675 and the TFE fluorocarbon shall conform with ASTM E-911. Main graduations shall be marked in Arabic numerals directly above the referenced graduation line. Shall be marked with the calibration method at a specified temperature. Shall be in accordance with the specified tolerance as listed in Table I. Shall be tested in accordance with Method A or B of ASTM C-148. Burets shall be uniformally symmetrical about the longitudinal center line. All markings shall be permanent and legible. Each buret shall have the manufacturer's name or registered trademark.

Burets shall have none of the following which will be considered defects:

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.

Cracks or chips
Improperly marked
Cloudiness or distortion
Unfirepolished or rounded edges
Internal foreign matter not easily removed.
Grease or dirt not easily removed in normal lab procedures
Loose or unretained stopcock
Leaking buret.

Workmanship. The buret shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One buret, of the size specified, constitutes the unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspections. Inspections, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

SAmpling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more test characteristics. Unless otherwise specified, sampling for examination shall be conducted at inspection level S-II, AQL 1.0, and sampling for tests (when applicable) shall be conducted at inspection level S-II, AQL 1.0.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

The following National Stock Numbers are covered by this document:

6640-00-409-8500	BURET, LABORATORY, Straight, 10 ml
6640-00-409-9000	BURET, LABORATORY, Straight, 25 ml
6640-00-410-0000	BURET, LABORATORY, Straight, 50 ml

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTVITY:
VA-OSS

Project No. 6640-1357

Location S3-8-17

U.S GOVERNMENT PRINTING OFFICE: 1984 - 609-034/44804

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.