A-A-51252
9 December 1985

COMMERCIAL ITEM DESCRIPTION

LENS SET, TRIAL, OPHTHALMIC, Large

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers one type of large trial lens set suitable for ocular refraction. The trial lens frame is not to be furnished with the set. Unless otherwise specified in the solicitation document, the carrying case is not require for each lens set.

Salient characteristics:

The lenses shall be first quality corrected curve of ophthalmic crown glass with a refraction index of 1.5230 ± 0.0015. The lenses shall be additive vertex power type. Lenses shall be encased in anodized aluminum, 1-1/2 inch outside diameter and color coded to differentiate between plus (black) and minus (red) powers. The encased lenses shall be large aperture type with typical lens aperture diameter not less than 15/16 inch from plano to 5.5D. Others can be made to smaller apertures.

The mounting shall be engraved with dioptric (and axis cylinder) markings in such a position that the markings are visible when the lenses are in their proper position within the tray or on the trial frame.

The tray shall be made with impact resistant ABS plastic or molded styrene, properly slotted to accept individual lens and imprinted with dioptric markings.

The lens set shall be of the following grades, as specified:

Grade A - 266 lenses total
Grade B - 364 lenses total

The set consists of 266 lenses total, which shall subdivide into the following:

(36) pairs of plus spheres
+0.12 thru +0.75 in 0.12 increments
+1.00 thru +4.00 in 0.25 increments
+4.50 thru +12.50 in 0.50 increments
(7) single plus spheres
+13.00, +13.50, +14.00, +15.00, +16.00 +18.00, +20.00
(31) pairs minus spheres
-0.12 thru -0.75 in 0.12 increments
-1.00 thru -4.00 in 0.25 increments
-4.50 thru -8.00 in 0.50 increments
-9.00 thru -12.00 in 1.00 increments
(5) single minus spheres
-13.00, -14.00, -16.00, -18.00, -20.00
(23) pairs each plus and minus cylinders
0.12 thru 0.75 in 0.12 increments
1.00 thru 3.00 in 0.25 increments

FSC 6540

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil - Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.

3.50 thru 6.00 in 0.50 increments
7.00 thru 8.00 in 1.00 increments
(6) pairs prisms
0.50 thru 1.50 in 0.50 increments
2.00 thru 4.00 in 1.00 increments
(7) single prism diopters
5.00, 6.00, 8.00, 10.00, 12.00, 15.00, 20.00
(8) accessory lenses, 1 each
Maddox Rod - Red Multiple, Stenopaic Slit, Double Prism (Maddox), Red Glass lens, 1 pair ± .50 cross cyllinder, plano, blank occluder, pin hole.

The set consists of 364 lenses total, which shall subdivide into the following:

(50) pairs of plus shperes
+0.25 thru +12.50 in 0.25 increments
(7) single plus spheres
+1300, +13.50, +14.00. +15.00,
+16.00, +18.00, +20.00
(48) pairs of minus spheres
-0.25 thru -12.00 in 0.25 increments
(5) single minus spheres
-13.00, -14.00, -16.00, -18.00, -20.00
(32) pairs of plus and minus cylinders
+0.25 thru +8.00 in 0.25 crements
(6) pairs of prism diopters
0.50 thru 1.50 in 0.50 crements
2.00 thru 4.00 in 1.00 increments
(7) single prism diopters
5.00, 6.00, 8.00, 10.00,
12.00, 15.00, 20.00
(8) accessory lenses, 1 each
Maddox Rod - Red Multiple, Stenopaic Slit,
Double Prism (Maddox), Red Glass Lens, 1 pair
+.25 cross cylinder, piano, blank occluder,
pin hole.

Workmanship. The trial lens set shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Set (SE). One lens set, as specified, constitutes one unit.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6540-01-056-7377

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:
VA-OSS

Project No. 6540-1146
S3-5-2

U.S. GOVERNMENT PRINTING OFFICE: 1986 - 605-036/44548

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.