A-A-51254
17 December 1985

COMMERCIAL ITEM DESCRIPTION

GAUZE, ABSORBENT, IODOFORM IMPREGNATED

The General Services Administration has authorized the use of this commerical item description.

This Commercial Item Description covers Iodoform Impregnated Gauze.

Salient characteristics:

Shall be sterile, Iodoform impregnated, absorbent gauze packing strips.

For size 1, shall be suitable for use in oral surgery, surgery of the nose and sinuses, nasal hemostasis, etc., and as drainage wicks in the treatment of boils, abcesses, fistulas and other draining wounds. The gauze shall be approximately 1/2 inch wide by 5 yards long and shall be of plain ribbon loom weave. In addition, the gauze shall be fabricated of bleached cotton gauze, having a nominal thread count of 30 per full width in the warp direction and 56 per inch in the filling direction.

For size 2, shall be suitable for use in oral surgery of the nose and sinuses, nasal hemostasis, etc., and as drainage wicks in the treatment of boils, abcesses, fistulas and other draining wounds. The gauze shall be approximately 1 inch wide by 5 yards long and shall be of plain ribbon loom weave. In addition, the gauze shall be fabricated of bleached cotton gauze, having a nominal thread count of 52 per full width in the warp direction and 56 per inch in the filling direction.

For size 3, shall be suitable for use as hemostatic packing, in gynecological surgery and as an antiseptic dressing. The gauze shall be approximately 2 inches wide by 5 yards long and shall be of plain ribbon loom weave. In addition, the gauze shall be fabricated of bleached cotton gauze, having a nominal thread count of 96 per full width in the warp direction and 56 per inch in the filling direction.

For size 4, shall be suitable for use in oral surgery; surgery of nose and sinuses; nasal hemostatis, etc., and as drainage wicks in the treatment of boils, abcesses, fistulas and other draining wounds. The gauze shall be approximately 1/4 inch by 5 yards long and shall be of plain ribbon loom weave. In addition, the gauze shall be fabricated of bleached cotton gauze, having a nominal thread count of 22 per full width in the warp direction and 56 per inch in the filling direction.

For sizes 1, 2, 3, and 4, the gauze shall have selvage edges, be free from loose threads and ravelings and be of one continuous piece. The gauze shall comply with the following USP requirements for Absorbent Gauze:

Ignited Residue, Acid or Alkali and Dextrin or Starch in water extract except the limits of the weight of the residue shall not exceed 0.50 percent and the limits of ignited residue shall not exceed 0.15 percent.

Residue on Ignition except the limits shall not exceed 0.50 percent.

FSC 6510

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:58Z
Check the source to verify that this is the current version before use.

Fatty Matter
Alcohol Soluble Dyes
Absorbency, except a 5 gram sample shall not exceed 30 seconds when tested before sterilization in accordance with USP test method for cotton

The impregnated gauze shall be subjected to a suitable sterilization process which shall render the gauze sterile.

The distribution of the Iodoform shall be even and uniform throughout the gauze and the gauze shall contain not less than 3.7 percent Iodoform based on dry weight of the gauze when determined by a suitable, accurate and reproducible method. The Iodoform shall be of the highest pharmaceutical grade.

Not more than 6 months shall have elapsed from the date of manufacture to the date of delivery to the government.

Testing and examination shall be conducted to determine compliance with all specification requirements. Test methods shall be suitable, accurate and reproducible.

Workmanship. The impregnated gauze strips shall be free from defects which detract from their appearance or impair their serviceability.

Unit.
Jar (JR).
For sizes 1 (6510-00-559-3209) and 2, one jar containing one gauze strip as specified, constitutes one unit. Gauze strip shall be supplied pleat-folded in the suitably sealed jar capable of maintaining sterility of contents, unless damaged or opened.

Package (PG).
For sizes 1 (6510-01-153-4534), 3, and 4, each impregnated strip shall be supplied pleat-folded in a suitable sealed immediate container (bottle), capable of maintaining sterility of contents unless damaged or opened. One package containing 12 immediate containers of one size only, as specified, constitutes one unit.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

6510-00-559-3209 - Size 1 (1/2 inch wide x 5 yards long)
6510-00-559-3211 - Size 2 (1 inch wide x 5 yards long)
6510-01-003-7696 - Size 3 (2 inches wide x 5 yards long)
6510-01-003-7697 - Size 4 (1/4 inch wide x 5 yards long)
6510-01-153-4534 - Size 1 (1/2 inch wide x 5 yards long)

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITY:
VA-OSS

Project No. 6510-0724

Location H2-23-30

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.