A-A-51256
28 November 1985

COMMERCIAL ITEM DESCRIPTION

CONTROL SOLUTION SET, BLOOD GAS ANALYZER

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers ampuled liquid/gas mixtures that provide three levels of control references for blood gas instruments.

Salient characteristics:

Shall be a control set of reference ampules for monitoring blood gas instruments (e.g. Instrument Laboratory Part No. 113-S2 and Corning Model Nos 165/2, 168 and 175).

Shall provide for three (3) control levels for the following assays: pH, PCO_2, PO_2, and total CO_2.

Shall include the following:

A. Ten (10) ampules providing assay values simulating acidosis conditions in blood.

B. Ten (10) ampules providing assay values simulating normal conditions in blood.

C. Ten (10) ampules providing assay values simulating alkalosis conditions in blood.

D. A package insert listing the mean assay values and ranges next to the corresponding ampules.

Shall be hermetically sealed ampules containing acetic acid buffered with triethanolamine, sodium bicarbonate, and a dye. As an alternate, a phosphate buffer may be supplied instead of triethanolamine.

Ampules for each assay level, that are supplied under the same contract or purchase order number, shall be manufactured from a single lot for which the ampules (for each assay level) have the same manufacturer's lot or control number identification.

The minimum number of months of the expiration dating period that shall remain at time of delivery to the Government shall be specified by the ordering activity.

Product labeling, or a package insert, shall give complete instructions for use and storage, including temperature equilibration prior to test, to assure optimum precision.

Workmanship. The control set shall be free from defects which detract from its appearance or impairs its serviceability.

Unit. Package (PG). One package containing ten ampules for each of three levels (30 ampules) as specified, constitutes one unit.

DISTRIBUTION STATEMENT A. Approved for public relase; distribution is unlimited.

FSC 6550

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

For information only. The following product, as of the date of this document, is purported by the following companies to meet the salient characteristics.

General Diagnostics	Product No. 30194
Curtin Matheson Scientific	Catalog No. 072-728

The following National Stock Number is covered by this document:

6550-01-119-7680

Ordering data (intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

The ordering activity shall specify the minimum number of months of the expiration dating period that shall remain at time of delivery to the Government.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS

Project No. 6550-0029

S3-4-27

U.S. GOVERNMENT PRINTING OFFICE: 1986 - 605-036/44570

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.