A-A-51259
27 November 1985

COMMERCIAL ITEM DESCRIPTION

BURNER, GAS, LABORATORY

The General Services Administration has authorized the use of this commercial item description in preference to Federal Specification NNN-B-795.

This commercial item description covers one type of laboratory gas burner used in laboratory procedures.

Salient characteristics:

Laboratory gas burners shall be constructed of a corrosion resisting metal or a zinc base alloy. As an alternate the base may be constructed of a suitable die cast material. Laboratory gas burners shall be of the following type with performance criteria listed in TABLE I.

TABLE I

Type	Design	Fuel	BTU per hr. (minimum rating)
1	Meker	Natural Synthetic Liquified	10,000
2	Tirrill (Straight Tube)	Natural Synthetic Liquified	5,000
3	Micro	Natural only	1,000

Laboratory gas burners shall have built or auxiliary gas supply control valves and burner heads. The micro design shall have no gas supply control valve. Shall have a cylindrical or octogonal base sufficiently large enough to prevent the burner from falling over when inclined to an angle of 20° from the horizontal. Shall be serrated or taper serrated at the gas inlet port for holding 3/8 inch inner diameter rubber tubing.

Meker design shall have an adjustable sleeve type shutter to regulate air flow and a brass needle adjustment for gas regulation flow. Shall produce 300-1000 BTU per cubic foot on natural, synthetic or liquified fuels.

FSC 6640

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.

Tirill type shall have a threaded control sleeve to regulate air flow and a brass needle valve for gas regulation flow. Shall produce 300-1000 BTU per cubic foot on natural, synthetic or liquified fuels.

Micro design shall have a natural gas efficiency of 1000 BTU per cubic foot. Shall be used with natural gas only.

Laboratory gas burners shall not wobble when placed on a flat surface. Each item shall have the manufacturer's name or registered trademark. All markings shall be legible and permanent. Laboratory gas burners shall have none of the following which will be considered defects.

Unbored orifice
Wobble on a flat surface
Nonuniform appearance or construction
Missing components
Chips or cracks
Sharp edges
Cross threaded air or gas regulator components or insufficient gas control

Workmanship. The laboratory gas burner shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One burner, of the type specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspections. Inspections, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more test characteristics. Unless otherwise specified, sampling for examination shall be conducted at inspection level S-II, AQL 2.5, and sampling for tests (when applicable) shall be conducted at inspection level S-II, AQL 1.0.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

The following National Stock Numbers are covered by this document:

6640-00-410-6000	BURNER, GAS, LABORATORY, Type 1, Meker Design
6640-00-410-7000	BURNER, GAS, LABORATORY, Type 2, Tirrill Design (Straight Tube)
6640-00-411-4000	BURNER, GAS, LABORATORY, Type 3, Micro Design

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

CIVIL AGENCY COORDINATING ACTIVITY:
VA-OSS

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

Project No. 6640-1359

Location S3-8-18

U.S. GOVERNMENT PRINTING OFFICE: 1986 - 605-036/45203

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:53Z
Check the source to verify that this is the current version before use.