MIL-T-36015A
5 June 1970
SUPERSEDING
MIL-T-36015
5 May 1959

MILITARY SPECIFICATION

TABLE, BEDSIDE, METAL, FOLDING LEGS

1. SCOPE

1.1 Scope. This specification covers one type of metal bedside table with folding legs, suitable for use in field military medical treatment facilities.

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposal, form a part of the specification to the extent specified herein:

SPECIFICATIONS

Federal

TT-E-527	Enamel, Alkyd, Lusterless.
TT-P-645	Primer, Paint, Zinc-Chromate, Alkyd Type.
PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard.

Military

MIL-L-10547	Liners, Case, and Sheet, Overwrap; Water-Vaporproof or Waterproof, Flexible.

STANDARDS

Federal

FED-STD-595	Colors.

FSC 6530

Source: https://assist.dla.mil -- Downloaded: 2026-10-22T17:15Z
Check the source to verify that this is the current version before use.

Military

MIL-STD-105 Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 Marking for Shipment and Storage.

DRAWING

Defense Personnel Support Center (DPSC)

20107 Table, Bedside, Metal, Folding Legs

(Copies of specifications, standards, and drawings, required by suppliers in connection with specific procurement functions should be obtained from the procuring agency or as directed by the contracting officer).

3. REQUIREMENTS

* 3.1 Material. Material shall be of the highest quality throughout, and suitable in all respects for the purpose intended.

3.1.1 Sheet and strip. Sheet and strip shall be cold-rolled carbon steel.

3.1.2 Tubing. Tubing shall be cold-rolled, continuously butt-welded or seamless carbon steel.

3.1.3 Hardware. Spacers,rivets and catch bolt shall be cold-rolled carbon steel.

* 3.2 Construction. Construction shall be in accordance with DPSC Drawing 20107, and as hereinafter specified. The table shall be the conventinnal cross bar type in which the top is permanently hinged to one pair of legs in such a manner that it can be compactly folded. Table shall consist of a flat top, to the underside of which, shall be welded leg guides and locks, and two pair of hinged folding legs. Moving parts shall function freely and evenly and the table shall facilitate secure locking in the open position and ease of disengagement for folding when tested as specified in 4.4.1.

Source: https://assist.dla.mil -- Downloaded: 202610-22T17:15Z
Check the source to verify that this is the currenversion before use.

3.2.1 Top shall be of one-piece steel sheet, with a raised lip around all sides, corners rounded, and all edges turned down. Top shall be formed with welded corners or shall be die drawn, and the edges shall be properly dressed to remove sharp edges. Cutouts for legs shall be made in the turned-down edge, where necessary, to facilitate flat folding. Leg guide channels shall be formed of steel sheet as shown on the drawing, and shall be welded to the undersides of the table for pivoting outside legs and guiding their closure. Slotted locking angle shall be formed of steel sheet as shown on the drawing, and shall be welded opposite the pivoted legs for readily and securely engaging locking pins of inside legs when table is in the open position.

3.2.2 Legs. Legs shall have the lower ends closed by securely welding thereto, stamped, rounded, steel caps. There shall be two pairs of legs, one pair nesting beside the other to conserve space. The legs of each pair shall be braced with a stretcher, formed of steel strip, formed as shown in DPSC Drawing 20107, which shall be securely welded to each leg. The stretcher shall be formed with a center ridge and flanged edges, or otherwise formed to add rigidity. The pairs of legs shall be joined by rivets where they intersect, forming freely pivoting connections. One pair of legs shall be securely hinged to one end of the underside of the table top. The other pair of legs shall be equipped with a suitable catch bolt to conveniently and firmly engage the slot in the locking angle.

3.3 Finish.

3.3.1 Cleaning and preparation. All metal shall be thoroughly cleaned of all grease, oil, dirt, chips and other foreign material prior to priming.

3.3.2 Primer. After cleaning, the parts shall be given one coat of primer conforming to TT-P-645. The primer shall be baked on as recommended by the paint supplier and approved by the contracting officer. The primer shall be applied so as to level out and produce a uniform surface free from pebbles or wrinkles.

3.3.3 Enamel. After the primer has cooled, all surfaces shall be covered with two coats of enamel conforming to TT-E-527. Each coat shall then be baked on separately for not less than 45 minutes at a temperature of 250° F., or for an equivalent baking period at a higher or lower temperature, as recommended by the paint supplier and approved by the contracting officer. The enamel coats shall be applied so as to level out and produce a final, smooth, uniform surface without runs, blisters, wrinkles, orangepeel or grit, and without areas of thin or no film. The total dry-film thickness shall be not less than 0.003 inch at any point.

3.3.3.1 Color. Color shall be olive drab conforming with color X34087 of FED-STD-595.

Source: https://assist.dla.mil -- Downloaded: 2026-10-22T17:15Z
Check the source to verify that this is the current version before use.

* 3.4 Welding. Welding shall be thorough, non-porous, free of pits, cracks, or other imperfections. Welds shall be cleaned.

3.5 Identification marking. The underside of each table shall be permanently marked, using either a suitable black waterproof stenciling ink or black waterproof paint, with the following information. The marking shall be clear, distinct, and legible. Marking secured by adhesives shall not be used.

Federal stock number
Manufacturer's name or registered trademark
Contract or Purchase order number
U.S.

3.6 Workmanship. Workmanship shall be first class throughout. Tables shall be free from defects which detract from their appearance or may impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified, the supplier may utilize his own facilities or any commercial laboratory acceptable to the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure that supplies and services conform to prescribed requirements.

4.1.1 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual, and auditory, investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.2 Certificates of quality. When available, certificates of quality, supplied by the manufacturer of the component or material, listing the specified test method and test results obtained, may be furnished in lieu of actual performance of such testing by the contractor.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, and Table I.

Source: https://assist.dla.mil -- Downloaded: 2026-10-22T17:15Z
Check the source to verify that this is the current version before use.

Table I.

	Inspection level	AQL (defects per hundred units)
For visual examination		
Major defects	S-4	1.0
Total defects (Major and Minor combined)	S-4	4.0
For dimensional examination	S-1	2.5

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 for an Inspection level of S-1 and an AQL of 6.5.

4.3 Examination. The tables shall be examined to determine compliance with all requirements contained in this specification.

4.3.1 Classification of defects. Examination shall be conducted in accordance with the following classification of defects, as shown in Table II.

Table II.

Categories	Defects*
Major	
101	Table not complete with all component parts.
102	Table fails to bear evenly on a flat surface without rock or wobble.
103	Components not properly assembled.
104	Component parts not free of cracks, tears, and ruptures or deformed areas.
105	Welds not sound, not secure, or not free of flux deposits or splatter.
106	Fasteners not properly and firmly secured.
107	Metal parts not free of burrs and sharp edges.
Minor	
201	Identification marking not present, not complete, not legible.
202	Color does not match specified standard.
203	Painted finish not free of foreign matter, wrinkles, orangepeel, blisters, peels, chipped areas, and bare or unfinished areas.

* Examination shall not be restricted to the classified possible defects listed above.

Source: https://assist.dla.mil -- Downloaded: 2026-10-22T17:15Z
Check the source to verify that this is the current version before use.

4.3.2 Dimensional examination. The tables shall be examined for defects in dimensions. Any dimension not within the tolerances specified herein shall be classified as a defect.

4.4 Tests. Tests shall be conducted, as necessary, to determine compliance with specification requirements.

4.4.1 Performance. Tables shall be opened, set up, and closed to the folded flat position.

5. PREPARATION FOR DELIVERY

5.1 Packaging.

5.1.1 Unit of issue. One table, as specified, constitutes, one unit of issue.

5.1.2 Packaging quantities. The number of units of issue indicated in Table III shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in the procurement document.

Table III. Packaging Quantities

Unit package	Intermediate package	Exterior container
1 unit	Not Required	6 units

5.1.2.1 Packing variation permitted. If the required number of units in the entire shipment is less than the number of units specified to be overpacked in an exterior container, such units may be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

5.1.3 Level A.

5.1.3.1 Unit package. Unit package shall be a box of appropriate size constructed in accordance with PPP-B-636, type CF, class domestic. Box shall be designed with suitable interior fittings or shall contain sufficient cushioning material to insure against shifting of, or damage to, the contents. Closure shall be adequate to prevent accidental opening.

Source: https://assist.dla.mil -- Downloaded: 202610-22T17:15Z
Check the source to verify that this is the current version before use.

5.1.4 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing.

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type; PPP-B-621, class 2; or PPP-B-636, class weather-resistant, grade V3. Closure and strapping shall be as specified in the appendix of the applicable box specification.

5.2.1.2 Case liner. Each Level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, domestic type; PPP-B-621, class 1; or PPP-B-636, class domestic. Closure shall be as specified in the appendix of the applicable box specification.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Interior packages. Each unit package shall be marked in accordance with MIL-STD-129. When labels are utilized, waterproofing shall be required only when applicable carton is fabricated of water-resistant material. Contract or purchase order number and name of contractor shall be shown.

5.3.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

Source: https://assist.dla.mil -- Downloaded: 2026-10-22T17:15Z
Check the source to verify that this is the current version before use.

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

(a) Number and date of this specification.
(b) Selection of applicable levels of packing (see 5.1 and 5.2).

6.2 Availability of color chips. The individual 3 by 5 inch chips of the color referenced in 3.3.3.1 are available from General Services Administration, Business Service Center, Region 3, Seventh and D Streets, S. W., Washington 25, D. C. Price 25 cents.

* 6.3 This specification covers the following item appearing in the Federal Supply Catalog:

Item	Federal Stock Number
Table, Bedside, Metal, Folding Legs	6530-708-9060

6.4 This specification does not cover all types, classes, grades, or sizes of the commodity indicated by the title of this specification, or those which are commercially available, but is intended to cover the types which are normally procured to meet requirements.

6.5 The margins of this specification are marked with an asterisk to indicate where changes (additions, modifications, corrections, deletions) from the previous issue were made. This was done as a convenience only and the Government assumes no liability whatsoever for any inaccuracies in these notations. Bidders and contractors are cautioned to evaluate the requirements of this document based on the entire content irrespective of the marginal notations and relationship to the last previous issue.

Custodians:
Army - MD
Navy - MS
Air Force-03

Review activities:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-22T17:15Z
Check the source to verify that this is the current version before use.

SPECIFICATION ANALYSIS SHEET

Form Approved
Budget Bureau No. 22-R255

INSTRUCTIONS: This sheet is to be filled out by personnel, either Government or contractor, involved in the use of the specification in procurement of products for ultimate use by the Department of Defense. This sheet is provided for obtaining information on the use of this specification which will insure that suitable products can be procured with a minimum amount of delay and at the least cost. Comments and the return of this form will be appreciated. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments and suggestions submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or serve to amend contractual requirements .

SPECIFICATION

MIL-T-36015A FSN:

ORGANIZATION

CITY AND STATE | CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE SPECIFICATION CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY SPECIFICATION REQUIREMENT CONSIDERED TOO RIGID

3. IS THE SPECIFICATION RESTRICTIVE?

☐ YES ☐ NO *(If "yes", in what way?)*

4. REMARKS *(Attach any pertinent data which may be of use in improving this specification. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity)*

SUBMITTED BY *(Printed or typed name and activity - Optional)* | DATE

DD FORM 1 JAN 66 1426 REPLACES EDITION OF 1 OCT 64, WHICH MAY BE USED.

Source: https://assist.dla.mil -- Downloaded: 20261O-22T17:15Z
Check the source to verify that this is the current version before use.

CAUTION

NOTICE TO BIDDERS/OFFERORS

DO NOT CONDITION OR BASE YOUR BID/OFFER ON ANY CURRENT PROCUREMENT ON THE INFORMATION SUBMITTED ON THIS FORM SINCE ANY CHANGES OR DELETIONS IN THE SPECIFICATIONS MAY RENDER YOUR BID/OFFER NON-RESPONSIVE IN WHICH CASE IT CANNOT BE CONSIDERED FOR AWARD.

Fold

DEFENSE PERSONNEL SUPPORT CENTER
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

DEFENSE SUPPLY AGENCY
OFFICIAL BUSINESS

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

Fold

DPSC FORM 2418
MAR 68

EDITION AUG 66, WILL BE USED UNTIL EXHAUSTED

Source: https://assist.dla.mil -- Downloaded: 202610-22T17:15Z
Check the source to verify that this is the current version before use.