FED. SUP CLASS
6640

REVIEW ACTIVITY: ARMY - ND

PART NUMBER	DIAMETER mm	PACKAGE SIZE lb
MS36259-3	0.177-0.250	1
*MS36259-1	0.5	5
*MS36259-2	5	1

* Nominal Diameter

MATERIAL: MS36259-1 shall be of barium crown glass.

MS36259-2 and MS36259-3 shall be of borosilicate glass in accordance with Federal Specification DD-G-541, Type I, Class A.

DESIGN: Beads shall be solid spheres.

NOTES:

1. Procurement document for MS36259-1 is Military Specification MIL-B-36282.
2. Procurement document for MS36259-2 is Military Specification MIL-B-3473.
3. Items differing from this Military Standard by minor variations in design or dimensions are considered standard under this document, provided such variations do not affect performance and are not in conflict with the material requirements as specified herein.
4. For design feature purposes this standard takes precedence over procurement documents referenced herein.
5. Referenced documents shall be of the issue in effect on date of invitations for bid.

Ⓐ ENTIRE STANDARD REVISED

APPROVED 15 Dec 1961 REVISED Ⓐ 12 MAR 1969

This military standard is approved by the Department of Defense and is mandatory on all activities. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

P.A. MU	TITLE	MILITARY STANDARD
Other Cust MS	BEAD, GLASS, LABORATORY	MS 36259
Procurement Specification See Notes 1 & 2	SUPERSEDES:	SHEET 1 OF 1

DD FORM 1 SEP 68 672 (Coordinated) EDITION OF 1 SEP 67 MAY BE USED.

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:37Z
Check the source to verify that this is the current version before use.