NOTICE—When Government drawings, specifications, or other data are used for any purpose other than in connection with a definitely related Government procurement operation, the United States Government thereby incurs no responsibility, nor any obligation whatsoever; and the fact that the Government may have formulated, furnished, or in any way supplied the said drawings, specifications, or other data is not to be regarded by implication or otherwise as in any manner licensing the holder or any other person or corporation, or conveying any rights or permission to manufacture, use, or sell any patented invention that may in any way be related thereto.

This military standard is approved by the Department of Defense and is mandatory on all activities. Suggestion for all new engineering and design applications and for repetitive use shall be made from this document.

FED. SUP CLASS
6640



CANCELLED AFTER 16 NOV 1983 NO SUPERSEDING STANDARD

PART NUMBER	...H ...x ...m	D JACKET LENGTH approx mm	E OVERALL LENGTH mm
MS362...	80	200	350
...4	80	300	450

...on DD-C-541, Type I, Class A.

... joint conforming to U. S. Department of ..., taper portion only.

... joint top diameter is - 0.15 mm.

...E. ... MS36060-1 and MS36060-2 to form Distilling Apparatus which ...ed water of high purity.

... ...eature purposes, this standard takes precedence over procurement ...referenced herein.

3. ...ced documents shall be of the issue in effect on date of invitations for bid.

(A) CANCELLATION

APPROVED 15 DEC 1961 REVISED (A) 16 NOV 1983

P.A. CmlC Other Cust BuMed DAY	TITLE: CONDENSER, LABORATORY Graham type	MILITARY STANDARD MS36264	
PROCUREMENT SPECIFICATION ...-D-380	SUPERSEDES:	SHEET	OF

DD FORM 1 SEP 67 672-1 (Coordinated) PREVIOUS EDITIONS OF THIS FORM ARE OBSOLETE.

Cancellation
Proj. No. 6640-1211

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:37Z
Check the source to verify that this is the current version before use.