FED. SUP CLASS
6640

USER ACTIVITIES:
ARMY - MD
NAVY - AS, YD

REVIEWER ACTIVITIES:

This military standard is approved for use by all Departments and Agencies of the Department of Defense. Selection for all new engineering and design applications and for repetitive use shall be made from this document



PART NUMBER	DIAMETER OF TUBE ACCOMMODATED in.	NOMINAL CHIMNEY DIMENSIONS in.		
		HEIGHT	TOP OUTSIDE DIAMETER	BOTTOM OUTSIDE DIAMETER
MS36270-1	1/2	2	1-1/2	2

(B) MATERIAL: Chimney and support shall be stainless steel in accordance with Federal Specification QQ-S-766.

DESIGN: Chimney shall be conical and attached to a support bracket with a clamp or a set screw for attachment to a plain burner tube for desired height.

(B) NOTE: 1. Referenced Government documents of the issue listed in that issue of the Department of Defense Index of Specifications and Standards (DoDISS) specified in the solicitation form a part of this standard to the extent specified herein.

APPROVED 22 JUN 1962 REVISED (A) 30 JUN 1969 (B) 7 JUL 1987

P. A.	INTERNATIONAL INTEREST	TITLE	MILITARY STANDARD
Other Cust EA / MS		CHIMNEY AND SUPPORT, LABORATORY GAS BURNER	MS36270
PROCUREMENT SPECIFICATION: NONE	SUPERSEDES:		SHEET 1 OF 1

DD FORM 1 MAY 73 672-1 (COORDINATED) PREVIOUS EDITIONS OF THIS FORM ARE OBSOLETE PROJECT NUMBER 6640-1434-2

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:37Z
Check the source to verify that this is the current version before use.