FED. SUP CLASS
6640



MATERIAL: Adapter shall be made of borosilicate glass in accordance with Federal Specification, DD-G-541, type 1 class A.

DESIGN: Made in two parts. Consists of a special West Condenser (500 mm jacket length) and a special Claisen connecting tube, both with ⅗ joints. The side neck on the connecting tube has an outer ⅗ 10/30 joint at the top for a thermometer. The tubulation size on the condenser is approximately 10 mm O.D.

NOTE: Items differing from this Military Standard by minor variations in design or dimensions are considered standard under this document, provided such variations do not affect performance and are not in conflict with the material requirements specified herein.

USER ACTIVITIES: ARMY - MD, DM
NAVY - MS

REVIEWER ACTIVITIES: ARMY - ME
NAVY - OS
AIR FORCE - 82

This military standard is approved for use by all Departments and Agencies of the Department of Defense. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

APPROVED 5 JAN 1968 REVISED

P.A. EA	INTERNATIONAL INTEREST	TITLE ADAPTER TUBE, DISTILLING	MILITARY STANDARD MS36280
Other Cust MS 99			
Procurement Specifications NONE	SUPERSEDES:		SHEET 1 OF 1

DD FORM 1 MAY 73 672 (COORDINATED) PREVIOUS EDITIONS OF THIS FORM ARE OBSOLETE

PROJECT NUMBER 6640-1132

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:37Z
Check the source to verify that this is the current version before use.