MIL-W-17316A(AS)
11 July 1980
Superseding
MIL-W-17316(AER)
28 July 1953

MILITARY SPECIFICATION

WIND MEASURING SET AN/PMQ-3( )
(COMPONENTS OF)

This specification is approved for use by the Naval Air Systems Command, Department of the Navy, and is available for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers design and performance requirements for procurement of a wind measuring set.

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposal, form a part of this specification to the extent specified herein.

SPECIFICATIONS

Federal

QQ-S-571	Solder, Tin Alloy, Lead-Tin and Lead Alloy
LLL-R-626	Resin, Gum, Wood, and Tall Oil

Military

MIL-L-3891	Luminous-Material and Equipment (Non-Radioactive)
MIL-A-8625	Anodic Coatings, for Aluminum and Aluminum Alloys
MIL-D-1000	Drawings, Engineering and Associated Lists
MIL-M-10304	Meter, Electrical Indicating, Panel Type, Ruggedized, General Specification for

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Commanding Officer, Naval Air Engineering Center, ESSD, Lakehurst, NJ 08733, by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document, or by letter.

☆U.S. GOVERNMENT PRINTING OFFICE: 1981—703-023/4923 FSC 6660

Source: https://assist.dla.mil -- Downloaded: 2026-10-18T09:57Z
Check the source to verify that this is the current version before use.

SPECIFICATIONS (Continued)

Military (Continued)

MIL-T-15659	Terminal, Lug, Solder, Copper and Phosphor Bronze
MIL-T-16366	Terminal, Electrical Lug and Conductor Splices, Crimp Style

STANDARDS

Military

MIL-STD-129	Marking of Shipments
MIL-STD-130	Identification Marking of U.S. Military Property
MIL-STD-143	Specifications and Standards Order of Precedence for the Selection of
MIL-STD-794	Parts and Equipment, Procedures for Packaging and Packing of
MIL-STD-889	Dissimilar Metals

BUREAU OF AERONAUTICS MANUALS

NAVAER 50-30FR-504	Handbook of Operation and Service Instructions for Wind Measuring Set AN/PMQ-3

(Copies of specifications, standards, drawings, and publications required by contractors in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 First article. When specified (see 6.2), the contractor shall furnish sample unit(s) for first article inspection and approval (see 4.3.1 and 6.2).

3.2 Component parts. The set shall consist of the following component parts:

a. Case CY-1067( )/PMQ-3

b. Anemometer-Wind Vane ML-446( )/PMQ-3

c. Transmitter (spare) T-321( )/PMQ-3

d. Wind Vane (spare) ML-447( )/PMQ-3

Source: https://assist.dla.mil -- Downloaded: 202610-18T09:57Z
Check the source to verify that this is the current version before use.

3.3 Materials. Materials shall conform with the applicable specifications and shall be as specified herein. Materials required in the execution of this contract but not specifically provided for herein shall be as used in the standard sample except that whenever a suitable material part or process is covered by specifications or standards used by the government, these shall be selected by the contractor in accordance with MIL-STD-143 for use in lieu of materials not so covered.

3.3.1 Substitution. Other substitutions of materials deemed desirable by the contractor may be considered subject to the approval of the procuring agency. In general all such materials shall be as light weight as practicable, consistent with required strength and durability, shall be in accordance with commercial standards regularly utilized in the manufacture of high quality precision equipment, shall be wholly suitable for the intended purpose and in no case shall materials of lesser quality than those of the standard sample be used.

3.3.2 Metals. Metals shall be of the corrosion-resisting type or suitably treated to resist corrosion by salt spray, or atmospheric conditions to which the equipment may be subjected when in storage or during normal service use.

3.3.2.1 Dissimilar metals. Unless suitably protected against electrolytic corrosion, dissimilar metals shall not be used in intimate contact with each other. Dissimilar metals are defined in MIL-STD-889.

3.3.3 Fungus-proof materials. Materials which are not nutrients for fungi shall be used to the greatest extent practicable. Where materials that are nutrients for fungi must be used, such materials shall be treated with a fungicidal agent acceptable to the procuring activity.

3.3.4 Aluminum parts. All aluminum parts shall be anodized in accordance with MIL-A-8625.

3.4 Design and construction. Design and construction shall be in accordance with the standard sample except as otherwise provided under detail component requirements below.

3.5 Interchangeability. All parts having the same manufacturer's part number shall be directly and completely interchangeable with each other with respect to installation and performance. Changes in manufacturer's part numbers shall be governed by the drawing number requirements of MIL-D-1000. All components and individual parts, not specifically modified in conformance with the express provisions of this specification shall be completely interchangeable in all respects with the corresponding parts of the standard sample without any deterioration in its performance whatsoever.

Source: https://assist.dla.mil -- Downloaded: 2026-10-18T09:57Z
Check the source to verify that this is the current version before use.

3.6 Detail requirements of components.

3.6.1 Case CY-1067( )/PMQ-3. Any features of the standard sample case contrary to the following shall be appropriately modified.

3.6.1.1 The case shall be so designed and constructed as to provide adequate protection to the instrument when tested in accordance with the shock test of paragraph 4.3.2.1.

3.6.1.2 All materials and structures utilized as retaining and/or shock absorbing supports for the instrument in the case shall be so selected, designed and installed as to assure long lasting adequate protection throughout the service life of the equipment.

3.6.1.3 Handles and latches shall be rugged and durable, capable of withstanding moderately rough handling without failure or significant impairment of serviceability.

3.6.1.4 The case shall incorporate appropriate means for retaining the desiccant required by paragraph 5.2. The method used shall entirely prevent any contact of the desiccant with the equipment or any shifting of the desiccant within the case and shall be satisfactory to the procuring agency.

3.6.2 Anemometer-Wind Vane ML-446( )/PMQ-3). The anemometer-wind vane shall consist of one complete operating unit of equipment including the following components and parts.

3.6.3 Transmitter T-321( )/PMQ-3. Construction and assembly shall be such as to assure maximum durability and sensitivity consistent with the basic turbine design. A shield or baffle shall be employed between the turbine and the generator to prevent entry of rain or snow into the generator.

3.6.3.1 The rotor shall turn freely on jewel bearings. Pivots and bearings shall neither bind nor shake and shall be as near frictionless as possible.

3.6.3.2 The rotor-generator unit shall generate a voltage substantially proportional to the wind velocity and such that 4.46 $\pm$.06 volts is produced at a true wind speed of 40 knots.

3.6.3.3 Average starting speed shall not exceed 2.0 knots and average stopping speed shall not exceed 1.5 knots when tested in accordance with paragraph 4.3.2.3.2.

3.6.4 Wind Vane ML-447( )/PMQ-3. The vane shall be carefully balanced and shall rotate freely on stainless steel ball bearings when the brake is released.

Source: https://assist.dla.mil -- Downloaded: 2026-10-18T09:57Z
Check the source to verify that this is the current version before use.

3.6.5 Wind speed meter. The meter shall be physically and electrically interchangeable with the standard sample. It shall have the same range and at least equal accuracy. It shall conform to all requirements of MIL-M-10304(SigC), not conflicting with the foregoing. It shall be mounted so as to be readily replaceable by field personnel.

3.6.5.1 Luminous pointer and dial. Luminous material shall conform to Navy Department Specification 17L18(INT) Type P, Class H1B material, color after irradiation yellow, size and shape of numerals such as to be readable at 20 inches.

3.6.6 Support, handle and housing. The housing shall be so constructed and assembled as to assure the maximum possible, long lasting protection of interior electrical and mechanical parts from the deleterious effects of salt spray and exposure to marine atmosphere. It shall be as light weight as practicable consistent with required strength, rigidity and durability.

3.6.6.1 The vane locking and release mechanism shall be smoothly operating and positively acting such that the accuracy of the indications are in no way affected and readings may be retained as long as desired during moderately careful handling of the equipment incident to making observations aboard ships.

3.6.7 Miscellaneous parts and processes.

3.6.7.1 Finishes. Finishes and finishing materials which will crack, chip, lift, or peel under normal use and aging or when subjected to changes of temperature and humidity incident to storage, transportation or use shall not be employed.

3.6.7.2 Wear surfaces of moving parts which cannot be protected with finishes shall be made of materials not appreciably subject to surface deterioration and shall be given an adequate degree of hard polish whenever improved wearing qualities would result.

3.6.7.3 Screws. All screws, nuts, and other similar parts used in the assembly of the instrument shall conform to the Handbook H28, and shall be class 2 fit unless otherwise specified herein. In general screws in hard material shall be fine thread series. All screws, nuts, bolts and washers shall be protected against corrosion.

3.6.7.4 Set screws. Set screws shall be of the hardened cup-pointed type.

3.6.7.5 Wire terminals. Solderless wire terminals in accordance with MIL-E-16366 are preferred, however, solder-type terminals in accordance with MIL-T-15659 may be used.

Source: https://assist.dla.mil -- Downloaded: 202640-18T09:57Z
Check the source to verify that this is the currentversion before use.

3.6.7.6 Soldering. Solder shall be in accordance with QQ-S-571. All soldering shall be effected in a neat and workmanlike manner and the surfaces surrounding the soldered connections shall be free from splattering of solder and rosin residues. All soldered connections shall be of such character and quality that the bonding between the soldered items may be determined by visual means. There shall be no evidence of "cold soldering", and the use of excessive amounts of solder will not be acceptable.

3.6.7.6.1 Strength of solder joint. In general, soldering alone shall not be depended upon for a satisfactory connection. At the point when wires are terminated by soldering, the wires shall be wrapped around (360° minimum) or otherwise mechanically secured to the terminals prior to soldering.

3.6.7.6.2 Soldering flux. Only pure rosin, LLL-R-626, or a mixture of alcohol and rosin shall be used as a flux.

3.6.8 Running spare parts. Running spare parts shall consist of one each of the components: Transmitter T-321( )/PMQ-3 and Wind Vane ML-447( )/PMQ-3. These items shall be mounted in a readily accessible and adequately protected manner within Case CY-1067( )/PMQ-3 and are in addition to those incorporated in the complete anemometer-wind vane ML-446( )/PMQ-3 unit.

3.7 Performance. The equipment shall meet the performance test requirements of Section 4 and shall be capable of continuing normal, satisfactory operation; and of storage or transportation without damage or deterioration of subsequent performance, when subjected to any possible combination of the service conditions listed below.

	Condition	Equipment Operating	Storage or in Transit
(a)	Wind Velocity at Full Scale	x	
(b)	Temperatures in the Range	-20°F to $\pm 120^{\circ}$F	-50°F to $\pm 150^{\circ}$F
(c)	Relative Humidity 0-100% Including condensation due to temperature change	x	x
(d)	Exposure to Marine atmosphere aboard ship	x	x
(e)	Fungi and Insects as encountered in Tropics	x	x
(f)	Wind driven precipitation	x	x
(g)	Vibration aboard ship	x	x
(h)	Vibration in aircraft		x

Source: https://assist.dla.mil -- Downloaded: 2026-10-18T09:57Z
Check the source to verify that this is the current version before use.

3.8 Special tools. Any special tools required for replacement of field replaceable parts as defined in Bureau of Aeronautics Manual NAVAER 50-30FR-504, shall be provided within the carrying case mounted in shake proof fasteners. Requirements for special tools shall be kept to a minimum.

3.9 Markings. Dial markings and luminous sight shall be in accordance with the Standard Sample.

3.10 Nameplates. Nameplates shall conform to MIL-STD-130 and be similar to the Standard Sample in any non-conflicting features. The manufacturer shall apply to the procuring agency for the exact equipment nomenclature, stock numbers, and any other nameplate data required for a particular procurement which is not otherwise available to the manufacturer.

3.11 Workmanship. Workmanship shall be first class in all respects and in accordance with the best commercial practices for the manufacture of precision instruments. Dimensional tolerances shall be determined by the contractor as necessary to assure interchangeability and precision fit of components and submitted to the procuring agency for approval. Approved tolerances shall be rigidly maintained for all equipment produced under these specifications.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.2 Classification of inspection. The inspection of the indicator shall be classified as follows:

a. First article inspection (4.3)

b. Quality conformance inspection (4.4)

4.3 First article inspection. First article inspection shall consist of all the examinations and tests of this specification.

4.3.1 First article samples. Unless otherwise specified, as soon as practicable after award of a contract or order, the manufacturer shall submit three Wind Measuring Sets of each part number for first article testing. The samples shall be representative of the construction,

Source: https://assist.dla.mil -- Downloaded: 202610-18T09:57Z
Check the source to verify that this is the current version before use.

workmanship, components, and materials to be used during production. When a manufacturer is in continuous production of these indicators from contract to contract, submission of further first article samples on the new contract may be waived at the discretion of the procuring activity (see 6.2). Approval of the first article samples or the waiving of the first article inspection does not preclude the requirements of submitting to the quality conformance inspection. The first article samples shall be furnished to the Government as directed by the contracting officer (see 6.2). Each sample size shall be plainly identified by a securely attached durable tag marked with the following information:

Sample for first article
Submitted by (name) (date) for
tests in accordance with the
requirements of specification
MIL-W-17316 under authorization
(reference letter authorizing
test)
Wind Measuring Set AN/PMQ-3( )
Manufacturer's Part No.
MS Part No.
Name of manufacturer

4.3.1.1 Upon completion of the first article inspection, all the applicable inspection reports and when applicable, recommendations and comments pertinent for use in monitoring production will be forwarded to the Government quality assurance representative. The samples may be consumed or destroyed in the first article inspection and shall not be considered as part of the quantity to be delivered under contract (see 6.2).

4.4 Quality conformance inspection. The quality conformance inspection of the indicator shall consist of the individual tests and the sampling plans of this specification.

4.5 First article tests.

4.5.1 Shock. The first article sample, packaged in accordance with paragraph 5.2, shall be tumbled end over side over end for four complete revolutions on a concrete surface. The box shall then be rotated 90° about its vertical axis and tumbled side over side over side for four complete revolutions. The same packaged equipment shall then be dropped on each of its eight corners from a height of 30 inches. The box shall then be opened and the equipment examined. There shall be no evidence of damage to the equipment as a result of this test. All other tests of the first article equipment shall be made subsequent to this test.

Source: https://assist.dla.mil -- Downloaded: 202640-18T09:57Z
Check the source to verify that this is the currentversion before use.

4.5.2 Temperature test. The anemometer rotor and vane shall be checked at -20°F and at $+120^{\circ}$F to determine that moving parts turn freely without binding at these temperatures.

4.5.3 Wind tunnel tests.

4.5.3.1 Speed calibration. Wind speed calibration shall be checked over the entire range on both high and low speed scales and for both ascending and descending valves at intervals not to exceed 5 knots on the high range and 2 knots on the low range. Accuracy on both high and low ranges shall be + 1 knot up to 10 knots, + 1.5 knots between 11 and 40 knots, + 2 knots 40 to 55 knots and + 3 knots above 55.

4.5.3.2 Starting and stopping speed. Three determinations shall be made of starting and stopping speed of the rotor. The average starting speed shall not exceed 2 knots and no valve shall exceed 2.3 knots. The average stopping speed shall not exceed 1.5 knots, and no valve shall exceed 1.7 knots.

4.5.3.3 Vane response. The vane shall align itself to within + 2 degrees of the wind stream when displaced 5 degrees there from at a velocity not to exceed 5 knots. The vane shall exhibit no tendency to oscillate at any wind speed from 0 to 60 knots.

4.5.3.4 Inclination. The lengthwise axis of the instrument shall be inclined at 5° and 15° respectively both to the right and to the left of vertical while maintained in a plane at right angles to the wind stream and the calibration checked at 3 substantially equally spaced points on both the high and the low speed scales. Accuracy shall be as specified in paragraph 4.3.2.3.1 and the vane shall re-align itself parallel with the airstream to within $\pm$ 2 degrees when displaced 5° therefrom.

4.5.4 Approval. Final approval of the first article sample shall be granted by the procuring agency and shall be accomplished prior to the manufacture of additional items of the contract. Approval of the first article sample shall not be construed as authorizing deviations from this specification except as may be authorized by the procuring officer.

4.5.4.1 Correction of defects. Any defects or unapproved deviations from this specification, revealed by the first article tests shall be fully corrected by the contractor to the satisfaction of the procuring agency, without further charge to the government and such corrections shall be incorporated in all units of the equipment delivered under the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-18T09:57Z
Check the source to verify that this is the current version before use.

4.5.4.2 Retest. When deemed necessary by the procuring agency, resubmission and retest of the first article model may be required.

4.5.4.3 Disposition of model. The finally approved model shall be returned via the cognizant inspector. All changes and corrections required as a result of first article tests shall be made and it shall then be retained for use by the inspector as a standard of design, construction, performance and workmanship during the course of the contract. It shall be currently modified in accordance with any subsequently approved changes and may then be delivered as the final item on the contract provided that its appearance, performance and serviceability has in no way been impaired.

4.6 Quality conformance inspections. The quality conformance tests shall consist of the following tests and are applicable to all units on order under the contract.

4.6.1 Unless otherwise specified, contractor's records of all inspection work and tests, giving the results of tests required to determine compliance with the requirements and tests specified herein, shall be kept complete and shall be available to the Government representative at all times. The record or report of inspection and tests shall be signed or approved by a responsible person specifically assigned by the contractor. Contractors not having laboratory testing facilities satisfactory to the Government shall engage the services of a commercial testing laboratory capable of conducting tests to determine compliance with all the requirements and inspection tests in the specification, and acceptable to the Government.

4.6.2 The following tests shall be conducted in accordance with detailed test procedures to be established by the contractor and approved by the cognizant inspector and the procuring agency. Similarly *approved test equipment entirely suitable and appropriate for the purpose* shall be used.

4.6.2.1 General visual mechanical inspection. Each of the components or assemblies listed below shall be given an individual inspection for physical dimensions, workmanship, finish, markings, soldering, absence or sticking or binding of moving parts, etc. as applicable to determine compliance with the materials, design and construction requirements of this specification, in addition to the specific tests noted thereafter.

4.6.2.2 Meter and resistor assembly. Check meter for balance, the pointer shall not deviate more than 0.026" from its rest point, that is, with instrument face in vertical upright position, when placed in any other inverted or rotated position.

Source: https://assist.dla.mil -- Downloaded: 2026-10-18T09:57Z
Check the source to verify that this is the current version before use.

4.6.2.2.1 Meter roll and friction. By applying proper voltages, accelerate and decelerate pointer movement slowly through the scale. There shall be no sticking or bouncy meter movement. At full scale tap meter; point shall not shift more than 0.013" in either direction; return meter pointer to zero and repeat tapping test.

4.6.2.3 Vane assembly. Visual mechanical inspection shall be given.

4.6.2.4 Trigger assembly. Check resistance from contact switch to ground. Resistance shall not be less than 0.2 megohms.

4.6.2.5 Housing assembly. (Including Meter and Vane) Finish shall not be burred or scratched in any way and luminescent coating on pointers, sight, and graduations shall be uniform and even, free from breaks and other defects.

4.6.2.5.1 Vane balance. With brake released, vane shall rotate freely about its axis, when the instrument is held in any position from vertical to horizontal, and vane shall hold any given azimuth regardless of the inclination of the axis from the vertical.

4.6.2.5.1.1 Brake clearance. Check brake clearance with feeler, clearance shall be in accordance with approved drawings.

4.6.2.5.1.2 Direction dial orientation. When the vane centerline is parallel to the sight centerline, the dial index shall read $0^{o} \pm 1^{o}$ when the vane nose points away from the eyepiece of the sight and $180^{o} \pm 1^{o}$ when pointing toward the eyepiece.

4.6.2.5.2 Meter calibration. With test equipment capable of measuring and controlling output voltages to within ± 0.010 volts the meter calibration shall be checked at the following values:

	Applied Voltage	Meter Reading
Low Scale	1.27V	12 knots $\pm$ 0.25
	.69V	7 knots $\pm$ 0.5
High Scale	4.46V	40 knots $\pm$ 0.5
	2.18V	20 knots $\pm$ 1.1

4.6.2.5.2.1 Meter overshoot. With voltage applied to read approximately 45 knots interrupt voltage intermittently and overshoot of pointer shall not extend beyond 55 knots.

4.6.2.5.3 Insulation test. Resistance from either phone jack connector to metal frame of housing shall not be less than 0.2 megohms.

Source: https://assist.dla.mil -- Downloaded: 2026-10-18T09:57Z
Check the source to verify that this is the current version before use.

4.6.2.5.4 Operational check. With transmitter plugged in place, check low and high scale ranges respectively for meter operation when rotor is turning.

4.6.2.6 Transmitter assembly. With transmitter mounted on a rotating arm or other equivalent device for achieving an accurately measurable low velocity air flow past the rotor, the rotor shall start to turn when air flow is increased from zero to 2 knots and rotor shall continue to rotate for at least 1.5 minutes as air flow is reduced to and maintained at 1.5 knots.

4.6.2.6.1 Clearances. Check clearance between lower edge of the rain shield and the base plate of the cage. Clearance shall be in accordance with the approved drawings.

4.6.2.6.2 Calibration. Calibration of the rotor-generator shall be checked at 120 cycles and shall produce a voltage output of 4.5V $\pm$ .555%.

4.6.2.6.3 Insulation test. Resistance from the plug tip to ground shall not be less than 0.2 megohms.

4.6.2.7 Carrying case. A visual inspection shall be given.

4.6.3 Rejection and retest. Items which have been rejected may be reworked or replaced to correct the defects, and resubmitted for acceptance. Before resubmitting, full particulars concerning previous rejection and the action taken to correct the defects found in the original shall be furnished the inspector. Units rejected after retest shall not be resubmitted without the specific approval of the procuring activity.

5. PACKAGING

5.1 Application. The packaging, packing and marking requirements specified herein apply only to direct purchases by and direct shipments to the Government.

5.2 Packaging. The wind measuring set in its carrying case shall be packaged in accordance with Group VII, Type I of MIL-STD-794 except that the desiccant shall be placed inside the carrying case.

5.3 Packing. The interior containers shall be packed for domestic or overseas shipment in accordance with MIL-STD-794 as required by the Contract or Order.

5.4 Marking. All marking of interior and exterior containers shall conform to MIL-STD-129. The marking shall include the date of manufacture of the instrument in addition to date of preservation.

Source: https://assist.dla.mil -- Downloaded: 202640-18T09:57Z
Check the source to verify that this is the currentversion before use.

6. NOTES

6.1 Intended use. This equipment is intended for use in determining wind speed and direction in meteorological applications requiring light weight and portability.

6.2 Ordering data. Procurement documents shall specify the following.

a. Title, number and date of this specification.

b. Quantity desired.

c. Whether first article inspection is required and where the samples should be delivered (see 4.3.1 and 4.3.1.1).

d. Whether first article inspection is waived (see 4.3.1).

e. Whether any special markings are required (see 5.4).

6.3 First article sample. Contracts and orders should specify that the submission and final approval of the first article sample is required prior to the manufacture of additional units on order.

6.4 First article test. Initial first article tests will generally be conducted at a government testing facility at no expense to the contractor. Upon application therefore and approval by the procuring activity, appropriate representatives of the contractor may be present during the first article testing. In the event that repeated first article testing is required to effect conformance with this specification, appropriate charges therefore may be made to the contractor.

6.5 When a first article is required, it shall be tested under the appropriate provisions of 7-104.55 of the Defense Acquisition Regulations. The first article shall be a sample selected from the initial procurement items. The contracting officer should include specific instructions in all procurement instruments, regarding arrangements for examinations, test and approval of the first article.

6.6 Drawing requirements. The attention of the contractors is invited to the fact that Government contracts for equipment, whether procured directly by the Government or through another contractor, require that all drawings submitted, must be in accordance with the requirements of MIL-D-1000.

Custodians:
Navy - AS

Preparing Activity
Navy - AS
(Project No. 6660-N052)

Source: https://assist.dla.mil -- Downloaded: 2026-10-18T09:57Z
Check the source to verify that this is the current version before use.

FOLD

POSTAGE AND FEES PAID
DEPARTMENT OF THE NAVY
DoD-316

U.S.MAIL

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

NAVAL AIR ENGINEERING CENTER
ESSD (CODE 9312)
LAKEHURST, NJ 08733

FOLD

Source: https://assist.dla.mil -- Downloaded: 202610-18T09:57Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS: This form is provided to solicit beneficial comments which may improve this document and enhance its use. DoD contractors, government activities, manufacturers, vendors, or other prospective users of the document are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity. A response will be provided to the submitter, when name and address is provided, within 30 days indicating that the 1426 was received and when any appropriate action on it will be completed.

NOTE: This form shall not be used to submit requests for waivers, deviations or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

DOCUMENT IDENTIFIER *(Number)* AND TITLE

MIL-W-17316A

NAME OF ORGANIZATION AND ADDRESS OF SUBMITTER

☐ VENDOR ☐ USER ☐ MANUFACTURER

1. ☐ HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE? ☐ IS ANY PART OF IT TOO RIGID, RESTRICTIVE, LOOSE OR AMBIGUOUS? PLEASE EXPLAIN BELOW.

A. GIVE PARAGRAPH NUMBER AND WORDING

B. RECOMMENDED WORDING CHANGE

C. REASON FOR RECOMMENDED CHANGE(S)

2. REMARKS

SUBMITTED BY *(Printed or typed name and address — Optional)*

TELEPHONE NO.

DATE

DD FORM 1426 1 OCT 76

Replaces edition of 1 Jan 72 which may be used.

S/N 0102-LF-001-4260

Source: https://assist.dla.mil -- Downloaded: 202640-18T09:57Z
Check the source to verify that this is the current version before use.