
FED. SUP CLASS
6640



MATERIAL: Rack shall be molded of white high-impact polystyrene plastic.

DESIGN: The rack shall be 216 x 149 x 19 mm high. It shall have a transverse recess to provide space for bulbs on blood pipets and for making gripping easy. It shall also have grooves for keeping the pipets separated, protected and visible. The rack shall hold a minimum of 12 pipets.

NOTE: 1. Items differing from this Military Standard by minor variations in design or dimensions are considered standard under this document, provided such variations do not affect performance and are not in conflict with the material requirements specified herein.

USER ACTIVITIES: ARMY - MD, DM
NAVY - MS

REVIEWER ACTIVITIES: ARMY - ME
NAVY - OS
AIR FORCE - 03

This military standard is approved for use by all Departments and Agencies of the Department of Defense. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

APPROVED 26 JUL 1982 REVISED

P.A. EA	INTERNATIONAL INTEREST	TITLE	MILITARY STANDARD
Other Cust MS 99		RACK, PIPET	MS36392
Procurement Specifications NONE	SUPERSEDES:		SHEET 1 OF 1

DD FORM 1 MAY 73 672 (COORDINATED) PREVIOUS EDITIONS OF THIS FORM ARE OBSOLETE

PROJECT NUMBER 6640-0797
STANDARD DATE

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:39Z
Check for updates to verify that this is the current version before use.