
FED. SUP CLASS
6640



MATERIAL: The funnel holder shall be made of linear polyethylene or polypropylene plastic.

DESIGN: MS36406-1 shall be a single holder. MS36406-2 shall be a double holder. The holders shall have a friction connection for clamping to a laboratory stand rod from 8 mm to 14 mm diameter. The holder shall have a removable inner ring that will support 25 mm to 76 mm funnels and an outer ring that will support 76 mm to 152 mm funnels.

NOTE: 1. Items differing from this Military Standard by minor variations in design or dimensions are considered standard under this document, provided such variations do not affect performance and are not in conflict with the material requirements specified herein.

USER ACTIVITIES: ARMY - MD, OM
NAVY - MS

REVIEWER ACTIVITIES: ARMY - ME
NAVY - OS
AIR FORCE - 03

APPROVED 26 JUL 1982 REVISED

This military standard is approved for use by all Departments and Agencies of the Department of Defense. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

P.A. EA	INTERNATIONAL INTEREST	TITLE	MILITARY STANDARD
Other Cust MS 99		HOLDER, FUNNEL, LABORATORY	MS36406
Procurement Specifications NONE	SUPERSEDES:		SHEET 1 OF 1

DD FORM 672 (COORDINATED) PREVIOUS EDITIONS OF THIS FORM ARE OBSOLETE. PROJECT NUMBER 6640-0797

Source: https://assist.dla.mil -- Downloaded: 2026-10-28T17:40Z
Check this source to verify that this is the current version before use.