MIL-O-16485C
AMENDMENT-4
14 JUNE 1966
SUPERSEDING[1]
MIL-O-16485C
AMENDMENT-3
18 FEBRUARY 1966

MILITARY SPECIFICATION

OHMMETERS, INSULATION RESISTANCE-INDICATING, PORTABLE

This amendment forms a part of Military Specification MIL-O-16485C, dated 21 March 1962, and is mandatory for use by all Departments and agencies of the Department of Defense.

Page 1, paragraph 2.1, Under Specifications, Military: Delete reference to MIL-M-16616 and add the following:

"MIL-P-80 — Plastic Sheet, Acrylic Base, Antielectrostatic, Transparent (For Indicating - Instrument Windows).

MIL-G-3787 — Glass, Laminated, Flat (Except Aircraft).

MIL-M-15071 — Manuals, Equipment and Systems."

Page 2, paragraph 3.1: Delete and substitute:

"**3.1 Preproduction sample.** Prior to production, a sample equipment shall be tested as specified in 4.3."

Page 2, paragraph 3.4.3: Delete and substitute:

"**3.4.3** *Windows.* Meter windows shall be of flat laminated safety glass in accordance with class II of MIL-G-3787 or of transparent plastic material not less than 3/32 inch thick, conforming to MIL-P-80, except that the anti-electrostatic compound shall be applied to both sides of the window. The windows shall be secured by clamps or other suitable means. Cement alone is not considered sufficient to hold the window in place."

Page 3, paragraph 3.4.8, first sentence: Delete and substitute:

"Generators shall be smooth running and shall not require starting and running torques to exceed 20 pound inches. The torque may be calculated by positioning the crank parallel to the floor and suspending a known weight from the handle of the

[1] **CHANGES FROM PREVIOUS ISSUE.** The outside margins of this document have been marked "#" to indicate where changes (deletions, additions, etc.) from the previous issue have been made. This has been done as a convenience only and the government assumes no liability whatsoever for any inaccuracies in these notations. Bidders and contractors are cautioned to evaluate the requirements of this document based on the entire content as written irrespective of the marginal notations and relationship to the last previous issue.

FSC 6625

Source: https://assist.dla.mil -- Downloaded: 2026-10-18T09:10Z
Check the source to verify that this is the current version before use.

crank. Knowing the crank distance (L), weight (F) and angle (Ø), the torque (T) is then T = FL sin Ø."

Page 3, paragraph 3.4.10: Add: The insulation shall be color coded; a red sleeve for the positive lead and a black sleeve for the negative lead."

Page 4, paragraph 3.7.6: Delete.

Page 4, paragraph 3.8, second sentence; Delete and substitute "When the ohmmeter is tested as indicated on the instruction plate (see 4.6.1.1 and 4.6.1.2), the voltage at the connection terminals shall be not less than 400 volts for the high resistance test or 300 volts for the low resistance test, nor more than 600 volts, d.c."

Page 4, paragraph 3.10, line 2: Delete "1.5 percent" and substitute "2 percent".

Page 4: Add new paragraph 3.17:

"**3.17 Vibration.** The ohmmeter shall be capable of withstanding the vibration test specified in 4.6.10."

Page 4: Add new paragraph 3.18:

"**3.18 Manuals.** Manuals shall be furnished in accordance with MIL–M–15071."

Page 4, paragraph 4.1: Delete and substitute:

"**4.1 Responsibility for inspection.** Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified, the supplier may utilize his own facilities or any commercial laboratory acceptable to the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements."

Page 5, paragraph 4.3: Delete and substitute:

"**4.3 Preproduction inspection.** Preproduction inspection shall be performed on the preproduction sample in accordance with 4.4 and 4.6. Any other tests deemed necessary to determine conformance with this specification shall also be performed."

Page 5, paragraphs 4.6.1.1 and 4.6.1.2: Line 6, after "microammeter" add "or in parallel with an electrostatic voltmeter approved by the bureau or agency concerned"; line 9, after "resistance" add "or the voltage as indicated by the electrostatic voltmeter".

Page 8: Add new paragraph 4.6.10:

"**4.6.10** *Vibration.* The ohmmeter shall be securely mounted on a vibration machine. It shall be subjected to a vibration of simple harmonic motion with an amplitude of 1/32 inch (1/16-inch total excursion) at a frequency varied uniformly between the limits of 10 to 35 cycles per second (c.p.s.). The frequency range shall be traversed from 10 to 35 c.p.s. and shall return in approximately 1 minute. Unless otherwise specified (see 6.2), this vibration shall be applied for a period of 3 hours in each of the three axes. Upon completion of the test the ohmmeter shall be examined for damage and tested for accuracy."

Page 8, paragraph 6.2, item (e): Delete and substitute:

"(e) Whether the duration of the vibration test should be other than specified (see 4.6.10).

(f) Selection of level of packaging and level of packing (see 5.1)."

Source: https://assist.dla.mil -- Downloaded: 2026-10-18T09:10Z
Check the source to verify that this is the current version before use.

Page 8: Add new paragraphs 6.4 and 6.4.1:

"**6.4 Preproduction.** Invitations for bids should provide that the Government reserves the right to waive the requirement for preproduction samples as to those bidders offering a product which has previously been procured or tested by the Government.

"**6.4.1** When a manufacturer can present proof of a previous approval of identical ohmmeters, preproduction inspection may be waived by the procuring activity for the next 2500 ohmmeters purchased or a period of 5 years, whichever occurs first. After this period the manufacturer will be required to submit a preproduction sample."

Custodians:

Army—EL

Navy—SH

Air Force—11

Review activities:

Army—EL, MI, MU

Navy—SH

Air Force—67

User activities:

Army—MO

Navy—MC

Preparing activity:

Navy—SH

(Project 6625-0434)

Source: https://assist.dla.mil -- Downloaded: 2026-10-18T09:10Z
Check the source to verify that this is the current version before use.