INCH-POUND
A-A-59131D
20 JULY 2020
SUPERSEDING
A-A-59131C
1 JUNE 2019

COMMERCIAL ITEM DESCRIPTION

DETECTOR METAL FLAW, EDDY CURRENT, IMPEDANCE PLANE

The General Services Administration has authorized the use of this Commercial Item Description, for all federal agencies.

1. SCOPE.

This commercial item description (CID) defines the minimum performance and technical requirements for a lightweight microprocessor-based eddy current flaw detection instrument suitable for Nondestructive Inspection (NDI) use in a military industrial maintenance environment, which may be at fixed military bases, deployed to austere forward operating locations, or naval vessels afloat.

2. CLASSIFICATION.

There are six types of metal flaw detectors.

Type I	Basic Kit with case, charger, and accessories.
Type II	Intermediate Kit with case, charger, and accessories, with fastener hole scanner.
Type III	Intermediate Kit with case, charger, and accessories, with fastener hole scanner, aluminum fastener hole standard, and probes.
Type IV	Intermediate Kit with case, charger, and accessories, with fastener hole scanner, steel fastener hole standard, and probes.
Type V	Intermediate Kit with case, charger, and accessories, with fastener hole scanner, aluminum, steel, and titanium fastener hole standards, and probes.
Type VI	Intermediate Kit with case, charger, and accessories, with aluminum, steel, and titanium fastener hole standards.

Beneficial comments, recommendations, additions, deletions, clarifications, etc. and any other data that may improve this document should be addressed to AFLCMC/EZSS, Bldg. 28, 2145 Monahan Way, Wright-Patterson AFB, OH 45433-7017 or emailed to ENGINEERING.STANDARDS@US.AF.MIL. Since contact information can change, you may want to verify the currency of this address information using the ASSIST Online database at https://assist.dla.mil.

AMSC N/A FSC 6635

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T10:04Z
Check the source to verify that this is the current version before use.

3. SALIENT CHARACTERISTICS.

3.1 General.
The instrument shall have a microprocessor-based, impedance plane display and have dual frequency with frequency mixing capability.

3.2 Foreign object damage.
The instrument shall be designed to prevent loss of components, which could generate foreign object damage to weapon systems or engines.

3.3 Size and weight.
The instrument size shall be a maximum 12 inches high, 12 inches wide, 6.5 inches deep and the weight shall not exceed 8 pounds with batteries installed (not including scanners).

3.3.1 Instrument support.
The instrument shall be equipped with high impact supports (e.g., feet, legs, pads, rails, etc.) on the bottom surface normally in direct contact with the supporting surface. The supports shall be fabricated from a material that will not scratch or mar aircraft surfaces, yet will provide sufficient friction to prevent the instrument from slipping off a bare aluminum, as rolled, surface tilted 20 degrees from horizontal. The supports shall be securely fastened to the test set to preclude loosening or detachment during field operations. An adjustable support shall be incorporated to allow tilting adjustment of the instrument to a least 45 degrees from vertical.

3.4 Power requirements.
The instrument shall operate when supplied with either 100-250 Volt Alternating Current – 50 or 60 Hertz (100-250 VAC – 50/60 Hz) automatically selected and connected directly, through means of a battery charger connected to 100-250VAC-50/60Hz, or batteries/battery pack(s) as described below.

3.4.1 Battery charger.
The instrument shall be supplied with a separate battery charger(s) that is/are able to simultaneously charge two batteries/battery pack(s) and provide power to operate the instrument and charge installed batteries/battery pack.

3.4.2 Battery duration.
The cells in the rechargeable battery pack(s) shall not have a charge “memory” limitation that causes the instrument to operate for less than the specified time, and provide at least 6 hours of continuous use without scanner or display backlighting, and at least 3 hours of continuous use with both scanner and backlighting at 68 °F. The instrument must incorporate a visible means of indicating the charge condition of the battery pack(s) and shall provide a low battery warning.

3.4.3 Battery quantity.
A minimum of two complete sets of batteries/battery pack(s) shall be provided.

3.4.4 Battery charging.
The batteries shall be easily rechargeable without tools and require no longer than 4 hours to charge from 50% to 100% capacity, nor more than 8 hours for a complete charge.

3.4.5 Battery replacement.
The instrument shall be capable of continuous 24-hour per day - 7 day per week operation, by swapping rechargeable battery packs by the user without tools and within 90 seconds. If designed for proprietary battery packs, the government shall acquire the design specifications and a limited license for manufacture

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T10:04Z
Check the source to verify that this is the current version before use.

after contract expiration.

3.4.6 Battery safety.
The instrument shall incorporate Underwriter Laboratories (UL) certified batteries.

3.4.6.1 Air transport.
The extra sets of batteries/battery pack shall have a protective case to permit airline transport (see http://www.faa.gov/ for additional information).

3.4.6.2 Lithium battery.
Any lithium batteries used shall be in accordance with NAVSEA TM S9310-AQ-SAF-010. Detailed instructions for the approval process including testing requirements are contained in this technical manual. In some cases, the testing requirements can be waived if particular conditions are met, per NAVSEA TM S9310-AQ-SAF-010.

3.5 Controls.
The instrument function controls shall be menu driven, not to exceed two sub-menu levels on the instrument. As a minimum, the instrument shall accommodate control of the following functions.

3.5.1 Frequency.
The operating frequency shall be selectable from 60 Hz to 6 MHz (minimum range), with a precision of two significant digits (i.e., 1 Hz steps from 60-100 Hz; 10 Hz steps from 100 Hz-1 kHz; 100 Hz steps from 1-10 kHz; 1 kHz steps from 10 kHz-100 kHz; 10 kHz from 100 kHz-1 MHz; and 100 kHz steps from 1-6 MHz). The frequency synthesizer shall be accurate to within 5% of the selected value. Sampling rate shall be a minimum of 500 Hz.

3.5.2 Gain.
The instrument shall be capable of at least 90 dB of amplification range for both the vertical (y) and horizontal (x) axes.

3.5.3 Probe drive.
The instrument probe drive circuit shall allow the equipment operator to vary the signal driving the probe independent of gain and frequency adjustment. The instrument shall operate low frequency spot, slider, ring style, or wide-field ribbon absolute, differential, or reflection probes without use of an external pre-amplifier. Low frequency operation, 1000 Hz or less, should be capable of providing a signal response through 0.25 inch of aluminum without preamplification.

3.5.4 Phase rotation.
The instrument shall be able to rotate the eddy current impedance vector, as viewed on the display, through a full 360 degrees in one degree or smaller increments in either the clockwise or counter-clockwise directions.

3.5.5 Filters.
The instrument shall be equipped with high pass, low pass, and bandpass filters. The low pass filter shall be selectable from 10 Hz to 1500 Hz minimum range, and the high pass filter shall be selectable from 0 to 500 Hz minimum range.

3.5.6 Alarm function.
The instrument shall incorporate an alarm (gate) system that will trigger an audible and/or visual alarm when the test signal enters or leaves a designated alarm zone. The alarm function shall be adjustable for size, position, and shape, and useable in both the impedance plane and sweep modes. The audio alarm

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T10:04Z
Check the source to verify that this is the current version before use.

frequency shall be lower than 2000 Hz and have an intensity of 50-70 dB at three feet.

3.5.7 Programmable test procedures.
The instrument shall have the capability to program, store, and recall a minimum of 100 instrument test parameter settings. The instrument memory shall enable alphanumeric labeling of programs. Programmed instrument settings shall be adjustable after recall. The memory shall retain the settings for a minimum of 168 hours after electrical power is turned off and primary batteries are removed. The instrument shall have a "load default" feature with the following minimum settings:

Gain, horizontal and vertical	60 dB
Probe drive	Mid-range setting
Frequency 1	200 kHz
Frequency 2	Off
Low pass filter	100 Hz
High pass filter	Off
Null	Manual mode
Horizontal null position	80%
Vertical null position	10%
Angle	70 degrees
Display erase	Off
Display persistence	Off
Screen	10 × 10 grid

3.5.8 Data capture.
The instrument shall have the capability to freeze, store, and export a minimum of 100 screen displays, instrument settings or both for subsequent review or download to a computer or printer. Files shall export to a common image file format (e.g, .tif, .jpg, etc.) or PDF. The instrument shall have the capability to capture a minimum of 60 seconds of impedance plane trace data (x/y versus time) for subsequent review or download to a computer. The instrument shall have the capability to store a minimum of 20 trace data files. External PC software shall not be required to export screen displays or instrument settings to an external memory device.

3.6 Instrument display.
The instrument shall have a color display with a minimum resolution of VGA (640 × 480 pixels) and no vertical/horizontal dimension less than 3.4 inches. The display shall be easily seen both in bright light (i.e., 500 foot-candles or greater) or low light (i.e., 10 foot-candles or less) conditions. The brightness level shall be adjustable. The refresh rate shall be a minimum of 60 Hertz. The display shall have a minimum of 10 × 10 grid selectable graticule for both major and minor divisions. Where applicable, displays shall have options of displaying numerical values in inch-pound units.

3.7 Persistence and erase.

a. The display persistence shall be variable and adjustable from 0.1 to 10 seconds, in 0.1 second

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T10:04Z
Check the source to verify that this is the current version before use.

increments, minimum range or until cleared by the operator.

b. The display erase shall variable and adjustable from 0.1 to 30 seconds, in 0.1 second increments, minimum range or until cleared by the operator.

c. The instrument shall have a tactile means to erase the screen trace.

3.8 Display modes.
All instruments shall display eddy current inspections signals in the impedance plane (flying dot) and the vertical amplitude versus time (sweep) as a minimum. These display modes shall be user selectable as individual displays or simultaneous as single or dual views. Basic and intermediate instruments shall have dynamic fastener hole scanner display.

3.8.1 Impedance plane mode.
The impedance plane mode shall display the eddy current signal as an impedance vector representing the vector sum of the test coil inductive reactance and resistance. The impedance plane mode shall be available for both manual and fastener hole scanner modes. The display shall incorporate separate controls for horizontal and vertical positioning of the balance point anywhere on the screen.

3.8.2 Sweep mode.
The sweep mode shall display the vertical impedance plane component of the eddy current signal as a function of time. The instrument shall be capable of synchronizing with an external signal.

3.8.3 Dynamic fastener hole scanner mode.
The display shall have a dynamic mode that is compatible with the fastener hole scanner defined in 3.11.1. The dynamic display shall automatically initiate a sweep trace at the start of each scanner probe rotation. Display options shall include showing only the current sweep (y versus t) while automatically erasing the previous sweep information, showing impedance plane and sweep, and a waterfall type display to show consecutive sweeps if connected to an indexing scanner. The full 360-degrees of rotation shall be displayed regardless of any nominal changes in rotation speed of the scanner probe.

3.9 Conductivity and coating thickness measurement.
The instrument shall be capable of measuring electrical conductivity and non-conductive coating thickness as specified below. The instrument shall come equipped with appropriate probes, cables, non-conductive shims, and standard for conductivity and thickness measurement.

3.9.1 Electrical conductivity.
The instrument shall be capable of measuring electrical conductivity of materials in the minimum range from 0.9 to 65% IACS (International Annealed Copper Standard). The instrument shall be capable of using 60 ±10 kHz and 500 ±20 kHz for conductivity measurement. The measurement accuracy for the range from to 65% IACS shall be within ±0.5% IACS from 32 to 100 °F with liftoff up to 0.010 inch. Measurements shall be displayed digitally.

3.9.2 Non-conductive coating thickness.
The instrument shall be capable of measuring non-conductive coatings up to 0.025 inch thick. The measurement accuracy from 0.0 to 0.015-inch, shall be within ±0.001 inch of the reading. Measurements shall be displayed digitally.

3.10 External connections.
The instrument shall accommodate the following external connections.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T10:04Z
Check the source to verify that this is the current version before use.

3.10.1 Probe interface.
The instrument shall have a connector to interface with commercial probes noted in 3.11.4 and scanners listed in 3.10.3 which use 16-pin Fischer Lemo FGG.2B connector. Scanner pin out wiring is available upon request.

3.10.2 Probe compatibility.
The instrument shall be electronically compatible with all currently available commercial off-the-shelf absolute, reflection, and differential-type probes.

3.10.3 Scanner interface.
The instrument shall interface and be compatible with fastener hole scanners including MiniMite, Spitfire 2000 and RA 2000. The instrument shall have a minimum scanner speed from at least 300 to 2500 revolutions per minute (RPM). As a minimum, the eddy current instrument shall provide the following: (1) scanner power; (2) sweep coordination (external triggering); (3) control of scanner probe revolutions per minute (RPM); (4) compatibility with remote null switching on the scanner; and (5) capability to change probes in the field. The scanner control capability shall be included with the instrument and be ready for use. Types II through V shall include a high-speed fastener hole scanner in the kit.

3.10.4 Analog output.
The instrument shall have analog outputs to allow for horizontal and vertical component test signal monitoring.

3.10.5 Computer interface.
The kit shall include Microsoft Windows (10 or later version) compatible software and cables to enable communication with a personal computer. Software capabilities shall include: (1) software version updates; (2) uploading and downloading instrument program settings; and (3) data reporting utilities such as downloading screen captures or traces. While open architecture is preferred, if the instrument does not store the data in a standard software format (e.g, .tif, .jpg, .txt, etc.), software shall be provided to convert the file format. The instrument shall have USB or secure digital (SD)/compact flash card connectivity. Software/firmware updates shall be capable directly through the interface without computer connection.

3.10 Wireless communication.
Wireless communication capabilities shall be hardware and/or software disabled at the factory and a certificate of compliance provided with each instrument unless otherwise specified in the contract.

3.10.6 Printer output.
The instrument shall be capable of printing screen presentations and test procedures. This may be accomplished either directly to a USB printer or through a Microsoft Windows (10 or later version) compatible computer, with supplied software.

3.10.7 Video output.
The instrument shall have a video output of the instrument display and supplied with cable.

3.11 Accessories.

3.11.1 Fastener hole scanner.
Types II through V shall contain a high speed fastener hole scanner. Fastener hole scanners shall have a minimum speed range of 300 to 2500 revolutions per minute and discrete speeds shall include 1000, 1200, and 1500 revolutions per minute, if not infinitely variable. Scanner size shall not exceed 6 × 3 × 2 inches, excluding the probe. Fastener hole scanner shall have sufficient torque to rotate a 0.750 inch diameter probe at 1500 revolutions per minute through the 0.875 inch thick aluminum plate assembly of the universal

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T10:04Z
Check the source to verify that this is the current version before use.

eddy current standard, NSN 6635-01-587-1418. The scanner shall operate with a probe using a Fischer 4-pin connector operating with a reflection differential coil arrangement.

3.11.2 Protective carry cases.
The instrument shall include soft-shell and hard-shell carrying cases. The soft shell case shall be ergonomically designed and protect the instrument from damage during use while still allowing access to all external connections and have appropriate hand/neck straps to help prevent accidental loss. The hard-shell case shall be suitable for storing and shipping the instrument with accessories. If the hard-shell case with instrument and accessories weighs in excess of 25 pounds, the hard shell case shall include rollers and collapsible handle typically found on luggage. The hard-shell case size and total weight shall be limited to airline checked baggage restrictions of 50 pounds and dimensions of 62 total inches (length plus width plus height). The cases shall be designed to protect the instrument from an 18 inch drop in soft case and a 36 inch drop in the hard case.

3.11.3 Cables.
The instrument shall have a cable to connect absolute "Microdot" type probes to the instrument. The instrument shall have a cable to connect Fischer Triax probes to the instrument. Types II through V instruments shall have a cable to connect the fastener hole scanner to the instrument. The instrument shall have a cable to connect to a USB computer port. Cable lengths shall be between 6 and 10 feet.

3.11.4 Probes.
All types shall contain a shielded surface probe kit, all with Fischer Triax connectors and internal balance coils. Probe kit contents: two each 1/8 inch diameter stainless-steel shaft, straight pencil probes with 4 inches overall length with a frequency range of 50-500 kHz; two each 1/8 inch diameter stainless-steel shaft, 0.50 inch drop, 90 degree pencil probes with 6 inch overall length with a frequency range of 50-500 kHz; two each 1/8 inch diameter flexible copper shaft, straight pencil probes with 6 inches overall length with a frequency range of 50-500 kHz; two each 1/8 inch diameter stainless-steel shaft, straight pencil probes with 4 inches overall length with a frequency range of 1-3 MHz; and two each 1/8 inch diameter stainless-steel shaft, 0.50 inch drop, 90 degree pencil probes with 6 inches overall length with a frequency range of 1-3 MHz. The instrument shall have a conductivity/coating thickness probe to meet the requirements in 3.9. Conductivity probe diameter shall not exceed 0.50 inch.

3.11.4.1 Probes – Types II through V.
Types II through V shall include a plastic fastener hole probe kit for the scanner containing 21 fastener hole probes with an adjustable diameter feature in sizes 1/8 inch to 3/4 inch in 1/32 inch increments, with its own separate ruggedized container. Working length shall be a minimum of 1.0 inch below 0.250 inch diameter and a minimum of 2.0 inches at or above 0.250 inch diameter. Fastener hole probes shall have an operational frequency range suitable for the included reference standards for each type.

3.11.5 Reference standards.
All instruments, except Type IV, shall include aluminum conductivity standards with conductivity of 30 ±1% and 60 ±1% IACS. Applicable fastener hole and surface reference standard meeting the requirements of NSN 6635-01-587-1418, P/N NRK-3A or NSN 6635-01-092-5129, P/N 7947479-10, shall be provided according to type. Types V and VI includes NSN 5280-01-352-1336, P/N NRK-3AST. Alloys are 7075-series aluminum, 4340 steel, and 6AL-4V titanium, unless otherwise specified. Drawings are available upon request.

3.12 Environmental conditions.
The instrument shall be operable in the following environmental conditions.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T10:04Z
Check the source to verify that this is the current version before use.

3.12.1 Explosive atmospheres.
The instrument shall be tested and certified by a reputable laboratory for safe use, while operating on batteries, in an explosive atmosphere as defined by Class 1, Division 2, Group D, in the NFPA 70 Chapter 5, Article 500. The testing method shall be in accordance with MIL-STD-810H, Method 511.7, Explosive Atmosphere, Procedure I, using a Fischer Triax eddy current manual surface probe and Fischer Lemo cable attached, in a typical operating condition.

3.12.2 Operating conditions.
The instrument and accessories shall perform to the specification in this CID throughout temperatures from 15 to 120 °F at up to 90% relative humidity and from sea level to 5280 feet above sea level. For use below 50 °F, the instrument shall start at an ambient temperature of 50 °F, and then shall operate on battery power for a minimum of 30 minutes at temperatures of 15 to 32 °F, and similarly for a minimum of 60 minutes at temperatures from 32 to 50 °F.

3.12.3 Storage conditions.
The instrument and accessories shall perform to the specifications in this CID following a seven day exposure to temperatures from 0 to 150 °F at up to 95% relative humidity with instruments in shipping cases.

3.12.3 Ingress protection (IP) rating.
The instrument shall have a minimum IP65 rating.

3.12.4 Drop test.
The instrument shall meet MIL-STD-810H Transit Drop Test, Method 516.8, Procedure IV.

3.12.5 Vibration test.
The instrument shall meet MIL-STD-810H, Method 514.8, Annex C, Table 514.8C-IX General Exposure, one hour each axis, instrument outside case and powered.

3.12.6 Shock test.
The instrument shall meet MIL-STD-810H, Method 516.8, Procedure I functional shock test, half-sine at 15G, 11ms with 6 positive/negative shocks per axis, instrument outside case and powered.

NOTE: Equipment certified to earlier versions of MIL-STD-810 may be acceptable upon review by a Government official.

3.13 Calibration.
The instrument shall be user calibrated without the need for manufacturer's periodic calibration. No factory recalibration decals shall be applied to the instrument.

3.14 Software/firmware revisions.
Software/firmware revisions shall be user installable from a Microsoft Windows-based (10 or later versions) personal computer or memory device direct to the instrument. All software/firmware revisions made during the life of the contract shall be included at no additional charge. A copy of revisions, and any necessary executable software, released during the life of the contract shall be provided to AFLCMC/EZPT-NDIO Tinker AFB, OK 73145-3318, or as specified in the solicitation or contract, to ensure proper configuration management.

3.15 User's documentation.
One copy, printed and on CD/DVD of the commercial operation and maintenance instructions, including calibration procedures and adjustments, as applicable, following MIL-PRF-32216 shall be shipped with

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T10:04Z
Check the source to verify that this is the current version before use.

each item. The documentation shall include information for battery safety and known environmental restrictions on battery disposal. The manual shall provide instructions for users to replace internal battery for clock and memory functions, if applicable.

3.16 Provisioning.
Two paper and two electronic copies of the operation and maintenance instructions (above), illustrated parts breakdown, diagnostics, and servicing instructions, shall be sent to the Contracting Officer for delivery to the Equipment Specialist and cognizant Engineer. The documentation shall be releasable for third-party repairs after expiration of the contract, subject to standard government non-disclosure agreements between original equipment manufacturers (OEMs) and third-party providers. Recommendations for long-term provisioning shall be sent to the Contracting Officer for delivery to the Program Management Specialist, Inventory Management Specialist, Production Management Specialist, Equipment Specialist, and cognizant Engineer. For subsequent delivery orders, any changes to the instructions or recommended provisioning shall be sent to the Contracting Officer for similar distribution.

3.17 Warranty.
A minimum one-year warranty, to include parts, labor, and two-way shipping, shall be included against manufacturing defects.

4. REGULATORY REQUIREMENTS.

The offeror/contractor is encouraged to use recovered materials to the maximum extent practicable, in accordance with 23.403 of the Federal Acquisition Regulation (FAR).

4.1 Metric products.
Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, provided they fall within specified tolerances using conversion tables contained in the latest revision of FED-STD-376, and all other requirements of this CID are met. If a product is manufactured to metric dimensions and those dimensions exceed tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

5. PRODUCT CONFORMANCE.

5.1 Product conformance.
The products provided shall meet the salient characteristics of this CID, conform to the producer's own drawings, specifications, standards, and quality assurance practices, and be the same product offered for sale in the commercial marketplace. The government reserves the right to require proof of such conformance.

5.2 Market acceptability (MA).
All of the following market acceptance criteria are necessary to document the quality, reliability, and supportability of the product to be provided under this CID:

1. The company producing the item must have been producing eddy current instruments for at least the past three years;
2. The company must have sold the commercial instrument in the industrial market for a minimum of two years.
3. The company must have sold at least 100 instruments similar to the instrument being offered, with no major defects within the last year. A major defect is a defect, which affects the performance of the instrument and requires an engineering change to the instrument's original configuration in order for the instrument to perform the requirements of this CID.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T10:04Z
Check the source to verify that this is the current version before use.

4. The instrument shall be at least 50% repairable in the continental US with an average repair time of two weeks or less. A list of authorized repair centers shall be supplied with the offer.

5.3 Conformance inspection.
A production article shall be tested to conform to characteristics in section 3. Specifications not readily measurable can be verified by manufactures' certified documentation or independent third-party compliance documentation. Conformance inspection may be performed by a Government official, Government authorized contractor, or independent agent.

6. PACKAGING.
Preservation, packing, and marking shall be as specified in the contract or order.

7. NOTES.
(This section contains information of a general or explanatory nature which may be helpful, but is not mandatory.)

7.1 Part or identification number (PIN).
The following PIN procedure is for government purposes and does not constitute a requirement for the contractor.

This example describes a part numbering system for CID A-A-59131D.

AA59139D 1 Example of reference part number:

AA59131D-1

Type:
- 1- Type I - Basic Kit with case, charger, and accessories
- 2- Type II - Intermediate Kit with case, charger, and accessories, with fastener hole scanner.
- 3- Type III - Intermediate Kit with case, charger, and accessories, with fastener hole scanner, aluminum fastener. hole standard, and
- 4- Type IV - Intermediate Kit with case, charger, and accessories, with fastener hole scanner, steel fastener hole standard, and
- 5- Type V - Intermediate Kit with case, charger, and accessories, with fastener hole scanner, aluminum, steel, and titanium fastener hole standards, and probes.
- 6- Type VI - Intermediate Kit with case, charger, and accessories, with aluminum, steel, and titanium fastener hole standards.

7.2 Source of documents.

7.2.1 Copies of government documents.
Copies of government documents are available online at https://quicksearch.dla.mil.

7.2.2 Copies of the Federal Acquisition Regulations (FAR).
Copies of the FAR documents are available online at http://acquisition.gov/far/index.html.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T10:04Z
Check the source to verify that this is the current version before use.

7.2.3 Copies of the National Fire Protection Association (NFPA).
Copies of NFPA document are available online at http://www.nfpa.org or from NFPA, 1 Batterymarch Park, Quincy, MA 02169-7571.

7.2.4 Copies of Naval Sea Systems Command Technical Manual (NAVSEA TM).
Copies of NAVSEA TM are available online at https://www.public.navy.mil/navsafecen/Documents/afloat/Surface/CS/Lithium_Batteries_Info/LithBattSafe.pdf

7.3 National stock numbers (NSNs).
The following is a list of NSNs assigned that correspond to this CID. The list may not be indicative of all possible NSNs associated with the CID.

NSN	Type	Part Number	CAGE CODE
TBD	Type I	A-A-59131D-1	81348
6635-00-463-1574RN	Type II	A-A-59131D-2	81348
TBD	Type III	A-A-59131D-3	81348
TBD	Type IV	A-A-59131D-4	81348
TBD	Type V	A-A-59131D-5	81348
TBD	Type VI	A-A-59131D-6	81348

7.4 Ordering data.
The contract or order should specify the following:

a. Title, number, and date of this document.

b. Product conformance provisions.

c. The solicitation will provide for any required independent-laboratory conformance inspection, or government testing, as appropriate.

d. Packaging requirements.

7.5 Key Words.

Dual frequency
Electrical conductivity
Electromagnetic testing
Fastener hole scanner

7.6 Changes from previous issue.
The margins of this commercial item description are marked with vertical lines to indicate where changes from the previous issue were made. This was done as a convenience only and the Government assumes no liability whatsoever for any inaccuracies in these notations. Bidders and contractors are cautioned to evaluate the requirements of this document based on the entire content irrespective of the marginal notations and relationship to the previous issue.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T10:04Z
Check the source to verify that this is the current version before use.

CONCLUDING MATERIAL

MILITARY INTERESTS:

Custodians:
Army - GL
Navy - AS
Air Force - 11
DLA - GS

CIVIL AGENCY COORDINATING ACTIVITY:
GSA-FAS

Preparing activity:
Air Force - 11

(Project 6635-2020-001)

NOTE: The activities listed above were interested in this document as of the date of this document. Since organizations and responsibilities can change, you should verify the currency of the information above using the ASSIST Online database at https://assist.dla.mil.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T10:04Z
Check the source to verify that this is the current version before use.