FED. SUP CLASS
6640



MS36249-1 MS36249-2 MS36249-3 MS36249-4 MS36249-5 MS36249-6

PART NUMBER	TOP O.D. MM	BOTTOM O.D. MM	HEIGHT MM	WALL THICKNESS MM
MS36249-1	22	10	20	2.5
MS36249-2	28	16	20	2.5
MS36249-3	37	22	25	3.0
MS36249-4	46	28	30	3.0
MS36249-5	58	37	35	3.0
MS36249-6	70	46	40	4.0
MS36249-7	Set, consisting of one each of -1 thru -6			

MATERIAL: Adapter shall be made of neoprene rubber. Item shall withstand steam pressure sterilization.

DESIGN: Adapter side walls shall taper at an angle of approximately 17°.

USE: For supporting gooch-type crucibles, funnels, micro filter tubes, etc., in vacuum filtration assemblies. Particularly suitable for buchner and fritted glass funnels, porcelain filters without headpieces and for reverse flow methods of cleaning filter crucibles. Adapters can be used singly or in combination with adjacent sizes.

NOTE: Items differing from this Military Standard by minor variations in design or dimensions are considered standard under this document, provided such variations do not affect performance and are not in conflict with the material requirements specified herein.

USER ACTIVITIES: ARMY - MD, DM
NAVY - MS

REVIEWER ACTIVITIES: ARMY - ME
NAVY - OS
AIR FORCE - 82

This military standard is approved for use by all Departments and Agencies of the Department of Defense. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

APPROVED 5 JAN 1963 REVISED

P.A. EA	INTERNATIONAL INTEREST	TITLE	MILITARY STANDARD
Other Cust MS 99		ADAPTER, FILTER	MS36249
Procurement Specifications NONE	SUPERSEDES:		SHEET 1 OF 1

DD FORM 1 MAY 73 672 (COORDINATED) PREVIOUS EDITIONS OF THIS FORM ARE OBSOLETE

PROJECT NUMBER 6640-1132

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:37Z
Check the source to verify that this is the current version before use.