MILITARY SPECIFICATION

APPLICATOR, DISPOSABLE, POVIDONE-IODINE IMPREGNATED

This specification is approved for use by the Defense Personnel Support Center, Defense Supply Agency, and is available for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers Applicator, Disposable, Povidone-Iodine Impregnated (See 6.2).

2. APPLICABLE DOCUMENTS

2.1 Issues of documents. The following documents, of the issue in effect on date of invitation for bids or request for proposals, form a part of this specification to the extent specified herein.

SPECIFICATIONS

FEDERAL

PPP-B-566	- Boxes, Folding, Paperboard.
PPP-B-636	- Boxes, Shipping, Fiberboard.
PPP-B-676	- Boxes, Setup.

STANDARDS

MILITARY

MIL-STD-105	- Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	- Marking for Shipment and Storage.

FSC 6510

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:34Z
Check the source to verify that this is the current version before use.

(Copies of specifications and standards required by contractors in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposals shall apply.

U. S. DEPARTMENT OF HEALTH, EDUCATION, AND WELFARE, FOOD AND DRUG ADMINISTRATION

Federal Food, Drug, and Cosmetic Act and Regulations Promulgated Thereunder.

(Application for copies should be addressed to the Food and Drug Administration, Washington, DC 20204.)

U. S. PHARMACOPEIAL CONVENTION, INC.

The United States Pharmacopeia.

(Application for copies should be addressed to the Mack Publishing Company, Easton, PA 18042.)

National Formulary.

(Application for copies should be addressed to the Mack Publishing Company, Easton, PA 18042.)

3. REQUIREMENTS

3.1 Material. Shall be a disposable rayon-tipped applicator with a povidone-iodine solution impregnated tip.

3.1.1 Applicator. Shall be 4-1/4 inches long and 1/8 inch in diameter and shall be symmetrical and smooth throughout. Shall be made of a suitable material sufficiently flexible for the purpose, and of such compatibility that neither it nor the povidone-iodine solution are altered in any way by each other.

3.1.2 Tip. Shall be Purified Rayon, U.S.P., attached to the applicator by use of a non-toxic adhesive suitable for the purpose.

3.1.3 Impregnating solution. Shall be Povidone-Iodine Solution, U.S.P., 10%. Each packette of 3 applicators shall contain not less than 7.2 ml of solution when tested as specified in 4.5.1.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:34Z
Check the source to verify that this is the current version before use.

3.1.4 Suitability. Shall be suitable for applying the solution to hard-to-reach skin areas including folds and creases, for use in preparing skin areas prior to injections, aspirations, surgery, catheterization, disinfecting wounds, pre-operative oral degerming, and preparing for dressing changes.

3.2 Finished product.

3.2.1 pH. The povidone-iodine solution expressed from the swab sticks shall have a pH of not more than 6.0 when determined potentiometrically at 25°C, using the U.S.P. method.

3.2.2 Available iodine. The povidone-iodine solution expressed from the swab sticks shall contain not less than 0.85 Gram and not more than 1.20 Gram of available iodine per 100 ml of solution when assayed as specified in 4.5.2.

3.2.3 Identification. The infrared absorption spectra of a dried sample of povidone-iodine solution expressed from the swab sticks in a potassium bromide dispersion at a concentration of 1 in 200 shall exhibit maxima and minima only at the same wavelengths as that of a previously submitted standard infrared curve.

3.2.3.1 A copy of the standard infrared curve (graph) shall be submitted by the bidder or offeror, and shall be kept on file in the Technical and Quality Assurance Division, Directorate of Medical Materiel, of the Defense Personnel Support Center for use if samples are submitted.

3.3 Ingredients.

3.3.1 Povidone-iodine solution. The Povidone-iodine solution used in the impregnation of the applicators shall comply with the tests, standards, and requirements of the U.S.P., including any supplements or revisions thereto.

3.3.2 Purified rayon. The Purified rayon used in the manufacture of the applicators shall comply with the tests, standards, and requirements of the U.S.P., including any supplements or revisions thereto.

3.3.3 Other ingredients. All other ingredients used in the manufacture of the product shall be of N.F. or U.S.P. quality, or, if not included in either of these compendia, the ingredients shall be suitable for use in this preparation.

3.4 Delivery. Not more than 6 months shall have elapsed from the date of manufacture of the product to the date of delivery to the Government.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:34Z
Check the source to verify that this is the current version before use.

3.5 **Workmanship**. The material and its container shall be free from defects which detract from their appearance or may impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Supplier responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and, (i) as to any expiration dated item, for the expiration dating period (potency period) specified by the contractor for such item, or such longer period as may be required by regulation of any federal agency; or (ii) as to non-dated items, for not less than 3 years after delivery of the item to the Government.

4.1.2 No company supplying any ingredient(s) to the contractor will be considered an acceptable facility for the performance of any inspection requirements specified herein.

4.2 Lot. For purposes of this specification, a lot, batch, or control is that single, uniform, and homogeneous quantity of impregnated applicators produced from one formulation, subjected to the same compounding and manufacturing operation, and filled into final containers.

4.3 Sampling. Sampling shall be conducted in accordance with MIL-STD-105.

TABLE I. Sampling

For visual examination	Inspection level	AQL (percent defective)	Unit of product
Major	S-2	1.0	Unit of issue
Minor	II	2.5	Unit of issue

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:34Z
Check the source to verify that this is the current version before use.

4.3.1 For raw material and end item testing. Noncompliance with any of the following is cause for rejection.

TABLE II. Raw material and end item testing.

Characteristic	Requirement paragraph	Test paragraph	Inspection level 1/	Minimum sample (packettes)
Impregnating Solution (Volume)	3.1.3	4.5.1	-	10
pH	3.3.1	U.S.P.	S-2	10 } Composite
Available Iodine	3.3.2	4.5.2	-	10 }
Identification	3.3.3	-	S-2	10 }

1/ Where possible, the sample shall be used for two or more tests.

4.4 Examination. The finished disposable applicator shall be examined to determine compliance with all requirements of this specification. Nonconformance will be permitted to the extent indicated in 4.3.

4.4.1 Classification of defects. Examination will be conducted in accordance with the following classification of defects.

TABLE III. Classification of defects.

Categories and defects 1/	
Major	
101	Applicator stick not smooth.
102	Applicator stick broken.
103	Dimensions not as specified in 3.1.1.
104	Applicator stick not symmetrical.
105	Foil packette incompletely sealed or torn.
106	Rayon tip not essentially free from foreign matter.
107	Rayon tip loose.
Minor	
201	Exterior of immediate container is not free from content or foreign matter.
202	Label not free from defacing marks or tears.

1/ Inspection is not restricted to the classified possible defects listed above.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:34Z
Check the source to verify that this is the current version before use.

4.5 Tests.

4.5.1 Impregnating solution. Open one finished foil packette and carefully pour the povidone-iodine solution into a 10 ml graduated cylinder. Also express the solution from each of the three impregnated applicators into the same graduated cylinder. The total volume of the solution in the cylinder shall be not less than 7.2 ml. A minimum of ten (10) packettes shall be tested.

4.5.2 Available iodine. Squeeze the povidone-iodine solution out of 10 packettes (30 swab sticks) into a beaker and take a 5 ml aliquot of the solution and assay by the method specified in the U.S.P. for povidone-iodine solution.

4.6 Packaging inspection. The inspection of the unit package, packing and marking for shipment and storage shall be in accordance with quality assurance provisions of the applicable container specification and the marking requirements of MIL-STD-129.

4.7 Labeling inspection. Sampling shall be in accordance with MIL-STD-105 and the unit of product shall be the unit of issue. The inspection lot shall be the number of units of issues of each batch, lot, control number, etc., offered for inspection at one time. The inspection level for labeling shall be S-2 and the acceptance number shall be zero.

5. PACKAGING

5.1 Packaging requirements. Packaging shall be level A or C as specified (see 6.1).

5.1.1 Unit of issue. One box containing fifty packettes of three applicators per packette as specified, constitutes one unit of issue.

5.1.2 Level A.

5.1.2.1 Immediate container. Three applicators shall be packaged in a hermetically sealed foil packette.

5.1.2.2 Unit package. Fifty packettes (150 applicators) shall be packaged in a box conforming to PPP-B-566 or PPP-B-676. Closure shall be as specified in the appendix of the applicable box specification.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:34Z
Check the source to verify that this is the current version before use.

5.1.3 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing.

5.2.1 Level A. Ten unit packages shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-636, class weather-resistant. Closure, strapping and waterproofing shall be as specified in the appendix of the box specification.

5.2.2 Level B. Ten unit packages shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-636, class domestic.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Labeling. Labeling shall be in accordance with the requirements of the Federal Food, Drug and Cosmetic Act, and shall include the information required below.

5.3.1.1 Immediate container. Each immediate container label shall bear the following information:

(a) label information in accordance with commercial practice.

5.3.1.2 Unit package. Each unit package label shall bear the following:

(a) labeling information in accordance with commercial practice.
(b) the national stock number.
(c) the date of manufacture prefixed by "MFD".

NOTES: (1) The national stock number shall not be preceded by any prefix.
(2) Strip labeling may be used.
(3) Label adhesion test shall not apply.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:34Z
Check the source to verify that this is the current version before use.

5.3.2 Intermediate package. Each intermediate package shall be marked in accordance with MIL-STD-129 except that the date of manufacture shall be shown in lieu of the date packed. Lot (control) number and the legend: "STORE AT CONTROLLED ROOM TEMPERATURE (59° - 86°F)" shall be shown.

5.3.3 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129 except that the date of manufacture shall be shown in lieu of the date packed. Lot (control) number and the legend: "STORE AT CONTROLLED ROOM TEMPERATURE (59° - 86°F)" shall be shown.

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

(a) Number and date of this specification.
(b) National Stock Number.
(c) Selection of applicable levels of packaging and packing (See 5.2 and 5.3).

6.2 This specification covers the following item(s):

National Stock No.	Item Identification
6510-01-008-7917	APPLICATOR, DISPOSABLE, POVIDONE-IODINE IMPREGNATED, 4-1/4 inches long, 150s

6.3 This specification does not cover all types, classes, grades, or sizes of the commodity indicated by the title of this specification, or those which are commercially available, but is intended to cover the types which are normally procured to meet Military requirements.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6510-0616

☆ U.S. GOVERNMENT PRINTING OFFICE: 1976— 603-766:2919

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:34Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval No. 22-R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/ vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE

MIL-A-37382(DSA-DM) APPLICATOR, DISPOSABLE, POVIDONE-IODINE IMPREGNATED

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

   A. GIVE PARAGRAPH NUMBER AND WORDING.

   B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?

   ☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO.

DATE

DD FORM 1 JAN 72 1426

REPLACES EDITION OF 1 JAN 66 WHICH MAY BE USED

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:34Z
Check the source to verify that this is the current version before use.