MILITARY SPECIFICATION

TEST KIT, SYPHILIS DETECTION, 100 tests

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers a test kit suitable for use in performing the Rapid Plasma Reagin Tests (RPR CARD TEST) for the presence of syphilis and other treponematoses by means of specifically prepared disposable cards and antigen. The kit is a self-contained unit with no additional equipment required in order to run the test (See 6.2).

2. APPLICABLE DOCUMENTS

2.1 Issues of documents. The following documents, of the issue in effect on date of invitation for bids or request for proposals, form a part of this specification to the extent specified herein:

SPECIFICATIONS

FEDERAL

PPP-C-186 - Containers, Packaging and Packing for Drugs, Chemicals, and Pharmaceuticals. (See 6.3)

STANDARDS

MILITARY

MIL-STD-105 - Sampling Procedures and Tables for Inspection by Attributes.

MIL-STD-129 - Marking for Shipment and Storage.

(Copies of specifications and standards required by contractors in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

FSC 6505

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:30Z
Check the source to verify that this is the current version before use.

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposals shall apply.

U.S. DEPARTMENT OF HEALTH, EDUCATION, AND WELFARE, FOOD AND DRUG ADMINISTRATION.

Federal Food, Drug, and Cosmetic Act and Regulations Promulgated Thereunder.

(Application for copies should be addressed to the Food and Drug Administration, Washington, DC 20204.)

U. S. PHARMACOPEIAL CONVENTION, INC.

The United States Pharmacopeia.

(Application for copies should be addressed to the Mack Publishing Company, Easton, PA 18042.)

National Formulary.

(Application for copies should be addressed to the Mack Publishing Company, Easton, PA 18042.)

3. REQUIREMENTS

3.1 Material. Shall be as specified hereinafter for the specified component. The kit shall consist of the following components necessary for performing 100 Rapid Plasma Reagin (RPR) tests.

3.1.1 Components. Each kit shall contain the following components:

(a) 10 packages, each package containing 10 Brewer Plasma Collection Slides (disposable), 10 sterile, disposable blood lancets, and directions for use. (See 3.2.1)

(b) 10 Brewer Diagnostic Cards with 10 depressions each, plus one Reading Guide Card. (See 3.2.2)

(c) 200 Wooden stirrers (toothpicks). (See 3.2.3)

(d) 100 (minimum) capillary tubes with at least three rubber nipples to fit. (See 3.2.4)

(e) 1 Plastic needle hub with needle attached. (See 3.2.5)

(f) 2 Ampuls, approximately 1 ml, RPR Card Test Antigen. (See 3.2.6)

(g) 1 Flexible plastic dispensing vial for antigen. (See 5.2.2.1.4)

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:30Z
Check the source to verify that this is the current version before use.

(h) 1 Directions for use pamphlet. (See 3.6)

(i) Venereal Disease Research Laboratory RPR "approval" note. (See 3.3)

3.2 Description of components.

3.2.1 Brewer plasma collection slides. The collection slides shall be plastic-coated boards approximately 2 inches wide and 5 inches long, having a thickness of 0.026 inch (plus or minus 0.0015 inch). The slides shall have a one-inch keyhole shaped depression with a maximum depression of 0.015 inch, and a minimum depression of 0.006 inch. Each card shall contain a perforated line one inch from the top which shall enable the top section to be bent down to form a support which will allow the plasma to drain. The depressed area of the slide shall be coated with a suitable anticoagulant and a lectin which shall remain stable at room temperature. The lectin shall cause both red and white blood cells to agglutinate, leaving the plasma free to flow down into the collection slot. Ten (10) packages of ten (10) slides each shall be furnished in each kit.

3.2.1.1 Blood lancets. The blood lancets shall be of metal with a thickness of 0.003 to 0.005 inch and shall have an overall length of 1-1/4 inches with the length of the inserting point 3/32 to 3/16 inch. One package of ten individually sterile wrapped lancets shall be packaged in each package of ten plasma slides.

3.2.1.2 Directions for plasma collection. Directions for use shall be included in each package of ten plasma slides. The directions shall describe and illustrate the method of collecting the plasma using the Brewer plasma collection slides and lancets.

3.2.2 Brewer diagnostic card. Shall be plastic-coated board, approximately 3 inches wide and 5 inches long, with ten (10), "teardrop" depressions. Ten (10) such diagnostic cards shall be furnished in each kit. The depression shall be sufficient to hold liquid used in the test.

3.2.2.1 Reading guide information. The information shall illustrate the control for a reactive test; a reactive (minimal) test; and a non-reactive test. One reading guide containing the information shall be furnished with each kit. If the reading guide card is not supplied, the information on the card shall be contained in the directions for use pamphlet.

3.2.3 Wooden stirrers. A box of 200 flat wooden toothpicks shall be included with each kit.

3.2.4 Capillary tubes. Capillary tubes shall be glass and shall dispense 0.060 ml $\pm$ 0.015 ml. A mark designating 0.03 ml shall appear in the middle of the capillary tube. At least 100 capillaries shall be furnished with each kit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:30Z
Check the source to verify that this is the current version before use.

3.2.4.1 Rubber nipples. Rubber nipples, when used with the capillary tubes, shall be suitable for withdrawal of 0.03 ml of blood plasma from the collection slide. At least 3 nipples shall be furnished with each kit.

3.2.5 Needle and needle hub. Shall be a 21 gauge, 3/4 inch, burr-free, square-cut needle capable of dispensing 65 to 70 drops per ml. The needle shall be attached to a plastic needle hub and shall be shielded with a suitable polyethylene cover. One needle unit shall be furnished with each kit.

3.2.6 Antigen. Each kit shall be furnished with two (2) ampuls, each containing approximately 1 ml of antigen charcoal suspension, suitable for use with the RPR Card Test.

3.2.6.1 Conformity to standards. The reactivity of each lot of RPR Card antigen suspension with a standard control serum shall be compared to a standard pattern of reactivity for conformity.

3.3 Approval of Venereal Disease Research Laboratory. A sample of each lot of RPR Card antigen suspension used in the Syphilis Detection Test Kit, shall be submitted to the Venereal Disease Research Laboratory, Communicable Disease Center, Atlanta, Georgia 30322, for testing and approval before release of the product. Each kit shall contain one (1) insert stating that the Venereal Disease Research Laboratory of the United States Public Health Service has tested the lot number of antigen suspension to reproduce reactivity of its standard test reagents. If the insert is not supplied, the information on the insert shall be on the outside of the unit package.

3.4 Ingredients. The chemicals, including sodium chloride, and the distilled water entering into the finished antigen shall comply with the Reagent Control section of the Manual of Tests for Syphilis, 1969, U.S. Department of Health, Education, and Welfare, Public Health Service. The end item shall be approved and released by the Center for Disease Control, Atlanta, Georgia 30333. In addition, the manufacturer's protocols for purity of these ingredients are acceptable.

3.5 Expiration dating period. The antigen shall have an expiration dating period of 12 months from the date of manufacture. The date of manufacture is ~~defined as the date of filling of the final containers, or the date of~~ manufacturer's or contractor's final quality approval of the final containers, which shall be not later than one month after the date of filling.

3.5.1 Delivery. Not less than 10 months of the expiration dating period (potency period) shall remain at time of delivery to the Government.

3.6 Directions for use pamphlet. One (1) pamphlet which describes and illustrates complete performance of the RPR Card Test shall be furnished as a component of each kit.

3.7 Sterility. The disposable lancets shall be sterile when tested in accordance with the U.S.P. method.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:30Z
Check the source to verify that this is the current version before use.

3.8 Workmanship. Workmanship shall be first class throughout. The kit and its components shall be free from defects which detract from their appearance or may impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the conrract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and, (i) as to any expiration dated item, for the expiration dating period (potency period) specified by the contractor for such item, or such longer period as may be required by regulation of any federal agency; or (ii) as to non-dated items, for not less than 3 years after delivery of the items to the Government.

4.1.2 A company supplying ingredient(s) to the contractor will not be considered an acceptable facility for the performance of any inspection requirements specified herein.

4.2 Lot. For purposes of this specification, a lot, batch, or control is that single, uniform, and homogeneous quantity of antigen suspension produced from one formulation, subjected to the same compounding and manufacturing operation, and filled into final containers.

4.3 Sampling. Sampling shall be conducted in accordance with the procedures set forth in MIL-STD-105.

TABLE I. Sampling

For visual examination	Inspection level	AQL (percent defective)	Unit of product
Major	S-2	1.0	Unit Package
Minor	II	2.5	Unit Package

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:30Z
Check the source to verify that this is the current version before use.

4.3.1 Classification of defects. Examination will be conducted in accordance with the following classification of defects:

TABLE II. Classification of defects.

Categories	Defects 1/
Major	
101	Ten Brewer Diagnostic Cards not included with each kit.
102	Reading Guide Information not included with each kit.
103	Plastic needle-hub with needle attached not included with each kit.
104	Less than 200 Wooden stirrers, 100 glass capillary tubes and 3 rubber nipples in each kit.
105	Two ampuls of 1 ml size and an RPR Card antigen suspension not included with each kit.
107	One directions for use pamphlet not included with each kit.
108	Venereal Disease Research Laboratory RPR "approval" note not included with each kit.
109	Less than 10 packages of 10 collection slides each in each kit.
110	Ten wrapped lancets in each package of collection slides not included.
Minor	
201	Exterior of kit not free from contents or foreign matter.
202	Label not free from defacing marks and tears.

1/ Inspection is not restricted to the classified possible defects listed above.

4.4 Tests.

4.4.1 Tests. Tests shall be conducted in accordance with the requirements set forth in this document.

4.5 Packaging Inspection. An examination shall be made to determine whether the packaging, packing, and marking comply with the requirements of section 5 for packaging, packing and marking. Defects shall be scored as specified in table III. Sampling shall be in accordance with MIL-STD-105. The sample unit shall be one container fully prepared for delivery. The lot shall be the number of containers offered for inspection at one time. The inspection level shall be S-2 and the acceptable quality level shall be 4.0 defects per hundred units. For labeling, the acceptance number shall be zero.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:30Z
Check the source to verify that this is the current version before use.

TABLE III. Examination of packaging. 1/

Examination	Defect
Containers	Not as specified.
Quantity	Quantity in interior packages and shipping containers not as specified.
Materials	Any component missing or damaged.
Workmanship	Inadequate application of components such as incomplete closure of container flaps, tape improperly applied, or containers distorted.
Markings	Omitted, Incorrect, Illegible, improper size, location, sequence, or method of application.

1/ The above examination is only applicable to intermediate and exterior container.

5. PACKAGING

5.1 Packaging requirements. The requirements for packaging shall be in accordance with all applicable requirements of PPP-C-186 (see 6.3), and as specified herein.

5.2 Packaging. Packaging shall be level A or C, as specified (see 6.1).

5.2.1 Unit of issue. Each test kit, as specified, constitutes one unit of issue.

5.2.2 Level A.

5.2.2.1 Immediate containers.

5.2.2.1.1 Antigen. Immediate container shall be a commercially available immediate container considered acceptable by the Food and Drug Administration (FDA) for the product contained therein and shall be of such compatibility that neither the contents nor the container are altered in any way by each other.

5.2.2.1.2 Capillary tubes and rubber nipples. Immediate container shall be as specified in 5.2.2.1.1. In addition, two (2) polyethylene discs shall be placed in the vial for protection of the capillary tubes. One disc shall be on the bottom of the vial under the capillary tubes and the other disc shall be placed on top of the capillary tubes. The three rubber nipples shall be placed between the upper disc and the closure. The vial and discs shall be of an appropriate size and design to prevent movement of contents when shaken.

5.2.2.1.3 Needle. The needle-hub component shall be suitably protected by a polyethylene shield.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:30Z
Check the source to verify that this is the current version before use.

5.2.2.1.4 Dispensing bottle for the antigen. Shall be a non-rigid ophthalmic type bottle having a threaded finish and an ophthalmic type dispensing tip inserted in the neck of the bottle. The dispensing tip shall fit the plastic needle hub. Closure shall be a rigid polystyrene screwcap with an integral plug conforming to the orifice of the bottle.

5.2.2.1.5 Stirrers. Stirrers shall be furnished in a set-up folding box of suitable size and shall be so designed as to permit ready dispensing of the stirrers.

5.2.2.1.6 Diagnostic cards. Shall be wrapped in a transparent polyvinylchloride film.

5.2.2.1.7 Plasma collection slide and blood lancets. Shall be a heat-sealed transparent polyethylene envelope of appropriate size to hold 10 plasma collection slides, 10 blood lancets, and the directions for their use. The lancets shall be sealed between one sheet of foil and one sheet of glazed paper in such a manner that each lancet shall be contained in an individual tear-off packet.

5.2.2.2 Unit packaging. Each unit shall be packaged as specified in 5.2.5 of PPP-C-186.

5.2.2.3 Intermediate packaging. Twelve (12) units of issue shall be packaged as specified in 5.2.8 of PPP-C-186.

5.2.3 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.3 Packing. Packing shall be level A, B or C as specified (see 6.1).

5.3.1 Level A. Twenty-four (24) units of issue shall be packed as specified in 5.4.4 of PPP-C-186.

5.3.2 Level B. Twenty-four (24) units of issue shall be packed as specified in 5.4.4 of PPP-C-186.

5.3.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.4 Marking.

5.4.1 Labeling. Labeling shall be in accordance with the requirements of the Federal Food, Drug and Cosmetic Act, and shall include the information required below.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:30Z
Check the source to verify that this is the current version before use.

5.4.1.1 Immediate container. Each immediate container label shall bear the following information:

(a) label information in accordance with commercial practice.

5.4.1.2 Unit package. Each unit package label shall bear the following information:

(a) label information in accordance with commerical practice.
(b) the national stock number.
(c) the expiration date.

NOTES: (1) The national stock number shall not be preceded by any prefix.
(2) Strip labeling may be used.
(3) Label adhesion test shall not apply.

5.4.2 Intermediate package. Each intermediate package shall be marked as specified in 5.5.3 of PPP-C-186. Type II shelf life markings shall be shown as specified in MIL-STD-129. Marking shall include the legend: "STORE BETWEEN 2°-8°C. (35°-46°F.)".

5.4.3 Exterior container. Exterior container shall be marked as specified in 5.5.4 of PPP-C-186. Type II shelf life marking shall be shown as specified in MIL-STD-129. Marking shall include the legend: "STORE BETWEEN 2°-8°C. (35°-46°F.)".

NOTE: The unrefrigerated shipping time for this item shall be not more than 7 days (168 hours).

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

(a) Title, number and date of this specification.
(b) the National Stock No.
(c) Selection of applicable levels of packaging and packing (see 5.2 and 5.3).

6.2 This specification covers the following item:

National Stock Number	Item Identification
6505-00-985-7224	TEST KIT, SYPHILIS DETECTION, Brewer Type, 100 Tests

6.3 All references to PPP-C-186 refer to Interim Federal Specification PPP-C-00186B which is mandatory for use until superceded by a revision.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:30Z
Check the source to verify that this is the current version before use.

6.4 This specification does not cover all types, classes, grades, or sizes of the commodity indicated by the title of this specification, or those which are commonly indicated by the title of this specification, or those which are commonly available, but is intended to cover the types which are normally procured to meet Military requirements.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6505-1497

 ☆U.S. GOVERNMENT PRINTING OFFICE: 1976-603-766/3651

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:30Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 South 20th Street
Philadelphia, PA 19101

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY
DoD - 304

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:30Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/ vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE

MIL-T-37353 TEST KIT, SYPHILIS DETECTION, 100 TESTS

NAME OF ORGANIZATION AND ADDRESS	CONTRACT NUMBER
	MATERIAL PROCURED UNDER A ☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

   A. GIVE PARAGRAPH NUMBER AND WORDING.

   B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?

   ☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*	TELEPHONE NO.
	DATE

DD FORM 1 JAN 72 1426 REPLACES EDITION OF 1 JAN 66 WHICH MAY BE USED

☆ GPO: 1974—603-680/6126 2-1

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:30Z
Check the source to verify that this is the current version before use.