
FED. SUP CLASS
6640



PART NUMBER	DIMENSIONS (cm) LENGTH	WIDTH	HEIGHT
MS51426-1	183	76	94
MS51426-2	213	76	94
MS51426-3	244	76	94
MS51426-4	272	76	94
MS51426-5	302		94
MS51426-6	333		94
MS51426-7	363		94
MS51426-8			94
MS51426-9			94
MS51[illegible]			94

CANCELLED

CANCELLED AFTER 21 AUG 1991

NO SUPERSEDING STANDARD

MATERIAL: ... one of the ...

... layers integrally mixed ... pigments and fillers.

... less steel.

DESIGN: The wall table shall consist of different types of base units chosen from MS36390 and one of the table tops as specified.

Type A top shall be formed under hydraulic pressure to make a homogeneous monolithic sheet. All exposed surfaces and edges shall be finished smoothly and all exposed edges shall be rounded. Top shall be 32 mm thick. Top shall be charcoal gray color.

Type B top shall not be less than 32 mm thick. All exposed surfaces shall be hone finished.

Type C laminated black plastic table top and curb, shall be made of a special laboratory grade plastic resin and fiber in accordance with NEMA specification. Top surface shall be 1.5 mm thick, glued under pressure with moisture resistant adhesive to a 32 mm thick flakeboard core. Underside of flakeboard shall be fitted with 0.5 mm thick melamine backing. Back and end curbs shall be 32 mm thick.

Type D top and curb shall be made from a modified epoxy resin. Material shall be uniform mixture throughout the full thickness, non-glaring and black in color. All exposed edges shall be radiussed. Curbs, when finished, shall be bonded to the surface of the top to form a square joint. Tops shall be 32 mm thick.

(A) CANCELLATION

APPROVED 27 JUL 1983 REVISED (A) 21 AUG 1991

USER ACTIVITIES: ARMY - MD, OM
NAVY - MS

REVIEWER ACTIVITIES: ARMY - ME
NAVY - OS
AIR FORCE - 82

AMSC N/A
DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

This military standard is approved for use by all Departments and Agencies of the Department of Defense. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

P.A. EA	INTERNATIONAL INTEREST	TITLE	MILITARY STANDARD
Other Cust MS 99		TABLE, WALL, LABORATORY	MS51426
Procurement Specifications NONE	SUPERSEDES:		SHEET 1 OF 2

DD FORM 1 MAY 73 672 (COORDINATED) PREVIOUS EDITIONS OF THIS FORM ARE OBSOLETE

CANCELLATION

Project Number 6640-1726-04

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T18:01Z
Check the source to verify that this is the current version before use.

Type E top and curb shall be made of stainless steel and shall have steel channel reinforcing members welded in place and the underside shall be coated with sound-deadening material. Exposed ends shall be of double-wall stainless steel construction. Factory-made joints, exposed ends and corners shall be continuous heliarc welded, ground and polished smooth. The resultant exposed surface shall have a No. 4 finish. Where required, hairline field joints shall be provided for bolting together at installation site.

The curbs at the rear of all table tops are 102 mm high.

NOTES:

1. Complete part number for the table top shall include the part number as shown in the table, to specify the size, and the type letter to specify the top; such as, MS51426-1A for a table 183 cm long with a type A top.

2. To specify a complete wall table, the table top number and numbers for all the base units needed must be included.

3. Items differing from this Military Standard by minor variations in design or dimensions are considered standard under this document, provided such variations do not affect performance and are not in conflict with the material requirements specified herein.

4. For design feature purposes, this standard takes precedence over procurement documents referenced herein.

5. Referenced documents shall be of the issue in effect on date of invitations for bid.

6. FSCMs: 03534, 14095, and 22527.

USER ACTIVITIES:

REVIEWER ACTIVITIES:

APPROVED REVISED

This military standard is approved for use by all Departments and Agencies of the Department of Defense. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

P.A. EA	INTERNATIONAL INTEREST	TITLE	MILITARY STANDARD
Other Cust MS 99		TABLE, WALL, LABORATORY	MS51426
Procurement Specifications NONE	SUPERSEDES:		SHEET 2 OF 2

DD FORM 1 MAY 73 672 (COORDINATED) PREVIOUS EDITIONS OF THIS FORM ARE OBSOLETE

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T18:01Z
Check the source to verify that this is the current version before use.