A-A-53927
22 June 1989

COMMERCIAL ITEM DESCRIPTION

INFUSOR, PRESSURE, BLOOD-INTRAVENOUS BAG

The General Services Administration has authorized the use of this commercial item description.

This commerical item description covers expendable pressure infusors suitable for use with intravenous bags of 500 ml and 1000 ml capacity.

Salient characteristics:

General. Shall be an expendable pressure infusor bag for use with invasive blood pressure monitoring and intravenous infusion procedures. Infusor shall be suitable for use in an operating room environment.

Style, design and dimensions. The infusor bag shall be in accordance with Figure 1.

Material. Infusor shall be a polyurethane coated nylon bag with an attached white netting (mesh) pocket of polyester material to enclose a blood or medication bag of the capacity shown on Figure 1.

Construction. Netting shown on Figure 1 shall allow visualization of enclosed bag contents. The braided polyester strap shall be provided to suspend the infusor. The piston type pressure gauge shall be provided at the bottom of the bag with a pressure range of 0 to 400 mm of Hg for measuring the bladder air pressure. The gauge shall be color-coded and numerically marked to indicate infusor air pressure. Pressure markings shall be at least three in number and shall be provided at 0 mm Hg, 150 mm Hg and 300 mm Hg. The gauge shall be clearly visible from 360 degrees. Gauge body shall be fabricated of ABS plastic with an acetal piston.

The bladder and inflation blub shall be connected with approximately 15 inches of vinyl tubing. The inflation bulb shall be fabricated of rubber material and shall be furnished with metal valves fabricated of corrosion-resisting material to allow the bag to be inflated. The stopcock shall be used to lock and release the bladder pressure and shall be fabricated of a polycarbonate body, polyethylene stem and an ABS cap. The air system shall be designed to withstand an air pressure of 600 mm Hg without leakage or other damage. The strap shall withstand a force of 6 lbs. without breaking, fraying, pulling free from the bag or other damage. Force shall be applied to the strap away from and on the longitudinal axis of the bag.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:26Z
Check the source to verify that this is the currentversion before use.

Workmanship. The infusors shall be free from defects which detract from their appearance or impair their serviceability. Infusors shall be free from cuts, tears, discoloration or other damage.

Unit. Each (EA) One infusor, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for Inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one infusor. Sampling shall be inspection level S-3 and an AQL of 2.5.

Tests. Test shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:26Z
Check the source to verify that this is the currentversion before use.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and or order.

NOTE: The following National Stock Numbers are covered by this document:

NSN	ITEM IDENTIFICATION
6515-01-280-8162	INFUSOR, Pressure, Blood-Intravenous Bag, Expendable, 500 ml

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

6515-01-280-8163

INFUSOR, Pressure, Blood-Intravenous Bag, Expendable, 1000 ml

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM


CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-4613

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:26Z
Check the source to verify that this is the current version before use.

NOTE: DIMENSIONS ARE FOR INFLATED INFUSOR (APPROXIMATE)



* 6" Deflated

FIGURE 1 INFUSOR, PRESSURE, BLOOD-INTRAVENOUS

A-A-53927

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2 DOCUMENT TITLE
A-A-53927	INFUSOR, PRESSURE, BLOOD-INTRAVENOUS BAG

3a. NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
b ADDRESS *(Street, City, State, ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)* _____

5 PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6 REMARKS

7a NAME OF SUBMITTER *(Last, First, MI)* – Optional	b WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8 DATE OF SUBMISSION *(YYMMDD)*

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

DD FORM 1426 PREVIOUS EDITION IS OBSOLETE

(FOLD ON THIS LINE)

INSTRUCTIONS In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO. 4966 Alexandria, VA

POSTAGE WILL BE PAID BY

Commander
Defense Personnel Support Center
ATTN: DPSC-RSTS
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

NO POSTAGE
NECESSARY
IF MAILED
IN THE
UNITED STATES

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:26Z
Check the source to verify that this is the current version before use.