A-A-53934
24 July 1989

COMMERCIAL ITEM DESCRIPTION

BAG, DENTAL PROSTHESIS (Plastic)

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for Military Specification MIL-P-37928 (DM), which is cancelled.

Salient characteristics:

This commercial item description covers polyethylene bags suitable for storage and shipment of dental prostheses. The bags shall be supplied in quantities of 100 as size 1 or in quantities of 1000 as size 2.

Material. The plastic bags shall be made from polyethylene film material in accordance with ASTM Designation D2103. Material shall be clean, non-colored (essentially colorless) and transparent. The film shall be manufactured by a properly controlled extrusion process to a minimum thickness of 0.0030 inches.

Design and construction. The plastic bags shall be used for storage and shipment of dental prostheses in a moisture-resistant closure to protect them from dehydration. The plastic bags shall be formed by folding one sheet and shall consist of two side seams. The edges of the side seams shall be parallel and all seams shall be parallel with the outer edges of the bag. The seams shall be sealed by normal heat sealing at a 400 to 600 degree Fahrenheit (^{0}F) temperature range. The opening shall be on the short axis of the bag and shall be identified by a red colored stripe of permanent polyethylene extrusion located on the edge of the interlocking seal flap. The bag shall have a transparent surface which will permit ready identification of its contents.

Interlocking seal. The interlocking seal on the bags shall be a single or double track zipper. The interlocking seal shall be positioned not less than 1/2 inch and not more than 3/4 inch, from the top or opening and shall be heat sealed to the plastic bag film by the same heat sealing process specified for the side seams. As an alternate, the zipper may be an integral part of the plastic film forming the bag. The seal shall be straight and continuous and shall be effected by applying pressure with the fingers to the seal area. The plastic interlocking seal shall permit positive repeated opening and closing operations without any resulting impairment to the interlocking feature.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

Dimensions. The plastic bag shall have uniform inside dimensions, exclusive of seals or seams, of four (4) inches in width and six (6) inches in length, held to tolerances of plus 1/8 inch or minus 1/16 inch.

Seam strength. Seam specimens shall withstand a static load of 3.5 pounds without peel back or slipping of seams when tested as specified herein.

Instructions for use. Instructions for use shall be included with each unit box. The instructions shall be printed on a separate sheet of white bond or offset paper, measuring 6 ± 1/2 inches by 4 ± 1/2 inches. The printing shall show no smearing or bleeding and shall be clearly legible. The following instruction shall be printed on the paper:

"Designed to accommodate delivery of finished dentures in a moist atmosphere. Bag may be repeatedly opened and sealed by simple finger pressure. No heat sealing is required. Adequate moisture may be provided by inserting a damp cotton ball or other absorbent material, or at most, a teaspoon of water. Avoid excess water as this may cause slight leaking at the corners of the seal".

Workmanship. The bags shall be uniformly constructed and shall be free from tears, cracks, cuts, creases, blisters, wrinkles or other imperfections which may impair their usefulness or interfere with the purpose for which the bags are intended.

Unit. Package (PG).

For size 1. One package containing one hundred bags, as specified, constitutes one unit.

For size 2. One package containing one thousand bags, as specified, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item. For purposes of sampling, an inspection lot for examination and tests shall consist of all bags made by the same process from the same component material by one manufacturer and submitted for delivery at one time.

Certificates of quality. When available, certificates of quality, supplied by the manufacturer of the component or material, listing the specified test method and test results obtained, may be furnished in lieu of actual performance of such testing by the contractor.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105. Unit of product for examination purposes shall be one bag except where specified otherwise.

Examination of count per unit container. The sample unit for this examination shall one unit package (unit of issue) of bags. The average count per unit package shall be not less than the specified or indicated quantity.

End item examination. The bags shall be examined for defects including, but not limited to those listed below:

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

Defects

Bag not fabricated from material specified.
Dimensions not as specified.
Tears, cuts, or holes. Presence of scuff marks, chafed spots, sharp wrinkles or creases in body or sealed areas.
Bags stick together to the extent that opening causes tearing or injury to any surface.
Seams not uniform and in alignment; not completely sealed throughout their entire area.
Seams other than specified. Not clean.
Not of uniform construction; open end not along one shorter side. Seams not straight or continuous.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105. The sample unit shall be one bag.

Tests. Tests shall be conducted to determine compliance with this document and shall be performed on samples randomly selected from each lot of bags offered for Government acceptance.

Seam strength. One hour after sealing, one 1-inch wide specimen shall be cut perpendicular to the line of the seam, from each seam of each sample bag. One leg of the specimen shall be so clamped that the specimen hangs freely. A load of 3-1/2 pounds shall be applied to the free end in such a manner as to avoid impact loading the seam. The load shall be allowed to act for 5 minutes, after which the seam shall be examined for slipping or peel back. The test shall be conducted at room temperature. (For reference purposes, room temperature shall be defined at 70° to 80°F). Peel back shall be defined as any separation of the seam from the inside line of contact of the sealer.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producer's own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch/pounds units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

National Stock Number	Item Identification
6520-00-926-9041	BAG, DENTAL PROSTHESIS, Plastic, 100s
6520-01-181-6972	BAG, DENTAL PROSTHESIS, Plastic, 1000s

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

Ordering data. Intermediate/exterior package quantities, labeling and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS

PROJECT NO. 6520-2287
LOCATION: NCR DISK 2

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-52034	BAG, DENTAL PROSTHESIS (PLASTIC)

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)*: _____

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE. This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
Attn: DPSC-BSTE
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.