A-A-53946
31 August 1989

COMMERCIAL ITEM DESCRIPTION

SCISSORS, ORAL SURGICAL (Goldman-Fox)

The General Services Administration has authorized the use of this Commercial Item Description.

Salient characteristics.

This commercial item description covers the following two oral surgical scissors suitable for cutting mucoperiosteal tissue in periodontal surgery.

No. 16

No. 16M

General. The scissors shall be in accordance with the applicable requirements as specified herein.

Material. The scissors shall be fabricated of class 4, type 420X martensitic corrosion-resisting steel, as specified in ASTM F899 for composition only.

Hardness. The scissors, after proper stress relieving, shall have a Rockwell hardness of C50-C55, as specified in ASTM E18.

Style, design and dimensions. The scissors shall be Goldman-Fox No. 16 and a modified Goldman-Fox No. 16M design, with one blade serrated with a number 2 file cut. The style, design and dimensions shall be in accordance with Defense Personnel Support Center (DPSC) Drawing No. 24578. The scissors shall be of a three piece construction consisting of two scissor sections locked together by a screw, which is permanently set in a fixed position. The screw lock and scissor sections shall be properly assembled to provide a cutting portion which shall be free of both excessive stiffness and looseness. The inside of the cutting portion of each scissor shall be ground to properly mate with the opposing scissor during cutting procedure, and the cutting edges of the blades shall be properly sharpened.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:27Z
Check the source to verify that this is the current version before use.

Finish. The scissors shall have a satin finish, except for the ground inside portion of the blades which shall have a smooth finish, with polishing marks in one direction, perpendicular to cutting edge. Finish shall be equivalent to that produced with grit No. 240.

Corrosion resistance. The scissors shall meet the boil and copper sulfate test requirements as specified in ASTM F1089.

Identification markings. Scissors shall be legibly and permanently marked with the name or registered trademark of the manufacturer. Instruments of foreign manufacture shall be permanently and legibly marked with the name of the country of origin, together with the name or registered trademark of the contractor and manufacturer. The instruments shall have the word "Stainless" or "SS" marked on the handle.

Cutting edges. The scissors shall cut a continuous length of test material of at least 3/4 inch, as specified herein.

Workmanship. The scissors shall be free from pits, marks, burrs, sharp edges, cracks or other imperfections which may detract from their appearance or impair their serviceability, proper functioning, and intended use.

Unit. Each (EA). One scissors, as specified, constitutes one unit.

Unit package. Each pair of scissors shall be wrapped in anti-corrosive tissue paper, then enclosed in a waterproof paper bag.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as pecified herein. Except as otherwise specified in the contact or purchase order, the contractor may use his own or ny facilities suitable for the performance of the inspection reqirements specified herein, unless disapproved by the Governent. The Government reserves the right to perform any of th inspections set forth in the specification where such inspectins are deemed necessary to assure supplies and services conform o prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:27Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection characteristics. The unit of product for inspection purposes shall be one pair of scissors. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, and sampling for tests (when applicable) shall be conducted at inspection level S-1. For tests, acceptance number shall be zero (0) for all sample sizes.

Examination. The scissors shall be examined to determine compliance with requirements contained in this commercial item description. Examination shall include but shall not be limited to the following:

Defects.

Style and design not as specified.
Cutting action of scissors not free of excessive stiffness or looseness.
One blade not uniformly filed cut.
Screw in lock not permanently set in a fixed position.
Screw lock not properly assembled.
Cutting edges not free of nicks, feathers, and burrs.
Scissors not free of corrosion, discoloration, pits, and burrs.
Scissors not free of cracks.
Cutting edge of scissors not uniformly sharp.
Finish not satin (excluding crocus - finished sharpened surfaces).
Finish not free of coarse grind marks or scratched surface areas.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:27Z
Check the source to verify that this is the current version before use.

Defects (continued).

Identification marking not permanent, legible, or complete.
Dimensions not within tolerance shown in drawing.

Tests.

Chemical analysis. In the absence of certificates of quality for the raw material, chemical analysis shall be conducted in accordance with method 111 or method 112 of FED-STD-151. In case of dispute, method 111 shall govern.

Scissors performance test. The test material shall be dental rubber dam material, medium thickness (0.007 to 0.009 inch) and in accordance with Federal Specification ZZ-R-690. Scissors to be tested shall be cleaned prior to test performance. Three separate, consecutive tests shall be performed for the scissors. Each test shall consist of making one cut through three plies of test material. Each cut shall be made using the scissors in the normal manner (with the examiner's fingers in the finger rings of the instrument), except that no lateral pressure shall be exerted.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch/pound units, providing they fall within the tolerance specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerance specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:27Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquistion Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

National Stock Number	Item Identification
6520-01-046-8418	SCISSORS, ORAL SURGICAL, Half Curved, No. 16, 5 inch
6520-01-292-1058	SCISSORS, ORAL SURGICAL, Double Curved, No. 16M, 5 inch

Ordering data. Intermediate/exterior package quantities and labeling and marking must be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:27Z
Check the source to verify that this is the current version before use.

A-A-53946

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

CIVIL AGENCY COORDINATING ACTIVITIES:
VA-OSS
PHS
FDA-MPQAS

48995

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

PROJECT NO. 6520-2079

LOCATION: NCR Disk No. 7

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:27Z
Check the source to verify that this is the current version before use.