A-A-53948
7 August 1989

COMMERCIAL ITEM DESCRIPTION

TRIPOD, LABORATORY APPARATUS

The General Services Administration has authorized the use of this commercial item description in preference to Federal Specification NNN-T-698.

This commercial item description covers iron ring tripods suitable for use in laboratory procedures.

Salient characteristics.

The tripods shall consist of an outer ring whose under surface shall be provided with three equally spaced threaded bosses which shall accommodate three threaded detachable legs. The tripods shall be available in Style A, consisting of a single ring and Style B, which shall consist of an outer ring plus a designated number of loose concentric inner rings.

Style A tripods shall be available in five sizes with the following nominal dimensions in inches:

Outside Diameter of Ring	Inside Diameter of Ring	Overall Height	Diameter of Legs
3 5/8	2 1/2	5 to 6	5/16
5	3	9	5/16
6	4 1/2	9	5/16
8	6	9	5/16
10	8	9	5/16

AMSC N/A FSC 6640

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:27Z
Check the source to verify that this is the current version before use.

Style B tripods shall be available in four sizes with the following nominal dimensions in inches, plus the number of concentric inner rings available for each size:

Outside Diameter of Outer Ring	Inside Diameter of Smallest Ring	Overall Height	Diameter of Legs	Number of Loose Rings
5	2	9	5/16	1
6	2	9	5/16	2
8	2	9	5/16	3
10	2	9	5/16	4

The rings shall be fabricated of a suitable malleable or cast iron material. The top surface shall be flat and level with a 1/8 to 9/32 inch inset flange along the inner border to accept loose concentric rings when required. The under surface of the outer ring only, shall be provided with three equally spaced threaded bosses, bored and tapped to accommodate the threaded detachable legs. Each boss shall have at least five complete threads. The 3 5/8 inch Style A tripod is not required to have the specified inset flange.

The legs shall be fabricated of coppered steel, wrought iron, zinc plated steel or comparable material, nominally 5/16 inch in diameter and threaded on one end with at least six full threads to fit the threaded bosses of the outer rings. The legs shall extend from the bosses at such an angle to insure dependable stability.

The rings shall have a black baked enamel, grey hammertone, or other suitable durable finish. Legs shall have a black baked enamel finish or a cadmium, copper, zinc or other suitable plating for a durable pleasing appearance.

Workmanship. The tripods shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Each (EA). One tripod, as specified, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:27Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for inspections and tests. Sampling for inspections and tests shall be conducted in accordance with commercial practice. The unit of product for test purposes shall be one tripod.

Inspections and tests. Inspections and tests shall be conducted in accordance with commercial practice to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:27Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/ contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6640-00-445-1000

Ordering data. (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:27Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS

Project No. 6640-1575

Location: CID/TRIPOD/S38/17

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:27Z
Check the source to verify that this is the current version before use.