A-A-53951
28 July 1989

# COMMERCIAL ITEM DESCRIPTION

## TOURNIQUET SYSTEM, PNEUMATIC

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a complete tourniquet system that shall maintain a bloodless field or controlled flow of blood to extremities during orthopedic, plastic or general surgery.

Salient characteristics.

Shall be a surgical, automatic, electrically powered, pneumatic tourniquet system that uses ambient air for tourniquet cuff inflation. For use in maintaining a bloodless field or controlled flow of blood to extremities during orthopedic, plastic or general surgery.

The system shall consist of the following items:

One microprocessor controlled pneumatic tourniquet machine.

Two dual hose assemblies.

Five dual port, single bladder cuffs - one each of the following lengths: 8 inch, 12 inch, 18 inch, 24 inch and 34 inch.

Two dual port, dual bladder cuffs - one each of the following lengths: 18 inch and 24 inch.

One telescoping tourniquet stand with a five leg base.

One calibration kit.

Two Operator and Service Manuals.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:27Z
Check the source to verify that this is the current version before use.

The microprocessor controlled pneumatic tourniquet machine shall have the following characteristics:

Operating voltage of 110/220 volts, 50/60 Hz. Minimum 10 foot electric power cord with hospital grade plug. An external auxilliary transformer as required for dual voltage operation.

A cuff pressure range of 50-475 mm Hg.

Pressure accuracy of ± 5 mm Hg (T=15-25 degree C).

Pressure regulation of ± 6 mm Hg of Set Point. (10 second average under not transient conditions without external leaks).

Time alarm set range of 0-240 minutes at 1 minute increments.

Timer accuracy of 0.25% ± 1 second.

Rechargeable 12 volt battery with a minimum discharge time of 15 minutes, with a fully charged battery.

Controls - power switch, pressure switches, time switches, cuff switches and alarm silence switch.

Displays - digital pressure display, digital time display, battery charging indicator, main cuff indicator, second cuff indicator and alarm silence indicator.

Alarm conditions, both visual and audible - Over Pressure, Under Pressure, Exceeded Set Time, Sustained Leak, Battery Voltage Low, Cuff Not Deflated and Hardware Failure.

Sequential Instructions for use clearly printed on the face of the machine.

The dual air hose assemblies shall be a minimum of 10 feet long, and have male Luer-lock connections to cuffs. May be color coded for easy identification.

The dual port, single bladder tourniquet cuffs shall be contoured to fit a patients limb for patient safety and better cuff performance. Connectors shall be female Luer-lock type for secure connection without leakage of air. Shall contain both ties and hook-and-pile closures.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:27Z
Check the source to verify that this is the current version before use.

The dual port, dual bladder tourniquet cuffs shall be contoured to fit a patients limb for patient safety and better cuff performance. Shall be clearly marked cuff "A" and cuff "B" for correct connection to the dual air hose assemblies. Connections shall be female Luer-lock type for secure connections without leakage of air. Shall contain both ties and hook-and-pile closures.

Shall have a metal telescoping stand with a five leg base capable of holding the tourniquet machine. The stand shall telescope to at least six feet high. Shall be fabricated of a suitable corrosion-resistant metal alloy that has been properly hardened and passivated. The base shall have all direction castors.

Shall contain a calibration kit compatible with the machine.

Shall contain two Operator and Service Manuals.

Workmanship. The tourniquet system shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One tourniquet system, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:27Z
Check the source to verify that this is the current version before use.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Inspections and tests. Inspections and tests shall be conducted in accordance with commercial practice to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for inspections and tests. Sampling for inspections and tests shall be conducted in accordance with commercial practice. The unit of product for test purposes shall be one complete tourniquet system.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:27Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6515-01-287-0607

Ordering data. (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:27Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-4625

Location: CID/TOURNIQUET/SYSTEM/S38/16



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:27Z
Check the source to verify that this is the current version before use.