A-A-53977
25 August 1989

COMMERCIAL ITEM DESCRIPTION

PIPET, VOLUMETRIC (Micro, Lang-Levy)

The General Services Administration has authorized the use of this Commercial Item Description in preference to Federal Specification NNN-P-392 as applicable to type I pipets.

This commercial item description covers borosilicate glass Lang-Levy design volumetric pipets.

Salient characteristics.

Shall be volumetric micropipets of the Lang-Levy design for use in laboratory procedures.

Shall be fabricated from borosilicate glass conforming to requirements of ASTM Designation E438.

Shall be of the following single volume microliter capacities (see "Ordering Data"): 10, 20, 50.

Shall be calibrated "to deliver" rated capacity with a ± 5% volumetric accuracy at 20 degrees centigrade.

Shall be constructed so that bottom end includes a tapered angle bend delivery tip, and the bore includes a constriction. The bore constriction shall be located at the rated capacity level, so that rated capacity is contained between the constriction and delivery tip. The capacity level denoted by the constriction shall be calibrated "to deliver" at 20 degrees centigrade when a pipet is held in a vertical position and blown.

Shall have an overall length of 145 to 150 millimeters (mm).

Shall be permanently and legibly marked with the manufacturer's name or trademark and the rated capacity.

AMSC N/A FSC 6640

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:27Z
Check the source to verify that this is the current version before use.

Shall be tested for volumetric accuracy to meet the requirements of the College of American Pathologists (see Quality Assurance Provisions).

Workmanship. The pipets shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Package (PG). One package containing twelve pipets of the size specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for inspections and tests. Sampling for inspections and tests shall be conducted in accordance with commercial practice.

Inspections and tests. Inspections and tests shall be conducted in accordance with commercial practice to determine compliance with specification requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:27Z
Check the source to verify that this is the current version before use.

Volumetric accuracy testing. To meet requirements for certification of the College of American Pathologists, the pipets shall be tested in accordance with ASTM Designation E542 to determine conformance to calibration and capacity tolerance requirements. A Certificate of Quality may be furnished in lieu of actual performance of individual testing provided that lot/sample identities can be maintained. The Certificate of Quality shall include the lot number, lot size, sample size and test method.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/ contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:27Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Other publications. The following documents form a part of this specification to the extent that is specified herein. Unless otherwise indicated, the effect on date of invitation for bids or request for proposals shall apply.

AMERICAN SOCIETY FOR TESTING AND MATERIALS (ASTM)

ASTM E438 STANDARD SPECIFICATION FOR GLASSES IN LABORATORY APPARATUS

ASTM E542 PRACTICE FOR CALIBRATION OF VOLUMETRIC WARE

(Application for copies should be addressed to American Society for Testing and Materials, 1916 Race Street, Philadelphia, PA 19103)

NOTE: The following National Stock Numbers are covered by this document:

NSN	MICROLITER CAPACITY	ITEM DESCRIPTION
6640-00-127-0134	10	PIPET, VOLUMETRIC, Micro, Lang-Levy, 10 Microliter, 12's
6640-00-127-0135	20	PIPET, VOLUMETRIC, Micro, Lang-Levy, 20 Microliter, 12's
6640-00-127-0136	50	PIPET, VOLUMETRIC, Micro, Lang-Levy, 50 Microliter, 12's

Ordering data. (Capacity required, intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:27Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6640-1580

Location: CID/PIPET/2/S31/17

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:27Z
Check the source to verify that this is the current version before use.