A-A-53978
15 September 1989

COMMERCIAL ITEM DESCRIPTION

POINT, ENDODONTIC, CORE, Gutta-Percha

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for MIL-P-0037232A(DM), 16 October 1985 which is cancelled.

This commercial item description covers core (standardized) gutta-percha endodontic points.

Salient Characteristics:

General. The core gutta-percha endodontic points shall be in accordance with American National Standards Institute (ANSI)/American Dental Association (ADA) No. 57, Type 1 Class 2 for Endodontic Filling Materials with the following additions and exceptions. The points shall be of the following sizes: 25, 30, 35, 40, 45, 50, 55, 60, 70, 80, 90, 100, 110, 120, 130, 140.

Material. The polymeric material shall be high grade gutta-percha or balata which shall be washed and treated to remove impurities and provide a uniform, clean, non-toxic point. A filler containing barium sulfate for radiopacity shall be included in the point. All ingredients, including color additive and the filler shall be uniformly distributed throughout the point.

Style. The points shall be designed for use as a core point in filling pulp canals in endodontic precedures. The points shall have an overall length of 30 mm ± 2 mm. The non-operative end of the point shall be flattened and shall have a serrated or grooved surface. The flattened end shall be 3 ± 1 mm long and shall have a thickness equal to or greater than 1/2 the diameter of the point 3 mm from the tip. The point shall be machine rolled or hand rolled.

Working properties. The gutta-percha points furnished shall be of such formulation and process as to provide a complete sealing of the prepared root canal when used with an accepted root canal sealer as defined by ANSI/ADA No. 57 in endodontic procedures. The working properties of the points shall exhibit conhesive stength, toughness and flexibility without brittleness.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:27Z
Check the source to verify that this is the current version before use.

Certification-acceptability. The points shall be certified or acceptable by the American Dental Association. At time of award of contact, the manufacturer shall furnish, upon request, a certificate and data showing that the brand of material to be furnished is the same in quality and brand as that appearing on the list of acceptable or certified materials. Compliance with this certification requirement does not absolve the manufacturer from complete compliance with the other requirements of this specification.

Color. The color of the points shall be pink.

Radiopacity. The radiopacity requirement of ANSI/ADA No. 57 shall apply. Testing for radiopacity shall be as specified under tests.

Delivery schedule. No more than 2 months shall have elasped from date of manufacturer of the points to date of delivery to the Government.

Shelf-life. The points shall comply with requirements of this specification for no less than 18 months from date of delivery, when stored at 35 to 46 degrees fahrenheit. The inspection test date shall be 18 months.

Toxicity. The manufacturer shall submit a statement that the points have no toxic effects on the normal, healthy person when used for the purpose intended.

Workmanship. Points shall be free of curled or twisted portions and shall have a smooth surface free from foreign matter and defects. The points shall show no defects which detract from their appearance or impair their serviceability.

Unit. Package (PG). One package containing one hundred endodontic points of one size only, as specified, constitutes one unit.

Unit package. The points shall be packaged in a rigid container designed to prevent damage during shipment or storage.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:27Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 at an inspection level of S-4. The unit of product for examination shall be one point for each package sampled.

Examination. The points shall be examined to determine compliance with the requirements of this specification herein. Points shall be examined for defects, including but not limited to, those listed below:

Defects

Points are not clean.
Style and design not as specified.
Color not as specified.
Tapered length surface not smooth rolled finish.
Laminations cracks or voids present.
Twisted or curled sections present.
Size not as specified.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:27Z
Check the source to verify that this is the currentversion before use.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 at an inspection level of S-1. The unit of product for test purposes shall be one point from each package sampled.

Tests. Tests to include dimensions shall be conduct to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics. There shall be two or more tests determinations per lot or batch number. Each determination shall be an average of specimens when averaging is specified in testing procedure.

Dimensions. The points shall be tested for dimensions in accordance with ANSI/ADA No. 57.

Radiopacity. The points shall be tested for radiopacity in accordance with ANSI/ADA No. 57. The test for radiopacity shall be performed on the first lot or batch of a contact or purchase order. For the remaining lots or batches of the same contract or purchase order, the contractor may provide certification of compliance with the radiopacity requirement specified herein, provided that the remaining lots or batches are composed of the same formulation and manufactured identically as the first lot or batch originally tested for radiopacity.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:27Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

NATIONAL STOCK NUMBER	Item Identification
6520-00-583-3388	POINT, ENDODONTIC, CORE, Gutta-Percha, No. 25, 100s
6520-00-257-3030	POINT, ENDODONTIC, CORE, Gutta-Percha, No. 30, 100s
6520-00-596-7290	POINT, ENDODONTIC, CORE, Gutta-Percha, No. 35, 100s
6520-00-257-3031	POINT, ENDODONTIC, CORE, Gutta-Percha, No. 40, 100s
6520-00-596-7292	POINT, ENDODONTIC, CORE, Gutta-Percha, No. 45, 100s
6520-00-257-3032	POINT, ENDODONTIC, CORE, Gutta-Percha, No. 50, 100s
6520-00-596-7311	POINT, ENDODONTIC, CORE, Gutta-Percha, No. 55, 100s
6520-00-257-3033	POINT, ENDODONTIC, CORE, Gutta-Percha, No. 60, 100s
6520-00-257-3036	POINT, ENDODONTIC, CORE, Gutta-Percha, No. 70, 100s
6520-00-257-3037	POINT, ENDODONTIC, CORE, Gutta-Percha, No. 80, 100s
6520-00-583-3389	POINT, ENDODONTIC, CORE, Gutta-Percha, No. 90, 100s

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:27Z
Check the source to verify that this is the current version before use.

6520-00-583-3402	POINT, ENDODONTIC, CORE, Gutta-Percha, No. 100, 100s
6520-00-061-6448	POINT, ENDODONTIC, CORE, Gutta-Percha, No. 110, 100s
6520-00-061-6449	POINT, ENDODONTIC, CORE, Gutta-Percha, No. 120, 100s
6520-01-061-6450	POINT, ENDODONTIC, CORE, Gutta-Percha, No. 130, 100s
6520-01-061-6451	POINT, ENDODONTIC, CORE, Gutta-Percha, No. 140, 100s

Ordering data. Intermediate/exterior package quantities and labeling and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

CUSTODIANS:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No.: 6520-2294
Location: NCR Disk 1

05151

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:27Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE: This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO. 4966 WASHINGTON D. C.
POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC- RSTE
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:27Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-53978	POINT, ENDODONTIC, CORE, Gutta-Percha

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)*:

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil – Downloaded: 2020-10-20T09:27Z
Check the source to verify that this is the current version before use.