GG-B-001294 (DSA-DM)
13 February 1967

# INTERIM FEDERAL SPECIFICATION

## BURNISHER, DENTAL

This interim Federal specification was developed by the Defense Personnel Support Center, 2800 South 20th Street, Philadelphia, Pennsylvania 19101, based upon currently available technical information. It is recommended that Federal agencies use it in procurement and forward recommendations for changes to the preparing activity at the address shown above.

## 1. SCOPE AND CLASSIFICATION

1.1 Scope. This specification covers burnishers suitable for use in burnishing plastics, inlay wax, and gold.

1.2 Classification. The burnishers covered by this specification shall be the following types and styles, as specified (see 6.1).

Type I. - Single end.

Style 1. - No. 29.
Style 2. - No. 25.
Style 3. - No. 27.

Type II - Double end.
Style 4. - No. 15 or "J".

FSC 6520

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

2. APPLICABLE DOCUMENTS

2.1 Specifications, standards, and drawings. The following specifications, standards, and drawings of the issue in effect on date of invitation for bids or request for proposal, form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

| | |
|---|---|
| GG-I-526 | Instruments, Dental and Surgical; General Specifications. |
| PPP-B-566 | Boxes, Folding, Paperboard. |
| PPP-B-585 | Boxes, Wood, Wirebound. |
| PPP-B-601 | Boxes, Wood, Cleated-Plywood. |
| PPP-B-621 | Boxes, Wood, Nailed and Lock-Corner. |
| PPP-B-636 | Box, Fiberboard. |
| PPP-B-676 | Box, Setup. |

Military

| | |
|---|---|
| MIL-P-116 | Preservation, Methods of. |
| MIL-L-10547 | Liners, Case, and Sheet, Overwrap; Water-Vaporproof, or Waterproof, Flexible. |

STANDARDS

Federal

| | |
|---|---|
| FED-STD-123 | Marking for Domestic Shipment (Civilian Agencies). |

Military

| | |
|---|---|
| MIL-STD-105 | Sampling Procedures and Tables for Inspection by Attributes. |
| MIL-STD-129 | Marking for Shipment and Storage. |

DRAWINGS

DEFENSE PERSONNEL SUPPORT CENTER

| | |
|---|---|
| 20088 | BURNISHER AND CARVER, DENTAL |
| 20540 | BURNISHERS, DENTAL. |



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

(Activities outside the Federal Government may obtain copies of Federal Specifications and Standards as outlined under General Information in the Index of Federal Specifications and Standards and at the prices indicated in the Index. The Index, which includes cumulative monthly supplements as issued, is for sale on a subscription basis by the Superintendent of Documents, U.S. Government Printing Office, Washington, D.C., 20402.

(Single copies of this specification and other product specifications required by activities outside the Federal Government for bidding purposes are available without charge at the General Services Administration Regional Offices in Boston, New York, Atlanta, Chicago, Kansas City, Mo., Dallas, Denver, San Francisco, Los Angeles, Seattle, and Washington, D.C.

(Federal Government activities may obtain copies of Federal Specifications and Standards and the Index of Federal Specifications and Standards from established distribution points in their agencies.)

(Copies of Military specifications, standards, and drawings required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 General requirements. The burnishers shall be in accordance with applicable requirements of GG-I-526 and as specified herein.

3.2 Material. The burnishers shall be Class 4, Type 440A stainless steel.

3.3. Hardness. After proper stress relieving, the operative ends shall have a Rockwell hardness of C52 to C56. The handles shall have a minimum Rockwell hardness of C20.

3.4 Style, design, and dimensions. Style, design, and dimensions of Types I and II burnishers shall conform to Defense Personnel Support Center Drawings 20088 and 20540, as applicable. Burnishers shall be suitable for burnishing plastic materials, waxes, and dental restorative materials.

3.5 Finish. Shanks and operative ends of burnishers shall have a mirror finish. Type I handles shall have a bright polished finish. Type II handles shall have a rough emery grind or fine knurled finish.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

3.6 Handle serrations. Handles of Type I burnisher shall have 40 ± 10 fillet or notch type serrations per inch. Depth of serrations shall be 0.002 to 0.004 inch.

3.7 Identification marking. Burnishers shall be permanently and legibly marked in accordance with GG-I-526. Type I, styles 1 through 3 burnishers shall be marked with the respective instrument number at approximately 10 millimeters from the proximal end. Type II, style 4 burnishers shall be marked with the letter "J" or number "15".

3.8 Workmanship. Workmanship shall be first class throughout. The burnishers shall be free from defects which detract from their appearance or may impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified, the supplier may utilize his own facilities or any commercial laboratory acceptable to the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties).

4.1.2 Certificates of quality. When available, certificates of quality, supplied by the manufacturer of the component or material, listing the specified test method and test results obtained, may be furnished in lieu of actual performance of such testing by the contractor.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105. The Acceptable Quality Levels (AQL's) shall be as indicated in table I.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

Table I. Acceptable Quality Levels

| Type Defect | Inspection Level | AQL (Percent defective) |
|---|---|---|
| Major | II | 1.0 |
| Minor A | II | 2.5 |
| Minor B | S-2 | 4.0 |

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105, and as indicated in table II. The unit of product for inspection purposes shall be one burnisher.

Table II. Sampling Criteria

| Characteristic | Requirement | Test procedure | Insp. level | AQL (Percent defective) |
|---|---|---|---|---|
| Hardness | 3.3 | GG-I-526 | S2 | * |
| Copper sulfate test | 3.2 | GG-I-526 | S2 | * |
| Boil test | 3.2 | GG-I-526 | S2 | * |

* Acceptance number shall be zero for all sample sizes.

4.2.3 Raw materials. A sufficient quantity of each lot of each metal used shall be selected for chemical analysis.

4.2.4 Where feasible, the same unit of product may be used for the determination of two or more test characteristics.

4.3 Examination. The burnishers shall be examined to determine compliance with requirements contained in this specification. Examination shall be conducted as specified in 4.2.1. Any burnisher in the sample containing one or more defects shall be rejected, and if the number of defective burnishers in any sample exceeds the acceptance number for that sample, the lot represented by that sample shall be rejected. Inspection shall not be restricted to the defects in table III.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

Table III. Classification of defects in accordance with MIL-STD-105.

| Categories | Defects |
|---|---|
| Critical: | None defined |
| Major: | |
| 101 | Burnishers not of appropriate design (as shown in drawing for appropriate type specified, see 3.4). |
| 102 | Burnisher working surfaces not free of nicks, feathers, and burrs. |
| 103 | Labels or printed marking on unit container not correct (particularly stock number and item identification). |
| 104 | Burnishers not free of cracks. |
| Minor A: | |
| 201 | Finish not as specified (see 3.5). |
| 202 | Finish not free of coarse grind marks or scratched surface areas. |
| 203 | Identification marking not present or not permanent, legible and complete. |
| 204 | Burnishers not free from dirt, oil, polishing compound or foreign matter. |
| Minor B: | |
| 251 | Dimensions not within tolerances shown in drawing. |

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

5. PREPARATION FOR DELIVERY

5.1 Packaging.

5.1.1 Unit of issue. Each burnisher, as specified, constitutes one unit of issue.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

5.1.2 Packaging quantities. The number of units of issue indicated in the following table shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in the procurement document.

Table IV. Packaging Quantities

| Unit package | Intermediate package | Exterior container |
|---|---|---|
| 1 unit | 12 units | 288 units |

5.1.2.1 Packing variation permitted. If the required number of units to be shipped is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

5.1.3 Level A.

5.1.3.1 Cleaning and drying. Each unit shall be thoroughly cleaned and dried in accordance with the applicable method as outlined in MIL-P-116.

5.1.3.2 Preservative. Each unit shall be preserved with type P-9 preservative in accordance with MIL-P-116.

5.1.3.3 Intimate wrap. Each unit shall be wrapped with barrier material as specified in MIL-P-116.

5.1.3.4 Unit package. Each unit shall be packaged in accordance with MIL-P-116, method 1C-2.

5.1.3.5 Intermediate package. Intermediate package shall be a box of appropriate size constructed in accordance with PPP-B-566 or PPP-B-676, except commercial colors will be acceptable, or PPP-B-636, type CF, class domestic. Closure shall be adequate to prevent accidental opening under normal handling.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

5.1.4 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing.

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for type 2 load, and constructed in accordance with PPP-B-585, style 3, for class 3 use; PPP-B-601, table II, using type III, class 1 plywood; PPP-B-621, table III; or PPP-B-636, type CF, class weather-resistant.

5.2.1.2 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes.

5.2.1.3 Closure. Closure of wood boxes shall be in accordance with appendix of applicable box specification. Closure of each fiberboard box shall be as specified in the appendix, and waterproofing shall conform with 30.4, of PPP-B-636.

5.2.1.4 Strapping. Strapping, when required, shall be in accordance with appendix of applicable box specification.

NOTE: Strapping shall not be required for shipments forwarded to a receiving activity within the continental limits of the United States for storage and redistribution.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall conform to PPP-B-601, table I, using type III, class 1 plywood; or shall be designed for type 2 load, and constructed in accordance with PPP-B-585, style 3, for class 2 use; PPP-B-621, table I; or PPP-B-636, type CF, class domestic and method I closure shall be utilized.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.



Source: https://assist.dla.mil -- Downloaded: 202640-26T11:53Z
Check the source to verify that this is the currentversion before use.

5.3 Marking.

5.3.1 Civil agencies. Shipments shall be marked in conformance with FED-STD-123.

5.3.2 Military agencies.

5.3.2.1 Interior packages. Each unit package and each intermediate package shall be marked in accordance with MIL-STD-129. When labels are utilized, waterproofing shall be required only when applicable carton is fabricated of water-resistant material. Contract or purchase order number and name of contractor shall be shown.

5.3.2.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

(a) Number and date of this specification.
(b) Type and style required. (see 1.2).
(c) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.2 This specification does not cover all types, classes, grades, or sizes of the commodity indicated by the title of this specification, or those which are commercially available, but is intended to cover the types which are normally procured to meet requirements.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Review activities:
Army - MD
Navy - MS
Air Force - 03

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

# SPECIFICATION ANALYSIS SHEET

Form Approved
Budget Bureau No. 22-R255

INSTRUCTIONS: This sheet is to be filled out by personnel, either Government or contractor, involved in the use of the specification in procurement of products for ultimate use by the Department of Defense. This sheet is provided for obtaining information on the use of this specification which will insure that suitable products can be procured with a minimum amount of delay and at the least cost. Comments and the return of this form will be appreciated. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments and suggestions submitted on this form do not constitute or imply authorization to waive any portion of the referenced document (s) or serve to amend contractual requirements.

SPECIFICATION: GG-B-001294 (DSA-DM)

ORGANIZATION:

CITY AND STATE:

CONTRACT NUMBER:

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE SPECIFICATION CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY SPECIFICATION REQUIREMENT CONSIDERED TOO RIGID

3. IS THE SPECIFICATION RESTRICTIVE?

☐ YES ☐ NO *(If "yes", in what way?)*

4. REMARKS *(Attach any pertinent data which may be of use in improving this specification. If there are additional papers attach to form and place both in an envelope addressed to preparing activity)*

SUBMITTED BY *(Printed or typed name and activity - Optional)*

DATE

DD FORM 1 JAN 66 1426

REPLACES EDITION OF 1 OCT 64, WHICH MAY BE USED.

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:53Z
Check the source to verify that this is the current version before use.

**CAUTION**

**NOTICE TO BIDDERS/OFFERORS**

DO NOT CONDITION OR BASE YOUR BID/OFFER ON ANY CURRENT PROCUREMENT ON THE INFORMATION SUBMITTED ON THIS FORM SINCE ANY CHANGES OR DELETIONS IN THE SPECIFICATIONS MAY RENDER YOUR BID/OFFER NON-RESPONSIVE IN WHICH CASE IT CANNOT BE CONSIDERED FOR AWARD.

Fold

DEFENSE PERSONNEL SUPPORT CENTER
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

DEFENSE SUPPLY AGENCY
OFFICIAL BUSINESS

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

Fold

DPSC FORM 2418
AUG 66

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:53Z
Check the source to verify that this is the current version before use.