GG-C-001304 (DSA-DM)
1 February 1967

# INTERIM FEDERAL SPECIFICATION

## CHISEL, DENTAL, SURGICAL

This interim Federal specification was developed by the Defense Personnel Support Center, 2800 South 20th Street, Philadelphia, Pennsylvania 19101, based upon currently available technical information. It is recommended that Federal agencies use it in procurement and forward recommendations for changes to the preparing activity at the address shown above.

## 1. SCOPE AND CLASSIFICATION

1.1 Scope. This specification covers dental chisels, suitable for surgical procedures in the oral cavity.

1.2 Classification. The chisel shall be of Stout design and the following sizes, as specified (see 6.1).

Size 1. - Stout No. 1
Size 2. - Stout No. 2
Size 3. - Stout No. 3

## 2. APPLICABLE DOCUMENTS

2.1 Specifications, standards and drawing. The following specifications, standards and drawing of the issue in effect on date of invitation for bids or request for proposal, form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

| | |
|---|---|
| GG-I-526 | Instruments, Dental and Surgical; General Specifications. |
| PPP-B-566 | Boxes, Folding, Paperboard. |
| PPP-B-585 | Boxes, Wood, Wirebound. |
| PPP-B-601 | Boxes, Wood, Cleated-Plywood. |

FSC 6520

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:53Z
Check the source to verify that this is the current version before use.

PPP-B-621 Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636 Box, Fiberboard.
PPP-B-676 Box, Setup.

Military

MIL-P-116 Preservation, Methods of.
MIL-L-10547 Liners, Case, and Sheet, Overwrap; Water-Vaporproof, or Waterproof, Flexible.

STANDARDS

Federal

FED-STD-123 Marking for Domestic Shipment (Civilian Agencies).

Military

MIL-STD-105 Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 Marking for Shipment and Storage.

DRAWING

DEFENSE PERSONNEL SUPPORT CENTER

23083 CHISELS, DENTAL, STOUT.

(Activities outside the Federal Government may obtain copies of Federal Specifications and Standards as outlined under General Information in the Index of Federal Specifications and Standards and at the prices indicated in the Index. The Index, which includes cumulative monthly supplements as issued, is for sale on a subscription basis by the Superintendent of Documents, U.S. Government Printing Office, Washinton, D.C., 20402.

(Single copies of this specification and other product specifications required by activities outside the Federal Government for bidding purposes are available without charge at the General Services Administration Regional Offices in Boston, New York, Atlanta, Chicago, Kansas City, Mo., Dallas, Denver, San Francisco, Los Angeles, Seattle, and Washington, D.C.



Source: https://assist.dla.mil -- Downloaded: 202640-26T11:53Z
Check the source to verify that this is the current version before use.

(Federal Government activities may obtain copies of Federal Specifications and Standards and the Index of Federal Specifications and Standards from established distribution points in their agencies.)

(Copies of Military specifications, standards, and drawings required by supplier in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS.

3.1 General. The chisels shall be in accordance with all applicable requirements of GG-I-526, and as specified herein.

3.2 Material. The material shall be Class 4, Type 440A, stainless steel, as specified in GG-I-526.

3.3 Hardness. After proper heat treatment, the operative end of the instrument shall have a Rockwell hardness of C52 to C56. The hardness of the shank shall be suitable for the purpose intended. The handle shall have a Rockwell hardness of C20 minimum.

3.4 Design and dimensions. The chisel shall be of Stout design, of one-piece construction, and shall conform to Defense Personnel Support Center (DPSC) Drawing 23083 for design and dimensions of the applicable size. The cutting instrument shall be suitable for cutting bone or tooth tissue in the oral cavity during surgical procedures.

3.5 Finish. The operative end shall have a mirror or highly polished finish, except for cutting edges and sharpened surfaces of blades which shall have a fine honed or ground finish such as produced by using grade FF emery on a leather wheel. The shank shall have a mirror or highly polished finish. The handle shall have a mirror or satin finish and shall have fillet or notch type serrations in accordance with DPSC Drawing 23083. The serrations shall be smoothly cut or formed and free of rounded edges.

3.6 Corrosion resistance. The cutting instrument shall be properly passivated and shall meet corrosion resistance requirements of GG-I-526 for Class 4, Type 440A stainless steel.



Source: https://assist.dla.mil -- Downloaded: 202610-26T11:53Z
Check the source to verify that this is the current version before use.

3.7 Identification marking. The chisel shall be marked in accordance with GG-I-526 and DPSC Drawing 23083. The marking shall also include Stout's number, corresponding to the size of instrument ordered, which shall be located approximately 1/4 inch from the proximal end.

3.8 Workmanship. In addition to the requirements of GG-I-526, the serrations on the handle shall be free of metal slivers, barbs or flakes produced by manufacturing.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified, the supplier may utilize his own facilities or any commercial laboratory acceptable to the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties).

4.1.2 Certificates of quality. When available, certificates of quality, supplied by the manufacturer of the component or material, listing the specified test method and test results obtained, may be furnished in lieu of actual performance of such testing by the contractor.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105. The Acceptable Quality Levels (AQL's) shall be as indicated in Table I.

Table I. Acceptable Quality Levels

| Type defect | Inspection level | AQL (Percent defective) |
|---|---|---|
| Major A | II | 1.0 |
| Major B | S-1 | 2.5 |
| Minor defects | II | 4.0 |



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

5.3.2.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

(a) Number and date of this specification.
(b) Size required (see 1.2).
(c) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.2 This specification does not cover all types, classes, grades, or sizes of the commodity indicated by the title of this specification, or those which are commercially available, but is intended to cover the types which are normally procured to meet requirements.

Custodians:
Army -MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Review activities:
Army - MD
Navy - MS
Air Force - 03

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

*U.S. GOVERNMENT PRINTING OFFICE: 1967-251-516/4314



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

**CAUTION**

**NOTICE TO BIDDERS/OFFERORS**

DO NOT CONDITION OR BASE YOUR BID/OFFER ON ANY CURRENT PROCUREMENT ON THE INFORMATION SUBMITTED ON THIS FORM SINCE ANY CHANGES OR DELETIONS IN THE SPECIFICATIONS MAY RENDER YOUR BID/OFFER NON-RESPONSIVE IN WHICH CASE IT CANNOT BE CONSIDERED FOR AWARD.

---

*Fold*

**DEFENSE PERSONNEL SUPPORT CENTER**
2300 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

DEFENSE SUPPLY AGENCY
OFFICIAL BUSINESS

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

---

*Fold*

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:53Z
Check the source to verify that this is the current version before use.

# SPECIFICATION ANALYSIS SHEET

Form Approved
Budget Bureau No. 22-R255

INSTRUCTIONS: This sheet is to be filled out by personnel, either Government or contractor, involved in the use of the specification in procurement of products for ultimate use by the Department of Defense. This sheet is provided for obtaining information on the use of this specification which will insure that suitable products can be procured with a minimum amount of delay and at the least cost. Comments and the return of this form will be appreciated. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments and suggestions submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or serve to amend contractual requirements.

SPECIFICATION

GG-C-001304 (DSA-DM)

ORGANIZATION

CITY AND STATE

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE SPECIFICATION CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES.

2. COMMENTS ON ANY SPECIFICATION REQUIREMENT CONSIDERED TOO RIGID

3. IS THE SPECIFICATION RESTRICTIVE?

☐ YES ☐ NO *(If "yes", in what way?)*

4. REMARKS *(Attach any pertinent data which may be of use in improving this specification. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity)*

SUBMITTED BY *(Printed or typed name and activity - Optional)*

DATE

DD FORM 1 JAN 66 1426

REPLACES EDITION OF 1 OCT 64, WHICH MAY BE USED.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.