GG-C-1265C
July 19, 1976
SUPERSEDING
Int. Fed. Spec. GG-C-001265b (DSA-DM)
March 6, 1970

FEDERAL SPECIFICATION

CARVERS, DENTAL, AMALGAM AND WAX

This specification was approved by the Commissioner, Federal Supply Services, General Services Administration, for the use of all Federal agencies.

1. SCOPE AND CLASSIFICATION

1.1 Scope. This specification covers carvers suitable for carving amalgam restorations and wax patterns.

1.2 Classification. Carvers shall be of the following styles, as specified (see 6.1):

Style 1 - Wall No. 3
Style 2 - Tanner No. 5T
Style 3 - Tanner No. 6T
Style 4 - Black No. 89-92
Style 5 - Hollenback No. 1/2
Style 6 - Hollenback No. 3

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposal, form a part of this specification to the extent specified herein.

FSC 6520

Source: https://assist.dla.mil -- Downloaded: 202640-26T11:53Z
Check the source to verify that this is the current version before use.

Federal Specifications:

GG-I-526	- Instruments, Dental and Surgical; General Specifications.
PPP-B-566	- Boxes, Folding, Paperboard.
PPP-B-636	- Boxes, Shipping, Fiberboard.
PPP-B-676	- Boxes, Setup.

Federal Standards:

Fed. Std. No. 66	- Steel, Chemical Composition and Hardenability.
Fed. Std. No. 123	- Marking for Shipment (Civil Agencies).

(Activities outside the Federal Government may obtain copies of Federal Specifications, Standards, and Handbooks as outlined under General Information in the Index of Federal Specifications and Standards and at the prices indicated in the Index. The Index, which includes cumulative monthly supplements as issued, is for sale on a subscription basis by the Superintendent of Documents, U.S. Government Printing Office, Washington, DC 20402.

(Single copies of this specification and other Federal Specifications required by activities outside the Federal Government for bidding purposes are available without charge from Business Service Centers, at the General Services Administration Regional Offices in Boston, New York, Washington, DC, Atlanta, Chicago, Kansas City, MO, Fort Worth, Denver, San Francisco, Los Angeles, and Seattle, WA.

(Federal Government activities may obtain copies of Federal Specifications, Standards, and Handbooks and the Index of Federal Specifications and Standards from established distribution points in their agencies.)

Military Specification:

MIL-P-116	- Methods of Preservation.

Military Standards:

MIL-STD-105	- Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	- Marking for Shipment and Storage.

Source: https://assist.dla.mil -- Downloaded: 202640-26T11:53Z
Check the source to verify that this is the current version before use.

Drawings:

Defense Personnel Support Center (DPSC)

24377 - Carvers, Dental, Hollenback.
24378 - Carvers, Dental.
24379 - Carvers, Dental, Tanner.

(Copies of Military Specifications, Standards, and Drawings required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposal shall apply.

American Society for Testing and Materials (ASTM) Standards:

E18 - Tests for Rockwell Hardness and Rockwell Superficial Hardness of Metallic Materials.

E92 - Tests for Vickers Hardness of Metallic Materials.

(Application for copies should be addressed to the American Society for Testing and Materials, 1916 Race Street, Philadelphia, PA 19103.)

3. REQUIREMENTS

3.1 General requirements. The carvers shall be in accordance with the applicable requirements of GG-I-526 with the options and additional requirements as specified herein.

3.2 Standard sample. The finish and workmanship of the operative ends shall be equal to or better than the standard sample (see 3.6.1, 3.9 and 6.2).

3.3 Materials. Carvers shall be fabricated of stainless steel specified in table I.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

TABLE I. Materials

Instruments	Operative end	Handle
One piece	Class 4, type 440A (GG-I-526)	Class 4, type 440A (GG-I-526)
Three piece	Class 4, type 440A, 440B, 440C (GG-I-526)	AISI 430F (Fed. Std. No. 66 for composition only) or Class 3, type 303 (GG-I-526)

3.4 Hardness. Operative ends shall have a Rockwell hardness of C52 to C56 (type 440A), C54 to C58 (type 440B), C56 to C60 (type 440C) after proper stress relieving. Handles and shanks shall be properly tempered and shall have a hardness suitable for the purpose intended. Hardness shall be determined as specified in 4.3.1.

3.5 Style, design, and dimensions. Carvers shall be double ended. Carvers shall be suitable for carving amalgam restorations and wax patterns. The style, design, and dimensions of the carvers shall conform to Defense Personnel Support Center (DPSC) Drawings listed in table II for the applicable style. Carvers furnished shall be of the style specified herein.

TABLE II. Drawing numbers

Style	Name	Drawing	Design
1	Wall No. 3	24378	One piece
2	Tanner No. 5T	24379	
3	Tanner No. 6T	24379	
4	Black No. 89-92	24378	
5	Hollenback No. 1/2	24377	One or three piece
6	Hollenback No. 3	24377	One piece

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

3.5.1 Drawings. The drawings forming a part of this specification are not production drawings. These drawings establish dimensional limitations and control design and configuration. The supplier shall be responsible for preparing his own shop drawings. The supplier shall select dimensions and tolerances within the limitations prescribed on the drawings to provide a product with proper design, configuration and function. No deviation from the dimensional limitations prescribed on the drawings is permissible without prior approval of the contracting officer.

3.6 Finish.

3.6.1 Operative ends and shanks. Carvers shall be finished as to provide a sharp operative edge suitable for carving amalgam and wax patterns. The finish of the operative ends of the carvers shall be as indicated in table III. The shanks shall be smooth and shall have a mirror or bright polished finish. Operative ends shall be free of defects such as coarse grind marks, cracks, burrs, deep pits, or feathers when viewed as specified in 4.2.1. The finish of the operative ends shall be equal to or better than the standard sample (see 3.2, 3.9 and 6.2).

TABLE III. Finish of operative ends

Style	End shape	Finish
1	Spoon end, Outer face Spoon end, Inner face Chisel end	Mirror Fine hone Fine hone
2	Spoon ends, Outer face Spoon ends, Inner face	Mirror Fine hone
3	Side faces of blade Top face of blade	Fine polished Mirror or fine polish
4	Cleoid and Discoid ends, Outer face	Mirror
	Cleoid and Discoid ends, Inner face	Fine hone
5 & 6	(See applicable drawing)	Fine hone

The fine honed finish specified herein shall be a finish such as that product by using grade FF emery on a leather wheel.

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:53Z
Check the source to verify that this is the current version before use.

3.6.2 Handles.

3.6.2.1 Octagonal handles. Octagonal handles of one piece carvers shall have a mirror or bright polished finish. Handles shall have 40 + 10 fillet or notch-type serrations per inch on on seven faces and shall have one smooth face, as shown in the applicable DPSC drawing specified in 3.5. Depth of serrations shall be 0.002 to 0.004 inch.

3.6.2.2 Round handles. Round handles of the three piece carvers (style 5) shall have a mirror or bright polished finish except for the finger grip portion which shall be finely serrated or fine diamond knurled.

3.7 Corrosion resistance. Carvers shall meet the requirements of GG-I-526 for resistance to coppper sulfate and resistance to boil for the applicable material. The corrosion tests as specified in GG-I-526 for type 430 shall apply to type 430F material.

3.8 Identification marking. Carvers shall be permanently and legibly marked in accordance with GG-I-526 and as specified herein. Carvers shall be marked with the identification marking as specified in table IV on the smooth surface of the handle. Instrument number for style 4 shall be placed on the handle no less than 5 mm nor more than 12 mm from the junction of the handle and the shank of the applicable operative end. The word "stainless" shall also be marked on instruments of foreign manufacture. As an alternate to the word "stainless", the use of a registered trade name identifying the steel as stainless shall be acceptable, provided prior approval is received from the cognizant contracting officer (see 6.1).

TABLE IV. Identification marking

Style	Identification marking
1 (Wall No. 3)	Wall 3
2 (Tanner No. 5T)	5T or T5
3 (Tanner No. 6T)	6T
4 (Black No. 89 & 92)	89 and 92
5 (Hollenback No. 1/2)	Hollenback 1/2
6 (Hollenback No. 3)	Hollenback 3

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:53Z
Check the source to verify that this is the currentversion before use.

3.9 Workmanship. In addition to the requirements of GG-I-526, the operative end shall be free of discolorations, pits or rounded edges. Workmanship shall be equal to or better than the standard sample (see 3.2 and 6.2).

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure that supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.2 Certificates of quality. Certificates of quality, supplied by the manufacturer of the stainless steel, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The Certificate shall include the name of the contractor, the contract number, the lot size, the sample size, the date of testing, the test method, individual test results, the specification requirements, the name of the manufacturer or supplier of the stainless steel, the NSN, the Item Identification, the name of the material and the lot number.

4.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item. Sampling for inspection shall be performed in accordance with MIL-STD-105, except where otherwise indicated hereinafter.

Source: https://assist.dla.mil -- Downloaded: 202640-26T11:53Z
Check the source to verify that this is the current version before use.

4.2.1 Examination. The carvers shall be examined for defects including, but not limited to, those listed in table V. Examination for defects 103 and 104 shall be under four power (4X) magnification.

TABLE V. Classification of defects

Category	Defect
Major A:	
101	Style and design not as specified for the applicable style of instrument.
102	Finish of operative end and shank not as specified.
103	Carving edges not properly sharpened.
104	Operative end not free from coarse grind marks, cracks, burrs, deep pits, feathers and other imperfections.
Major B:	
151	Dimensions of operative ends not as specified on applicable drawing.
Minor A:	
201	Finish of handle not as specified.
202	Finish of handle not free of coarse grind marks.
203	Carvers not free from dirt, oil, polishing compound, and foreign matter.
204	Identification marking not present, not properly located, illegible, or incomplete.
205	Packaging and packing not as specified.
Minor B:	
251	Dimensions (other than 151) not as specified on applicable drawing.
252	Handle serrations missing, not uniform, or not complete.
253	Handle serration dimensions not as specified.

The inspection levels and acceptable quality levels (AQLs) shall be as listed in table VI. The unit of product for examination purposes shall be one carver.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

TABLE VI. Examination criteria

Type defect	Inspection level	AQL (percent defective)
Major A	II	1.0
Major B	S-2	1.5
Minor A	II	2.5
Minor B	S-2	4.0

4.2.2 Testing. Carvers shall be tested as specified in table VII and 4.3.1. Inspection levels and acceptable quality levels (AQLs) shall be as listed in table VII. The unit of product for test purposes shall be one carver. Lots for hardness testing (heat treated operative ends) shall be as specified in 4.2.2.1.

TABLE VII. Tests

Characteristic	Requirement	Test procedure	Inspection level	AQL
Chemical composition	3.3	GG-I-526	S-1 *	**
Hardness	3.4 and GG-I-526	4.3.1	S-2	1.5
Copper sulfate	3.7 and GG-I-526	GG-I-526	S-2	1.0
Boil test	3.7 and GG-I-526	GG-I-526	S-2	1.0

* A sufficient quantity of each lot of metal used in the construction of the carvers shall be selected for composition analysis in lieu of sample carvers, provided lot identity has been maintained and can be demonstrated to the Government.

**Acceptance number shall be zero for all sample sizes.

4.2.2.1 Lots for hardness criteria. For finished instrument (operative end) hardness testing, a lot or batch of instruments shall consist of operative ends (entire instrument for one piece carvers) produced by one manufacturer at one plant, from the same material (same raw material heat and heat treatment) and from the same heat treatment lot or batch (instruments undergoing heat treatment at the same time and in the same heat treatment equipment).

Source: https://assist.dla.mil -- Downloaded: 202640-26T11:53Z
Check the source to verify that this is the currentversion before use.

4.2.2.2 Sample unit usage. Where feasible, the same unit of product shall be used for the determination of two or more test characteristics.

4.2.3 Inspection for preparation for delivery. An examination shall be made to determine compliance with the packaging, packing and marking requirements of section 5. Defects shall be scored as specified in table VIII. Sampling shall be in accordance with MIL-STD-105. The sample unit shall be one container fully prepared for delivery. The lot shall be the number of containers fully prepared for delivery and offered for inspection at one time. The inspection level shall be S-2 with AQL of 4.0 expressed in terms of percent defective.

TABLE VIII. Examination of preparation for delivery

Examine	Defect
Contents	Not as specified.
Container	Not as specified.
Marking	Omitted, incorrect, illegible; improper size, location, sequence or method of application.
Packaging and packing materials	Any component missing or damaged.
Workmanship	Bulging or distortion of containers.

4.2.3.1 Examination of closure, waterproofing and banding of container. When the shipping container is required to comply with PPP-B-636, the examination for defects in closure, waterproofing or banding shall be in accordance with the appendix of the box specification.

4.3 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.3.1 Hardness. Hardness shall be determined in accordance with ASTM Designation E 18 (Rockwell) or E 92 (Vickers or Diamond Pyramid).

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

5. PREPARATION FOR DELIVERY

5.1 Packaging. Packaging shall be level A or C as specified (see 6.1).

5.1.1 Level A.

5.1.1.1 Unit package. Each carver shall be thoroughly cleaned and dried in accordance with the applicable method in MIL-P-116. Each carver shall be wrapped in neutral tissue or other non-corrosive material and packaged in accordance with MIL-P-116, method IC-2. A preservative shall not be applied to the carver.

5.1.1.2 Intermediate package. Twelve unit packages, containing carvers of one type and style only, shall be packaged in a snug-fitting box constructed in accordance with PPP-B-566, PPP-B-636, class domestic or PPP-B-676. Closure of the box shall be adequate to prevent accidental opening.

5.1.2 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing. Packing shall be level A, B or C as specified (see 6.1).

5.2.1 Level A. Twenty-four intermediate packages (288 units) containing carvers of one type and style shall be packed in an exterior container designed for a type 2 load and constructed in accordance with PPP-B-636, class weather-resistant. Closure, strapping and waterproofing shall be as specified in the appendix of the box specification.

5.2.2 Level B. Twenty-four intermediate packages (288 units) of one type and style only shall be packed in an exterior container designed for a type 2 load and constructed in accordance with PPP-B-636, class domestic. Closure shall be as specified in the appendix of the box specification.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers, of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

5.3 Marking.

5.3.1 Civil agencies. In addition to any special marking required by the contract or order, the unit, intermediate, and shipping containers shall be marked in accordance with Fed. Std. No. 123.

5.3.2 Military agencies.

5.3.2.1 Interior packages. Each unit package and each intermediate package shall be marked in accordance with MIL-STD-129. When labels are utilized, waterproofing shall be required only when applicable carton is fabricated of water-resistant material. Contract or purchase order number and name of contractor shall be shown.

5.3.2.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Ordering data. Purchasers should select the preferred options permitted herein and include the following information in procurement documents:

a. Title, number and date of this specification.
b. Type and style required (see 1.2).
c. Selection of applicable levels of packaging and packing (see 5.1 and 5.2).
d. Approval of trade name, if applicable (see 3.8).

6.2 Standard sample. For access to standard sample, address the procuring activity issuing the invitation for bids or request for proposal (see 3.2, 3.6.1 and 3.9).

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:53Z
Check the source to verify that this is the current version before use.

MILITARY CUSTODIANS:

Army - MD
Navy - MS
Air Force - 03

CIVIL AGENCY COORDINATING ACTIVITIES:

GSA - FSS
HEW - FDA, HSM
VA - DMS

Project No. 6520-1404

PREPARING ACTIVITY: DSA - DM

Orders for this publication are to be placed with General Services Administration, acting as an agent for the Superintendent of Documents. See section 2 of this specification to obtain extra copies and other documents referenced herein. Price 35 cents each.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 South 20th Street
Philadelphia, PA 19101

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY
DoD - 304

U.S.MAIL

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/ vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE

GG-C-1265C CARVERS, DENTAL, AMALGAM AND WAX

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?

☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO.

DATE

DD FORM 1 JAN 72 1426 REPLACES EDITION OF 1 JAN 66 WHICH MAY BE USED

☆ GPO. 1974— 603-680/ 6126 2-1

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:53Z
Check the source to verify that this is the current version before use.