GG-C-1282C
August 14, 1985

(SUPERSEDING)
GG-C-1282B
November 9, 1976

FEDERAL SPECIFICATION

CURETTES, ALVEOLAR (Molt and Miller)

This specification is approved by the Commissioner, Federal Supply Service, General Services Administration, for the use of all Federal agencies.

1. SCOPE AND CLASSIFICATION

1.1 Scope. This specification covers curettes of Molt's and Miller's design suitable for use in oral surgical procedures.

1.2 Classification.

1.2.1 Types, styles, and numbers. Curettes shall be of the following types, styles and numbers, as specified (see 6.1):

Type I - Single-end

Style A - Straight

Molt No. 1
Molt No. 2
Molt No. 4

Style B - Angular

Molt No. 5L
Molt No. 6R
Molt No. 9L
Molt No. 10R

Type II - Double-end

Miller No. 10
Miller No. 11
Miller No. 12

2. APPLICABLE DOCUMENTS

2.1 The following documents, of the issues in effect on date of invitation for bids or request for proposal, form a part of this specification to the extent specified herein.

FSC 6520

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

Federal Specifications:

GG-I-526	- Instruments, Dental and Surgical; General Specifications.
QQ-N-290	- Nickel Plating (Electrodeposited).
PPP-B-566	- Boxes, Folding, Paperboard.
PPP-B-601	- Boxes, Wood, Cleated-Plywood.
PPP-B-621	- Boxes, Wood, Nailed and Lock-corner.
PPP-B-636	- Boxes, Shipping, Fiberboard.
PPP-B-676	- Boxes, Setup.

Federal Standards:

Fed. Std. No. 123 - Marking for Shipment (Civil Agencies).

(Activities outside the Federal Government may obtain copies of Federal specifications, standards, and commercial item descriptions as outlined under General Information in the Index of Federal Specifications, Standards and Commercial Item Descriptions. The Index, which includes cumulative bimonthly supplements as issued, is for sale on a subscription basis by the Superintendent of Documents, US Government Printing Office, Washington, DC 20402.

(Single copies of this specification, other Federal specifications, and commercial item descriptions required by activities outside the Federal Government for bidding purposes are available without charge from General Services Administration Business Service Centers in Boston, MA; New York, NY; Philadelphia, PA; Washington, DC; Atlanta, GA; Chicago, IL; Kansas City, MO; Fort Worth, TX; Houston, TX; Denver, CO; San Francisco, CA; Los Angeles, CA; and Seattle, WA.

(Federal Government activities may obtain copies of Federal standardization documents and the Index of Federal Specifications, Standards and Commercial Item Descriptions from established distribution points in their agencies.)

Military Specifications:

MIL-P-116	- Preservation, Methods of.
MIL-L-10547	- Liners, Case and Sheet, Overwrap; Water-Vaporproof, or Waterproof, Flexible.

Military Standards:

MIL-STD-105	- Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	- Marking for Shipment and Storage.
MIL-STD-147	- Palletized Unit Loads.
MIL-STD-794	- Parts and Equipment, Procedures for Packaging of.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

Defense Personnel Support Center (DPSC) Drawings:

20282 - Curettes, Alveolar Molt No. 1, 2, 4, 5L, 6R, 9L, and 10R.
24449 - Curettes, Alveolar, Miller No. 10, 11, and 12.

(Copies of Military specifications, standards and drawings required by manufacturers in connection with specific acquisition functions should be obtained from the contracting activity or as directed by the contracting activity.)

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless a specific issue is identified, the issue in effect on date of invitation for bids or request for proposal shall apply.

American Society For Testing and Materials (ASTM) Standards:

B 135 - Specification for Seamless Brass Tube.
E 18 - Standard Methods of Test for Rockwell Hardness and Rockwell Hardness of Metallic Materials.
E 92 - Standard Method of Test for Vickers Hardness of Metallic Materials.

(Application for copies should be addressed to the American Society for Testing and Materials, 1916 Race Street, Philadelphia, PA 19103.)

(Nongovernment standards and other publications are normally available from the organizations which prepare or which distribute the documents. These documents also may be available in or through libraries or other informational services.)

2.3 Order of precedence. In the event of a conflict between the text of this specification and the references cited herein, the text of this specification shall take precedence. Nothing in this specification, however, shall supersede applicable laws and regulations unless a specific exemption has been obtained.

3. REQUIREMENTS

3.1 General requirements. The curettes shall be in accordance with applicable requirements of GG-I-526 with the options and additional requirements specified herein.

3.2 Material.

3.2.1 Operative end and shanks. Operative ends and shanks snall be Class 4, Type 440 A stainless steel.

3.2.2 Handles. Type I curette handles shall conform to 3.2.2.1 or 3.2.2.2 for composition only. Type II curette handles shall conform to 3.2.2.1 for composition only.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

3.2.2.1 Brass. Handles shall conform to one of the following:

a. Class 8, alloy number 260 (GG-I-526).
b. Copper alloy number 330 (ASTM B135).
c. Copper alloy having 62 to 71 percent copper; 0.5 percent lead, maximum; 0.2 percent iron, maximum; 0.2 percent other elements, maximum; and the remainder zinc.

Brass handles shall be plated as specified in 3.7.

3.2.2.2 Stainless steel. Handles shall conform to class 3, type 304 except with a chrome content range of 17 to 20 percent, or class 4, type 416, stainless steel.

3.3 Hardness. After proper stress relieving, operative ends shall have a Rockwell hardness of C52 to C56. The hardness of the shanks shall be suitable for the purpose intended. Hardness shall be determined as specified in 4.3.2.

3.4 Style, design, and dimensions. Curettes shall be of the following types, styles, and curette numbers, as specified herein. Type I curettes shall be single-ended curettes of Molt design and Type II curettes shall be double-ended curettes of Miller design. Type I curettes shall be in accordance with DPSC drawing 20282. Type II curettes shall be in accordance with DPSC drawing 24449. Curettes shall be alveolar bone dental surgical curettes suitable for use in oral surgical procedures. Curettes shall be suitable for use as exploratory instruments and to remove tooth particles or debris from sockets to enucleate cysts, dental granulomas or cystic neoplasms and to remove small sequestra from sockets.

3.4.1 Drawings. The drawings forming a part of this specification are not production drawings. Drawings establish dimensional limitations and control design and configuration. The supplier shall be responsible for preparing his own shop drawings. The supplier shall select dimensions prescribed on the drawings to provide a product with proper design, configuration and function. No deviations from the dimensional limitations prescribed on the drawings are permissible without prior approval of the contracting officer.

3.5 Construction. Type I single-ended curettes shall be of two piece construction. Type II double-ended curettes shall be of three piece construction. Single-ended curettes shall consist of an operative end (blade), shank, and handle. Double-ended curettes shall consist of a handle with an operative end (blade) and shank at each end. The shanks shall be permanently secured in the handle by brazing. There shall be no movement, loosening, or separation between the operative end and handle when the operative end is subjected to an 85 pound (minimum) pull out force when tested as specified in 4.3.3.

Source: https://assist.dla.mil -- Downloaded: 202640-26T11:53Z
Check the source to verify that this is the currentversion before use.

3.6 Finish. Operative ends and shanks shall have a bright finish except for the inside surface and curetting edge of the spoon which shall have a fine honed finish. The operative ends shall have a sharp edge around the periphery of the blade and shall be free of defects such as coarse grind marks, cracks, burrs, deep pits and turned over edges, when viewed as specified in 4.2.1. Type I brass handles shall have a longitudinal scratch, sandblast, or satin finish. Type I and Type II stainless steel handles shall have a satin finish. Type II brass handles shall have a bright finish except for the serrated or knurled portion. The serrated or knurled portion of the handle shall be free of barbs and shall provide adequate finger grip during use.

3.7 Plating. Plating shall be in accordance with GG-I-526 for Class b material. As an alternate to the plating specified in GG-I-526, the handles shall be nickel-plated only in accordance with QQ-N-290 with a minimum thickness of 0.001 inch, and a maximum of 0.002 inch.

3.8 Corrosion resistance. Components fabricated from martensitic stainless steel (class 4, type 440A or 416) shall meet the copper sulfate requirements of GG-I-526. The complete instrument shall meet the boil test requirements of GG-I-526. The salt spray test shall not apply.

3.9 Identification marking. Curettes shall be permanently and legibly marked on the front face of the instrument. In addition to the requirements of GG-I-526, each curette shall be marked with the applicable Holt or Miller number. Marking shall be on the shank of the instrument (the area near the junction of the handle) for Type I (Holt curettes) and on the smooth front face of the handle for Type II (Miller curettes).

3.10 Workmanship. In addition to the requirements of GG-I-526, the operative ends shall be free of discoloration and rounded edges.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in this specification where such inspections are deemed necessary to assure that supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 202640-26T11:53Z
Check the source to verify that this is the currentversion before use.

4.1.2 Certificates of quality. Certificates of quality, supplied by the manufacturer of the raw materials, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The Certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier of the raw materials, the NSN, the Item Identification, the name of the material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item. Sampling for inspection shall be performed in accordance with MIL-STD-105, except where otherwise indicated hereinafter.

4.2.1 Examination. The curettes shall be examined for defects, including but not limited to, those listed in table I. Examination for defects 104 and 105 shall be under four power (4X) magnification.

TABLE I. Classification of defects.

Categories	Defects
Critical:	None defined.
Major A:	
101	Style and design not as specified for the applicable type, style, and number of curette.
102	Construction of curette not as specified.
103	Finish of operative end not as specified.
104	Operative end edge not sharp or has turned over edges.
105	Operative end and shank not free from coarse grind marks, cracks, burrs, and deep pits.
106	Operative end (front face of spoon) not properly oriented with respect to the broad face of handle (Type I).
Major B:	
151	Operative end and shank dimensions not as specified.
Minor:	
201	Handle dimensions not as specified.
202	Shank and handle finish not as specified.
203	Edges and corners of handle not free of burrs and sharpness.
204	Overall length of instrument not as specified.
205	Identification marking not present, not properly located, illegible, or incomplete.
206	Instrument not free from dirt, oil, compound, or foreign matter.
207	Plating of handle (brass handle only) not uniform in appearance.
208	Plating of handle (brass handle only) not free of blisters, visible porosity, or other defects.

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:53Z
Check the source to verify that this is the currentversion before use.

The inspection levels and acceptable quality levels (AQLs) shall be listed in table II. The unit of product for examination purposes shall be one curette.

TABLE II. Examination criteria.

Type defective	Inspection level	AQL (percent defective)
Major A	II	1.0
Major B	S-1	2.5
Minor defects	II	4.0

4.2.2 Testing. Curettes shall be tested as specified in table III and 4.3. Inspection levels and acceptable quality levels (AQLs) shall be listed in table III. The unit of product for test purposes shall be one curette.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

TABLE III. Tests.

Component	Characteristic	Requirement	Test procedure	Inspection level	AQL
Entire curette	Composition	3.2	GG-I-526	S-1*	**
Handle (Brass handle only see 3.2.2.1	Thickness of plating	3.2.2.1 and 3.7	4.3.1	S-2	1.5
Operative end	Hardness	3.3 and GG-I-526	4.3.2	S-2	1.5
Curette	Pull test	3.5	4.3.3	S-2	**
Operative end, shank, and handle (class 4 material only, see 3.2.2.2)	Copper sulfate test	3.8	GG-I-526	S-2	1.0
Entire curette	Boil test	3.8	GG-I-526	S-2	1.0

* A sufficient amount from each lot of each metal used in the fabrication of the curette may be selected for chemical analysis in lieu of sample curettes, provided lot identity has been maintained and can be demonstrated to the Government.

** Acceptance number shall be zero for all sample sizes.

4.2.2.1 Sample unit usage. Where feasible, the same sample unit shall be used for the determination of two or more test characteristics.

4.3 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.3.1 Thickness of plating. Thickness of plating shall be determined in accordance with requirements of QQ-N-290.

4.3.2 Hardness. Hardness shall be determined in accordance with ASTM E18 (Rockwell) or E92 (Vickers or Diamond Pyramid).

Source: https://assist.dla.mil -- Downloaded: 202640-26T11:53Z
Check the source to verify that this is the current version before use.

4.3.3 Pull test. The handle and shank of the operative end shall be securely held in separate chucks or other holding devices. A slowly increasing pull out force, increasing from zero to 85 pounds (minimum), shall be applied to the shank of the operative end. The force shall be maintained at the high value for a minimum of 1 second. The instrument shall then be examined for movement, looseness, or separation of the operative end with respect to the handle. The test shall be repeated for each operative end of the double-ended instrument.

4.4 Inspection of packaging. The sampling and inspection of the preservation and interior package marking shall be in accordance with groups A and B quality conformance inspection requirements of MIL-P-116. The sampling and inspection of the packing and marking for shipment and storage shall be in accordance with the quality assurance provisions of the applicable container specification and the marking requirements of MIL-STD-129.

5. Packaging

5.1 Preservation. Preservation shall be level A or C as specified (see 6.2).

5.1.1 Level A.

5.1.1.1 Cleaning. Curettes shall be cleaned in accordance with MIL-P-116.

5.1.1.2 Drying. Curettes shall be dried in accordance with MIL-P-116.

5.1.1.3 Preservative application. Preservatives shall not be used.

5.1.1.4 Unit package. Each curette shall be wrapped in neutral tissue or other suitable anticorrosive material. Each protected curette shall be packaged in accordance with MIL-P-116, method IC-1 utilizing a bag conforming to MIL-P-117, type I or II, class C, style 1 only. As an alternate, each curette shall be packaged in a rigid plastic tube of appropriate size having a friction or plug type closure. The tube shall contain suitable cushioning material to immobilize and protect the curette from damage.

5.1.1.5 Intermediate package. Twelve curettes of one type, style, and number only shall be packaged in a box conforming to PPP-B-566; PPP-B-636, class domestic, or PPP-B-676. Closure shall be as specified in the applicable box specification.

5.1.2 Level C. The level C preservation for curettes shall conform to MIL-STD-794 requirements for this level.

5.2 Packing. Packing shall be level A, B, or C as specified (see 6.2).

5.2.1 Level A. Twenty-four intermediate packages containing curettes of one type, style, and number only shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-601, overseas type; PPP-B-621, class 2, or PPP-B-636, class weather-resistant. Fiberboard boxes shall be waterproofed in accordance with PPP-B-636. Closure and strapping shall be as specified in the applicable box specification.

Source: https://assist.dla.mil -- Downloaded: 202640-26T11:53Z
Check the source to verify that this is the currentversion before use.

5.2.1.1 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall be as specified in the liner specification.

5.2.2 Level B. Twenty-four intermediate packages shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-601, domestic type; PPP-B-621, class 1, or PPP-B-636, class domestic. Closure shall be as specified in the applicable box specification.

5.2.3 Level C. The level C packing for curettes shall conform to MIL-STD-794 requirements for this level.

5.2.4 Unitized loads. Unitized loads, commensurate with the level of packing specified in the contract or order, shall be used whenever total quantities for shipment to one destination equal 40 cubic feet or more. Quantities less than 40 cubic feet need not be unitized. Unitized loads shall be uniform in size and quantities to the greatest extent practicable.

5.2.4.1 Levels A and B. Curettes packed as specified in 5.2.1 and 5.2.2, shall be unitized on pallets conforming to MIL-STD-147.

5.2.4.2 Level C. Curettes packed as specified in 5.2.3, shall be unitized as specified in MIL-STD-794.

5.3 Marking.

5.3.1 Military activities.

5.3.1.1 Levels A, B, and C. Each unit pack, intermediate pack, exterior container and unitized load shall be marked as specified in MIL-STD-129.

5.3.2 Civil agencies. Marking shall be as specified in FED-STD-123.

5.4 General.

5.4.1 Exterior container. Exterior container (see 5.2.1, 5.2.2, and 5.2.3) shall be of minimum tare and cube consistent with the protection required and shall contain equal quantities of identical stock numbered items to the greatest extent practicable.

5.4.2 Packaging inspection. The inspection of these packaging requirements shall be in accordance with 4.4.

6. NOTES

6.1 Intended use. The curettes are intended for use as exploratory instruments and to remove tooth particles or debris from sockets to enucleate cysts, dental granulomas or cystic neoplasms and to remove small sequestra from sockets.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

6.2 Ordering data.

6.2.1 Acquisition requirements. Acquisition documents should specify the following:

(a) Title, number, and date of this specification.
(b) Type, style and number required (see 1.2).
(c) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.3 National Stock Number (NSN) coverage. This specification covers the following items appearing in the Federal Supply Catalog:

National Stock Number	Item Identification
6520-00-519-6550	CURETTE, ALVEOLAR, Molt, No. 1
6520-00-519-6600	CURETTE, ALVEOLAR, Molt, No. 2
6520-00-519-6700	CURETTE, ALVEOLAR, Molt, No. 4
6520-00-519-6740	CURETTE, ALVEOLAR, Molt, No. 5L
6520-00-519-6770	CURETTE, ALVEOLAR, Molt, No. 6R
6520-00-519-6800	CURETTE, ALVEOLAR, Molt, No. 9L
6520-00-519-6900	CURETTE, ALVEOLAR, Molt, No. 10R
6520-01-125-6602	CURETTE, ALVEOLAR, Miller, No. 10
6520-01-125-6603	CURETTE, ALVEOLAR, Miller, No. 12
6520-01-126-4028	CURETTE, ALVEOLAR, Miller, No. 11

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS

Project No. 6520-1830

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.