GG-C-1320B
2 October 1985
SUPERSEDING
GG-C-1320A
OCTOBER 21, 1976

FEDERAL SPECIFICATION

CUP, POLISHING, DENTAL HANDPIECE (SCREW STEM)

This specification is approved by the Commissioner, Federal Supply Service, General Services Administration, for the use of all Federal agencies.

1. SCOPE AND CLASSIFICATION

1.1 Scope. This specification covers screw stem polishing cups suitable for use with angle prophylaxis handpieces in cleaning and polishing natural teeth.

1.2 Classification. Polishing cups shall be furnished in the following styles as specified (see 6.1):

Style 1 - With 8 ribs.
Style 2 - With 22 serrations.

2. APPLICABLE DOCUMENTS

2.1 Government documents.

2.1.1 Specifications and standards. The following specifications and standards form a part of this specification to the extent specified herein. Unless otherwise specified, the issues of these documents shall be those listed in the issue of the Department of Defense Index of Specifications and Standards (DoDISS) and supplement thereto, cited in the solicitation.

Federal Specifications:

PPP-B-566 - Boxes, Folding, Paperboard.
PPP-B-636 - Boxes, Shipping, Fiberboard.
PPP-B-676 - Boxes, Setup.

Federal Standards:

Fed. Std. No. 123 - Marking for Shipment (Civil Agencies).

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-ATT, 2800 South 20th Street, Philadelphia, PA 19101, by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document or by letter.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

(Activities outside the Federal Government may obtain copies of Federal specifications, standards, and commercial item descriptions as outlined under General Information in the Index of Federal Specifications, Standards and Commercial Item Descriptions. The Index, which includes cumulative bimonthly supplements as issued, is for sale on a subscription basis by the Superintendent of Documents, US Government Printing Office, Washington, DC 20402.

(Single copies of this specification, other Federal specifications, and commercial item descriptions required by activities outside the Federal Government for bidding purposes are available without charge from General Services Administration Business Service Centers in Boston, MA; New York, NY; Philadelphia, PA; Washington, DC; Atlanta, GA; Chicago, IL; Kansas City, MO; Fort Worth, TX; Houston, TX; Denver, CO; San Francisco, CA; Los Angeles, CA; and Seattle, WA.

(Federal Government activities may obtain copies of Federal standardization documents and the Index of Federal Specifications, Standards and Commercial Item Descriptions from established distribution points in their agencies.)

Military Specifications:

MIL-H-36809	- Handpiece, Dental, Right Angle, Prophylaxis, Round Sleeve, Screw Stem.
MIL-P-116	- Preservation, Packaging, Methods of

Military Standards:

MIL-STD-105	- Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	- Marking for Shipment and Storage.
MIL-STD-147	- Palletized Unit Loads.
MIL-STD-794	- Parts and Equipment, Procedures for Packaging of.

Drawing:

Defense Personnel Support Center (DPSC)
20675 - Cup, Polishing, Dental Handpiece

(Copies of specifications, standards, and drawings required by contractors in connection with specific acquisition functions should be obtained from the contracting activity or as directed by the contracting activity.)

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise specified, the issues of the documents which are DoD adopted shall be those listed in the issue of the DoDISS specified in the solicitation. Unless otherwise specified, the issues of documents not listed in the DoDISS shall be the issue of the nongovernment documents which is current on the date of the solicitation.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

AMERICAN SOCIETY FOR TESTING AND MATERIAL (ASTM)

ASTM D 297	Standard Methods for Rubber Products - Chemical Analysis.
ASTM D 572	Standard Test Method for Rubber Deterioration by Heat and Oxygen Pressure.
ASTM D 2240	Standard Test Method for Rubber Property - Durometer Hardness.

(Application for copies of ASTM publications should be addressed to the American Society for Testing and Materials, 1916 Race Street, Philadelphia, PA 19103.)

(Nongovernment standards and other publications are normally available from the organizations which prepare or which distribute the documents. These documents also may be available in or through libraries or other informational services.)

2.3 Order of precedence. In the event of a conflict between the text of this specification and the references cited herein, the text of this specification shall take precedence. Nothing in this specification, however, shall supersede applicable laws and regulations unless a specific exemption has been obtained.

3. REQUIREMENTS

3.1 Material.

3.1.1 Rubber (style 1). The elastomeric material used in style 1 polishing cup shall be either silicone rubber or a rubber containing 38 to 48 percent natural rubber and 8 to 12 percent (Styrene-Butadiene) synthetic rubber when tested as specified in 4.4.3 (see 3.5).

3.1.2 Rubber (style 2). The elastomeric material used in style 2 polishing cup shall contain 60 to 70 percent natural rubber (crepe or smoked rib sheet) when tested as specifed in 4.4.3 (see 3.5).

3.1.3 Screw stem. Screw stem shall be fabricated of free machining brass, suitable for adhesion to the rubber. As an alternate, the finish on the screw shall be bright mechanical nickel plate of 0.0001 inch maximum thickness.

3.2 Style, design, and dimensions. The cup shall be suitable for use as prophylaxis polishing cup. The cup shall conform to the design and dimensions of Defense Personnel Support Center (DPSC) Drawing 20675 for the style specified. The polishing cup shall be suitable for use with screw stem angle prophylaxis handpiece conforming to MIL-H-36809. The cup cavity shall be shaped essentially as a truncated rone, and shall have the ribs (style 1) or serrations (style 2) evenly spaced on the sloping inner walls. The screw shall be securely imbedded in the cup. The cup shall be sufficiently elastic to spread out under normal hand pressure without qtearing, and to conform to the surface of the teeth when in operation. The cup shall be gray, white, blue or amber in color. The cup shall not discolor the teeth.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

3.2.1 Drawings. The drawings forming a part of this specification are not production drawings. These drawings establish dimensional limitations and control design and configuration. The supplier shall be responsible for preparing his own shop drawings. The supplier shall select dimensions and tolerances within the limitations prescribed on the drawings to provide a product with proper design, configuration and function. No deviations from the dimensional limitations prescribed on the drawings are permissible without prior approval of the contracting officer.

3.3 Accelerated aging. Cup shall not become hard or brittle or soft and tacky, when tested as specified in 4.4.2.

3.4 Hardness requirements.

3.4.1 Style 1. Rubber used in style 1 polishing cup shall have a Shore Durometer Scale A range of hardness of 70 ± 3 International hardness degrees when tested as specified in 4.4.1.

3.4.2 Style 2. Rubber used in style 2 polishing cup shall have a Shore Durometer Scale A range of hardness of 50 ± 5 International hardness degrees when tested as specified in 4.4.1.

3.5 Irritation. The manufacturer shall furnish, prior to delivery, a statement directly to the contracting officer stating that the polishing cup does not normally irritate oral tissues, and does not contain poisonous ingredients in sufficient concentration to be harmful to human beings when used as directed.

3.6 Workmanship. Polishing cup shall be free from defects such as cuts, tears, splits, and flash, which may detract from its appearance or impair its serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:53Z
Check the source to verify that this is the currentversion before use.

4.1.2 Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.3 Certificates of qualtity. Certificates of quality, supplied by the manufacturer of the materials, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The Certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the Item Identification, the name of the material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Quality conformance inspection.

4.2.1 Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105. The unit of product for examination shall be one individual cup. Inspection and acceptable quality levels shall be as indicated in table I.

TABLE I. Examination.

Type defect	Inspection level	AQL (Percent defective)
Major A	II	1.0
Major B	S-2	1.5
Minor A	II	2.5
Minor B	S-2	4.0

4.2.2 Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105, inspection level S-1. Acceptance number shall be zero for all sample sizes. The unit of product for accelerated aging test shall be one individual cup. When tests are performed on raw material, sufficient quantity from each lot or batch of rubber used shall be tested.

4.3 Examination. Rubber cup shall be examined for defects including, but not limited to, those listed in table II.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

TABLE II. Classification of defects

Categories	Defects
Major A:	
101	Style not as specified.
102	Cup cavity not shaped essentially as a truncated cone.
103	Cup not sufficiently elastic to spread out under normal hand pressure, without tearing.
104	Screw stem not securely imbedded in cup.
105	Screw threads incomplete or stripped.
Major B:	
151	Screw dimensions not as specified.
152	Rubber cup dimensions affecting fit into handpiece not as specified.
Minor A:	
201	Color not as specified.
202	Cup not free from cuts, tears, splits, flash, and other imperfections.
Minor B:	
251	Rubber cup dimensions (other than 151 and 152) not as specified.
252	Number of ribs (style 1) or serrations (style 2) not as specified.

4.3.1 Inspection of packaging. The sampling and inspection of the preservation and interior package marking shall be in accordance with groups A and B quality conformance inspection requirements of MIL-P-116. The sampling and inspection of the packing and marking for shipment and storage shall be in accordance with the quality assurance provisions of the applicable container specification and the marking requirements of MIL-STD-129.

4.4 Tests.

4.4.1 Hardness test. The vulcanized rubber used in the fabrication of the cups shall be subjected to the durometer hardness test in accordance with American Society for Testing and Materials (ASTM) Designation D 2240, Test Method for Rubber Property-Durometer Hardness. (Cycle time for testing specimen, 5 seconds).

4.4.2 Accelerated aging test. Polishing cups shall be subjected to the aging test in accordance with American Society for Testing and Materials (ASTM) Designation D 572, Test Method for Rubber Deterioration by Heat and Oxygen.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

4.4.3 Rubber composition. Rubber composition shall be determined by using American Society for Testing and Materials (ASTM) Designation D 297, Methods for Chemical Analysis of Rubber Products, section 10 to 12 or section 52.

5. PACKAGING

5.1 Preservation. Preservation shall be level A or C, as specified (see 6.1).

5.1.1 Level A.

5.1.1.1 Preservative applications. Preservatives shall not be used.

5.1.1.2 Unit packages Thirty-six cups of one style only shall be packaged in accordance with MIL-P-116, method III within a box or sealed plastic bag of an appropriate size and design.

5.1.1.3 Intermediate packages. Four or twelve unit packages containing cups of one style only shall be placed in intermediate containers conforming to PPP-B-566 or PPP-B-676. Intermediate containers shall be uniform in size, shape, and quantities, shall be of minimum tare and cube and shall contain multiples of five unit packages, not to exceed 100 unit packages. No intermediate packages are required when the total quantity shipped to a single destination is less than 100 unit packages.

5.1.2 Level C. The level C preservation for cups shall conform to the MIL-STD-794 requirements for this level.

5.2 Packing. Packing shall be level A, B, or C, as specified (see 6.1).

5.2.1 Level A. Seventy-two intermediate packages containing twelve unit packages each or two hundred and sixteen intermediate packages containing four unit packages each (864 unit packages) containing cups of one style only, shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-636, class weather-resistant. Closure, strapping, and waterproofing shall be as specified in the appendix of PPP-B-636.

5.2.2 Level B. Seventy-two intermediate packages containing twelve unit packages each or two hundred and sixteen intermediate packages containing four unit packages each (864 unit packages) containing cups of one style only shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-636, class domestic. Closure shall be as specified in the appendix of the box specification.

5.2.3 Level C. The level C packing for cups shall conform to the MIL-STD-794 requirements for this level.

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:53Z
Check the source to verify that this is the currentversion before use.

5.2.4 Unitized loads. Unitized loads, commensurate with the level of packing specified in the contract or order, shall be used whenever total quantities for shipment to one destination equal 40 cubic feet or more. Quantities less than 40 cubic feet need not be unitized. Unitized loads shall be uniform in size and quantities to the greatest extent practicable. The cups shall be unitized on pallets in accordance with MIL-STD-147.

5.3 Marking.

5.3.1 Levels A, B, and C. In addition to any special or other identification marking required by the contract (see 6.1) each unit package, intermediate and exterior container and unitized load shall be marked in accordance with MIL-STD-129.

5.4 General.

5.4.1 Exterior containers. Exterior containers (see 5.2.1, 5.2.2, and 5.2.3) shall be of a minimum tare and cube consistent with the protection required and shall contain equal quantities of identical stock numbered items to the greatest extent practicable.

5.4.2 Packaging inspection. The inspection of these packaging requirements shall be in accordance with MIL-P-116.

6. NOTES.

6.1 Ordering data.

6.1.1 Acquisition requirements. Acquisition documents should specify the following:

a. Title, number and date of this specification.
b. Style required (see 1.2).
c. Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.2 National Stock Number coverage. The following National Stock Number is covered by this document:

6520-00-982-9377 CUP, POLISHING, DENTAL HANDPIECE, Rubber, 36s.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

Review activities:

Army-MD
Navy-MS
Air Force-03

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

Project No. 6520-1821

Location S2-9-4

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE: This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO. 4966 WASHINGTON D. C.
POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Headquarters
Defense Personnel Support Center
ATTN: Directorate of Medical Materiel
DPSC-ATT
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
GG-C-1320B	CUP, POLISHING, DENTAL HANDPIECE (SCREW STEM)

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)*:

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.