GG-E-1230C
15 June 1987
SUPERSEDING
GG-E-1230B
13 July 1976

FEDERAL SPECIFICATION

EXCAVATORS, DENTAL (Corrosion-Resisting Steel)

This Specification is approved by the Commissioner, Federal Supply Service, General Services Administration, for the use of all Federal agencies.

1. SCOPE AND CLASSIFICATION

1.1 Scope. This specification covers dental excavators suitable for use in dental operative procedures.

1.2 Classification.

1.2.1 Styles. The excavators shall be of the following styles, as specified (see 6.1):

Style I - Black's, modified, double-ended.
Enamel Hatchet, No. 15, 16, (see 3.4 and 3.7)
Spoon, No. 36, 37, (see 3.4 and 3.7).
Discoid, No. 38, 39.

Style II - Wedelstaedt's, modified, double-ended.
Enamel Hatchet No. 17, 18, (see 3.4 and 3.7)

AMSC N/A FSC 6520

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be use in improving this document should be addressed to: Headquarters, Defense Personnel Center, Directorate of Medical Materiel, DPSc-RST, 2800 South 20th Street, Philadelphia, PA 19101, by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document or by letter.

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

Style III - Darby Perry, Double-ended.
No. 5, 6.

Style IV - Endodontic, double ended.
No. 31.
No. 32L.

Style V - Excavator, double-ended
No. 18.
No. 19.

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposal form a part of this specification to the extent specified herein:

Federal Specifications:

GG-I-526	-	Instruments, Dental and Surgical; General Specifications.
PPP-B-566	-	Boxes, Folding, Paperboard.
PPP-B-601	-	Boxes, Wood, Cleated-plywood.
PPP-B-621	-	Boxes, Wood, Nailed and Lock-corner.
PPP-B-636	-	Boxes, Shipping, Fiberboard.
PPP-B-676	-	Boxes, Setup.

Federal Standards:

FED-STD No. 66	-	Steel, Chemical Composition and Hardenability.
FED-STD-123	-	Marking for shipment (civil agencies).

(Activities outside the Federal Govenment may obtain copies of Federal specifications, standards, and commercial item descriptions as outlined under General Information in the Index of Federal Specifications, Standards and Commercial Item Descriptions. The Index, which includes cumulative bimonthly supplements as issued, is for sale on a subscription basis by the Superintendent of Documents, US Government Printing Office, Washington, DC 20402).

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

(Copies of listed federal and military standards, specifications, Commercial Item Descriptions (CIDs), handbooks and associated documents listed in the Department of Defense Index of Specifications and Standards (DoDISS), should be obtained from the DoD Single Stock Point, Commanding Officer, Naval Publications and Forms Center, 5801 Tabor Avenue, Philadelphia, PA 19120. Copies of industry association documents should be obtained from the sponsor. Copies of all other listed documents should be obtained from the contracting activity or as directed by the contracting activity).

(Federal Government activities may obtain copies of Federal Standardization documents and the Index of Federal Specifications Standards and Commercial Item Descriptions from established distribution points in their agencies).

Military Specifications:

MIL-P-116	- Methods of Preservation.
MIL-L-10547	- Liners, Case and Sheet, Overwrap; Water-Vaporproof or Waterproof, Flexible.

Military Standards:

MIL-STD-105	- Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	- Marking for Shipment and Storage.
MIL-STD-147	- Palletized Unit Loads.
MIL-STD-794	- Parts and Equipment, Procedures for Packaging of.

Drawings:

Defense Personnel Support Center (DPSC)

24189	- Excavators, Dental, Corrosion-Resisting Steel.
24335	- Excavator, Dental, Endodontic, Corrosion-Resisting Steel, No. 31.
24369	- Excavator, Dental, Corrosion-Resisting Steel, Darby Perry.
24402	- Excavator, Dental No. 18 and No. 19
24509	- Excavator, Dental, Endodontic, No. 33L

(Copies of Military specifications, standards, and drawings required by contractors in connection with specific acquisition functions should be obtained from the contracting activity or as directed by the contracting activity).

Source: https://assist.dla.mil -- Downloaded: 202640-26T11:53Z
Check the source to verify that this is the currentversion before use.

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposal shall apply.

American Society for Testing and Materials (ASTM)

E 18	- Standard Methods of Test for Rockwell Hardness and Rockwell Superficial Hardness of Metallic Materials.
E 92	- Standard Method of Test for Vickers Hardness of Metallic Materials.

(Application for copies should be addressed to the American Society for Testing and Materials, 1916 Race Street, Philadelphia, PA 19103).

2.3 Order of precedence. In the event of a conflict between the text of this specification and the references cited herein, the text of this specification takes precedence. Nothing in this specification, however, supersedes applicable laws and regulations unless a specific exemption has been obtained.

3. REQUIREMENTS

3.1 General requirements. The excavators shall be in accordance with the applicable requirements of GG-I-526 with the options and additional requirements specified herein.

3.2 Material. One piece excavators shall be fabricated of class 4, martensitic stainless steel, type 440A or 440C. Operative ends of three piece instruments shall be class 4, martensitic stainless steel, type 440C. Handles of three piece instruments shall be class 3, austenitic stainless steel, type 303; class 4, martensitic stainless steel, type 416; or class 6, ferritic stainless steel conforming to the requirements of 430F in accordance with FED. STD. No. 66 for composition only. The excavators of three-piece construction shall be capable of being subjected to dry heat without breakdown of the adhesive or brazing of the handle and operative ends as specified in 4.3.2.

3.3 Hardness. After proper stress relieving, the Rockwell hardness of the operative ends shall be C52 to C56 for type 440A and C56 to C60 for type 440C. Handles and shanks shall be properly tempered and shall have a hardness suitable for the purpose intended. Hardness shall be determined as specified in 4.3.1.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

3.4 Style, design and dimensions. Style I, II, III, IV and V excavators shall be one or three piece (handle and two operative ends). Style I excavators shall be double-ended excavators of Dr. Black's design modified by Dr. Ferrier. Style II excavators shall be double-ended excavators of Dr. Weldelstaedt's design modified by Dr. Ferrier. Style III excavator shall be double-ended excavators of Darby-Perry design. Style IV shall be double-ended endodontic excavators. Style V shall be double-ended excavators. The style, design and dimensions for the style excavators shall be in accordance with the applicable Defense Personnel Support Center (DPSC) drawings as specified in table I. The handle of one piece instruments shall be octagonal in cross section and shall have serrations on seven faces and a smooth face. Handles of three piece instruments shall be one piece, solid, round handles. The finger grip portions of the round handles shall be finely serrated or fine diamond knurled. Styles I and II excavators shall conform to the Black's formula as specified in the applicable drawing and as specified in table IV. Excavator handles shall be straight. Excavators shall be suitable for use in hand excavating and forming of cavities, cleaving or splitting enamel and removal of dentine. Endodontic excavators shall be suitable for deep excavations during endodontic procedures.

3.4.1 Construction. The excavators of three piece construction, consisting of a handle and an operative end at each end shall be permanently secured to the handle. There shall be no movement, loosening, or separation between the operative end and the handle when the operative end is subjected to an 85 pound (minimum) pull out force, when tested as specified in 4.3.3.

TABLE I. Drawing numbers

Style	Instrument Number	Drawing
I	15, 16 (Alternate Number 10, 11)	24189
I	36, 37 (Alternate Number 42, 43)	24189
I	38, 39	24189
II	17, 18 (Alternate Number 8, 9)	24189
III	5, 6	24369
IV	31	24335
IV	33L	24509
V	18, 19	24402

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

3.4.2 Drawings. The drawings forming a part of this specification are not production drawings. These drawings establish dimensional limitations and control design and configuration. The supplier shall be responsible for preparing his own shop drawings. The supplier shall select dimensions and tolerances within the limitations prescribed on the drawings to provide a product with proper design, configuration and function. No deviations from the dimensional limitations prescribed on the drawings is permissible without prior approval of the contracting officer.

3.5 Finish.

3.5.1 Operative ends. All excavator blades shall have sharp cutting edges and shall be free of defects such as coarse grind marks, cracks, burrs, or feathers when viewed as specified in 4.2.1. Operative ends (blades), including sharpened edges of beveled portions, of the excavators listed in table II shall have a fine honed finish.

TABLE II. Fine honed finish.

Style	Instrument number
I	15, 16
II	17, 18

Operative ends (blades) of excavators listed in table III shall have a smooth mirror finish on the outer face and a fine honed finish on the inner face. The junction of the inner and outer faces (cutting edges) shall form a sharp edge around the periphery of the blade.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

TABLE III. Fine honed and mirror finish.

Style	Instrument number
I	36, 37
I	38, 39
III	5, 6
IV	31, 33L
V	18, 19

The fine honed finish specified herein on excavator faces shall be a finish such as produced by using grade FF emery on a leather wheel.

3.5.2 Handles and shanks. Handles and shanks shall have a mirror or bright polished finish except for the finely serrated or fine knurled finger grip portions of handles of three piece instruments. Handles of one piece instruments shall have 40 ± 10 fillet or notched type serrations per inch on seven faces. Depth of serrations shall be 0.002 to 0.004 inch.

3.6 Corrosion resistance. Excavators shall meet the requirements of GG-I-526 for resistance to copper sulfate and resistance to boil for the applicable material. The corrosion tests as specified in GG-I-526 for type 430 shall apply to type 430F material.

3.7 Identification marking. All markings shall be on the smooth face of the octagonal handle of one piece excavators and on the smooth portion of the round handle of three piece excavators. In addition to the requirements of GG-I-526, each excavator shall be marked with the respective instrument number, Black's formula (where applicable) and other identification marking specified, as listed in table IV. The word "Stainless" shall also be marked on instruments of foreign manufacture. As an alternate to the word "stainless", the use of a registered trade name identifying the steel as stainless shall be acceptable, provided prior approval is recceived from the cognizant contracting officer for one piece octagonal instruments. The use of "SS" in lieu of "stainless" is acceptable.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

TABLE IV. Identification marking.

Style	Instrument number	Black's Formula or other Identification marking
I	15, 16 (Alternate number 10, 11)	15-8-14
I	36, 37 (Alternate number 42, 43)	15-8-14
I	38, 39	11-1/2-7-14
II	17, 18 (Alternate number 8, 9)	10-6-14 (Alternate Black's Formula 10-7-14)
III	5, 6	DPE
IV	31, 33L	-
V	18, 19	-

3.8 Workmanship. In addition to the requirements of GG-I-526, the operative end shall be free of discolorations, pits and rounded edges.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure that supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

4.1.2 Certificates of quality. Certificates of quality, supplied by the manufacturer of the stainless steel, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The Certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier of the stainless steel, the NSN, the Item Identification, the name of the material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item. Sampling for inspection shall be performed in accordance with MIL-STD-105, except where otherwise indicated hereinafter.

4.2.1 Examination. The excavators shall be examined for defects including, but not limited to, those listed in table V. Examination for defects 102 and 103 shall be under four power (4X) magnification.

TABLE V. Classification of defects

Categories	Defects
Critical	None defined.
Major A:	
101	Style and design not as specified for the applicable type and style excavator (see 3.4).
102	Cutting edge of operative end not sharp or has rounded edges.
103	Operative end not free from coarse grind marks, cracks, burrs, deep pits, feathers and other imperfections.
104	Operative end finish not as specified.
Major B:	
151	Operative end dimensions not as specified.

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:53Z
Check the source to verify that this is the current version before use.

TABLE V. Classification of defects (Cont.)

Categories	Defects
Minor A:	
201	Handle or shank finish not as specified.
202	Handle or shank finish not free from coarse grind marks.
203	Handle not straight.
204	Identification marking not present, not properly located, illegible, or incomplete.
205	Excavators not free from dirt, polishing compound, and foreign matter.
Minor B:	
251	Dimensions (other than 151) not as specified on drawing.
252	Handle serrations not as specified.
253	Handle serrations missing, not uniform or not complete.

The inspection levels and acceptable quality levels (AQLs) shall be listed in table VI. The unit of product for examination purposes shall be one excavator.

TABLE VI. Examination criteria

Type defect	Inspection level	AQL (percent defective)
Major A	II	1.0
Major B	S-2	1.5
Minor A	II	2.5
Minor B	S-2	4.0

4.2.2 Testing. Excavators shall be tested as specified in table VII and 4.3. Inspection levels and acceptable quality levels (AQLs) shall be as listed in table VII. The unit of product for test purposes shall be one excavator.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

TABLE VII. Tests

Characteristic	Requirement	Test procedures	Inspection level	AQL
Composition	3.2 and GG-I-526	GG-I-526	S-1*	**
Hardness	3.3 and GG-I-526	4.3.1	S-2***	1.5
Pull test	3.4.1	4.3.3	S-2	1.0
Temperature test	3.2	4.3.2	S-2	1.0
Copper Sulfate Test	3.6 and GG-I-526	GG-I-526	S-2	1.0
Boil Test	3.6 and GG-I-526	GG-I-526	S-2	1.0

* A sufficient quantity of each lot of metal used in the construction of the excavators shall be selected for composition analysis in lieu of sample excavators, provided lot identity has been maintained and can be demonstrated to the Government.

** Acceptance number shall be zero for all sample sizes:

*** In lieu of sampling plans for normal inspection of MIL-STD-105, sampling plans for reduced inspection may be used when the inspection lot or batch complies with both of the below listed conditions:

a. consists of units manufactured from the same melt of metal, specified chemical composition, produced by the same manufacturer at the same plant, and using the same facilities at essentially the same time.

b. consists of units subjected to the same heat treatment processes simultaneously in the same equipment; except that when continuous belt automatic heat treatment equipment is used, a heat treatment lot shall be considered no more than one day's production under constant conditions.

4.2.2.1 Sampling unit usage. Where feasible, the same unit of product shall be used for the determination of two or more test characteristics.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

4.3 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.3.1 Hardness. Hardness shall be determined in accordance with ASTM Designation E 18 (Rockwell) or E 92 (Vickers or Diamond Pyramid).

4.3.2 Temperature test. The finished three-piece excavators shall be subjected to dry heat at 193 ± 5°C (360 ± 9°F) for 48 half-hour cycles. At the end of each cycle, with the instrument at room temperature the pull test must be performed.

4.3.3 Pull test. The handle and shank of the operative end shall be securely held in separate chucks or other holding devices. A slowly increasing pull out force, increasing from zero to 85 pounds (minimum), shall be applied to the shank of the operative end. The force shall be maintained at the high value for a minimum of 1 second. The instrument shall then be examined for movement, looseness, or separation of the operative end with respect to the handle. The test shall be repeated for each operative end of the double-ended instrument.

4.4 Inspection of packaging. The sampling and inspection of the preservation and interior package marking shall be in accordance with groups A and B quality conformance inspection requirements of MIL-P-116. The sampling and inspection of the packing and marking for shipment and storage shall be in accordance with the quality assurance provisions of the applicable container specification and the marking requirements of MIL-STD-129.

4.5 Metric Products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

5. PACKAGING

5.1 Preservation. Preservation shall be Level A or C, as specified (see 6.1).

5.1.1 Level A.

5.1.1.1 Cleaning and drying. Excavators shall be cleaned and dried in accordance with MIL-P-116.

5.1.1.2 Preservation applications. Preservatives shall not be used.

5.1.1.3 Unit package. Each excavator shall be packaged in accordance with MIL-P-116, method IC-1. As an alternate, each excavator shall be packaged in a rigid plastic tube having a metal or plastic screw cap or plastic plug type closure. The excavator shall be immobilized within the container and the operative ends protected with cork or other suitable material.

5.1.1.4 Intermediate package. Twelve unit packages containing excavators of one style and number only shall be packaged in a box conforming to PPP-B-566 or PPP-B-676. Closure shall be as specified in the applicable box specification.

5.1.2 Level C. The level C preservation for excavators shall conform to the MIL-STD-794 requirements for this level.

5.2 Packing. Packing shall be level A, B or C, as specified (see 6.1).

5.2.1 Level A. Twenty-four intermediate packages (288 units) containing excavators of one style and number only shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-601, overseas type, PPP-B-621, class 2 or PPP-B-636, class weather-resistant. Closure and strapping shall be as specified in the applicable box specification.

5.2.1.1 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall be as specified in the liner specification. Case liner shall not be required for fiberboard boxes. Each fiberboard box shall be waterproofed as specified in PPP-B-636.

5.2.2 Level B. Twenty-four intermediate packages (288 units) containing excavators of one style and number only shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-601, domestic type, PPP-B-621, class 1 or PPP-B-636, class domestic. Closure shall be as specified in the applicable box specification.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

5.2.3 Level C. The level C packing for excavators shall conform to the MIL-STD-794 requirements for this level.

5.2.4 Packing variation permitted. If the required number of units to be shipped is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which will insure safe delivery to destination.

5.2.5 Unitized loads. Unitized loads, commensurate with the level of packing specified in the contract or order, shall be used whenever total quantities for shipment to one destination exceed 250 pounds or 20 cubic feet. Loads shall be unitized on pallets conforming to MIL-STD-147. Quantities less than 20 cubic feet need not be unitized. Pallet loads including the pallet shall not exceed 54 inches in height, 48 inches in width, and 40 inches in length. Pallets shall conform to Type IV or Type V.

5.2.5.1 Levels A and B. Excavators packed as specified in 5.2.1 and 5.2.2 shall be unitized on pallets as specified in 5.2.5.

5.2.5.2 Level C. Excavators packed as specified in 5.2.3 shall be unitized as specified in MIL-STD-794.

5.3 Marking.

5.3.1 Civil agencies. Unless otherwise specified, markings for shipment shall be as specified in FED-STD-123.

5.3.2 Military requirements.

5.3.2.1 Levels A, B and C. Each unit pack, intermediate package, exterior container and unitized load shall be marked as specified in MIL-STD-129.

5.4 General.

5.4.1 Exterior container. Exterior container (see 5.2.1, 5.2.2, and 5.2.3) shall be of minimum tare and cube consistent with the protection required and shall contain equal quantities of identical stock numbered items to the greatest extent practicable.

5.4.2 Packaging inspection. The inspection of these packaging requirements shall be in accordance with 4.4.

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:53Z
Check the source to verify that this is the current version before use.

6. NOTES

6.1 Ordering data.

6.1.1 Acquisition requirements. Acquisition documents should specify the following:

(a) Title, number, and date of this specification.
(b) Style required (see 1.2.1).
(c) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).
(d) Approval of trade name, if applicable (see 3.7).

6.2 National Stock Number (NSN) coverage. This specification covers the following items appearing in the Federal Supply Catalog:

NSN	ITEM IDENTIFICATION	STYLE
6520-00-000-0071	EXCAVATOR, DENTAL, Darby-Perry, No. 5 and 6	III
6520-00-149-0278	EXCAVATOR, DENTAL, Endodontic, No. 31	IV
6520-00-330-4606	EXCAVATOR, DENTAL, Corrosion-Resisting Steel, No. 19	V
6520-00-935-7181	EXCAVATOR, DENTAL, Black's Formula 11-1/2, 7, 14; No. 38 and 39, Corrosion-Resisting Steel	I
6520-00-935-7182	EXCAVATOR, DENTAL, Black's Formula 15, 8, 14; No. 15 and 16, Corrosion-Resisting Steel	I
6520-00-935-7183	EXCAVATOR, DENTAL, Black's Formula 10, 6, 14; No. 17 and 18, Corrosion-Resisting Steel	II
6520-00-935-7184	EXCAVATOR, DENTAL, Black's Formula 15, 8, 14; No. 36 and 37, Corrosion-Resisting Steel	I
6520-01-024-7006	EXCAVATOR, DENTAL, Corrosion-Resisting Steel, No. 18	V
6520-01-126-4029	EXCAVATOR, DENTAL, Endodontic, No. 33L	IV

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

6.3 Subject term (key word) listing:
Black's Formula
Corrosion-Resisting Steel Excavators
Darby-Perry
Dental Excavators
Discoid No. 38, 39
Enamel Hatchet No. 15,16,17,18
Endodontic Excavators
Spoon No. 36, 37
Wedelstaedt's Modified

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Review Activities:

Army - MD
Navy - MS
Air Force - 03

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

05425

Project No. 6520-1947

Location: NCR Disk #6

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:53Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE: This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO. 4956 WASHINGTON D. C.
POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-ATTS
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 202640-26T11:53Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
GG-E-1230C	EXCAVATORS, DENTAL (Corrosion Resisting Steel)

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)*:

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T14:59Z
Check the source to verify that this is the current version before use.