GG-F-001240A(DSA-DM)
August 23, 1972
Interim revision of
Fed. Spec.
GG-F-1240
February 13, 1968

INTERIM FEDERAL SPECIFICATION

FORCEPS, TOOTH EXTRACTING

This interim Federal specification was developed by the Defense Personnel Support Center, 2800 South 20th Street, Philadelphia, Pennsylvania 19101, based upon currently available technical information. It is recommended that Federal agencies use it in procurement and forward recommendations for changes to the preparing activity at the address shown above. The General Services Administration has authorized the use of this interim Federal specification as a valid exception to Federal Specification GG-F-1240.

1. SCOPE AND CLASSIFICATION

1.1 Scope. This specification covers dental tooth extracting forceps.

1.2 Classification.

1.2.1 Types, sizes, classes, and styles. The forceps covered by this specification shall be of the following types, sizes, classes, and styles, as specified (see 6.1).

Type I. Maxillary (upper jaw) forceps.

Size A. Adult.

Class 1. Anterior.

Style A. No. 65.

Class 2. Posterior.

Style A. No. 53R.
Style B. No. 53L.

FSC 6520

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

Type I. (cont'd.)

Size A. Adult.

Class 3. Universal (see 6.5).

Style A. No. 150.
Style B. No. 150AS.
Style C. No. 210S.

Size B. Child.

Class 1. Universal (see 6.5).

Style A. No. 150S.

Type II. Mandibular (lower jaw) forceps.

Size A. Adult.

Class 1. Universal (see 6.5).

Style A. No. 13.
Style B. No. 17.
Style C. No. 22.
Style D. No. 151.
Style E. No. 151AS.
Style F. No. 222.

Size B. Child.

Class 1. Universal (see 6.5).

Style A. No. 151S.

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposal, form a part of the specification to the extent specified herein:

Source: https://assist.dla.mil -- Downloaded: 202640-26T11:53Z
Check the source to verify that this is the currentversion before use.

SPECIFICATIONS

Federal

GG-I-526	Instruments, Dental and Surgical; General Specifications.
PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard.
PPP-B-676	Boxes, Setup.

STANDARDS

Federal

FED-STD-66	Steel, Chemical Composition and Hardenability.
FED-STD-123	Marking for Domestic Shipment, (Civilian Agencies).
FED-STD-151	Metal, Test Methods.

(Activities outside the Federal Government may obtain copies of Federal Specifications and Standards as outlined under General Information in the Index of Federal Specifications and Standards and at the prices indicated in the Index. The Index, which includes cumulative monthly supplements as issued, is for sale on a subscription basis by the Superintendent of Documents, U. S. Government Printing Office, Washington, D. C. 20402.

(Single copies of this specification and other product specifications required by activities outside the Federal Government for bidding purposes are available without charge at the General Services Administration Regional Offices in Boston; New York; Atlanta; Chicago; Kansas City, Mo.; Fort Worth; Denver; San Francisco; Los Angeles, Seattle and Washington, D. C.)

(Federal Government activities may obtain copies of Federal Specifications and Standards and the Index of Federal Specifications and Standards from established distribution points in their agencies.)

Source: https://assist.dla.mil -- Downloaded: 202640-26T11:53Z
Check the source to verify that this is the currentversion before use.

SPECIFICATIONS

Military

MIL-P-116	Preservation, Methods of.
MIL-L-10547	Liners, Case, and Sheet, Overwrap; Water-Vaporproof or Waterproof, Flexible.

STANDARDS

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

DRAWINGS

Defense Personnel Support Center (DPSC)

20327	Forceps, Tooth Extracting, No. 150S and No. 151S.
20389	Forceps, Tooth Extracting, No. 150 and No. 151.
23019	Forceps, Tooth Extracting, No. 65.
24003	Forceps, Tooth Extracting, No. 53R and 53L.
24218	Forceps, Tooth Extracting, No. 13 and 22.
24219	Forceps, Tooth Extracting, No. 210S
24233	Forceps, Tooth Extracting, No. 150AS and 151AS.
24310	Forceps, Tooth Extracting, No. 222.
24311	Forceps, Tooth Extracting, No. 17.

(Copies of specifications, standards, drawings, and publications required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 OTHER PUBLICATIONS. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposal shall apply.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

AMERICAN SOCIETY FOR TESTING AND MATERIALS (ASTM)

E 18 Standard Methods of Test for Rockwell Hardness and Rockwell Superficial Hardness of Metallic Materials.
E 92 Standard Method of Test for Vickers Hardness of Metallic Materials.

(Application for copies should be addressed to the American Society for Testing and Materials, 1916 Race Street, Philadelphia, Pa. 19103.)

3. REQUIREMENTS

3.1 General requirements. The forceps shall be in accordance with the applicable requirements of GG-I-526 and as specified herein.

3.2 Standard sample. Unless otherwise specified, as a minimum requirement, the forceps shall be equal to the standard sample for the design characteristics (final contour and finish of beaks) for which the standard sample is referenced (see 3.8.1, 3.9 and 6.4).

3.3 Drawings. The drawings forming a part of this specification are specification control drawings. The supplier shall be responsible for preparing his own shop drawings. Where tolerances prescribed could cumulatively result in improper design, the supplier shall provide a product with dimensions within the tolerances prescribed on the drawing to insure proper function, fit, assembly or operation of the item. No deviation from the prescribed dimensions or tolerances is permissible without prior approval of the contracting officer.

3.4 Material. The forceps shall be made of class 4, type 410X stainless steel or type 420 stainless steel except the carbon content for type 420 shall be 0.15 to 0.22 percent. Pins shall conform to the composition requirements of AISI 430 stainless steel in accordance with FED-STD-66. The screw specified for type II, size A, class 1, style A (No. 13) and style C (No. 22) forceps shall be made of class 4, type 416 stainless steel.

3.5 Hardness . After proper stress relieving, the forceps (beaks and handles) shall have a Rockwell hardness of C42 to C49. The joint section of type 410X forceps shall have a Rockwell hardness of C34 to C37. The joint section of type 420 forceps shall have a Rockwell hardness of C42 to C49. Hardness shall be determined as specified in 4.4.2.

3.6 Deformation. The forceps shall withstand the minimum loads as listed in table I for the applicable type, size, class, and style of forceps when tested as specified in 4.4.3. During or as a result of testing, no chipping, cracking, fracturing, or permanent deformation shall occur on any portion of the forceps, and the beaks, pins, or screws shall not show any signs of weakness.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

Table I. Minimum deformation loads

Type	Size	Class	Style
100 Pounds			
I	A	1	A (No. 65)
I	A	3	A (No. 150), B (No. 150AS)
I	B	1	A (No. 150S)
II	A	1	A (No. 13), C (No. 22), D (No. 151), E (No. 151AS)
II	B	1	A (No. 151S)
150 Pounds			
I	A	3	C (No. 210S)
II	A	1	F (No. 222)
200 Pounds			
I	A	2	A (No. 53R), B (No. 53L)
II	A	1	B (No. 17)

3.7 Style, design, and dimensions. The forceps shall be extracting forceps suitable for the purpose specified in table II, and shall conform to the Defense Personnel Support Center (DPSC) drawing specified in table II for the applicable type, size, class, and style.

Table II

Drawing number and purpose of each type, size, class, and style of forceps

Type	Size	Class	Style	DPSC drawing	Primary purpose
I	A (No. 65)	1	A	23019	For all upper roots, fragments, and overlapping centrals and laterals.
I	A (No. 53R)	2	A	24003	For upper molars, first and second, right side of jaw.
I	A (No. 53L)	2	B	24003	For upper molars, first and second, left side of jaw.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

Table II
Drawing number and purpose of each type, size, class, and style of forceps (cont'd.)

Type	Size	Class	Style	DPSC drawing	Primary purpose
I	A (No. 150)	3	A	20389	For upper centrals, laterals, bicuspids and roots.
I	A (No. 150AS)	3	B	24233	For upper first and second bicuspids.
I	A (No. 210S)	3	C	24219	For upper third molars.
I	B (No. 150S)	1	A	20327	For upper deciduous centrals, laterals, bicuspids, and roots.
II	A (No. 13)	1	A	24218	For lower first and second bicuspids.
II	A (No. 17)	1	B	24311	For lower first and second molars.
II	A (No. 22)	1	C	24218	For lower first, second and third molars.
II	A (No. 151)	1	D	20389	For lower centrals, laterals, bicuspids, and roots.
II	A (No. 151AS)	1	E	24233	For lower first and second bicuspids.
II	A (No. 222)	1	F	24310	For lower third molars.
II	B (No. 151S)	1	A	20327	For lower deciduous centrals, laterals, bicuspids, and roots.

3.8 Construction. The handles shall have male diamond knurling with 10 to 20 teeth per inch. Knurling shall be deep and the raised points smooth. The handles shall provide a comfortable grip and shall be free from rough edges and sharp corners. The joints of the forceps shall be either lap joint (scissors) type or box lock type, except type II, size A, class 1, style A (No. 13) and C (No. 22), which shall be a screw joint type. The edges of the joints shall be properly beveled to prevent pinching of tissues. The forceps shall be free from excessive stiffness and shall not have excessive play or indication of looseness when the forceps are in the open or closed position.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

3.8.1 Beaks. Concave surfaces (grooves) of beaks shall be properly centered with respect to the long axis of the beaks. All grooves, surfaces, and tips shall be well defined and properly formed. The final contour of the beaks shall be equal to the standard sample (see 3.2 and 6.4).

3.9 Finish. The forceps shall have a mirror finish, except as otherwise specified herein. The gripping portion of the handle shall be knurled and shall have a bright finish. The concave surfaces of the grooves in the beaks shall have a fine milled or ground finish, except for type I, size A, class 3, style B (No. 150AS) and type II, size A, class 1, style E (No. 151AS). The beaks of type I, size A, class 3, style B (No. 150AS) and type II, size A, class 1, style E (No. 151AS) shall have longitudinal coarse markings (file type grooves) on the inside surfaces, serrations on the distal side edges (perpendicular to the longitudinal axis of the beaks), and notched ('V' shaped) grasping edges. The periphery of the distal side edges of the beaks of type I, size B, class 1, style A (No. 150S) and type II, size B, class 1, style A (No. 151S) forceps shall be serrated (perpendicular to the longitudinal axis of the beaks). The finish of the grooves and grasping edges of the forceps shall be equal to the standard sample (see 3.2 and 6.4).

3.10 Weight. The weight of the applicable type, size, class, and style forceps shall be in accordance with table III. The tolerance on weight shall be plus or minus 1/2 ounce.

Table III. Weight of forceps

Type	Size	Class	Style
4 ounces			
I	B	1	A (No. 150S)
II	B	1	A (No. 151S)
5-1/2 ounces			
I	A	1	A (No. 65)
II	A	1	A (No. 13)
II	A	1	C (No. 22)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

Table III. Weight of forceps (cont'd.)

6 ounces			
Type	Size	Class	Style
I	A	2	A (No. 53R), B (No. 53L)
I	A	3	A (No. 150)
I	A	3	B (No. 150AS)
I	A	3	C (No. 210S)
II	A	1	B (No. 17)
II	A	1	D (No. 151)
II	A	1	E (No. 151AS)
II	A	1	F (No. 222)

3.11 Corrosion resistance. Forceps shall meet corrosion resistance requirements (copper sulfate and boil test) of GG-I-526 when tested in accordance with 4.4.5.1 and 4.4.5.2.

3.12 Identification marking. Each forceps shall be legibly and permanently marked with the applicable forceps number on the outside portion of the handle (free end below the knurled portion) except where the knurled portion extends to the end of the handle (free end) then the applicable forceps number shall be marked on the inside portion of the handle. In addition to contractor marking, each forceps shall be marked with the name or trademark of the manufacturer, if different from that of the contractor. Forceps of foreign manufacture shall also be marked "stainless".

3.13 Workmanship. Workmanship shall be first class throughout. Forceps shall be free from defects which detract from their appearance or impair their serviceability, proper functioning, and intended use.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in this specification where such inspections are deemed necessary to assure that supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the raw material, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and demonstrated to the Government. The Certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the FSN, the Item Identification, the name of the raw material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Quality conformance inspection. The examinations and tests required to assure conformance of those items or lots of items to be offered for acceptance are classified as specified in tables IV, VI, and VII (see 4.2.1, 4.3 and 4.3.1) and table V (see 4.2.2).

4.2.1 Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105. The inspection levels and acceptable quality levels (AQL's) shall be as indicated in table IV. Unit of product for examination purposes shall be one forceps.

Table IV. Acceptable quality levels

Type defect	Inspection level	AQL (percent defective)
Major A	II	1.0
Major B	S-1	2.5
Minor defects	II	4.0
Preparation for delivery defects	S-2	4.0

4.2.2 Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and table V. Unit of product for test purposes shall be one forceps. Where feasible, the same sample unit shall be used for the determination of two or more test characteristics. Lots for hardness testing shall be as specified in 4.2.2.1.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

Table V. Tests

Characteristic	Requirement	Test procedure	Inspection level	AQL (percent defective)
Composition	3.4	4.4.1	See 4.2.3	*
Hardness	3.5	4.4.2	S-2	*
Deformation test	3.6	4.4.3	S-2	*
Weight	3.10	4.4.4	S-2	4.0
Copper sulfate test	3.11	4.4.5.1	S-2	*
Boil test	3.11	4.4.5.2	S-2	*

* Acceptance number shall be zero for all sample sizes.

4.2.2.1 Lots for hardness criteria. For finished forceps hardness testing, a lot or batch shall consist of forceps produced by one manufacturer, at one plant, from the same material (same raw material heat and heat treatment) and from the same heat treatment lot or batch (forceps undergoing heat treatment at the same time, in the same tray, rack or other containing device).

4.2.3 Sampling of raw materials. A sufficient amount from each lot of each metal used in the fabrication of the forceps shall be selected for chemical analysis.

4.3 Examination. The forceps shall be examined to determine compliance with specification requirements. Examination shall be in accordance with 4.2.1. The forceps shall be examined for defects, including but not limited to, those listed in Table VI.

Table VI. Classification of defects

Categories	Defects
Critical:	None defined.
Major A:	
101	Style and design not as specified for type, size, class, and style of instrument.
102	Instrument not constructed as specified.
103	Beak grooves not properly centered.
104	Beak surfaces and tips not well defined and properly formed.
105	Finish of beak, beak grooves, and grasping edges of forceps not as specified.
106	Beak serrations not as specified (where applicable).
107	Beak not free from cracks or burrs.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

Table VI. Classification of defects (cont'd.)

Categories	Defects
Critical:	None defined.
Major A: (cont'd.)	
108	Edges of the joint of the forceps not properly beveled.
109	Joint of forceps not free from excessive stiffness.
110	Joint of forceps has excessive play or is excessively loose.
111	Contour of beaks not as specified.
112	Interfaces of lock joint are not regular or have large crevices.
113	Beak grasping edge "V" notch missing or not formed properly (Type I, Size A, Class 3, Style B (No. 150AS) and Type II, Size A, Class 1, Style E (No. 151AS).
114	Beak serrations missing or incomplete (Type I, Size A, Class 3, Style B (No. 150AS), Type II, Size A, Class 1, Style E (No. 151AS), Type I, Size B, Class 1, Style A (No. 150S), and Type II, Size B, Class 1, Style A (No. 151S).
Major B:	
121	Beak dimensions not as specified.
Minor:	
201	Handle finish not as specified.
202	Instrument not permanently and legibly marked with name or trademark of the contractor.
203	Instrument number not present or properly located.
204	Handle, joint and overall dimensions not as specified.
205	Knurling or grooves on handle not as specified.

4.3.1 Examination for defects in preparation for delivery. An examination shall be made to determine that the packaging, packing, and marking comply with the requirements of section 5. Defects shall be scored in accordance with table VII. For examination of interior packaging, the sample unit shall be one container fully prepared for delivery, selected at random just prior to the closing operations. Defects of closure listed shall be examined on shipping containers fully prepared for delivery. The lot size shall be the number of shipping containers in the end inspection lot.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

Table VII. Classification of preparation for delivery defects

Examine	Defects
Marking (interior and exterior)	Marking omitted, incorrect, illegible, improper size, location, sequence, or method of application.
Quantity	Quantity in unit and intermediate packages and shipping container not as specified.
Materials	Any component missing or damaged. Intimate wrap not as specified.
Containers	Containers not as specified.
Workmanship	Inadequate application of components such as incomplete closure of container flaps, loose strapping, or distortion of container.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Chemical Composition. In the absence of certificates of quality for the metallic raw material, chemical analysis shall be conducted in accordance with method 111 or 112 of FED-STD-151. In case of dispute, method 111 shall govern.

4.4.2 Hardness. Hardness shall be determined in accordance with ASTM Designation E18 (Rockwell) or E92 (Vickers or Diamond Pyramid).

4.4.3 Deformation test. A brass plate shall be placed between the beaks of the forceps and the load, as specified in 3.6, shall be gradually applied over a period of 25 to 35 seconds, at points of greatest convexity of the handles. The test plate shall meet the requirements of class 8, alloy number 270, yellow brass for composition only. The test plate shall have a Rockwell hardness of B70 to B90. For forceps with a clearance of 0.007 inch or less between the tips of the beaks in the closed position, the minimum thickness of the plate shall be 1/8 inch. For a clearance of greater than 0.007 inch, the minimum thickness of the test plate shall be 1/4 inch. For all forceps, the test plate shall be of sufficient thickness so that the handle ends (free ends) and the beak tips shall not touch when the forceps are subjected to full load as specified.

4.4.4 Weight. The forceps shall be weighed on a calibrated scale of suitable sensitivity and accuracy.

4.4.5 Corrosion test.

4.4.5.1 Copper sulfate test. The copper sulfate test shall be conducted in accordance with GG-I-526.

4.4.5.2 Boil test. The boil test shall be conducted in accordance with GG-I-526.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

5. PREPARATION FOR DELIVERY

5.1 Packaging. Packaging shall be level A or C, as specified (see 6.1).

5.1.1 Unit of issue. Each forceps, as specified, constitutes one unit of issue.

5.1.2 Packaging quantities. The number of units of issue indicated in table VIII shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in the procurement document.

Table VIII. Packaging Quantities

Type	Size	Class	Style	Unit package	Intermediate package	Exterior container
I	A and B	1	A	1	6	96
I	A	2	B	1	6	96
I	A	3	A	1	6	96
I	A	3	B and C	1	6	144
II	A	1	B, D and F	1	6	96
II	B	1	A	1	6	96
II	A	1	A, C and E	1	6	144

5.1.2.1 Packing variation permitted. If the required number of units to be shipped is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

5.1.3 Level A.

5.1.3.1 Cleaning and drying. Each unit shall be thoroughly cleaned and dried in accordance with the applicable method as outlined in MIL-P-116.

5.1.3.2 Preservative. Not required.

5.1.3.3 Intimate wrap. Each unit shall be wrapped in neutral tissue or other suitable anticorrosive materials.

5.1.3.4 Unit package. Each unit shall be packaged in accordance with MIL-P-116, method IC-2.

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:53Z
Check the source to verify that this is the current version before use.

5.1.3.5 Intermediate package. Six forceps shall be intermediate packaged in a box of appropriate size constructed in accordance with PPP-B-676, except that commercial colors will be acceptable, or PPP-B-636, type CF, class domestic. Closure shall be adequate to prevent accidental opening under normal handling.

5.1.4 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing. Packing shall be level A, B, or C, as specified (see 6.1).

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type; PPP-B-621, class 2; or PPP-B-636, class weather-resistant, grade V3c. Fiberboard boxes shall conform to the special requirements specified in PPP-B-636.

5.2.1.2 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes.

5.2.1.3 Closure. Closure of wood boxes shall be in accordance with appendix of applicable box specification. Closure of each fiberboard box shall be as specified in the appendix, and waterproofing shall conform with 30.4 of PPP-B-636.

5.2.1.4 Strapping. Strapping, when required, shall be in accordance with appendix of applicable box specification.

NOTE: Strapping shall not be required for shipments forwarded to a receiving activity within the continental limits of the United States for storage and redistribution.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, domestic type; PPP-B-621, class 1; or PPP-B-636, class domestic. Closure shall be as specified in the appendix of the applicable box specification. Fiberboard box shall be closed by method II of PPP-B-636. Fiberboard boxes shall conform to special requirements specified in PPP-B-636.

Source: https://assist.dla.mil -- Downloaded: 202640-26T11:53Z
Check the source to verify that this is the currentversion before use.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Civil agencies. In addition to any special marking as specified in the contract or order, shipments shall be marked in accordance with Fed. Std. No. 123.

5.3.2 Military agencies.

5.3.2.1 Interior packages. Each unit package and each intermediate package shall be marked in accordance with MIL-STD-129. When labels are utilized, waterproofing shall be required only when applicable carton is fabricated of water-resistant material. Contract or purchase order number and name of contractor shall be shown.

5.3.2.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

(a) Title, number, and date of this specification.
(b) Type, size, class, and style required (see 1.2).
(c) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.2 Specification changes. Because of the extensiveness of changes to this document, the margins have not been marked with an asterisk to indicate where changes from the previous issue have been made.

6.3 Limitations. This specification does not cover all types, classes, grades, styles, or sizes of the commodity indicated by the title of this specification, or those which are commercially available, but is intended to cover those items which are normally procured to meet Federal Government requirements.

6.4 Standard sample. For access to standard sample, address the procuring activity issuing the invitation for bids or request for proposal (see 3.2, 3.8.1 and 3.9).

6.5 Universal class. The classification "Universal" designates forceps designed for extracting teeth in the right and left sides of the same jaw, but not for extracting teeth in the opposing jaw.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activities:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. - 6520-1330

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

☆ U. S. GOVERNMENT PRINTING OFFICE: 1972 -714-910/388

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.