GG-P-001367(DSA-DM)
24 July 1967

INTERIM FEDERAL SPECIFICATION

PLUGGERS, AMALGAM, DENTAL (Black)

This interim Federal specification was developed by the Defense Personnel Support Center, 2800 South 20th Street, Philadelphia, Pennsylvania 19101, based upon currently available technical information. It is recommended that Federal agencies use it in procurement and forward recommendations for changes to the preparing activity at the address shown above.

1. SCOPE AND CLASSIFICATION

1.1 Scope. This specification covers amalgam pluggers of Black's design, suitable for condensing amalgam in the tooth cavity.

1.2 Classification. Pluggers shall be furnished in the following shapes as specified (see 6.1);

Shape 1.	Black No. 1.
Shape 2.	Black No. 2.
Shape 3.	Black No. 3.
Shape 4.	Black No. 4.
Shape 5.	Black No. 5.

2. APPLICABLE DOCUMENTS

2.1 Specifications, standards and drawing. The following specifications, standards, and drawing of the issue in effect on date of invitation for bids or request for proposal, form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

GG-I-526	Instruments, Dental and Surgical; General Specifications.
MMM-A-250	Adhesive, Water Resistant, (For Sealing Fiberboard Boxes).
PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard.
PPP-T-76	Tape; Pressure-Sensitive Adhesive Paper, (For Carton Sealing).

FSC 6520

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:54Z
Check the source to verify that this is the current version before use.

STANDARDS

Federal

FED-STD-123 Marking for Domestic Shipment (Civilian Agencies).

(Activities outside the Federal Government may obtain copies of Federal Specifications and Standards as outlined under General Information in the Index of Federal Specifications and Standards and at the prices indicated in the Index. The Index, which includes cumulative monthly supplements as issued, is for sale on a subscription basis by the Superintendent of Documents, U.S. Government Printing Office, Washington, D.C., 20402.

(Single copies of this specification and other product specifications required by activities outside the Federal Government for bidding purposes are available without charge at the General Services Administration Regional Offices in Boston, New York, Atlanta, Chicago, Kansas City, Mo., Dallas, Denver, San Francisco, Los Angeles, Seattle, and Washington, D.C.

(Federal Government activitites may obtain copies of Federal Specifications and Standards and the Index of Federal Specifications and Standards from established distribution points in their agencies.)

SPECIFICATIONS

Military

MIL-P-116 Preservation, Methods of.

STANDARDS

Military

MIL-STD-105 Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 Marking for Shipment and Storage.

DRAWING

DEFENSE MEDICAL SUPPLY CENTER

20396 PLUGGER, AMALGAM, DENTAL, Black, No. 1, 2, 3, 4, & 5.

(Copies of Military specifications, standards, and drawing required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:54Z
Check the source to verify that this is the current version before use.

3. REQUIREMENTS

3.1 General requirements. The amalgam pluggers of Black's design shall be in accordance with applicable requirements of GG-I-526 and as specified herein.

3.2 Material. Pluggers shall bs class 4, type 440A martensitic stainless steel of GG-I-526.

3.3 Style design and dimensions. The amalgam pluggers shall be of Black's design and of one-piece construction, consisting of an operative end, shank, and handle. Pluggers shall be suitable for condensing amalgam in the tooth cavity. The pluggers shall conform to Defense Medical Supply Center (DMSC) Drawing 20396 for style, design, and dimensions for the applicable plugger number.

3.4 Hardness. After proper stress relieving, the operative end of plugger shall have a Rockwell hardness of C52 to C56. Hardness of the shank shall be suitable for the purpose intended. Handle shall have a Rockwell hardness of C20 minimum.

3.5 Finish. Pluggers shall have a mirror finish except for the serrated working faces.

3.5.1 Handles. Handles shall have fillet-type serrations. Depth of serrations shall be 0.002 to 0.004 inch, and there shall be 40 ± 10 serrations per inch.

3.5.2 Operative ends. The operative ends shall be properly serrated. The number of file cuts in serrating the working surface for the pluggers shall be as follows with a tolerance of plus 2 cuts in either direction.

Plugger No.	1	2	3	4	5
File cuts	4 X 4	7 X 7	9 X 9	2 X 10	10 X 2

3.6 Corrosion resistance. The copper sulfate and boil tests, as specified in GG-I-526, shall apply to the entire plugger.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:54Z
Check the source to verify that this is the current version before use.

3.7 Identification marking. In addition to the marking specified in GG-I-526, each plugger shall be permanently and legibly marked with Black's formula, (as shown in table I), and with the respective instrument (Black) number. The instrument number shall appear 1/4 ± 1/16 inch from the proximal end of the instrument.

Table I. Black's Formula

Instrument	Black's Formula
Black No. 1	15-7-12
Black No. 2	25-7-12
Black No. 3	35-7-12
Black No. 4	10-30-7-12
Black No. 5	30-10-7-12

3.8 Workmanship. Workmanship shall be first class throughout. Pluggers shall be free from defects which detract from their appearance or may impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified, the supplier may utilize his own facilities or any commercial laboratory acceptable to the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties).

4.1.2 Certificates of quality. When available, certificates of quality, supplied by the manufacturer of the component or material, listing the specified test method and test results obtained, may be furnished in lieu of actual performance of such testing by the contractor.

Source: https://assist.dla.mil -- Downloaded: 202640-26T11:54Z
Check the source to verify that this is the current version before use.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105. The Inspection and Acceptable Quality Levels (AQL's) shall be as indicated in table II.

Table II. Inspection and Acceptable Quality Levels

Type defect	Inspection level	AQL (Percent defective)
Major A	II	1.0
Major B	S-1	2.5
Minor	II	4.0

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as indicated in table III, except that the acceptance number shall be zero for all sample sizes.

Table III. Tests

Characteristic	Rqmt.	Test procedure	Unit of product	Inspection level
Hardness	3.4	GG-I-526	1-plugger	S-2
Copper sulfate	3.6	GG-I-526	1-plugger	S-2
Boil test	3.6	GG-I-526	1-plugger	S-2

4.2.3 Raw materials. A sufficient quantity of each lot of metal used in the construction of the pluggers shall be selected for chemical analysis.

4.2.4 Sample unit usage. Where feasible, the same sample unit shall be used for the determination of two or more test characteristics.

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:54Z
Check the source to verify that this is the current version before use.

4.3 Examination. Pluggers shall be examined for defects including, but not limited to, those listed in table IV.

Table IV. Classification of Defects

Categories	Defects
Critical:	None defined.
Major A:	
101	Shape not as specified.
102	Finish on operative end not as specified.
103	Serrations on operative end not as specified.
104	Plugger working surfaces not free of feathers, burrs, or cracks.
Major B:	
151	Dimensions of operative end not as specified.
152	Labels or printed marking on unit container not correct (particularly stock number and item identification).
Minor:	
201	Dimensions on handles not as specified.
202	Finish on handles not as specified.
203	Serrations on handles not as specified.
204	Edges and corners of handle not free of burrs and sharpness.
205	Black's formula and instrument number not present or properly located.
206	Instrument not identified with name or registered trademark of manufacturer.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

5. PREPARATION FOR DELIVERY

5.1 Packaging.

5.1.1 Unit of issue. One plugger, as specified, constitutes one unit of issue.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:54Z
Check the source to verify that this is the current version before use.

5.1.2 Packaging quantities. The number of units of issue indicated in table V shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in the procurement document.

Table V. Packaging Quantities

Unit package	Intermediate package	Exterior container
1 unit	12 units	288 units

5.1.2.1 Packing variation permitted. If the required number of units in the entire shipment is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

5.1.3 Level A.

5.1.3.1 Cleaning and drying. Each unit shall be thoroughly cleaned and dried in accordance with the applicable method as outlined in MIL-P-116.

5.1.3.2 Preservative. Not required.

5.1.3.3 Intimate wrap. Each unit shall be wrapped in neutral tissue or other suitable anticorrosive materials.

5.1.3.4 Unit package. Each unit shall be packaged in accordance with MIL-P-116, method 1A-8.

5.1.3.5 Intermediate package. Intermediate package shall be a box of appropriate size and design constructed in accordance with PPP-B-636, type CF, class weather-resistant. Boxes shall be closed by taping the top and bottom flaps with 2-inch wide tape conforming to PPP-T-76. Closing of bottom flaps with adhesive conforming to MMM-A-250 or metal stitches is permitted.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:54Z
Check the source to verify that this is the current version before use.

5.1.4 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing.

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for type 2 load, and constructed in accordance with PPP-B-585, style 3, for class 3 use; PPP-B-601, table II, using type I, class 1 plywood; PPP-B-621, table III; or PPP-B-636, type CF, class weather-resistant. Closure and strapping shall be in accordance with appendix of applicable box specification.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall conform to PPP-B-601, table I, using type III, class 1 plywood; or shall be designed for type 2 load and constructed in accordance with PPP-B-585, style 3, for class 2 use; PPP-B-621, table I; or PPP-B-636, type CF, class domestic. Closure shall be in accordance with appendix of applicable box specification.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Civil agencies. Interior and exterior containers shall be marked in accordance with FED-STD-123.

5.3.2 Military agencies.

5.3.2.1 Interior packages. Each unit package and each intermediate package shall be marked in accordance with MIL-STD-129. When labels are utilized, waterproofing shall be required only when applicable carton is fabricated of water-resistant material. Contract or purchase order number and name of contractor shall be shown.

Source: https://assist.dla.mil -- Downloaded: 202640-26T11:54Z
Check the source to verify that this is the currentversion before use.

5.3.2.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

(a) Number and date of this specification.
(b) Shape required (see 1.2).
(c) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.2 This specification does not cover all types, classes, grades, or sizes of the commodity indicated by the title of this specification, or those which are commercially available, but is intended to cover the types which are normally procured to meet Federal requirements.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparaing activity:
Defense Supply Agency - DM

Review activities:
Army - MD
Navy - MS
Air Force - 03

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

*U.S. GOVERNMENT PRINTING OFFICE: 1967-301-510-/548

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:54Z
Check the source to verify that this is the current version before use.

CAUTION

NOTICE TO BIDDERS/OFFERORS

DO NOT CONDITION OR BASE YOUR BID/OFFER ON ANY CURRENT PROCUREMENT ON THE INFORMATION SUBMITTED ON THIS FORM SINCE ANY CHANGES OR DELETIONS IN THE SPECIFICATIONS MAY RENDER YOUR BID/OFFER NON-RESPONSIVE IN WHICH CASE IT CANNOT BE CONSIDERED FOR AWARD.

Fold

DEFENSE PERSONNEL SUPPORT CENTER
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

DEFENSE SUPPLY AGENCY
OFFICIAL BUSINESS

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

Fold

DPSC FORM 2418
AUG 66

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:54Z
Check the source to verify that this is the current version before use.

SPECIFICATION ANALYSIS SHEET

Form Approved
Budget Bureau No. 22-R255

INSTRUCTIONS: This sheet is to be filled out by personnel, either Government or contractor, involved in the use of the specification in procurement of products for ultimate use by the Department of Defense. This sheet is provided for obtaining information on the use of this specification which will insure that suitable products can be procured with a minimum amount of delay and at the least cost. Comments and the return of this form will be appreciated. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments and suggestions submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or serve to amend contractual requirements.

SPECIFICATION: GG-P-001367(DSA-DM)

ORGANIZATION:

CITY AND STATE:

CONTRACT NUMBER:

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE SPECIFICATION CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY SPECIFICATION REQUIREMENT CONSIDERED TOO RIGID

3. IS THE SPECIFICATION RESTRICTIVE?

☐ YES ☐ NO *(If "yes", in what way?)*

4. REMARKS *(Attach any pertinent data which may be of use in improving this specification. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity)*

SUBMITTED BY *(Printed or typed name and activity - Optional)*

DATE

DD FORM 1 JAN 66 1426 REPLACES EDITION OF 1 OCT 64, WHICH MAY BE USED.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:54Z
Check the source to verify that this is the current version before use.