GG-R-001308a(DSA-DM)
April 28, 1969
SUPERSEDING
Revision of
Int. Fed. Spec. GG-R-001308(DSA-DM)
May 1, 1967
See Section 6(6.3)

INTERIM FEDERAL SPECIFICATION

RAZOR, SURGICAL PREPARATION

This interim Federal specification was developed by the Defense Personnel Support Center, 2800 South 20th Street, Philadelphia, Pennsylvania 19101, based upon currently available technical information. It is recommended that Federal agencies use it in procurement and forward recommendations for changes to the preparing activity at the address shown above.

The General Services Administration has authorized the use of this interim Federal specification.

1. SCOPE AND CLASSIFICATION

1.1 Scope. This specification covers razors and blades suitable for use in preparing patients for surgical procedures.

1.2 Classification. Razors and blades covered by this specification shall be of the following types and compositions as specified (see 6.1):

Type I. Razor, Surgical Preparation.

Type II. Blade, (for type I, Razor).

Composition 1. Corrosion-resisting steel, coated.
Composition 2. Corrosion-resisting steel, uncoated.
Composition 3. Carbon steel.

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 202640-26T11:53Z
Check the source to verify that this is the current version before use.

2. APPLICABLE DOCUMENTS

2.1 Specifications, standards and drawing. The following specifications, standards and drawing of the issue in effect on date of invitation for bids or request for proposal, form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

PPP-B-566	Boxes, Folding, Paperboard.
PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard.
PPP-B-676	Boxes, Setup.

STANDARDS

Federal

FED-STD-66	Steel, Chemical Composition and Hardenability.
FED-STD-123	Marking for Domestic Shipment (Civilian Agencies).
FED-STD-151	Metals; Test Methods.

(Activities outside the Federal Government may obtain copies of Federal Specifications and Standards as outlined under General Information in the Index of Federal Specifications and Standards and at the prices indicated in the Index. The Index, which includes cumulative monthly supplements as issued, is for sale on a subscription basis by the Superintendent of Documents, U.S. Government Printing Office, Washington, D.C., 20402.

(Single copies of this specification and other product specifications required by activities outside the Federal Government for bidding purposes are available without charge at the General Services Administration Regional Offices in Boston, New York, Atlanta, Chicago, Kansas City, Mo., Fort Worth, Denver, San Francisco, Los Angeles, Seattle, and Washington, D.C.

(Federal Government activities may obtain copies of Federal Specifications and Standards and the Index of Federal Specifications and Standards from established distribution points in their agencies.)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

SPECIFICATIONS

Military

MIL-L-10547	Liners, Case, and Sheet, Overwrap; Water-Vaporproof, or Waterproof, Flexible.

STANDARDS

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

DRAWING

Defense Personnel Support Center (DPSC)

20204	Blade and Razor, Safety, Straight Type.

(Copies of Military specifications, standards, and drawing required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposal shall apply.

AMERICAN SOCIETY FOR TESTING AND MATERIALS (ASTM)

ASTM Designation E 18	Standard Methods of Test for Rockwell Hardness and Rockwell Superficial Hardness of Metallic Materials.
ASTM Designation E 92	Standard Method of Test for Vickers Hardness of Metallic Materials.

(Application for copies should be addressed to the American Society for Testing and Materials, 1916 Race Street, Philadelphia, Pa. 19103).

3. REQUIREMENTS

3.1 Material.

3.1.1 Razor. The safety razor shall be fabricated from nickel silver containing 10.0 ± 1.0 percent (%) nickel. The sheath of the blade holder and the razor guard shall be 24-gage Brown and Sharpe (B & S) sheet metal (0.020 inch thick). The core of the blade holder shall be 9-gage B & S sheet metal (0.114 inch thick). The handle shall be 19-gage B & S sheet metal (0.036 inch thick). The pin rivet and washers shall be nickel silver.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

3.1.1.1 Alternate material for handle. As an alternate, the handle shall be fabricated from 3/4-hard AISI 301 corrosion-resisting steel sheet, 0.030 inch thick (U.S. Standard Gage No. 22).

3.1.2 Blade. The rib of the blade shall be fabricated from carbon steel conforming to C1008 of FED-STD-66. The cutting blade shall be fabricated from the materials specified in 3.1.2.1 through 3.1.2.3 for the composition indicated.

* 3.1.2.1 Composition 1, corrosion-resisting steel, coated. The cutting blade shall be fabricated from corrosion-resisting steel containing 0.90 to 1.05 percent carbon, 0.15 to 0.25 percent silicon, 0.90 to 1.10 percent manganese, 12.0 to 14.0 percent chromium, and maximums of 0.040 percent phosphorus and 0.040 percent sulfur. As an alternate, the corrosion-resisting steel shall contain 0.53 to 0.70 percent carbon, 0.15 to 0.50 percent silicon, 0.20 to 0.80 percent manganese, 12.0 to 14.7 percent chromium, and maximums of 0.025 percent phosphorus and 0.020 percent sulfur. The beveled edge shall be treated with a fluorocarbon-telomer coating such as sintered polytetrafluoroethylene.

* 3.1.2.2 Composition 2, corrosion-resisting steel, uncoated. The cutting blade shall be fabricated from either of the corrosion-resisting steels specified in 3.1.2.1, without a plastic coating.

* 3.1.2.3 Composition 3, carbon steel. The cutting blade shall be fabricated from high carbon steel containing 1.20 to 1.30 percent carbon, 0.15 to 0.30 percent silicon, 0.20 to 0.50 percent manganese, 0.15 to 0.50 percent chromium, and maximums of 0.035 percent phosphorus and 0.035 percent sulfur.

* 3.2 Style, design, and dimensions. The surgical razor and the blade shall conform to Defense Personnel Support Center (DPSC) Drawing 20204 for style, design, and dimensions. The razor and the blade shall be suitable for use in preparing patients for surgical procedures.

3.3 Construction.

* 3.3.1 Razor. The razor shall be of the conventional straight safety-razor type, and shall consist essentially of a blade holder, a handle, and a razor guard. The razor shall properly accommodate the removable and reversible blade (type II).

3.3.1.1 Blade holder. The razor blade holder shall consist of a core, a grooved insert, and a sheath.

3.3.1.1.1 Core. One end of the core shall be rounded off and the upper surface provided with a crescent-shaped, smooth grip for the fourth finger, and with a slightly depressed serrated thumb grip on the lower surface. The upper surface of the serrated thumb-grip end shall be grooved to receive the inner end of the grooved insert.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

3.3.1.1.2 Grooved insert. The lower exposed surface of the grooved insert shall have a full-length groove of adequate size in the center width, when assembled in the blade holder to hold the rib of the blade securely and to prevent any appreciable side or end movement of the blade. The inner end shall fit securely in the grooved end of the core and the outer end shall be securely held in place by the lip on the free end of the sheath.

3.3.1.1.3 Sheath. The width of the sheath shall be sufficient to fit tightly over the upper surface of the grooved insert and core, and extend down on each side flush with lower surface of the core, except that the sides between the serrated thumb grip and the free end shall be of uniform width, and one side section between the serrated thumb grip and the inner end shall be extended over the lower surface of the core to the opposite side edge. The center portion of the free end shall be provided with a lip to hold the grooved insert securely in place. The upper corners of the free end shall be slightly rounded. The inner ends shall be rounded off. The core and grooved insert shall be assembled in the sheath and shall be securely held by crimping.

* 3.3.1.2 Handle. The handle and the assembled blade holder shall be attached by a pin rivet and two 5/16-inch washers. The washers shall be placed on the pin rivet between the handle and the holder. The rivet shall hold the parts with sufficient tension for the blade holder to remain folded in the handle to protect the blade when closed, and to stay in an open position during use, but without excessive tension which would prevent the holder from being freely opened and closed.

* 3.3.1.3 Razor guard. The razor guard shall be a removable and reversible tooth guard, as shown in the drawing.

* 3.3.2 Blade. The blade shall be two-piece, consisting of a rib and a cutting blade. The rib shall securely hold the cutting blade. The cutting blade shall contain a single cutting edge. The central portion of the cutting blade may contain an optional rectangular hole, provided the hole is required during the manufacturing process, and does not exceed the maximum limits specified on the drawing.

3.4 Finish.

* 3.4.1 Razor. The razor components fabricated from nickel silver sheet metal shall be chromium plated. All exterior surfaces of the razor shall have a mirror or bright polished finish. The outer ends of the nickel silver pin rivet attaching the blade holder in the handle shall be smoothly peened.

* 3.4.2 Blade. The cutting edge of the cutting blade shall be bilaterally ground and honed to razor sharpness. The final honed bevels shall be flat, smooth, and continuous throughout the entire length. Except for the extreme 1/8 inch at either end, the cutting edge shall be free from micro-cracks, and not more than 5 percent of the edge length shall show nicks, notches, or unhoned portions when viewed microscopically at a magnification of 100 diameters. The remainder of the blade shall have a smooth, bright polished or mirror finish. The entire blade shall be free from rust and scale.

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:53Z
Check the source to verify that this is the currentversion before use.

* 3.5 Hardness. The cutting blade shall be uniformly hardened to the hardness range shown in table I for the applicable composition, when tested as specified in 4.4.2. For the coated blade (composition 1) the hardness reading shall be taken as close as possible to the cutting edge, but not on the beveled portion.

Table I. Hardness Values for Cutting Blade

Material of cutting blade (See 3.1.2)	Rockwell Superficial hardness number 30-N scale 1/	Vickers hardness number 2/
Composition 1, corrosion-resisting steel, coated		
Regular (0.90 to 1.05% carbon)	68 - 73	500 - 600
Alternate (0.53 to 0.70% carbon)	77 - 81	700 - 800
Composition 2, corrosion-resisting steel, uncoated		
Regular (0.90 to 1.05% carbon)	74 - 79	610 - 740
Alternate (0.53 to 0.70% carbon)	71 - 76	550 - 650
Composition 3, carbon steel	79 - 84	740 - 920

1/ using a 30-kilogram force (kgf) load and a superficial diamond penetrator

2/ using a 5-kgf load

* 3.6 Corrosion resistance. The razor shall be capable of being sterilized for surgical use. There shall be no evidence of injury to the razor, such as corrosion, scumming, or pitting when boil tested as specified in 4.4.3.

* 3.7 Identification marking. The razor, the removable tooth guard, and the blade rib shall be permanently and legibly marked with the name or trademark of the manufacturer or supplier.

3.8 Workmanship. Workmanship shall be first class throughout. Razors shall be free from defects which detract from their appearance or may impair their serviceability.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

4. QUALITY ASSURANCE PROVISIONS

* 4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification, where such inspections are deemed necessary to assure that supplies and services conform to prescribed requirements.

4.1.1 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties).

4.1.2 Certificates of quality. When available, certificates of quality, supplied by the manufacturer of the component or material, listing the specified test method and test results obtained, may be furnished in lieu of actual performance of such testing by the contractor.

* 4.2 Quality conformance inspection. The examinations required to assure conformance of those items or lots of items to be offered for acceptance are classified as specified in tables II and IV (see 4.2.1 and 4.3); the tests required for acceptance are specified in table III (see 4.2.2).

* 4.2.1 Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105. The inspection levels and acceptable quality levels (AQL) shall be as indicated in table II.

Table II. Inspection and Acceptable Quality Levels

Type defect	Inspection level	AQL (percent defective)
Major A	II	1.0
Major B	S-2	1.5
Minor A	II	2.5
Minor B	S-2	4.0

* 4.2.2 Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as indicated in table III, except that the acceptance number shall be zero for all sample sizes.

Table III. Sampling for Tests

Component	Characteristic	Requirement	Test Procedure	Sample Unit	Inspection Level
Blade	Hardness	3.5	4.4.2	1 blade	S-1
Razor	Boil test	3.6	4.4.3	1 razor	S-1

Source: https://assist.dla.mil -- Downloaded: 202640-26T11:53Z
Check the source to verify that this is the currentversion before use.

* 4.2.3 Raw materials. A sufficient quantity of each lot of raw material used in the construction of the razor and of the blade components shall be selected for chemical analysis.

* 4.3 Examination. The razors and blades shall be examined for defects, including, but not limited to, those listed in table IV.

Table IV. Classification of Defects

Category	Defect
Major A:	
101	Blade does not fit razor properly.
102	Razor component(s) missing or not properly assembled.
103	Workmanship defects such as nicks, notches, or unhoned portions present in blade.
104	Workmanship defects, such as burrs, nicks, notches, etc., present in razor.
105	Bevels of cutting edge of blade not flat, smooth, and continuous throughout entire length.
106	Rib not securely attached to cutting blade.
107	Blade holder not held in handle by proper amount of tension.
Major B:	
151	Thickness of blade rib and cutting edge not within range specified.
Minor A:	
201	Finish not as specified.
202	Outer ends of pin rivet not smoothly peened.
203	Identification marking not as specified.
204	Fewer than 5 blades in unit package.
205	Packaging not as specified.
206	Packing not as specified.
207	Marking on packaging and packing not as specified.
Minor B:	
251	Dimensions, other than those in 151, not as specified.

* 4.3.1 Dimensional examination. The examination of the dimensions shall be made using standard measuring instruments.

* 4.4 Test procedures.

* 4.4.1 Chemical composition. Chemical composition of the steel and of the nickel silver shall be determined in accordance with method 111 or 112 of FED-STD-151.

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:53Z
Check the source to verify that this is the current version before use.

* 4.4.2 Hardness. Hardness shall be determined in accordance with American Society for Testing and Materials (ASTM) Designation E 18 or E 92.

* 4.4.3 Boil test. The razor without the blade shall be boiled in distilled water in a glass beaker for 30 minutes, and the razor shall be allowed to remain submerged for 24 hours. The razor shall then be removed from the beaker and examined for evidence of injury.

5. PREPARATION FOR DELIVERY

5.1 Packaging. Packaging shall be level A or C, as specified (see 6.1).

5.1.1 Unit of issue. One razor (type I) or 5 blades (type II), as specified, constitutes one unit of issue.

5.1.2 Packaging quantities. The number of units of issue indicated in table V shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in the procurement document.

Table V. Packaging Quantities

Type	Unit package	Intermediate package	Exterior container
Type I, Razor	1 unit	6 units	144 units
Type II, Blade	1 unit	10 units	800 units

5.1.2.1 Packing variation permitted. If the required number of units to be shipped is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

5.1.3 Level A.

5.1.3.1 Unit package.

5.1.3.1.1 Type I, razor. Each unit shall be adequately wrapped or enveloped in cellophane or other suitable material and further packaged in a carton of appropriate size constructed in accordance with PPP-B-566 or PPP-B-676, except that commercial colors are acceptable. Carton closure shall be adequate to prevent accidental opening.

5.1.3.1.2 Type II, blade. Each blade shall be treated with a suitable preservative to prevent corrosion and wrapped with a suitable wax-impregnated paperboard blade-edge protector. Five blades so protected shall be wrapped in not less than 20-pound weight waxed paper, heat-sealed and further packaged in a carton of appropriate size constructed in accordance with PPP-B-566 or PPP-B-676 except that commercial colors are acceptable. Closure of carton shall be adequate to prevent spilling of contents under normal handling.

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:53Z
Check the source to verify that this is the currentversion before use.

5.1.3.2 Intermediate package. Intermediate package shall be a carton of appropriate size constructed in accordance with PPP-B-566 or PPP-B-676, except that commercial colors will be acceptable. Closure shall be adequate to prevent spilling of contents under normal handling.

5.1.4 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing. Packing shall be level A, B, or C, as specified (see 6.1).

5.2.1 Level A.

* 5.2.1.1 Exterior container. Exterior container shall be designed for type 2 load, and constructed in accordance with PPP-B-585, style 3, for class 3 use; PPP-B-601, table III, using type I, class 1 plywood; PPP-B-621, table III; or PPP-B-636, type CF, class weather-resistant. Closure and strapping shall be in accordance with appendix of applicable box specification. Each fiberboard box shall be waterproofed in accordance with the requirements of PPP-B-636.

5.2.1.2 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes.

5.2.2 Level B.

* 5.2.2.1 Exterior container. Exterior container shall conform to PPP-B-601, table II, using type III, class 1 plywood; or shall be designed for type 2 load and constructed in accordance with PPP-B-585, style 3, for class 2 use; PPP-B-621, table I; or PPP-B-636, type CF, class domestic. Closure shall be in accordance with appendix of applicable box specification.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, constructed so as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Civilian agencies. Shipments shall be marked in accordance with FED-STD-123.

5.3.2 Military agencies.

5.3.2.1 Unit packages. Each unit package shall be marked in accordance with MIL-STD-129. Contract or purchase order number, name of contractor, and (for type II only), composition (e.g. stainless steel, coated; stainless steel, uncoated; or carbon steel) shall be shown.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

5.3.2.2 Interior packages. Each unit package and each intermediate package shall be marked in accordance with MIL-STD-129. When labels are utilized, waterproofing shall be required only when applicable carton is fabricated of water-resistant material. Contract or purchase order number and name of contractor shall be shown.

5.3.2.3 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

(a) Title, number and date of this specification.
(b) Type required (see 1.2).
(c) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.2 This specification does not cover all types, classes, grades, or sizes of the commodity indicated by the title of this specification, or those which are commercially available, but is intended to cover the types which are normally procured to meet Federal Government requirements.

6.3 Supersession data. This specification includes the requirements of MIL-R-2601 dated 15 February 1951.

* 6.4 The margins of this specification are marked with an asterisk to indicate where changes (additions, modifications, corrections, deletions) from the previous issue were made. This was done as a convenience only and the Government assumes no liability whatsoever for any inaccuracies in these notations. Bidders and contractors are cautioned to evaluate the requirements of this document based on the entire content irrespective of the marginal notations and relationship to the last previous issue.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activities:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:53Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE: This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO. 4966 WASHINGTON D. C.
POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-LS
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 202640-26T11:53Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)*: _____

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

DD FORM 1426 82 MAR PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil - Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.