GG-S-00176C(DSA-DM)
15 October 1973
USED IN LIEU OF
GG-S-00176b (DSA-DM)
28 February 1967

INTERIM FEDERAL SPECIFICATION

SCREEN, X-RAY INTENSIFYING

This interim Federal specification was developed by the Defense Personnel Support Center, 2800 South 20th Street, Philadelphia, Pennsylvania 19101, based upon currently available technical information. It is recommended that Federal agencies use it in procurement and forward recommendations for changes to the preparing activity at the address shown above. The General Services Administration has authorized the use of this interim Federal specification as a valid exception to Federal Specification GG-S-176.

1. SCOPE AND CLASSFICATION

1.1 Scope. This specification covers intensifying screens for use in diagnostic medical radiography.

1.2 Classification.

1.2.1 Types and sizes. Intensifying screens shall be of the following types and sizes, as specified (see 6.2).

Type I. - High-speed screens.
Type II. - Medium-speed (general purpose) screens.
Type III. - High-resolution (detail) screens.

Size 1. - 5 by 7 inches.
Size 2. - 6-1/2 by 8-1/2 inches.
Size 3. - 8 by 10 inches.
Size 4. - 10 by 12 inches.
Size 5. - 11 by 14 inches.
Size 6. - 14 by 17 inches.
Size 7. - 9-1/2 by 9-1/2 inches.
Size 8. - Other, as specified.

FSC 6525

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:49Z
Check the source to verify that this is the current version before use.

2. APPLICABLE DOCUMENTS

2.1 Specification and standards. The following specifications and standards, of the issues in effect on date of invitation for bids or request for proposals, form a part of this specification to the extent specified herein.

SPECIFICATIONS

Federal

| | |
|---|---|
| L-F-310 | Film, Dental Radiographic; Film, Photofluorographic, Medical; Film, Radiographic, Medical. |
| PPP-B-585 | Boxes, Wood, Wirebound. |
| PPP-B-601 | Boxes, Wood, Cleated-Plywood. |
| PPP-B-621 | Boxes, Wood, Nailed and Lock-Corner. |
| PPP-B-636 | Box, Fiberboard. |

Military

| | |
|---|---|
| MIL-L-10547 | Liners, Case, and Sheet, Overwrap, Water Vaporproof or Waterproof, Flexible. |

STANDARDS

Federal

| | |
|---|---|
| FED-STD-123 | Marking for Domestic Shipment (Civilian Agencies). |

Military

| | |
|---|---|
| MIL-STD-105 | Sampling Procedures and Tables for Inspection by Attributes. |
| MIL-STD-129 | Marking for Shipment and Storage. |

(Activities outside the Federal Government may obtain copies of Federal Specifications and Standards as outlined under General Information in the Index of Federal Specifications and Standards and at the prices indicated in the Index. The Index, which includes cumulative monthly supplements as issued, is for sale on a subscription basis by the Superintendent of Documents, U. S. Government Printing Office, Washington, D. C., 20402.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:49Z
Check the source to verify that this is the current version before use.

(Single copies of this specification and other product specifications required by activities outside the Federal Government for bidding purposes are available without charge at the General Services Administration Regional Offices in Boston, New York, Atlanta, Chicago, Kansas City, Mo., Dallas, Denver, San Francisco, Los Angeles, Seattle, and Washington, D. C.

(Federal Government activities may obtain copies of Federal Specifications and Standards and the Index of Federal Specifications and Standards from established distribution points in their agencies.)

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposal shall apply.

American Society for Testing Materials: (ASTM)

ASTM Designation: C190 Tensile Strength of Hydraulic Cement Mortars.

(Copies may be obtained from the American Society for Testing Materials, 1916 Race Street, Philadelphia, Pa. 19103.)

American National Standards Institute (ANSI).

PH2.19 Diffuse Transmission Density.
PH3.19 Radiographic Intensifying Screens.

(Copies may be obtained from the American National Standards Institute, 1430 Broadway, New York, N. Y. 10018.)

3. REQUIREMENTS

* 3.1 Material. All parts entering into the construction of the screens shall be fabricated from materials which are entirely satisfactory for the purpose and of sufficient strength to maintain operational alignment for the life of the screens. Screens shall be coated on a plastic or cardboard base. Cardboard base shall be treated with a moisture and fungus-resistant preparation.

3.2 Description. Intensifying screens shall consist of a pair of screens; one to be used as a front screen and one to be used as a back screen.

3.3 Construction. Screens shall be so constructed as to lie essentially flat without warping or curling, and to insure close apposition of intensifying screens and film when placed in a standard cassette.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:49Z
Check the source to verify that this is the current version before use.

3.4 Dimensions. The dimensions, thickness, and squareness of screens shall be in accordance with ANSI Standard PH3.19.

3.5 Surface. The fluorescent surface of the screens shall be of the cleanable type, and the cleanable surface shall be an integral part of the screen.

3.5.1 Cleanability. The surface of the screens, when tested as specified in 4.4.1, shall be completely clean and free from staining. The radiograph shall show no sign of artefacts attributable to surface staining.

3.6 Resistance to abrasion. The resistance of the surface of the screen to abrasion and wear shall be such that, when tested as specified in 4.4.2, the surface of the screen shall show no permanent discoloration nor shall a film exposed with the screen show evidence of artefacts attributable to surface-damage to the screen.

3.7 Speed factor. The speed factor of screens covered by this specification shall be defined as the quotient of the speed of the film-screen combination divided by the speed of the film alone. The speed of the film-screen combination is defined as 100 times the reciprocal of the exposure in milliroentgens required to produce a density of 1.0 above base and fog density in a film exposed between the pair of screens under the standard conditions described in 4.4.3.1. The speed of the film alone is the speed of the film determined in accordance with L-F-310. The speed factor of screens covered by this specification shall be as follows, when determined as specified: Type I (high-speed screens), not less than 1.6; type II (medium speed screens) between 1.2 and 1.4. and type III (high resolution screens) between 0.35 and 0.50.

3.8 Uniformity of coating. When tested as specified in 4.4.4, the film exposed between the screens shall show no evidence of mottle or uneven density.

3.9 Persistence. When tested as specified in 4.4.5, the film shall show no evidence of shadows of the test objects.

3.10 Definition. When tested as specified in 4.4.6, type I (high-speed) screens shall resolve at least half the holes in Area I (0.024 inches); type II (medium-speed) screens shall resolve at least half the holes in Area II (0.0225 inches); and type III (high-resolution) screens shall resolve at least half the holes in Area III (0.021 inches). Resolution shall be determined by viewing the exposed x-ray film in a viewer without magnification.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:49Z
Check the source to verify that this is the current version before use.

3.11 Mounting material. Manufacturer shall include with each package (pair) sufficient mounting material, of tape, glue or gum to install screens in cassette, including instructions for using the material. Mounting materials shall be free from any constituents which will produce aerial fog or impair the fluorescent emission by X-ray absorption, chemical reaction, or staining of the active layers.

3.12 Marking for identification. Each screen shall be marked on the front (screen surface) with the name or registered trademark of the manufacturer, serial number, contract number, and identification as to "front" or "back" screen, whichever is applicable. If front and back screens are interchangeable, no marking regarding front or back identification is necessary. The marking shall be permanent and clearly visible after screens are mounted in the cassette and shall not encroach on the useful area of the screen. Screens shall be marked "High Speed", "Medium Speed" or "Detail", as applicable. Serial number shall consist of one digit for the year of manufacture, three digits for the lot number and two digits for the month of manufacture. As an alternate, the serial number shall consist of the month, day and year of manufacture.

3.13 Instructions. An instruction sheet for installation, care and cleaning the screens shall be furnished with each pair of screens. The instruction sheet shall also state that the enclosed pressure sensitive label shall be marked with the date of installation and applied to the back of the cassette. This pressure sensitive label with protective backing shall be furnished with each pair of screens. The label shall be printed with the screen type (high speed, medium speed or detail), "date of installation", with appropriate space for addition of the date, and the film recommended for use with the screens. Instruction sheet, label and mounting material shall be packaged so that they do not contact the screens.

3.14 Workmanship. Workmanship shall be first class throughout. Screens shall be free from imperfections which affect their serviceability or the diagnostic quality of radiographs.

3.15 Static Electricity. When tested as specified in 4.4.7, the screen shall record a static charge less than 1000 volts, positive or negative.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:49Z
Check the source to verify that this is the current version before use.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual, and auditory, investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the materials for components of the intensifying screen may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The Certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the Item Identification, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Sampling. Sampling for examination shall be conducted in accordance with MIL-STD-105 and Table I. Unit of inspection shall be one screen.

Table I

| | Inspection level | AQL (defects per hundred units) |
|---|---|---|
| For visual examination | | |
| Major defects | II | 1.0 |
| Minor defects | II | 4.0 |
| For dimensional examination | S-3 | 2.5 |



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:49Z
Check the source to verify that this is the current version before use.

4.2.1 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105. The acceptance level shall be zero. (see table II)

Table II

| Characteristic | Requiremnet | Test | Inspection level |
|---|---|---|---|
| Cleanability | 3.5.1 | 4.4.1 | S-3 |
| Abrasion resistance | 3.6 | 4.4.2 | S-1 |
| Speed factor | 3.7 | 4.4.3 | S-1 |
| Uniformity of coating | 3.8 | 4.4.4 | S-1 |
| Persistance | 3.9 | 4.4.5 | S-1 |
| Definition | 3.10 | 4.4.6 | S-1 |
| Static electricity | 3.15 | 4.4.7 | S-1 |

4.3 Examination. The screens shall be examined to determine compliance with all requirements contained in this specification.

4.3.1 Classification of defects. Examination shall be conducted in accordance with the following classification of defects.

Table III Categories and defects*

| | |
|---|---|
| Major | |
| 101 | Any component missing or unusable. |
| 102 | Screen fails to lie essentially flat without warp or curl. |
| 103 | Fluorescent surface not of cleanable type. |
| 104 | Cleanable surface not an integral part of the screen. |
| 105 | Mounting material, ie., tape, glue, gum, insufficient for the purpose intended. |
| 106 | Screen coating not free of scratches, base spots, or areas of visibly thin coating. |
| 107 | Cardboard base not treated with a moisture and fungus resistant preparation. |



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:49Z
Check the source to verify that this is the current version before use.

Table III Categories and defects* (cont'd)

| | |
|---|---|
| Minor | |
| 201 | Identification marking missing, not complete, or not legible. |
| 202 | Instructions for cleaning, and for mounting, missing, incomplete, illegible or not clear. |
| 203 | Unit not free of grease, oil, dirt, or other foreign matter. |

*Examination shall not be restricted to the classified possible defects listed above.

4.3.2 Dimensions. The screens shall be examined for defects in dimensions. Any dimension not within the tolerance specified herein shall be classified as a defect.

4.4 Tests. As an alternate to the L-F-310 film and manual processing specified herein, an equivalent polyester type film with automatic processing may be used. Automatic processor shall be operated in accordance with the manufacturer's directions, and test film shall be processed in one continuous run through the automatic processor. All test film shall be processed within 2 hours and not more than 24 hours after exposure.

4.4.1 Cleaning test. A few drops of the following test solution shall be placed on the surface of the screen and allowed to dry thoroughly. The spotted sample shall be placed in a darkened area for 24 hours and the dried residue cleaned with cotton, using the cleaner recommended by the manufacturer of the screen. The screen sample and a piece of film conforming to Type III, class B of L-F-310 shall then be subjected to X-ray using 80 KVP of suitable exposure to produce, in the processed film, a radiographic density of 0.6 to 0.8. The radiograph and screen sample shall be inspected and shall meet the requirements specified in 3.5.1.

Test solution

Sodium chloride----------1.0 gram.

Lamp black----------0.5 gram.

Distilled water----------100 ml.

Shake well before using



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:49Z
Check the source to verify that this is the current version before use.

4.4.2 Abrasion resistance. Test specimens shall be prepared by cutting pieces 1-1/2 inches square from the sample screen. Mask off a 1/4 inch edge all around the specimen by folding half-inch masking tape (do not use cellophane tape) over the edge. The prepared specimens shall be placed in a 1,000 ml. porcelain ball mill charged with 1,000 grams of Ottawa sand, ASTM designation C-190, 20 grams of carbon black, and one 2 inch porcelain ball. Not more than 10 specimens shall be placed in the mill at one time. The mill shall be operated at 35 to 40 r.p.m. for 30 minutes. Upon completion of the milling, the screen specimens shall be removed with a soft dry cloth. The masking tape then shall be removed, care being taken that none of the front or back coating of the screen is removed with the tape. The screen specimens shall be cleaned by the procedure recommended by the manufacturer until no further dirt can be removed. The central portion of the screen specimen shall be compared with the protected border. There shall be no discoloration of the central area. A radiograph of the screen shall be made as described in 4.4.1. The film shall be examined in the central area and shall meet the requirements specified in 3.6.

4.4.3 Speed factor.

4.4.3.1 Film-screen speed. A cassette shall be properly loaded with the screens under test and a strip of roentgenographic film conforming to the requirements of L-F-310 for type III, class B film. A graduated series of exposures shall be given using roentgen radiation having a quality equivalent to a half-value layer of 4 mm. of aluminum. The exposure for each step shall be 0.10-second duration, and the intensities shall be such as to result in a series of densities in the processed film ranging from approximately 0.6 to 1.6 above base and fog densities in not less than six steps. Each exposure shall be suitably measured in milliroentgens. The temperature of the screen and film at the time of exposure shall be $77^{\circ} \pm 2^{\circ}$ F. Following exposure, the film shall be processed according to L-F-310, washed, and dried. Read the densities of the exposed and unexposed portions of the film on a suitable densitometer calibrated to read diffuse transmission density in accordance with ANSI Standard PH2.19. Plot density against the logarithm of the exposure in milliroentgens. From this plot, determine the exposure required to produce a density of 1.0 above the base and fog density. The speed of the film-screen combination is determined as 100 times the reciprocal of this exposure.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:49Z
Check the source to verify that this is the current version before use.

4.4.3.2 Film speed. The speed of a strip of film of the same emulsion number (preferably cut from the same sheet of film as used in 4.4.3.1) shall be determined according to L-F-310.

4.4.3.3 Speed factor. The speed factor of the pair of screens shall be determined by dividing the film-screen speed determined in 4.4.3.1 by the film speed determined in 4.4.3.2.

4.4.4 Uniformity of coating. A suitable cassette shall be loaded, in accordance with the manufacturer's directions, with the screens under test and a piece of roentgenographic film conforming to the requirements for type III, class B film L-F-310. A roentgen exposure at approximately 70 KVP shall be given, sufficient to produce in the processed film a density of from 0.6 to 0.8. Following exposure, the film shall be processed according to L-F-310, washed, dried, and examined for uniform density.

4.4.5 Persistence. The screens shall be fitted in a suitable cassette, without film. Prior to test, the screens shall be left undisturbed for a minimum of 24 hours. The test shall be conducted in ambient temperature of 77 ± 3 degrees F. Several metal coins or other roentgen-opaque objects shall be placed on the front of the cassette. The cassette shall then be exposed to x-rays as indicated in Table IV.

Table IV.

| | KVP | MA | TARGET-SCREEN DISTANCE | TIME |
|---|---|---|---|---|
| Calcium tungstate screen | 85 | 5 | 20 inches | 12 seconds |
| Barium lead sulfate screen | 100 | 5 | 35 inches | 60 seconds |

The Cassette shall then be taken to the darkroom and exactly one minute after the termination of the exposure, for the calcium tungstate screens, or 30 seconds for the barium, lead sulfate screens, the Cassette shall be opened and a sheet of medium speed high contrast medical x-ray film inserted and the Cassette closed. The film shall be left in the Cassette for 10 minutes and then completely processed either manually or in an automatic film processor. The film shall then be examined for evidence of shadows of roentgen-opaque objects.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:49Z
Check the source to verify that this is the current version before use.

4.4.6 Definition. Two sheets of cast acrylate, Plexiglas or equal, with a density of between 1.16 and 1.20 grams per cubic centimeter shall be fabricated as follows: The first sheet shall be approximately 10 inches long by 8 inches wide by 4 inches thick. The 4-inch thickness may be made from sheets no less than one inch thick. The second sheet shall be approximately 10 inches long by 6 inches wide by 1/4 inch thick. The second sheet shall be divided into three areas by drilling holes of approximately 3/32 inch diameter around the perimeter of each area. Holes shall be spaced 1/4 inch on centers and each area shall then be drilled with holes spaced 1/4 inch on centers; Area I shall have holes 0.024 inches in diameter (Drill No. 73); Area II shall have holes 0.0225 inches in diameter (Drill No. 74); and Area III shall have holes 0.021 inches in diameter (Drill No. 75). The second sheet (1/4 inch thick) shall be attached to the first sheet (4 inches thick) such that a 2-inch by 10-inch area of the first sheet is exposed to face the X-ray tube. This 2-inch by 10-inch area shall be cut down one inch such that the thickness in this area is now 3 inches. In the center of this 2- by 10- by 3- inch area, a 2-inch by 2-inch by 2-inch area shall be cut out. In this cutout area shall be placed 8.5 millimeter of aluminum, Type 1100 alloy, such that one inch of plastic and 8.5 millimeters of aluminum face the tube side. The cut-down adjacent area of the first sheet (3 inch thick) and cutout area will determine the correctness of the contrast developed in the X-ray film and the quality of the radiation. The fabricated acrylate sheets shall then be placed over a bakelite front cassette with screens to be tested and loaded with a piece of type III, class B, X-ray Film (L-F-310). An initial exposure shall be made at approximately 80 PKV with 2.5 millimeters of aluminum filtration and 40 inches tube screen distance. Exposure shall be such as to produce a density above base and fog of between 1.25 and 1.75 in the region of the holes. The respective densities in the cutout area and adjacent area (3 inches thick) shall be equal and at least 0.35 but not more than 0.50 density units above the hole region. If the adjacent area density is greater than the cutout area density by more than 0.05 density units, increase PKV. If cutout area density is greater than adjacent area (3 inches thick) by 0.05 density units, decrease PKV. If the densities of the cutout area and adjacent area are equal but not .35 density units above the hole region, the developed contrast is too low. The developed radiograph shall include all areas for density determinations and shall meet the requirements specified.

4.4.7 Static electricity. The screen, isolated from ground, shall be rubbed vigorously for 10 seconds with a 12 by 12 inch cotton flannel cloth. Room conditions for the test shall be 70 $\pm$ 2° F. and 45 to 50 percent humidity. A static meter, such as a Keithley Model 250, shall be placed 3/4 inch above the screen surface. The static charge shall be read within one minute after rubbing. After discharging both the screen and meter, repeat the test using a sheet of film, Type III Class B in accordance with L-F-310, to apply the charge.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:49Z
Check the source to verify that this is the current version before use.

5. PREPARATION FOR DELIVERY

5.1 Packaging.

5.1.1 Unit of issue. One pair of screens, as specified, constitutes one unit of issue.

5.1.2 Packaging quantities. The number of units of issue indicated in table V shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in the procurement document.

Table V. Packaging Quantities

| Size | Unit package | Intermediate package | Exterior container |
|---|---|---|---|
| 1 | 1 unit | 24 units | 144 units |
| 2 | As specified by the procuring activity | | |
| 3 | 1 unit | Not required | 24 units |
| 4 | 1 unit | Not required | 12 units |
| 5 | 1 unit | Not required | 12 units |
| 6 | 1 unit | Not required | 12 units |

5.1.2.1 Packing variation permitted. If the required number of units to be shipped is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

5.1.3 Level A.

5.1.3.1 Unit package.

5.1.3.1.1 Screen. Each screen shall be wrapped in tissue or parchment paper, or packaged in an individual envelope fabricated from tissue or parchment paper.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:49Z
Check the source to verify that this is the current version before use.

5.1.3.1.2 Unit container. One piece of cardboard or parchment paper of approximately the same size as a screen shall be interposed between the two wrapped or enveloped screens. One pair of screens (1 unit) shall be packaged in a one-piece folder, envelope style, fabricated from chip-board or fiber-board not less than 0.033 inch in thickness. When a fiberboard folder is used, it shall be further packaged in a paper envelope of suitable size and design. Closure of unit container shall be adequate to prevent spillage of contents under normal handling.

5.1.3.2 Intermediate package. Intermediate package shall be a box of appropriate size constructed in accordance with Federal Specification PPP-B-636, Type CF, class domestic. Each box shall be adequately secured to prevent accidental opening.

5.1.4 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing.

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for type 2 load, and constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type; PPP-B-621, class 2; or PPP-B-636, class weather-resistant. Fiberboard boxes shall conform to the special requirements as specified in PPP-B-636.

5.2.1.2 Case liner. Each level A wood box shall be lined with a water-proof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes.

5.2.1.3 Closure. Closure of wood boxes shall be in accordance with appendix of applicable box specification. Closure of each fiberboard box shall be as specified in the appendix, and waterproofing shall conform with 30.4, of PPP-B-636.

5.2.1.4 Strapping. Strapping, when required, shall be in accordance with appendix of applicable box specification.

NOTE: Strapping shall not be required for shipments forwarded to receiving activity within the continental limits of the United States for storage and redistribution.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:49Z
Check the source to verify that this is the current version before use.

5.2.2 Level B.

* 5.2.2.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, domestic type; PPP-B-621, class 1; or PPP-B-636, class domestic. Closure of wood boxes shall be as specified in the appendix of the applicable box specification. Closure of fiberboard boxes shall conform to method II of PPP-B-636. In addition, fiberboard boxes shall conform to the special requirements as specified in PPP-B-636.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Civil agencies.

5.3.1.1 Interior packages and exterior containers. In addition to any marking specified in the contract or order, interior packages and shipping containers shall be marked in accordance with Federal Standard No. 123.

5.3.2 Military.

5.3.2.1 Interior packages. Each unit package and each intermediate package shall be marked in accordance with MIL-STD-129. Marking shall include the legend: "STORE AT A CONTROLLED ROOM TEMPERATURE (59° - 86°F)".

5.3.2.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129. Marking shall include the Legend: "STORE AT A CONTROLLED ROOM TEMPERATURE (59° - 86°F)".

6. NOTES

6.1 Intended use. Diagnostic medical radiography in medical facilities.

6.2 Ordering data. Procurement documents should specify the following:

(a) Number and date of this specification.
(b) Selection of applicable levels of packing. (see 5.1 and 5.2).
(c) Type and size required (see 1.2).



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:49Z
Check the source to verify that this is the current version before use.

6.3 This specification does not cover all types, classes, grades, or sizes of the commodity indicated by the title of this specification, or those which are commercially available, but is intended to cover the types which are normally procured to meet requirements.

6.4 The margins of this specification are marked with an asterisk to indicate where changes (additions, modifications, corrections, deletions) from the previous issue were made. This was done as a convenience only and the Government assumes no liability whatsoever for any inaccuracies in these notations. Bidders and contractors are cautioned to evaluate the requirements of this document based on the entire content irrespective of the marginal notations and relationship to the last previous issue.

6.5 The following X-ray intensifying screens listed in the Federal Supply Catalog are covered by this specification:

| | |
|---|---|
| 6525-00-721-9827 | High Speed, 8 x 10 inches |
| 6525-00-721-9828 | High Speed, 10 x 12 inches |
| 6525-00-721-9829 | High Speed, 11 x 14 inches |
| 6525-00-721-9830 | High Speed, 14 x 17 inches |
| 6525-00-612-8300 | Medium Speed, 8 x 10 inches |
| 6525-00-612-9300 | Medium Speed, 10 x 12 inches |
| 6525-00-612-9500 | Medium Speed, 11 x 14 inches |
| 6525-00-613-0300 | Medium Speed, 14 x 17 inches |

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activities:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6525-1113

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:49Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:49Z
Check the source to verify that this is the current version before use.

# STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE

GG-S-00176C(DSA-DM)

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?

☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO.

DATE

DD FORM 1 JAN 72 1426

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:49Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 South 20th Street
Philadelphia, Pa. 19101

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY
DoD - 304

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:49Z
Check the source to verify that this is the current version before use.