GG-S-00928c (DSA-DM)
February 16, 1971
Interim Revision of
GG-S-928b
May 2, 1958

INTERIM FEDERAL SPECIFICATION

SYRINGE, METAL, EAR

This interim Federal specification was developed by Defense Personnel Support Center, 2800 South 20th St., Philadelphia, Pennsylvania 19101, based upon currently available technical information. It is recommended that Federal agencies use it in procurement and forward recommendations for changes to the preparing activity at the address shown above.

The General Services Administration has authorized the use of this interim Federal specification as a valid exception to Federal specification GG-S-928b.

1. SCOPE AND CLASSIFICATION

1.1 Scope. This specification covers a metal syringe for use in irrigating, testing injuries and diseases of the human ear.

1.2 Classification.

1.2.1 Sizes. Metal ear syringes shall be of the following sizes as specified (see 6.2):

Size A. - Two-ounce capacity
Size B. - Four-ounce capacity

2. APPLICABLE DOCUMENTS

2.1 The following documents, of the issue in effect on date of invitation for bids or request for proposals, form a part of this specification to the extent specified herein:

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

SPECIFICATIONS

Federal

GG-I-526	Instruments, Dental and Surgical; General Specifications.
PPP-B-566	Boxes, Folding, Paperboard.
PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard.
PPP-B-676	Boxes, Set-Up, Paperboard.
PPP-T-76	Tape, Pressure-Sensitive Adhesive Paper, (for Carton Sealing).

STANDARDS

Federal

Fed-Std-No. 123	Marking for Domestic Shipment (Civilian Agencies).

(Activities outside the Federal Government may obtain copies of Federal Specifications and Standards as outlined under General Information in the Index of Federal Specifications and Standards and at the prices indicated in the Index. The Index, which includes cumulative montly supplements as issued, is for sale on a subscription basis by the Superintendent of Documents, U.S. Government Printing Office, Washington, D.C. 20204.

(Single copies of this specification and other product specifications required by activities outside the Federal Government for bidding purposes are available without charge at the General Services Administration Regional Offices in Boston; New York; Atlanta; Chicago; Kansas City, Mo.; Fort Worth; Denver; San Francisco; Auburn, Washington; and Washington, D.C.

(Federal Government activities may obtain copies of Federal Specifications and Standards and the Index of Federal Specifications and Standards from established distribution points in their agencies.)

SPECIFICATIONS

Military

MIL-L-10547	Liners, Case, and Sheet, Overwrap; Water-Vaporproof or Waterproof, Flexible.

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:53Z
Check the source to verify that this is the currentversion before use.

STANDARDS

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

(Copies of specifications, standards and drawings required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

DRAWINGS

Defense Personnel Support Center (DPSC)

24244 SYRINGE, METAL, EAR (2 OZ. and 4 OZ.)

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposals shall apply:

AMERICAN NATIONAL STANDARDS INSTITUTE (A.N.S.I.)

A.N.S.I. B1.1	Unified Screw Threads.
A.N.S.I. B1.2	Gages and Gaging for Unified Screw Threads.

(Application for copies should be addressed to the American National Standards Institute, 1430 Broadway, New York, New York 10018.)

3. REQUIREMENTS

3.1 Material.

3.1.1 Metal Parts. Metal parts shall be fabricated from Class 8, type 260, 270 or 360 material in accordance with Federal Specification GG-I-526. As an alternate, the barrel, cap and guard shall be fabricated from chromium plated over nickel plated brass with the following composition:

Copper:	62.0 percent minimum
Lead:	0.5 percent maximum
Iron:	0.5 percent maximum
Zinc:	Remainder (Total other elements: 0.5 percent maximum).

As another alternate, the remaining metal parts shall be fabricated from chromium plated over nickel plated brass with the following composition:

Copper:	56.5 percent minimum
Lead:	3.0 percent maximum
Zinc:	Remainder (Total other elements: 1.0 percent maximum).

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

The plating and corrosion requirements for the alternate brass compositions shall be in accordance with GG-I-526 for Class 8, Type 260 material. As a further alternate, the inside surfaces of the barrel and the plunger proper shall be nickel plated only. The corrosion requirements shall be the salt spray and boil tests in accordance with GG-I-526.

3.1.2 Sealing gaskets and "0" rings. Gaskets that are placed on the syringe tip, the proximal end of the guard assembly, and the "0" ring contained on the plunger assembly to Type II design, shall be fabricated from natural or synthetic rubber of the best quality, suitable for the purpose intended.

3.2 Threads. All threads shall be in accordance with A.N.S.I. Standard B1.1, except that the threads for barrel and cap shall be National Standard Thread, 32 threads per inch.

3.3 Style, design and dimensions. The ear syringe shall consist of the following components: Barrel, cap, plunger rod assembly, guard assembly, and two syringe tips (1 bulbous ear tip and 1 pomeroy ear tip). The syringe shall be compact and shall have a working capacity of 2 or 4 fluid ounces for sizes A and B, respectively. The ear syringes shall be suitable for irrigation of the human ear. The cap shall be readily removable and readily assembled to the proximal end of the barrel. The plunger assembly shall accurately fit into the barrel providing for free plunger travel throughout the full excursion of the barrel. External threads of the syringe tip and guard assembly shall be 5/16-28 UN-2A in accordance with A.N.S.I. Standard B1.1. Internal threads of the distal end of the barrel and guard assembly shall be 5/16-28 UN-2B in accordance with A.N.S.I. Standard B1.1.

The entire assembly shall be in accordance with Defense Personnel Support Center Drawing No. 24244, for appropriate size. Dimensions relating to barrel inside diameter and length as shown on drawing, shall be such as to provide the required fluid capacity as specified in 3.8.

3.3.1 Plunger assembly. The plunger proper shall be of either Type I or Type II design as specified below.

3.3.1.1 Type I design. The Type 1 plunger proper shall be of heavy wall construction, cup shaped, as shown on drawing. The piston component shall be accurately sized to conform to the inside diameter of the barrel. The distal edge of the piston shall be slightly beveled to permit ready insertion in the barrel. The plunger rod proper shall be securely attached to the piston component at its distal end, and securely attached to the thumb rest at its proximal end.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

3.3.1.2 Type II design. The Type II plunger proper shall be of the "O" ring design as shown on drawing. Two extra "O" - rings shall be furnished with each syringe that contains plunger assemblies of the Type II design. "O" rings shall be coated with a light film of an appropriate lubricant, having a viscosity of not less than 100 centistokes at 25° Centigrade (C.), and shall be furnished with instructions as specified in 3.11. The piston component shall be accurately sized to conform to the inside diameter of the barrel. The plunger rod proper shall be securely attached or threaded to the piston component at its distal end, and securely attached or threaded to the thumb rest at its proximal end. The piston shall be recessed to properly accommodate an "O" - ring of adequate size.

3.4 External threads of syringe tips and guard assembly. The 5/16-28UN-2A external threads of the syringe tips and guard assembly shall allow the "GO" thread ring gage to pass over their respective full length of threads, turn freely, and shall not allow passage of the "LO" and "NOT GO" thread ring gages when tested as specified in 4.4.1. Definitions of acceptance and rejection criteria of the gaging tests shall be as specified in A.N.S.I. Standard B1.2.

3.5 Internal threads of the distal end of the barrel and guard assembly. The 5/16-28 UN-2B internal threads of the barrel and guard assembly shall allow the "GO" thread plug gage to be inserted and turn freely, and shall not allow entry of the "HI" and "NOT GO" thread plug gages when tested as specified in 4.4.1.1. Definitions of acceptance and rejection criteria of the gaging tests shall be as specified in A.N.S.I. Standard B1.2.

3.6 Fluid tightness.

3.6.1 Barrel and plunger. The barrel and plunger shall be tested for fluid tightness as specified in 4.4.3. During the first and second tests, when pressure is applied as specified in 4.4.2, there shall not be leakage exceeding 3 cubic centimeters of water.

3.6.2 Tips and guard. The tips shall be tested for fluid tightness as specified in 4.4.2.1. During the first and second tests, when pressure is applied as specified in 4.4.2.1, there shall not be water leakage at the threaded end of the tips exceeding one falling drop.

3.7 Freedom from foreign matter. All internal cavities of the syringe shall be free from all foreign matter. Internal flaking of plating in any of the internal passages of the syringe components shall be cause for rejection.

3.8 Syringe capacity. The syringes shall deliver 2 $\pm$ 0.25 fluid ounces for Size A syringes and 4 $\pm$ 0.50 liquid ounces for Size B syringes in a leakproof manner.

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:53Z
Check the source to verify that this is the current version before use.

3.9 Graduation markings of plunger rod. The plunger rod shall have graduation markings, in ounces, as indicated on drawing for Sizes A and B. With the plunger proper fully inserted in the barrel, the "0" graduation line shall be even with the top edge of the guide portion of the cap for Sizes A and B.

3.10 Identification marking. Syringes manufactured in the U.S.A. shall be permanently and legibly marked with the name or registered trademark of the contractor. Syringes of foreign (other than U.S.A.), in addition to the above, shall be marked with the country of origin.

3.11 Instruction Sheet. The following instructions shall be furnished with each syringe containing Type II plunger assemblies:

"INSTRUCTIONS

Apply Dow Corning No. 200 fluid, with not less than 100 centistrokes viscosity grade at 25° C (or other suitable nontoxic lubricant compatible with solutions to be used) to the "0" - ring as required to allow piston to move freely throughout excursion of barrel after sterilization."

3.12 Workmanship. Workmanship shall be first class throughout. Syringes shall be free from defects which will detract from their appearance or impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specifications where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the brass and rubber, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the FSN, the item identification, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Sampling.

4.2.1 Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, and as indicated in Table I. Unit of product for purposes of examinations shall be one syringe.

Table I. Sampling Criteria for Examination

Type defect	Inspection level	AQL (Percent Defective)
Major	II	1.0
Minor	II	2.5

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105, and as indicated in Table II. Unit of product for test purposes shall be one syringe.

Table II. Sampling Criteria for Tests

Component	Characteristic	Reqmt	Test procedure	Inspec. level	AQL
Syringe	Material	3.1.1	GG-I-526	*	1.0
	Salt-Spray test (for all brass material)	3.1.1	GG-I-526	S-2	1.0
	Boil test (for all brass material)	3.1.1	GG-I-526	S-2	1.0
	Plating test	3.1.1	GG-I-526	S-2	1.0
	Dimensions	3.3	- - -	S-1	4.0
	External thread test	3.4	4.4.1	S-2	1.0
	Internal thread test	3.5	4.4.1.1	S-2	1.0
	Barrel and plunger fluid tightness tests	3.6.1	4.4.2	S-4	1.5
	Fluid tightness test for tips	3.6.2	4.4.2.1	S-4	1.5

* Four ounces of each lot of each metal used in the construction of the syringes shall be selected for chemical analysis.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

4.2.3 Sample unit usage. Where feasible, the same sample unit shall be used for the determination of two or more test characteristics.

4.3 Examination. The syringes shall be examined to determine compliance with all the requirements contained in this specification. Examination shall be conducted as specified in Tables I and III. Inspection, however, shall not be restricted to defects listed in Table III. Any syringes in the sample containing one or more defects shall be rejected and if the number of defective syringes in any sample exceeds the acceptance number for that sample, the lot represented by the sample shall be rejected.

Table III. Classification of Defects

Categories	Defects
Critical	None defined.
Major	
101	Syringe not of proper size.
102	Syringes do not deliver liquid quantities specified in 3.8 for appropriate size.
103	Plunger not easily operated (stroked).
104	All internal and external threads not well defined.
105	Cap not readily removed or assembled, from barrel.
106	Inside surfaces of barrel not plated.
107	Inside surfaces of barrel not free from pits, tool marks, burrs, or foreign material.
108	One or more components missing (including gaskets and extra "0" - rings).
109	Tip not readily assembled or removed from guard assembly.
110	Guard assembly not readily assembled or removed from distal end of barrel.
111	Zero (0) graduation marking not even with top edge of the guide portion of the cap.
112	Plunger not graduated as specified.
113	Syringes not free from burrs, sharp edges, pits or cracks.
114	Plating not uniform or not free from flaking.
115	Plunger rod not secured to piston, or thread adapter not permanently secured to guard.
Minor	
201	Contractor's identification marking missing or illegible.
202	Instruction sheet missing (as specified in 3.11).

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 External thread test The contractor shall provide the necessary thread ring gages to the Government Quality Assurance Representative (QAR) for checking the external threads of the tips and guard assembly in accordance with the required thread dimensions. The following dimensional notations shall be inscribed on the ring gages:

Thread Size:	5/16-18 UN-2A
"GO" PITCH DIAMETER:	0.2283
"LO" PITCH DIAMETER:	0.2849
"GO" MAJOR DIAMETER:	0.3115
"NOT GO" MAJOR DIAMETER:	0.3050

The gaging procedure to be used, dimensions, tolerances and gage identifications shall be in accordance with A.N.S.I. Standard B1.2. Fits of gages onto external threads shall be noted.

4.4.1.1 Internal thread tests. The contractor shall provide the necessary thread plug gages to the QAR for checking the internal threads of the guard assembly and the barrel (distal end) in accordance with the required thread dimensions. The following dimensional notations shall be inscribed on the plug gages:

Thread Size:	5/16-28 UN-2B
"GO" PITCH DIAMETER:	0.2893
"HI" PITCH DIAMETER:	0.2937
"GO" MINOR DIAMETER:	0.2740
"NOT GO" MINOR DIAMETER:	0.2820

The gaging procedure to be used, dimensions, tolerances and gage identifications shall be in accordance with A.N.S.I. Standard B1.2. Fits of gages into internal threads shall be noted.

4.4.2 Fluid tightness test of barrel and plunger. The syringes selected for testing shall be thoroughly cleaned with soap and water, or other appropriate cleaning solution, rinsed thoroughly and dried. The syringes shall then be attached to a hydrostatic testing device containing a water-pressure gage. A hydrostatic pressure of 10 pounds per square inch gage (psig) shall be applied in a leakproof manner. Water shall be used as the testing fluid and shall be applied through the tip of the syringe. The plunger shall be drawn to the 1-ounce-graduation mark and allowed to remain for a period of not less than 10 minutes. After this first test, any volume of water that may have passed the plunger shall be measured and recorded. This test shall be performed with the unit completely assembled, including the tip and guard gaskets and the piston "O"-ring. The plunger shall then be drawn to the 2-ounce graduation mark, where it shall be allowed to remain for a period of not less than 10 minutes with the pressure still applied. After this second test, any volume of water that may have passed through the plunger shall be measured and recorded. All of the above testing shall be performed with the syringe placed in a horizontal position.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

4.4.2.1 Fluid tightness test of the tips and guard. The fluid tightness test of the tips and guard shall be performed exactly as described in 4.4.2, except that the 10 psig hydrostatic pressure shall be applied for a period of not less than 1 minute, and water that may have formed at the threaded connections of the tip and guard after the first test, shall be removed or wiped away.

5. PREPARATION FOR DELIVERY

5.1 Packaging. Packaging shall be Level A or C as specified (see 6.2).

5.1.1 Unit of issue. One syringe, as specified, constitutes one unit of issue.

5.1.2 Packaging quantities. The number of units of issue indicated in the following table shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in the procurement document.

Table IV. Packaging Quantities

Unit package	Intermediate package	Exterior container
1 unit	Not Required	48 units

5.1.2.1 Packing variation permitted. If the required number of units in the entire shipment is less than the number of units specified to be overpacked in an exterior container, such units may be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

5.1.3 Level A.

5.1.3.1 Unit package. Unit package shall be a box of appropriate size constructed in accordance with PPP-B-566 or PPP-B-676, except that commercial colors will be acceptable. Box shall be designed with suitable interior fittings or shall contain sufficient cushioning material to insure against shifting of or damage to, the contents. Closure shall be adequate to prevent accidental opening.

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:53Z
Check the source to verify that this is the currentversion before use.

5.1.4 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing. Packing shall be Level A, B or C as specified (see 6.2).

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type; PPP-B-621, class 2; or PPP-B-636, class weather-resistant, grade V3. Closure and strapping shall be as specified in the appendix of the applicable box specification.

5.2.1.2 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, domestic type; PPP-B-621, class 1; or PPP-B-636, class domestic. Closure shall be as specified in the appendix of the applicable box specification.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Civil agencies. Marking for shipment shall be in accordance with Fed. Std. No. 123.

5.3.2 Military activities.

5.3.2.1 Interior packages. Each unit package shall be marked in accordance with MIL-STD-129. When labels are utilized, waterproofing shall be required only when applicable carton is fabricated of water-resistant material. Contract or purchase order number and name of contractor shall be shown.

5.3.2.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

6. NOTES

6.1 Intended use. The metal ear syringe shall be suitable for use in the irrigation of the human ear.

6.2 Ordering data. Procurement documents should specify the following:

(a) Number and date of this specification.
(b) Size of syringe procured (see 1.2).
(c) Selection of applicable levels of packing and packaging (see 5.1 and 5.2).
(d) Desired options offered herein.

6.3 This specification does not cover all types, classes, grades, or sizes of the commodity indicated by the title of this specification, or those which are commercially available, but is intended to cover the types which are normally procured to meet Federal Government requirements.

6.4 For Military Medical Supply Agency procurement, Pomeroy ear tip specified in 3.3 is not required.

6.5 This specification covers the following item listed in the Federal Supply Catalog:

6515-379-5850 SYRINGE, METAL, EAR, 2 oz.

NOTE: Changes to this specification are not indicated by asterisk since the technical data added are only clarifications of the previous requirement.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activities:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6515-1553

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

CAUTION

NOTICE TO BIDDERS/OFFERORS

DO NOT CONDITION OR BASE YOUR BID/OFFER ON ANY CURRENT PROCUREMENT ON THE INFORMATION SUBMITTED ON THIS FORM SINCE ANY CHANGES OR DELETIONS IN THE SPECIFICATIONS MAY RENDER YOUR BID/OFFER NON-RESPONSIVE IN WHICH CASE IT CANNOT BE CONSIDERED FOR AWARD.

Fold

DEFENSE PERSONNEL SUPPORT CENTER
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

DEFENSE SUPPLY AGENCY
OFFICIAL BUSINESS

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

Fold

DPSC FORM 2418
MAR 68

EDITION AUG 66, WILL BE USED UNTIL EXHAUSTED

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

SPECIFICATION ANALYSIS SHEET

Form Approved
Budget Bureau No. 22-R255

INSTRUCTIONS: This sheet is to be filled out by personnel, either Government or contractor, involved in the use of the specification in procurement of products for ultimate use by the Department of Defense. This sheet is provided for obtaining information on the use of this specification which will insure that suitable products can be procured with a minimum amount of delay and at the least cost. Comments and the return of this form will be appreciated. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments and suggestions submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or serve to amend contractual requirements.

SPECIFICATION

GG-S-00928c(DSA-DM) FSN:

ORGANIZATION

CITY AND STATE

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE SPECIFICATION CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY SPECIFICATION REQUIREMENT CONSIDERED TOO RIGID

3. IS THE SPECIFICATION RESTRICTIVE?

☐ YES ☐ NO *(If "yes", in what way?)*

4. REMARKS *(Attach any pertinent data which may be of use in improving this specification. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity)*

SUBMITTED BY *(Printed or typed name and activity - Optional)*

DATE

DD FORM 1 JAN 66 1426 REPLACES EDITION OF 1 OCT 64, WHICH MAY BE USED.

Source: https://assist.dla.mil -- Downloaded: 202640-26T11:53Z
Check the source to verify that this is the current version before use.