GG-S-743C
July 28, 1975
SUPERSEDING
Int. Fed. Spec. GG-S-00743b
July 30, 1971 and
Fed. Spec. GG-S-743
September 20, 1957

FEDERAL SPECIFICATION

STAND, SURGICAL INSTRUMENT

This specification was approved by the Commissioner, Federal Supply Service, General Services Administration, for use of all Federal agencies.

1. SCOPE

1.1 This specification covers an adjustable surgical instrument stand fabricated of corrosion-resisting steel, suitable for use in operating rooms.

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposal, form a part of this specification to the extent specified herein.

Federal Specification:

PPP-B-636 - Boxes, Shipping, Fiberboard.

Federal Standards:

Fed. Std. No. 66 - Steel, Chemical Composition and Hardenability.
Fed. Std. No. 123 - Marking for Shipment (Civil Agencies).
Fed. Std. No. 151 - Metals, Test Methods.

FSC 6530

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:52Z
Check the source to verify that this is the current version before use.

(Activities outside the Federal Government may obtain copies of Federal Specifications and Standards as outlined under General Information in the Index of Federal Specifications and Standards and at the prices indicated in Index. The Index, which includes cumulative monthly supplements as issu is for sale on a subscription basis by the Superintendent of Documents, U.S Government Printing Office, Washington, DC 20402.

(Single copies of this specification and other Federal specifications required by activities outside the Federal Government for bidding purposes are available without charge at the General Services Administration Region Offices in Boston, New York, Washington, DC, Atlanta, Chicago, Kansas C MO, Denver, San Francisco, Los Angeles and Seattle, WA.

(Federal Government activities may obtain copies of Federal Specificatio and Standards and the Index of Federal Specifications and Standards from established distribution points in their agencies.)

Military Standards:

MIL-STD-105 - Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 - Marking for Shipment and Storage.

Drawings:

Defense Personnel Support Center (DPSC)

20019 - Tray, Instrument, Corrosion-Resisting Steel.
20713 - Stand, Surgical Instrument, Adjustable, Corrosion-Resisting Steel.

(Copies of Military Specifications, Standards and DPSC drawings require by suppliers in connection with specific procurement functions should be obtained as directed by the contracting officer.)

2.2 Other publications. The following document forms a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposal shall apply.

National Fire Protection Association (NFPA)

56A Standard for the Use of Inhalation Anesthetics.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:52Z
Check the source to verify that this is the current version before use.

(Application for copies should be addressed to National Fire Protection Association, 60 Batterymarch Street, Boston, MA 02110.)

3. REQUIREMENTS

3.1 Material.

3.1.1 Corrosion-resisting steel. Corrosion-resisting steel shall conform to series 300 of FED-STD-66 and in conformance with 4.4.3.

3.1.2 Hardware. Except as indicated in the DPSC drawing, nuts, bolts, screws, washers and other hardware shall be corrosion-resisting steel or plated cold rolled steel; except the thumb-control release and bushing shall be aluminum or corrosion-resisting steel.

3.1.3 Tolerances. The thickness of the steel and the wall thickness and diameter of tubing referenced in this specification are normal standard gage sizes. Commercial tolerances shall apply.

3.2 Construction. The stand shall consist of a base, an upright, a telescoping extension with a frame, and a removable tray. Construction of the stand shall be in conformance with DPSC drawing 20713. The tray shall conform with DPSC Drawing 20019. Stand shall be equipped with an automatic locking device which permits the tray to be readily raised, lowered or locked at any desired height by one-hand control.

3.2.1 Base. Base shall be constructed as shown on the drawing. Base shall be formed from 0.109 inch thick No. 12 U.S. Standard gage, corrosion-resisting steel. Casters and legs shall be securely attached to the bottom. The legs shall be formed downward and in contact with the floor plane.

3.2.2 Upright. Upright shall be of the style, design, and dimensions shown on the drawing. It shall be fully secured as shown on the applicable drawing.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:52Z
Check the source to verify that this is the current version before use.

3.2.3 Telescoping extension and tray supporting frame. Style, design, and dimensions of the telescoping extension and tray supporting frame shall be as indicated on the referenced drawing. Extension shall be capable of being raised or lowered as desired, in a smooth, non-seizing operation. The frame shall be provided with six conductive rubber bumpers, two spaced along each long dimension and one in each short dimension of the frame. The bumpers shall fully support the tray to prevent rattling of the tray in the frame. The exposed ends of tubing shall be closed by welding. The extension shall slide smoothly to permit free extension of the tray from 39 to 62 inches above the floor. The total horizontal rotation of the frame shall be not more than 1-1/2 inches when tested as specified in 4.4.4.

3.2.4 Automatic locking device. The stand shall have an efficient automatic locking mechanism with one-hand control, which shall permit a free up-and-down movement of the telescoping upright. The telescoping upright or extension rod shall automatically and positively lock in place at the point desired, when the tray is being raised. To lower the tray it shall be necessary to depress the control until the desired height is reached. Control for operating the automatic locking device shall be located near the top of the extension as shown on the drawing (see 6.4).

3.2.5 Casters. Casters shall be swivel type, with 2 inch diameter, ball bearing, electrically conductive rubber-tired wheels. The conductivity of the casters shall be such that the indicated resistance between the tray and a metal plate placed alternately under each caster, but insulated from the floor, shall not exceed 250,000 ohms when tested as specified in 4.4.1.

3.3 Stability. Base and legs shall have sufficient weight and stability so that the stand shall not tip when a force of 1-1/2 pounds is applied to the vertical upright 40 inches above the floor when tested as specified in 4.4.2. The stand shall return to its original, stable position when the extended stand is tilted 13 degrees from the vertical and released when tested as specified in 4.4.2.

3.4 Welding. Welding rod shall be material similar in composition to the welded members. Welds shall be thorough, nonporous, free of pits, cracks, undercuts or other imperfections. All welds shall be cleaned and all surfaces shall be free from sharp edges. All shall be smooth to blend with adjoining surfaces except welds on underside of the base.

3.5 Identification marking. Each stand shall be marked with the contract or purchase order number. Marking shall be legible, and permanent.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:52Z
Check the source to verify that this is the current version before use.

3.6 Finish. All exposed corrosion-resisting steel surfaces shall be finished in a No. 4 commercial finish. Carbon steel hardware shall be cadmium or zinc plated. All surfaces shall be free from pitmarks and scale.

3.6.1 Alternate exterior finish. As an alternate, the exterior surface of the tray only shall have a No. 2 bright annealed finish, resulting in a uniform, bright, and shiny surface.

3.7 Workmanship. Workmanship shall be first class throughout. The finished stand shall be free from defects which detract from its appearance or may impair its serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual and auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the steel, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the Item Identification, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:52Z
Check the source to verify that this is the current version before use.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, and Table I.

Table I. Sampling for Examination

	Inspection level	AQL (defects per hundred units)
For visual examination		
Major defects	II	1.0
Minor defects	II	4.0
For dimensional examination	S-3	2.5

4.2.2 For tests. Sampling for tests shall be conducted in accordance with Inspection level S-3 of MIL-STD-105 and as shown in Table II.

Table II. Sampling for Tests

Component (Unit product)	Characteristic	Requirement	Test procedure	Insp. level	AQL (perc defective)
End Item	Electrical conductivity	3.2.5	4.4.1	100%	*
End Item	Stability	3.3	4.4.2	**	
Steel	Chemical Composition	3.1.1	4.4.3		*
End Item	Horizontal Rotation	3.2.3	4.4.4	S-3	4.0

* Acceptance number shall be zero, and not less than three samples shall be taken regardless of lot size.

** One complete stand per contract.

4.3 Examination. The stands shall be examined to determine compliance with all requirements contained in this specification.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:52Z
Check the source to verify that this is the current version before use.

4.3.1 Classification of defects. Examination shall be conducted in accordance with the classification of defects as shown in Table III.

Table III. Classification of Defects

Categories	Defects *
Major	
101	Stand not complete with all components.
102	Unit not free of burrs, rough surfaces, sharp edges.
103	Any component does not function as specified.
104	Any weld not continuous, as specified, not clean, not free of pitting, splatter, discoloration.
105	Extension cannot be raised or lowered without binding.
106	Locking control does not release easily; does not lock automatically.
Minor	
201	Unit not clean; not free of dirt, oil, and grease.
202	Finish not as specified.
203	Identification marking not present, not properly located, not legible.
204	Trays not free of rattle in frames.
205	Extension rod and upright not free of rattle.

* Examination shall not be restricted to the classified possible defects listed above.

4.3.2 Dimensional examination. The stands shall be examined for defects in dimensions. Any dimension not within the tolerances specified herein shall be classified as a defect.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Electrical conductivity. The assembled stand, complete with rubber casters, shall be tested using the equipment described in NFPA No. 56A. The test shall be conducted with the metal plate alternately under each caster.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:52Z
Check the source to verify that this is the current version before use.

4.4.2 Stability. The telescoping extension shall be extended to its full height and the wheels shall be blocked to prevent rolling. A gradually increasing force (up to 1-1/2 pounds) shall be applied to the vertical upright, 40 inches above floor level and in the direction opposite to the tray frame and legs. The stand shall be checked for stability. The fully extended stand shall be placed on a 13° inclined plane with the caster wheels parallel to the leg centerline, pointing fully away from the legs and blocked to prevent rolling. The stand shall be checked for stability.

4.4.3 Chemical composition. The composition of the corrosion-resisting steel used in the construction of the stand shall be determined by method 111.2 or 112.2 of FED-STD-151.

4.4.4 Rotation. With the tray support frame in the uppermost position, the telescoping extension shall be rotated with respect to the upright. Measurement of the horizontal rotation shall be made at the midpoint of the end of the frame farthest from the upright.

4.5 Examination for preparation for delivery. An examination shall be made to determine whether the packaging, packing, and marking comply with the requirements of Section 5. Defects shall be scored as specified in Table IV. Sampling shall be in accordance with MIL-STD-105. The sample unit shall be one container fully prepared for delivery. The lot shall be the number of containers offered for inspection at one time. The inspection level shall be S-2 and the acceptable quality level shall be 4.0 defects per hundred units.

Table IV. Examination for Preparation for Delivery

Examination	Defect
Containers	Not as specified.
Quantity	Quantity in interior packages and shipping containers not as specified.
Packaging and Packing Materials	Any component missing or damaged.
Workmanship	Inadequate application of components, such as incomplete closure of container flaps, tape improperly applied or containers distorted.
Markings	Omitted, incorrect, illegible, improper size, location, sequen or method of application.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:52Z
Check the source to verify that this is the current version before use.

5. PREPARATION FOR DELIVERY

5.1 Packaging. Packaging shall be level A or C as specified (see 6.2).

5.1.1 Unit of issue. One stand, as specified, constitutes one unit of issue.

5.1.2 Level A.

5.1.2.1 Unit preparation. All exterior finished surfaces shall be protected by wrapping with flexible corrugated or other suitable materials. Wrapping and cushioning material shall be secured by tying with cotton cord, tape or by sealing with kraft gummed paper tape. Work shall be accomplished in such a manner as to insure against marring, scratching, or damage to the unit.

5.1.3 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing. Packing shall be level A, B or C as specified (see 6.2).

5.2.1 Level A.

5.2.1.1 Exterior container. Each unit, prepared as specified in 5.1.2.1 shall be overpacked in an exterior container designed for a type 3 load constructed in accordance with PPP-B-636, class weather resistant. The box shall conform to the special requirements as specified in PPP-B-636. Closure, strapping and waterproofing shall be as specified in the appendix of the box specification.

5.2.2 Level B.

5.2.2.1 Exterior container. Each unit prepared as specified in 5.1.2.1 shall be overboxed in an exterior container designed for a type 3 load, and constructed in accordance with PPP-B-636, class domestic. The box shall conform to special requirements as specified in PPP-B-636. Closure shall be as specified in PPP-B-636.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:52Z
Check the source to verify that this is the currentversion before use.

5.3 Marking.

5.3.1 Civil agencies. Marking for shipment shall be in accordance with FED-STD-123.

5.3.2 Military activities.

5.3.2.1 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Intended use. An adjustable, stable, surgical instrument stand intended for use in medical treatment facilities.

6.2 Ordering data. Procurement documents should specify the following:

(a) Number and date of this specification.

(b) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.3 This specification covers the following item appearing in the Federal Supply Catalog.

NSN	Item Identification
6530-00-708-1590	STAND, SURGICAL INSTRUMENT, Adjustable, Corrosion-Resisting Steel.

6.4 An acceptable automatic locking device for that specified in 3.2.4 is available from the Deal Slide Lock Co. their Model BSRS-53 with D wedge, and from Brenner Metal Products Co., their Part No. 1274-23.

6.5 This specification does not cover all types, classes, grades, or sizes of the commodity indicated by the title of this specification, or those which are commercially available, but is intended to cover the types which are normally procured to meet requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:52Z
Check the source to verify that this is the current version before use.

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DSA-DM

Project No. 6530-1331

Civil Agency Coordinating Activity:

GSA-FSS

Orders for this publication are to be placed with General Services Administration, acting as an agent for the Superintendent of Documents. See section 2 of this specification to obtain extra copies and other documents referenced herein. Price 35 cents each.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:52Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:52Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE
GG-S-743C STAND, SURGICAL INSTRUMENT

NAME OF ORGANIZATION AND ADDRESS	CONTRACT NUMBER
	MATERIAL PROCURED UNDER A ☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?
   A. GIVE PARAGRAPH NUMBER AND WORDING.

   B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?
   ☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*	TELEPHONE NO.
	DATE

DD FORM 1426 REPLACES EDITION OF 1 JAN 66 WHICH MAY BE USED

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:52Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 South 20th Street
Philadelphia, PA 19101

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY
DoD - 304

U.S. MAIL

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:52Z
Check the source to verify that this is the current version before use.

5 0285 M

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:52Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:52Z
Check the source to verify that this is the current version before use.

NOTICE OF CANCELLATION

GG-S-743C
NOTICE 1
31 JANUARY 1996

FEDERAL SPECIFICATION

STAND, SURGICAL INSTRUMENT

GG-S-743C, dated 28 JULY 1975, is hereby cancelled without replacement.

Preparing Activity

DOD-MB

Project Number: 6530-2461

AMSC/NA FSC 6530

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:52Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:52Z
Check the source to verify that this is the current version before use.