FEDERAL SPECIFICATION

SUCTION AND PRESSURE APPARATUS, ELECTRIC (FOR MEDICAL USE)

This specification was approved on the above date by the Commissioner, Federal Supply Service, General Services Administration, for the use of all Federal agencies.

1. SCOPE AND CLASSIFICATION

1.1 Scope. — This specification covers an explosion-proof suction and pressure apparatus designed to incorporate an ether vaporizer.

1.2 Classification.—The suction and pressure apparatus covered by this specification shall be of one type.

2. APPLICABLE SPECIFICATIONS AND OTHER PUBLICATIONS

2.1 Specifications.—The following Federal Specifications, of the issues in effect on date of invitation for bids, form a part of this specification:

CC-M-636—Motors; Alternating-Current, Fractional - Horsepower, Single - Phase and Universal.

FF-C-86—Casters; Truck.

QQ-S-636 — Steel; Carbon (Low-Carbon), Sheets and Strips.

QQ-S-766 — Steel, Corrosion-Resisting, Plates, Sheets, Strips, and Structural Shapes.

TT-E-491—Enamel; Gloss, Synthetic (for Metal and Wood Hospital Furniture).

TT-P-659 — Primer-Surfacer; Synthetic, Tints and White (for Metal and Wood Hospital Furniture).

ZZ-R-601—Rubber Goods; General Specifications (Methods of Physical Tests and Chemical Analyses).

(Copies of Federal specifications and the Federal Specificaitons Index may be obtained upon application, accompanied by money order, coupon, or cash, to the Superintendent of Documents, Government Printing Office, Washingon 25, D. C. Prices may be obtained from the Superintendent of Documents.)

2.2 Other publication. — The following standard of the Underwriters' Laboratories, Inc., of the issue in effect on date of invitation for bids, forms a part of this specification:

Subject 73—Standard for Motor-Operated Appliances.

(Copies of standards of the Underwriters' Laboratories are available from Underwriters' Laboratories, Inc., 161 Sixth Ave., New York 13, N. Y., 207 East Ohio St., Chicago 11, Ill., and 500 Sansome St., San Francisco 11, Calif.)

2.3 Specications and other publications applicable only to individual departments are listed in section 7.

3. REQUIREMENTS

3.1 Underwriters' Laboratories, Inc.—The contractor shall submit to the purchasing agency proof that the commodity he furnishes under this specication is acceptable to the Underwriters' Laboratories, Inc., regarding operation in hazardous locations, Class I, Group C, as prescribed by the National Board of Fire Underwriters. The label, marker, listing, or certified report of the Underwriters' Laboratories, Inc., will be accepted as evidence of compliance with this requirement.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

Photolithographic Medical Film

3.1.1 In lieu of evidence of the Underwriters' Laboratories, Inc., the contractor may submit evidence from other sources that his commodity conforms to applicable published standards of the Underwriters' Laboratories, Inc. (See 2.2.)

3.1.2 Compliance with fire-and-casualty-hazard requirements of Underwriters' Laboratories, Inc., does not absolve the contractor from complete compliance with the requirements of this specification in order to secure the acceptance of his commodity. In the event of difference between any of the requirements of this specification and those of Underwriters' Laboratories, Inc., the requirements of this specification shall govern.

3.2 Material.

3.2.1 *Carbon steel.*—The carbon steel shall conform to the requirements of Federal Specification QQ-S-636 for standard commercial cold-rolled steel.

3.2.2 *Corrosion-resisting steel.*—The corrosion-resisting steel shall conform to the requirements of Federal Specification QQ-S-766 for Class 1 material.

3.2.3 *Glass.*—The bottles shall be of clear glass, well-annealed, free from cracks, flaws, or other defects, and shall show no corrosion, scumming, chipping, or cracking when tested as described in 4.3.1.

3.2.4 *Rubber tubing.*—The rubber tubing shall be fabricated of rubber having suitable characteristics for the intended purpose. The tubing shall be electrically conductive (see 3.4 and 4.3.3) and shall withstand the accelerated aging test described in 4.3.4 without cracking, without becoming tacky or brittle or decreasing in tensile strength by more than 25 percent.

3.2.5 *Primer surfacer.* — The primer surfacer shall conform to the requirements of Federal Specification TT-P-659.

3.2.6 *Enamel.*—The enamel shall be flexible, acid and alkali resistant, and shall conform to the requirements of Federal Specification TT-E-491, Class B. The color after application to the unit shall be as specified by the procuring agency.

3.2.7 *Other materials.* — Other materials shall be of high commercial quality throughout, shall be durable, suitable in all respects for the purpose intended, and as hereinafter specified.

3.3 Description.—The suction and pressure apparatus herein described shall be an explosion-proof unit of the cabinet type suitable for heavy-duty hospital work. It shall consist essentially of a cabinet, a motor, two eccentric rotor rotary-vane-type pumps, an anesthesia unit, a suction unit, and filtering devices, all properly connected. The unit shall be provided with the required accessories as specified hereinafter.

3.3.1 *Cabinet.*—The cabinet shall be of the table type, having a ventilated, noiseproofed compartment. The compartment shall be of the proper size and design for mounting the motor and compressor. The cabinet shall be provided with handles at the ends and shall have a drawer of suitable size to contain the accessories hereinafter specified. The overall dimensions of the cabinet shall not exceed 32 inches in width (including the handles), 17 inches in depth, and 35 inches in height. The cabinet shall be provided with four electrically conductive rubber casters not less than 4 inches in diameter, conforming to the requirements of Federal Specification FF-C-86, Type II, Class A, material 2 or 3, capacity 100 pounds. The conductivity of the casters shall be such that the resistance between the metallic frame of the cabinet and a metallic plate placed under one caster but insulated from the floor shall not exceed 0.25 megohm. An electrically conductive rubber bumper shall surround the lower periphery of the cabinet and shall be firmly secured thereto. The bumper shall ex-

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

tend beyond the outermost projection on the cabinet. The conductivity of the bumper shall be such that the resistance between the metallic frame of the cabinet and any point on the rubber bumper shall be not greater than 0.25 megohm.

3.3.1.1 *Cabinet finish.*

3.3.1.1.1 *Phosphate treatment and filling.*— After cleaning and degreasing, each cabinet shall be given a phosphate treatment (see 7.2.2). This shall be followed by a heavy coat of mineral filler baked and sanded to give a smooth, even finish. The filler shall cover the joints and blemishes caused by the welding.

3.3.1.1.2 *Primer.* — After the phosphate treatment and filling, the cabinets shall have a coat of neutral primer sprayed on and baked.

3.3.1.1.3 *Enamel.*—All interior and exterior surfaces of the cabinets shall be finished in enamel. Two coats of enamel shall be applied and baked to a complete polymerization. The first coat shall be sanded to a smooth finish before the second coat is applied. The final coat shall have a smooth, hard, and tough finish with a permanent luster. The color of the enamel shall be as specified by the purchasing agency. (See 7.2.3.)

3.3.1.1.4 *Top and hardware.*—The top shall be corrosion-resisting steel with exposed surfaces finished to conform to the requirements of Federal Specification QQ-S-766 for finish 4. The finish of the hardware, including the fittings and the fixtures, shall be chromium plated over nickel, polished to a mirror finish.

3.3.2 *Motor.* — Unless otherwise specified, each unit shall be provided with a motor of not less than 1/6-hp. capacity. The motor shall be of the heavy duty type, totally enclosed, explosion-proof, Class I, Group C locations, and shall conform to the requirements of Federal Specification CC-M-636. It shall operate on alternating current, 110 volts, 60 cycles. The motor shall be securely mounted within the cabinet in a manner to eliminate vibration and noise. The motor shall be provided with suitable means for lubrication and shall not require lubrication oftener than every 2,000 hours of operation. The motor shall be operated by an easily accessible explosion-proof switch which shall be marked to indicate the "off" and "on" positions.

3.3.3 *Pumps.*—Each unit shall be provided with two pumps of the eccentric-rotor rotary-vane type which shall have a capacity of not less than ½ cubic foot per minute. The pumps shall be either directly connected to the motor shaft or connected through a suitable flexible connection in a manner to insure quiet operation. One pump shall be arranged to create a positive pressure of at least 20 pounds per square inch, and the other shall be arranged to create a negative pressure of not less than 20 inches of mercury. The pump shall be provided with suitable means for lubrication and shall not require lubrication oftener than every 10 hours of operation.

3.3.4 *Anesthesia-unit assembly.* — The anesthesia-unit assembly shall consist essentially of an ether bottle and valves, a pressure gage, filters, an ether tip, and the necessary tubing to connect the apparatus suitably. The anesthesia unit shall be so designed that it shall be capable of delivering amounts of ether vapor to the ether tip as indicated below, when tested as specified in 4.3.5.

Flow in liters per minute at outlet of ether tip	Ounces (minimum) of ether evaporated per liter flow
4	0.055
8	.047
10	.045
12	.044
18	.043

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

3.3.4.1 *Ether bottle and valves.*—The ether bottle shall be of 32-ounce capacity and shall be of such construction as to withstand an internal hydrostatic pressure of 85 pounds per square inch without bursting. The bottle shall be suitably connected to the positive pressure pump and shall be securely mounted on the top of the cabinet in a manner to permit it to be removed easily. A pressure relief device shall be incorporated in the positive pressure system which shall act to vent the pressure in the ether bottle at a value not in excess of 30 percent of the hydrostatic test value specified above. The flow of air through the ether bottle shall be controlled by needle valves or other suitable devices which will permit fine adjustments and accurate control of flow from full open to full closed. The valves or devices shall be marked or calibrated to indicate their position relative to full open or full closed. The ether bottle shall be provided with a suitable closure which will render it airtight and which shall be designed so as to permit the bottle to be filled easily without removal.

3.3.4.2 *Pressure gage.*—The anesthesia unit shall be provided with a pressure gage calibrated to indicate pressures from 0 to 30 pounds per square inch. This gage shall be so connected in the anesthesia system as to plainly show the pressure within the pressure system. The gage shall be accurate to within plus or minus 10 percent at each scale reading.

3.3.4.3 *Anesthesia-system filters.*—A series of filters shall be provided in the anesthesia system to prevent dust particles or oil vapors from entering the air stream. The filters shall be of such size that they need not be replaced any oftener than every 50 hours of operation and shall be so designed that the filtering medium may be easily replaced.

3.3.5 *Suction assembly.* — The suction assembly shall consist essentially of a suction bottle and valves, a gage, and filters or traps.

3.3.5.1 *Suction bottle.*—The suction bottle shall be of flint glass or other suitable glass. The bottle shall be preferably of ½-gallon capacity; however, 1-gallon-capacity bottles will be acceptable. The bottle shall be securely mounted on top of the cabinet in a manner to permit it to be removed easily. The suction bottle shall be provided with an airtight closure.

3.3.5.2 *Suction-system gage.* — The suction unit shall be provided with a suction gage calibrated to indicate negative pressures from 0 to 30 inches of mercury. This gage shall be connected to the suction system to show the negative pressure in the suction bottle plainly. The gage shall be accurate to within plus or minus 10 percent at each scale reading.

3.3.5.3 *Suction-system filters and traps.*—Suitable filters or traps shall be provided in the suction system, interposed between the suction bottle and the suction pump, to protect the pump adequately from entrained moisture droplets or from fouling due to overflow of the suction bottle. The filters shall be designed so that the filtering medium may be easily replaced and traps shall be designed so that they may be easily cleaned.

3.4 **Rubber tubing and couplings.**—Each unit shall be provided with sufficient tubing to perform its intended function. The tubing shall be $\frac{5}{32}$ inch inside diameter by $\frac{3}{32}$ inch wall and shall be provided with couplings fabricated of brass or steel which shall be chromium-plated over nickel. The couplings shall be of a tapered slip type with positive screw lock. The tapered portion of the couplings shall be tapered 29/64 inch per inch from a diameter of $\frac{13}{32}$ inch decreasing for a distance of $\frac{13}{32}$ inch to a diameter of ¼ inch. The female portion of the coupling shall have a corresponding taper. The male part of the coupling shall be provided with a threaded portion ($\frac{7}{16}$ inch by 27 NC thread) which shall mate with a threaded collar on the female part of the coupling.

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:53Z
Check the source to verify that this is the current version before use.

The collar and threaded portion shall insure locking of the coupling to form a tight leakproof connection. The tubing shall be electrically conductive to the extent that the resistance of objects which are interconnected with the rubber tubing, unless they are electrically connected, shall not exceed 1 megohm when measured as specified in 4.3.3.

3.5 Performance at operative end of tubings.—The delivery at the pressure end of the tubing with valves full open shall be 0.5 cubic feet per minute with a minus tolerance of 5 percent. The intake at the suction end of the tubing shall be 0.5 cubic feet per minute with a minus tolerance of 5 percent.

3.6 Electric wiring. — The unit shall be wired with a 3-wire grounded electrical system. The unit shall be equipped with a flexible insulated three-wire cord not less than 10 feet long. No attachment plug shall be furnished.

3.7 Voltage withstand. — The unit shall withstand without failure the voltage-withstand test specified in 4.3.2.

3.8 Accessories. — The suction and pressure apparatus shall be completely equipped with all the required accessories including the following:

(a) Ether tip.
(b) Tonsil and abdominal suction tubes.

3.9 Identification marking. — Each unit shall be permanently and legibly marked with the name or registered trade-mark of the manufacturer, the identification symbol, serial number, voltage, amperage, type of current and number of cycles, and such other data as may be required by Federal and State laws.

3.10 Workmanship. — The workmanship shall be first class throughout. The suction and pressure apparatus shall be free from any imperfections which affect its appearance or which may impair its serviceability. The finished apparatus shall operate smoothly and noiselessly and shall efficiently perform its functions in a manner suitable for the purpose intended.

4. SAMPLING, INSPECTION, AND TEST PROCEDURES

4.1 Sampling.

4.1.1 *Unit.* — Unless otherwise specified, each apparatus shall be inspected and tested.

4.1.2 *Bottles.*—Unless otherwise specified, three bottles from each lot of 100 or fraction thereof delivered, shall be taken at random for inspection and tests.

4.1.3 *Enamel and primer.* — Unless otherwise specified, 1 quart each of the enamel and primer shall be furnished for tests. The manufacturer shall certify that the samples furnished are from the same lot as that used in the manufacture of the commodity.

4.2 Inspection. — Each unit shall be inspected for materials, workmanship, dimensions, finish, construction and presence of accessories, packaging, packing, and markings.

4.3 Tests.—Tests shall be conducted in accordance with the applicable specifications referenced in 2.1, and as specified hereinafter.

4.3.1 *Glass.*

4.3.1.1 *Shock test.*—The bottles taken for test shall be immersed in boiling water for 5 minutes, emptied, and then plunged bottom down into water at a temperature of 50° ± .1° C.

4.3.1.2 *Boiling test.*—The bottles taken for test shall be completely immersed in distilled water in a closed vessel having a small steam vent and boiled continuously for 6 hours, boiling distilled water being added occasion-

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

ally to compensate for evaporation losses. The bottles shall be suspended in water so that they will not come in contact with the walls of the containing vessel.

4.3.1.3 *Annealing.*—Unless otherwise specified, annealing shall be determined by polariscope. Samples showing more than a slight degree of ring strain or longitudinal strain shall be rejected.

4.3.2 *Voltage withstand.* — All electrical parts shall be tested for voltage withstand. The voltage withstand potential shall be 1,000 volts, 60 cycles, alternating current, applied for 1 minute between noncurrent-carrying metal parts and current-carrying metal parts.

4.3.3 *Electrical conductivity of rubber parts.*—The electrical conductivity of parts required to be conductive, as specified herein, shall be tested with an ohmmeter which will maintain a voltage between 90 and 500 volts across a resistance of 1 megohm.

4.3.4 *Accelerated aging tests.*—The tubing shall be subjected to either of the following aging tests:

4.3.4.1 *Oxygen bomb test.* — The tubing shall be subjected to the oxygen bomb test described in Federal Specification ZZ–R–601 for a period of 4 days at 70° ± 1° C. with oxygen pressure at 300 ± 10 pounds per square inch.

4.3.4.2 *Air oven test.*—The tubing shall be subjected to the air oven test described in Federal Specification ZZ–R–601 for a period of 7 days at 70° ± 1° C.

4.3.5 *Ether-vaporization rate.* — The ether bottle shall be filled with ether to cover the diffusing element of the bubbling tube to a height of 2 inches. The ether bottle including the contents and the bubbling tube shall be weighed and then chilled to such a temperature that at the beginning of the test the temperature will be 32° ± 1° F. The ether bottle shall be assembled to the apparatus and the apparatus operated at the prescribed rates of flow (see 3.3.4) for 4 minutes. At the end of each 4-minute flow test the ether bottle contents and bubbling tube shall be removed and weighed. The amount of ether vaporized per liter shall be calculated from the average of four tests for each flow rate.

5. PREPARATION FOR DELIVERY

5.1 Packaging. — Unless otherwise specified, each suction and pressure apparatus shall be packaged in accordance with accepted commercial practice.

5.2 Packing.—The subject commodity shall be prepared for shipment to permit acceptance by carrier for transportation at the lowest applicable rate, and to afford maximum protection from normal hazards of transportation.

5.3 Marking.

5.3.1 *Packages.* — Unless otherwise specified, packages shall be marked with the name of the commodity, the quantity contained therein, and the names of the manufacturer and the contractor.

5.3.2 *Shipping containers.* — Unless otherwise specified, shipping containers shall be marked with the name of the commodity as defined by the contract or order under which the shipment is made, quantity contained therein, name of the contractor, number of the contract or order, name and address of the consignee, and the gross weight.

6. NOTES

6.1 Ordering data.—Purchasers should exercise any desired options offered herein. (See 3.2.6, 3.3.1.1.3, 3.3.2, 4.1, 4.3.1.3, 5.1, 5.3, 6.3, and 7.2.9.)

6.2 This specification covers the following

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:53Z
Check the source to verify that this is the currentversion before use.

item appearing in the Armed Services Catalog of Medical Matériel:

3-753-100—Suction and Pressure Apparatus, 110 Volt, 60 Cycle, AC.

6.3 It is believed that this specification adequately describes the characteristics necessary to secure the desired material, and that normally no samples will be necessary prior to award to determine compliance with this specification. If, for any particular purpose, samples with bids are necessary, they should be specifically asked for in the invitation for bids, and the particular purpose to be served by the bid sample should be definitely stated, the specification to apply in all other respects.

6.4 This specification does not include all types, classes, grades, sizes, etc., of the commodity indicated by the title of the specification, or which are commercially available, but is intended to cover the type, etc., normally procured for Federal Government requirements.

Patent notice.—When Government drawings, specifications, or other data are used for any purpose other than in connection with a definitely related Government procurement operation, the United States Government thereby incurs no responsibility nor any obligation whatsoever; and the fact that the Government may have formulated, furnished, or in any way supplied the said drawings, specifications, or other data, is not to be regarded by implication or otherwise as in any manner licensing the holder or any other person or corporation, or conveying any rights or permission to manufacture, use, or sell any patented invention that may in any way be related thereto.

7. DEPARTMENTAL REQUIREMENTS

7.1 The following specifications and other publications, of the issues in effect on date of invitation for bids, and special requirements, form a part of this specification for purchases made under this specification by the respective departments.

7.2 Army, Navy and Air Force.

7.2.1 *Applicable specifications and other publications.*

Federal Specification:

NN-B-621—Boxes, Wood, Nailed and Lock-Corner.

Military Specifications:

JAN-P-106—Packaging and Packing for Overseas Shipment—Boxes; Wood, Nailed.

JAN-P-125—Packaging and Packing for Overseas Shipment—Barrier-Materials, Waterproof, Flexible.

MIL-M-3157 — Marking of Domestic Shipments by Contractors for Supplies and Equipment Consigned to the Medical Departments of the Armed Services.

U. S. Army Specification:

57-O-2—Finishes, Protective, for Iron and Steel Parts.

Navy Department Specification:

General Specifications for Inspection of Material.[1]

Container Corporation of America Publication:

Color Harmony Manual.

(Copies of specifications required for Army and Air Force purchases may be obtained upon application to the procuring agency. Both the title and identifying symbol or number should be stipulated when requesting copies.

(Copies of Federal and Military specifications required for Navy purchases, and Navy Department specifications, may be obtained without cost upon application to the Bureau of Supplies and Accounts, Department of the Navy, Washington 25, D. C., except that Naval Activities should make application to the Commanding Officer, Naval Supply Depot, Scotia 2, New York. Both the title and identifying symbol or number should be stipulated when requesting copies of specifications.

(Color Harmony Manual is published by the Container Corporation of America, 38 South Dearborn St., Chicago, Ill.)

[1] Applicable to Navy purchases only.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

7.2.2 *Phosphate treatment material.*—The phosphate treatment material (see 3.3.1.1.1) shall conform to the requirements of U. S. Army Specification 57–O–2, Type II, Class A.

7.2.3 *Color of enamel.*—The color after application to the unit shall be a light green, hue 20 ge, of the Color Harmony Manual, Container Corporation of America.

7.2.4 *Service data.*—The contractor shall enclose with each unit two copies of one or more booklets which contain complete instructions for installation, operation, and maintenance, and a list of spare parts. Wiring diagrams shall be included for electrical items or components. Each part other than common fasteners, etc., shall be identified by the manufacturer's part number. In addition, the contractor shall mail one copy of the service data to Specifications Branch, Armed Services Medical Procurement Agency, 84 Sands St., Brooklyn 1, N. Y., and one copy to Medical Technical Maintenance Division, Armed Services Medical Procurement Agency, St. Louis Medical Depot, 12th and Spruce Sts., St. Louis, Mo.

7.2.5 *Inspection procedures (Navy only).*—The general inspection procedures shall be in accordance with General Inspections for Inspection of Material.

7.2.6 *Packaging.*

7.2.6.1 *Unit packaging.*—One suction and pressure apparatus, as specified, constitutes one unit. Each unit shall be disassembled as far as is practicable and packaged as follows:

7.2.6.1.1 *Disassembled parts.* — Disassembled parts shall be adequately protected by wrapping with flexible corrugated fiberboard. Parts so wrapped shall then be packaged in a nailed wood box of suitable size and design. The box shall contain sufficient cushioning material (corrugated pads, air cells, or cellulose wadding) to insure against shifting and breakage.

7.2.6.1.2 *Motor and pumps.*—If motor and pumps are flexibly mounted (i.e., floating type), they shall be adequately blocked in a manner to prevent movement of the motor or pumps while in transit or storage.

7.2.6.1.3 *Cabinet.*—Doors and drawers of the cabinet shall be secured to prevent accidental opening while in transit or storage. The cabinet shall then be wrapped with kraft paper of suitable weight and padded with corrugated fiberboard or with excelsior pads. Wrapping and padding shall be firmly secured with kraft gummed paper tape, or with cotton twine or tape, or with any combination of the three. All work shall be accomplished in a manner to prevent scratching or marring of finished surfaces or damage to the cabinet.

7.2.7 *Packing.*

7.2.7.1 *Domestic.* — The cabinet, and the box containing disassembled components, shall be packed in a shipping container conforming to the requirements of Federal Specification NN–B–621. In addition to the regular specification requirements, the shipping container design shall include skids of appropriate size to facilitate handling and to maintain the unit in an upright position. Box and cabinet shall be secured within the shipping container by means of braces, blocks, or cushioning material in such a manner as to insure against shifting, damage, or marring of surfaces. The cabinet shall be blocked or braced in such a manner that the weight is not carried directly by the casters and so that any protruding parts of the cabinet do not contact the inner walls of the container.

7.2.7.2 *Overseas.*—The cabinet, and the box containing disassembled components, shall be packed in a shipping container designed for type 3 load and conforming to the requirements of Military Specification JAN–P 106. Closure and strapping shall conform to the requirements of the appendix thereto. Each shipping container shall be lined with

Source: https://assist.dla.mil -- Downloaded: 202640-26T11:53Z
Check the source to verify that this is the currentversion before use.

a sealed, waterproof bag, or its equivalent, made from material conforming to the requirements of Military Specification JAN-P-125. Barrier shall be sealed with an appropriate waterproof adhesive. In addition to the regular specification requirements, the shipping container design shall include skids of appropriate size to facilitate handling and to maintain the unit in an upright position. Box and cabinet shall be secured within the shipping container by means of braces, blocks, or cushioning material in a manner to insure against shifting, damage, or marring of surfaces. The cabinet shall be blocked or braced in such a manner that the weight is not carried directly by the casters and so that any protruding parts of the cabinet do not contact the inner walls of the container.

7.2.8 *Marking.*

7.2.8.1 *Packages.* — Packages shall be marked to show the contents contained therein.

7.2.8.2 *Shipping containers.*

7.2.8.2.1 *For domestic shipment.* — Each shipping container shall be marked in accordance with the requirements of Military Specification MIL-M-3157.

7.2.8.2.2 *For overseas shipment.* — Each shipping container shall be marked as specified by the procuring agency.

7.2.9 *Ordering data.* — Procurement documents should specify whether the subject commodity is to be packed and shipping containers marked for domestic or overseas shipment. (See 7.2.7 and 7.2.8.2.)

NAVY INTEREST: Md

This specification is a part of Section IV, Part 5, of the Federal Standard Stock Catalog.

For sale by the Superintendent of Documents, Washington 25, D. C. Price 10 cents.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.