COMMERCIAL ITEM DESCRIPTION

FLASK, DENTURE, CROWN AND BRIDGE

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for Military Specification, MIL-F-36585, which is cancelled.

This commercial item description covers a crown and bridge denture flask suitable for use in processing plastic crowns, inlays, and bridges during dental laboratory procedures.

Salient characteristics:

Material.

Flasks. The denture flasks shall be fabricated from manganese bronze C86500 as specified in ASTM Designation B584, for composition only.

Pins. The dowels (guide pins) shall be fabricated from Class 3, type 303 or type 304 austenitic stainless steel as specified in ASTM Designation F899, for composition only.

Style, design and dimensions. The style, design and dimensions of the crown and bridge denture flask shall be in accordance with Defense Personnel Support Center (DPSC) Drawing No. 24146. The flask shall be of four-piece construction and shall consist of a top cover, an upper section, a lower section, and a bottom cover. The internal draft allowance shall facilitate withdrawal of the hardened investment.

Dowels. Six dowels shall be force-fitted into the flask to ensure accurate alignment of the assembled components. The upper section shall contain two intrasectional dowels which shall maintain accurate registration with the lower section without any lateral play when opposing mating surfaces are assembled. The upper and lower sections shall contain two cover dowels for properly positioning the top or the bottom cover, as applicable, within the specified limits and with no more than 0.004 inch shake permissible. The placement of the cover dowels shall be such that the top and bottom covers cannot be interchanged. The dowels shall be tapered to facilitate assembly of the four components parts.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:27Z
Check the source to verify that this is the current version before use.

Relief grooves and holes. The top and bottom covers and the upper section shall each contain one small relief groove at the junction line of the covers and center sections to facilitate disassembling of the flask. The top and the bottom covers shall each contain two relief holes to permit exudation of excess investment.

Finish. The interior of the flask shall have an "as-cast" finish, and the exterior shall have a clean, smooth commercial finish. All mating surfaces shall be properly machined or ground to ensure positive metal to metal contact during dental laboratory processing. The dowels shall have a ground, polished finish.

Identification marking. Each crown and bridge flask shall be serialized and each of the four component parts shall be permanently and legibly stamped with the same serial number to assure ready identification and to facilitate accurate assembly by the user. The name or trademark of the manufacturer or supplier shall be permanently and legibly marked on the exterior grooved portion of each of the respective covers.

Workmanship. The flasks shall be free from cracks, bubbles, pits, or any other defects which detract from their appearance or may impair their serviceability.

Unit. Each (EA). One flask, as specified, constitutes one unit.

Unit package. The individual flask shall be packaged in a durable carton with a descriptive label. The dimensions of the carton shall be 4-1/2 inches by 3 inches by 2-1/2 inches nominal.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:27Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Certificates of quality. Certificates of quality for raw material, supplied by the manufacturer of the material, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection, examination and test characteristics. The unit of product for inspection, examination and test purposes shall be one flask. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, except for dimensions which shall be conducted at inspection level S-1, and sampling for tests (when applicable) shall be conducted at inspection level S-2 and acceptance number shall be zero (0) for all sample sizes.

Examination. Examination shall include, but shall not be limited to the following:

Defects

Dimensions not as specified.
All components of the same flask not stamped with any serial number.
Any component of the same flask missing.
Mating surfaces of any of the four components of the flasks neither ground nor machined properly.
Name or trademark of the manufacturer not present on exterior of top or bottom cover of flask,
Serial numbers on components of the same flask not identical.
Name or trademark of the manufacturer and serial numbers illegible, or incomplete.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:27Z
Check the source to verify that this is the current version before use.

Tests.

Chemical analysis. In the absence of certificates of quality, chemical analysis of the instrument shall be determined in accordance with method 111 or 112 of FED-STD-151. In the event of a conflict, method 111 shall govern.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producer's own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch/pounds units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:27Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

National Stock Number	Item Identification
6520-00-926-4770	FLASK, DENTURE, CROWN AND BRIDGE

Ordering data. Intermediate/exterior package quantities, labeling and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS

PROJECT NO. 6520-2280
LOCATION: NCR DISK 2

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:27Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1 DOCUMENT NUMBER	2 DOCUMENT TITLE
A-A-53980	FLASK, DENTURE, CROWN AND BRIDGE

3a. NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR ☐ USER ☐ MANUFACTURER ☐ OTHER *(Specify)* _____
b ADDRESS *(Street, City, State, ZIP Code)*	

5 PROBLEM AREAS

a. Paragraph Number and Wording

b Recommended Wording

c Reason/Rationale for Recommendation

6 REMARKS

7a. NAME OF SUBMITTER *(Last First, MI)* – Optional	b WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c MAILING ADDRESS *(Street City State ZIP Code)* – Optional	8 DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:27Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE. This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
Attn: DPSC-RSTE
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:27Z
Check the source to verify that this is the current version before use.