A-A-53985
28 February 1990

COMMERCIAL ITEM DESCRIPTION

ARTICULATOR, DENTAL, Galetti Type

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers a Galetti type dental articulator for use in preventive and corrective orthodontic cases and in orthognatical surgery.

Salient Characteristics:

Material. The upper and lower member assemblies shall be fabricated from cast aluminum. The following components shall be fabricated from anodized aluminum: cap nut, cap screws, ball joint lock, locknut, wingnut and moveable terminals . The spring shall be fabricated from steel. The central shaft, pin, shaft screw and latch screw shall be fabricated from chrome plated steel. The clamp thumbscrew shall be fabricated from chrome plated brass and chrome plated steel. The shaft pad shall be fabricated from bronze. The latch shall be fabricated from high impact rigid plastic. The keystone shall be fabricated from brass. All other components shall be fabricated from chrome plated brass.

Style and design. The articulator shall be a plasterless Galetti type non-arcon articulator capable of rapidly mounting dental casts of the maxillary and mandibular into a selected occlusal relation. The articulator shall be capable of relating the maxillary and mandibular model in centric occlusion or relation and executing protrusive, retrusive, lateral and vertical excursions. All mating parts shall fit properly. All hinges shall provide free operation and movement of the hinged parts. All locking screws and thumbnuts shall lock the parts securely in the required position.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:27Z
Check the source to verify that this is the current version before use.

The upper and lower member assemblies shall each have three circular grooved terminals. The terminals shall be designed to grasp the dental casts. Two of the 3 terminals shall be fixed to the members, one at each front corner. The third terminal on each member shall be moveable along a slot. A cast locking screw shall facilitate adjustment of the third terminal from the front to the back of each member assembly. The locking screw shaft shall enter through a hole in the front leg of the member assembly and screw into the underside of the moveable terminal. When the locking screw is tightened the cast shall be locked in place by the three terminals.

The locking screw shaft shall have a length of approximately 3 inches with a threaded length of at least 2 inches. The vertical dimension of occlusion shall be precisely adjustable so as to permit simulation of a freeway space not less than 3/4 inch and arrestable within the range of the freeway so as to produce a positive stop. The clamp thumbsrew and central shaft shall be used to lock and adjust the upper assembly. There shall be a ball joint which attaches to the central shaft at the rear of the upper member. The ball joint shall have a lock to adjust the incline of the upper member.

The upper member shall be detachable from the articulator by disconnecting the latches at the back of the device. There shall be a vertical stop mounted at the rear of the articulator designed to prevent the upper member from closing onto the lower member. The articulator shall weigh a minimum of 1 pound 5 ounces. The style and design shall be in accordance with Defense Personnel Support Center Drawing #24569.

Finish. The upper and lower members and keystone shall be protected with a corrosion-inhibiting, baked, epoxy finish. All other surfaces shall have a smooth, satin finish.

Instructions. Adequate and accurate instructions shall be furnished with each unit on the use and care of the articulator.

Workmanship. The instrument shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One articulator with instructions for use, as specified, constitutes one unit.

Unit package. The articulator shall be packaged in a container designed to prevent damage during shipment or storage.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:27Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, at an inspection level of II with an AQL of 2.5 (percent defective). Where feasible, the same sample shall be used for the determination of two or more inspection characteristics. The unit of product for examination purposes shall be one articulator. Examination shall include, but shall not be limited to the defects listed below.

Defects

Style and design not as specified.
All mating parts do not fit properly.
Operation not smooth.
Finish not as specified.
Components missing, cracked, deformed.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:27Z
Check the source to verify that this is the current version before use.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105, at an inspection level of S-1 with an AQL of 1.0 (percent defective). The unit of product for test purposes shall be one articulator.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch/pounds units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:27Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

National Stock Number	Item Identification
6520-01-162-1974	ARTICULATOR, DENTAL, Galetti, Type

Ordering data. Intermediate/exterior package quantities and labeling and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

53239

Project No. 6520-2052

Location: NCR Disk #4

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:27Z
Check the source to verify that this is the current version before use.

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-53985	ARTICULATOR, DENTAL, GALETTI TYPE

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)* _____

5. PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:27Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
Attn: DPSC-RSTE
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:27Z
Check the source to verify that this is the current version before use.