A-A-53991
30 November 1989

COMMERCIAL ITEM DESCRIPTION

VALVE, SURGICAL DRAIN, CHEST, HEIMLECH, Disposable

The General Services Administration has authorized the use of this commercial item description as a replacement for Military Specification MIL-V-36535A, which is cancelled.

This commercial item description covers a sterile, disposable, surgical drain valve of the Heimlech design suitable for controlling chest drainage.

Salient characteristics:

Materials.

The rubber flutter vent tubing shall be a single piece of formed pure gum rubber, amber colored and lightly powered with talc. Flutter vent tubing sleeve shall be fabricated from amber colored, latex rubber. The inlet port shall be fabricated from rigid plastic and shall be colored for indentification purposes. The body casing shall be fabricated from rigid transparent plastic.

Construction. The valve shall be designed to provide a safe, one-way chest drainage device. The valve assembly shall be in accordance with DPSC Drawing No. 24090. The vent tubing shall be open at the proximal end, flat in section, with internal surfaces in contact for a sufficient length from the distal end to prevent reverse flow, even during passage of fluid, air or clots. The open end shall be secured to the distal end of the plastic inlet port in a leakproof manner. The valve distal and proximal ends shall have external grooved tapers to accommodate catheters or plastic tubing. The distal end shall contain internal grooves to secure this end to the body casing. The proximal end of the body casing shall be open and have formed protusions which shall snap into grooves of the plastic inlet port to form a leakproof assembly after the rubber flutter vent tubing and the flutter vent tubing sleeve have been assembled to the plastic inlet port. The inlet and outlet ports shall be designed to accommodate catheters or tubing in a secure manner without the need of clamping devices.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:27Z
Check the source to verify that this is the current version before use.

Finish. The valve shall be smoothly finished. Surfaces shall be free from cracks, pits, excessive flash, holes, and foreign matter.

Security of assembly.

Rubber flutter vent tubing to plastic inlet port. The rubber flutter vent tubing shall not separate or loosen from the plastic inlet port when subjected to an axial force of 2 pounds for ten seconds either during production of the valve or on the end item (destructive testing).

Plastic inlet port to plastic casing. The plastic inlet port and plastic body casing junction shall not separate or loosen when subjected to a 10 pound load for ten seconds.

Leakage.

Plastic inlet port to body casing. The inlet port shall be plugged in a leakproof manner. The drain valve shall be placed in an upright position and a suitable liquid shall be admitted into the outlet port until the drain valve is completely filled. The junction of the plastic inlet port to body casing, when observed for 3 minutes, shall show no leakage.

Drain valve assembly. The complete drain valve assembly shall be subjected to a vacuum of 0.5 inches of mercury applied to the inlet port with the outlet port plugged in a leakproof manner. The assembly shall not lose total vacuum for 10 seconds. The assembly shall permit free air flow from the inlet port through the formed rubber vent tubing in normal use.

Pyrogens, toxicity and biological tests. All plastic and rubber components shall be free from pyrogens in accordance with USP requirements. Limulus Amebocyte Lysate test is an acceptable alternate method in lieu of live rabbit injection. All plastic shall meet the requirements of the U.S.P. for Class II plastics. Each lot of product need not to be tested provided evidence is furnished that samples of the formulation and identical method of processing have previously been tested, that they comply with USP requirements for Class II plastics and that they are the same as being offered on this contract. The inlet port and rubber flutter vent tubing components shall pass the safety (toxicity) test specified in the U.S.P.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:27Z
Check the source to verify that this is the current version before use.

Sterility. All components of the drain valve shall be sterile. The sterility assurance level shall be 1×10^{-6} in accordance with AAMI guidelines.

Workmanship. Workmanship shall be first class throughout. The surgical drain valves shall be free from defects which detract from their appearance or impair their serviceability. The flutter vent tubing shall be free from pits, holes on cuts.

Unit. Package (PG). One package containing ten individually sterile valves, as specified, constitutes one unit. The individual package shall be a sealed peel open type container capable of maintaining sterility of contents unless package is damaged or opened.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one surgical drain. Inspection level shall be S-2, AQL 0.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:27Z
Check the source to verify that this is the currentversion before use.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerance specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for all compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:27Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6515-00-926-9150	VALVE, SURGICAL DRAIN, Chest, Heimlech, Disposable, 10s

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-4677

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:27Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1 DOCUMENT NUMBER	2 DOCUMENT TITLE
A-A-53991	VALVE, SURGICAL DRAIN, CHEST, HEIMLECH, Disposable

3a NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
b ADDRESS *(Street, City, State, ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)* _____

5 PROBLEM AREAS

a Paragraph Number and Wording

b Recommended Wording

c Reason/Rationale for Recommendation

6 REMARKS

7a NAME OF SUBMITTER *(Last, First, MI)* – Optional	b WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8 DATE OF SUBMISSION *(YYMMDD)*

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:27Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE. This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 49pt Alexandria, VA

POSTAGE WILL BE PAID BY

Commander
Defense Personnel Support Center
ATTN: DPSC-RST
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:27Z
Check the source to verify that this is the current version before use.