A-A-53993
10 October 1989

COMMERCIAL ITEM DESCRIPTION

RACK, BUR, H-Shaped, 6 Inches by 4 Inches

The General Services Administration has authorized the use of this commercial item description.

Salient characteristics:

This commercial item description covers a bur rack which shall be H-shaped, suitable for holding, displaying, storing, and processing 44 mm to 68 mm long, 3/32 inch diameter burs. (See notes).

Material. The rack shall be fabricated from coated aluminum alloy 6061 in accordance with ASTM Designation B221 for chemical composition only.

Style, design and dimensions. The rack shall be 1/2 inch in depth when the swivel bar display on each end is folded in the downward position, and shall weigh 7 to 8 ounces. The rack shall be "H"-shaped with two rectangular-shaped bur holding segments mounted on each end. The rectangular bur holding segments shall be secured to the inner portion of the "H" via swivel type connections, allowing the burs to be displayed when in the up position and stored and processed when in the down position. One end of each of the bur holding segments shall be perforated to allow insertion of twelve, 3/32 inch diameter burs. The other end of each of the bur holding segments shall be notched to hold the proximal portion of twelve various length, 3/32 inch diameter burs. One segment shall accommodate five 44 mm long burs and seven 55 mm long burs. The other segment shall accommodate twelve 65 mm long burs. The rack shall conform to Figure 1 for design and dimensions.

Finish. The rack shall be machine finished to 64 microinches root mean square (RMS) average deviation in surface roughness.

Corrosion resistance. The rack shall not show evidence of corrosion after being subjected to 5 cycles of steam autoclaving at a temperature of 270° (^{0}F) + 5 ^{0}F and a pressure of 27 pounds per square inch (psi) $\pm$ 1 psi for a minimum of 20 minutes, or ethylene oxide sterilization.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:27Z
Check the source to verify that this is the current version before use.

Workmanship. The rack shall be free from defects which may detract from its appearance or impair its serviceability.

Unit. Each (EA). One rack, as specified, constitutes one unit.

Unit package. Each rack shall be suitably packaged in a box of appropriate size, with adequate cushioning to preclude damage to or shifting of the rack during shipping and handling.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Certificates of quality. Certificates of quality for raw material, supplied by the manufacturer of the aluminum, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the Item Identification, the name of the material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, at an inspection level of II.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:27Z
Check the source to verify that this is the current version before use.

Examination. The rack shall be examined to determine compliance with all requirements contained in this commercial item description. Examination for defects shall include but shall not be limited to those listed below:

Defects

Style, design and dimensions not as specified.
Finish not as specified.
Workmanship not as specified.
Rack does not hold 24 burs of the size specified.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product shall be one bur rack. Inspection level shall be S-1.

Tests. Tests shall be conducted to determine compliance with specification requirements.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producer's own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch/pounds units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:27Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTES:

1. The following National Stock Number is covered by this document:

National Stock Number	Item Identification
6515-01-167-2624	RACK, BUR, H-Shaped, 6 inches by 4 inches (15.24 cm by 10.16 cm)

2. The following National Stock Numbers are used with the bur rack:

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:27Z
Check the source to verify that this is the currentversion before use.

National Stock Number	Item Identification
6515-01-166-7786	BUR, 4.8 by 55 mm, 6 flutes
6515-01-166-7787	BUR, 3.2 by 55 mm, 6 flutes
6515-01-166-7788	BUR, 6.3 by 44 mm, 8 flutes
6515-01-167-2616	BUR, 5.5 by 44 mm, 8 flutes
6515-01-166-9005	BUR, 4.0 by 55 mm, 6 flutes
6515-01-167-2622	BUR, 4.0 by 44 mm, 6 flutes
6515-01-167-3788	BUR, 3.2 by 44 mm, 6 flutes
6515-01-167-3789	BUR, 4.8 by 44 mm, 6 flutes
6515-01-166-9006	BUR, 2.4 by 44 mm, 4 flutes
6515-01-193-5038	BUR, 1.85 by 55 mm, no flutes
6515-01-193-5039	BUR, 2.30 by 55 mm, no flutes
6515-01-193-5040	BUR, 1.5 by 55 mm, 3 flutes
6515-01-193-5041	BUR, 1.65 by 55 mm, no flutes
6515-01-193-5042	BUR, 1.5 by 68 mm, 4 flutes
6515-01-194-0383	DRILL, WIRE PASS, CRS 1.5 by 64 mm
6515-01-268-2535	BUR, 3.2 by 65 mm, 6 flutes
6515-01-267-9633	BUR, 4.0 by 65 mm, 6 flutes
6515-01-267-6110	BUR, 4.8 by 65 mm, 6 flutes

Ordering data. Intermediate/exterior package quantities, labeling and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:
Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:27Z
Check the source to verify that this is the current version before use.

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS

PROJECT NO. 6515-4718
LOCATION: NCR DISK 2

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:27Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-53993	RACK, BUR, H-Shaped, 6 Inches by 4 Inches

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)*: _____

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil – Downloaded: 2026-10-26T09:27Z
Check the source to verify that this is the current version before use.