A-A-53997
10 October 1989

COMMERCIAL ITEM DESCRIPTION

MOUTHPIECE, SALIVA EJECTOR WITH REFLECTOR BLADES, DENTAL

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for Military Specification MIL-M-36221A, dated 4 May 1987, which is cancelled.

This commercial item description covers a saliva ejector mouthpiece with reflector blades and an adjustable chin clamp, which can be attached to the saliva ejector tube of a dental operating unit during dental operative procedures.

Salient characteristics:

Material. The saliva ejector tube, shall be fabricated of brass in accordance with ASTM Designation B135 for composition only, and shall be chromium over nickel plated. The saliva ejector spring, fillister head spring screw, chin clamp assembly (clamp, plate and knurled head thumb screw) and the three reflector blades shall be fabricated of Class 3 chromium-nickel austenitic corrosion resistant steel, in accordance with ASTM Designation F899 for composition.

Style, design and dimensions. The saliva ejector mouthpiece set shall consist of a saliva ejector tube, saliva ejector spring and fillister head screw, chin clamp assembly (clamp, plate and knurled head thumb screw), and three sizes of interchangeable reflector blades. The saliva ejector mouthpiece, when attached to the saliva ejector tube assembly of the dental operating unit, shall be suitable for removing fluids from the mouth freely, without any obstructions. The reflector blades when assembled into the saliva ejector tube, with adjustable chin clamp, shall serve as a mouth mirror, light reflector, and tongue guard. The reflector blades shall be suitable for secure mounting in the intra-oral portion of the saliva ejector tube in a manner to permit ready interchange of reflector blades without the use of any tools. The saliva ejector mouthpiece and reflector blades, size 1, 2 and 3 shall conform to Defense Personnel Support Center (DPSC) Drawing No. 20733 for style, design and dimensions.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:27Z
Check the source to verify that this is the current version before use.

Finish. The mouthpiece components shall be smoothly finished. All edges shall be broken to prevent any injury to the soft tissue of the oral cavity. The finish of the reflector blades shall provide a surface of optical quality to serve as a mirror during operative procedures. The saliva ejector mouthpiece set shall withstand repeated use without loss of finish.

Corrosion resistance. The mouthpiece sets shall show no sign of corrosion when subjected to the boil test specified in ASTM Designation F1089.

Instructions. Instructions for use shall accompany each mouthpiece set.

Identification marking. The chin clamp of each mouthpiece set of domestic manufacture shall be marked with the name or registered trademark of the manufacturer. The chin clamp of each mouthpiece set of foreign manufacture shall be marked with the name of the country of origin, together with the name or registered trademark of the manufacturer. All markings shall be legible and permanent.

Workmanship. The mouthpiece shall be free from defects which may detract from its appearance or impair its serviceability.

Unit. Each (EA). One mouthpiece, as specified, constitutes one unit.

Unit package. Unit packaging shall be a carton of appropriate size, containing sufficient cushioning material to insure against shifting of or damage to the mouthpiece during shipping and handling. Closure shall be adequate to prevent spillage of contents under normal handling.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:27Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, at an inspection level of S-2. The unit of product for examination purposes shall be one mouthpiece.

Examination. The mouthpieces shall be examined for defects including, but not limited to those listed below:

Defects

Style and design not as specified.
Any component missing.
Reflector blades do not mount securely in intra-oral portion of saliva ejector tube.
Reflector blades not interchangeable without the use of tools.
Chin clamp not readily adjustable.
Saliva ejector tube not free of obstructions.
Finish of the reflector blades does not provide a surface of optical quality.
Mouthpiece not furnished clean.
Dimensions not as specified.
Identification markings on chin clamp not as specified.
Instructions absent, incomplete, illegible, or incorrect.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105, at an inspection level of S-2.

Tests. Tests shall be conducted to determine compliance with requirements specified herein.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:27Z
Check the source to verify that this is the current version before use.

Chemical analysis of the metals. In the absence of certificates of quality for the raw material, chemical analysis shall be conducted in accordance with method 111 or 112 of FED-STD-151. In case of dispute, method 111 shall be used for referee analysis.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producer's own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch/pounds units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:27Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

National Stock Number	Item Identification
6520-00-074-4947	MOUTHPIECE, SALIVA EJECTOR, WITH REFELCTOR BLADES, DENTAL

Ordering data. Intermediate/exterior package quantities, labeling and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

PROJECT NO. 6520-2297
LOCATION: NCR DISK 2

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:27Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
Attn: DPSC-RSTE
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:27Z
Check the source to verify that this is the current version before use.

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1 DOCUMENT NUMBER	2 DOCUMENT TITLE
A-A-53997	MOUTHPIECE, SALIVA EJECTOR WITH REFLECTOR BLADES, DENTAL

3a. NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b ADDRESS *(Street, City, State ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)* _____

5 PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6 REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8 DATE OF SUBMISSION *(YYMMDD)*

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:27Z
Check the source to verify that this is the current version before use.

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE